

REUSABLE JARS are equipped with screw cover with handle for easy grip, autoclavable silicone o-ring for tight seal, overflow valve system integrated in the cover, clear graduated scale in ml with 100ml or 200ml intervals, CA-MI branding. Suitable for central vacuum systems and CA-MI suction units. Available in two types of polycarbonate, Makrolon® and Apec®.

REUSABLE JARS in MAKROLON® – FOR STEAM STERILIZATION UP TO 121 °C



	Jar with: cover overflow valve o-ring	Jar	Cover with: overflow valve o-ring	O-ring	Suitable for:
400 ml	RE 210301	RE 210305	RE 210302	RE 210304	EMIVAC
1000 ml	RE 210001/02	RE 210003	RE 210352/01	RE 210354	ASPIRET ASKIR series (all except C30 series) Central Vacuum Plants
2000 ml	RE 210351/01	RE 210353	RE 210352/01	RE 210354	ASPIRET - ASKIR Series HOSPIVAC Series Central Vacuum Plants
4000 ml	RE 210006	RE 210007	RE 210008	RE 210306	HOSPIVAC Series Vuoto centralizzato
5000 ml	RE 210010	RE 210013	RE 210012	RE 210307	HOSPIVAC Series Central Vacuum Plants

REUSABLE JARS in APEC® 1745 – FOR STEAM STERILIZATION UP TO 143 °C



	Jar with: cover overflow valve o-ring	Jar	Cover with: overflow valve o-ring	O-ring	Suitable for:
1000 ml	RE 210009	RE 210002	RE 210352/02	RE 210354	ASPIRET ASKIR series (all except C30 series) Central Vacuum Plants
2000 ml	RE 210351/05	RE 210353/01	RE 210352/02	RE 210354	ASPIRET - ASKIR Series HOSPIVAC Series Central Vacuum Plants
5000 ml	RE 210010/01	RE 210013/01	RE 210012/01	RE 210307	HOSPIVAC Series Vuoto centralizzato

FLOVAC® DISPOSABLE LINERS equipped with polyethylene disposable liner, hydrophobic filter, antibacterial filter, overflow system, reusable container with clear graduated scale in ml with 50ml intervals. The gelling kit is a powder inside liners with germicidal function turning the sucked liquid into a semisolid mass, preventing cross-contamination risks of staff in charge for hygiene and waste disposal.

FLOVAC® DISPOSABLE LINERS & CONTAINERS



	Liner with: cover filter		Liner with: cover filter gelling kit		Reusable Container	Suitable for
1000 ml	31848		31858		31843**	ASPIRET ASKIR series (all except C30 series)
1000 ml	31845		31854		31840	Central Vacuum Plants
2000 ml	31846		31855		31841	ASPIRET - ASKIR Series HOSPIVAC Series Central Vacuum Plants
3000 ml	31847		31856		31842	HOSPIVAC Series Central Vacuum Plants

** Requires additional round spacer (SP.0220) when ordered for ASKIR Series



Certificate

No. Q5 063105 0045 Rev. 03

Holder of Certificate:



CA-MI S.R.L.

Via Ugo La Malfa, 13
Frazione Pilastro
43013 Langhirano (PR)
ITALY

Certification Mark:



Scope of Certificate:

Design and development, production, service and sale of medical equipments for surgery (electrical and manual suction pumps), electric breast pumps, medical devices for breathing (aerosol therapy equipments and thermal water inhaler) and related accessories, medical devices for monitoring of vital physiological parameters (pulse oximeters, thermometers, blood pressure monitors, sphygmomanometer and stethoscopes), incentive spirometers and medical devices for phlebology (graduated compression medical stockings). Distribution of active and non-active non implantable medical devices.

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned above has established and is maintaining a quality management system, which meets the requirements of the listed standard(s). All applicable requirements of the testing and certification regulation of TÜV SÜD Group have to be complied with. For details and certificate validity see: www.tuvsud.com/ps-cert?q=cert:Q5_063105_0045_Rev_03

Report No.: ITA1885389

Valid from: 2022-08-02

Valid until: 2025-08-01

Date, 2022-08-02

Christoph Dicks
Head of Certification/Notified Body

Certificate

No. Q5 063105 0045 Rev. 03

Applied Standard(s):

EN ISO 13485:2016
Medical devices - Quality management systems -
Requirements for regulatory purposes
(ISO 13485:2016)
DIN EN ISO 13485:2016

Facility(ies):

CA-MI S.R.L.
Via Ugo La Malfa, 13, Frazione Pilastro, 43013 Langhirano (PR),
ITALY

Design and development, production, service and sale of medical equipments for surgery (electrical and manual suction pumps), electric breast pumps, medical devices for breathing (aerosol therapy equipments and thermal water inhaler) and related accessories, medical devices for monitoring of vital physiological parameters (pulse oximeters, thermometers, blood pressure monitors, sphygmomanometer and stethoscopes), incentive spirometers and medical devices for phlebology (graduated compression medical stockings). Distribution of active and non-active non implantable medical devices.

CA-MI S.r.l.
Via Strada per Parma 34, Frazione Pilastro, 43013 Langhirano
(PR), ITALY

Warehouse of active and non-active non implantable medical devices and components used in production.

CA-MI S.r.l.
Via Ugo La Malfa 27, Frazione Pilastro, 43013 Langhirano (PR),
ITALY

Production of medical devices for surgery (electrical and manual suction pumps), warehouse of active and non-active non implantable medical devices and components used in production.

/



EC Certificate

Production Quality Assurance System
Directive 93/42/EEC on Medical Devices (MDD), Annex V
(Devices in Class IIa, IIb or III)

No. G2 063105 0047 Rev. 01

Manufacturer:

CA-MI S.R.L.

Via Ugo La Malfa, 13
Frazione Pilastro
43013 Langhirano (PR)
ITALY

**Product
Category(ies):**

**Aerosol Therapy Equipment, Kits for Aerosol Therapy,
Thermal Water Inhaler, Suction Unit, Surgical Suction
Equipment, Breast Pump, Kit Accessory for Electric
Breast Pump, Blood Pressure Monitor, Electronic
Thermometer, Infrared Thermometer, Tens Device,
Pulse Oximeter**

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for manufacture and final inspection of the respective devices / device categories in accordance with MDD Annex V. This quality assurance system conforms to the requirements of this Directive and is subject to periodical surveillance. For marketing of class IIb and III devices an additional Annex III certificate is mandatory. All applicable requirements of the testing and certification regulation of TÜV SÜD Group have to be complied with. For details and certificate validity see: www.tuvsud.com/ps-cert?q=cert:G2_063105_0047_Rev_01

Report No.:

ITA1626749

Valid from:

2021-02-09

Valid until:

2024-05-26

Date,

2021-02-09

Christoph Dicks
Head of Certification/Notified Body