

**EC Certificate**  
**Directive 93/42/EEC Annex V**  
**Production Quality Assurance**  
**Medical Devices**

**Registration No.:** DD 60135422 0001

**Report No.:** 15083140 006

**Manufacturer:** Xiantao Tongda Non-woven  
Products Co., Ltd.  
No. 28, Pengchang Road  
433018 Xiantao  
China

**Products:** Aspects of manufacture concerned with securing and  
maintaining sterile conditions:  
Face Masks, Surgical Gowns, Coveralls, Caps  
Replaces Approval, Registration No.: DD 60102531 0001

**Expiry Date:** 2024-01-16

The Notified Body hereby declares that the requirements of Annex V of the directive 93/42/EEC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex V, section 4 of the aforementioned directive. For placing on the market of class IIb and class III devices covered by this certificate an EC type-examination certificate according to Annex III is required.

**Effective Date:** 2019-07-16

**Date:** 2019-07-16



Notified Body

*M. Aihara*  
M.Sc. M. Aihara

**TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg**  
TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC  
concerning medical devices with the identification number 0197.

# 中华人民共和国医疗器械注册证

注册证编号：鄂械注准20202142956



注册人名称	仙桃市通达无纺布制品有限公司
注册人住所	仙桃市彭场镇仙沙公路南侧
生产地址	仙桃市彭场镇禾丰村胡家路南侧
代理人名称	不适用
代理人住所	不适用
产品名称	一次性使用医用口罩
型号、规格	平面耳挂式(长×宽)：175*95mm 非灭菌级。
结构及组成	口罩由两层聚丙烯纺粘无纺布夹一层聚丙烯熔喷布经折叠超声波复合而成。口罩上配有鼻夹及口罩带，鼻夹由可塑性材料聚丙烯镀锌铁线制成；口罩带为耳挂式，由涤纶和氨纶材料制成。
适用范围	供临床各类人员在非有创操作过程中佩戴，覆盖住使用者的口、鼻及下颌，为防止病原体微生物、颗粒物的直接透过提供一定的物理屏障。
附件	产品技术要求。
其他内容	无
备注	医疗器械注册证有效期1年。

审批部门：湖北省药品监督管理局

批准日期：2020年5月9日

有效期至：2021年5月8日





## Fiscal Year 2020 CERTIFICATION OF REGISTRATION

This certifies that:

**Xiantao Tongda Non-Woven Products Co., Ltd.**  
**South Of Xiansha Road, Pengchang Town, Xiantao, Hubei, China**

has completed the FDA Establishment Registration (as manufacturer) and Device Listing with the US Food & Drug Administration, through

U.S. Agent for FDA           SUNGO TECHNICAL SERVICE INC.  
Communications:           6050 W EASTWOOD AVE APT 201, CHICAGO,  
  ILLINOIS 60630, USA  
  Telephone: +1-855-957-7779 / E-mail: sungo.group@yahoo.com

**Registration Number:3011511692**  
**Device Listing#: See annex**

*SUNGO Technical Service Inc. will confirm that such registration remains effective upon request and presentation of this certificate until the end of the calendar year stated above, unless said registration is terminated after issuance of this certificate. SUNGO Technical Service Inc. makes no other representations or warranties, nor does this certificate make any representations or warranties to any person or entity other than the named certificate holder, for whose sole benefit it is issued. This certificate does not denote endorsement or approval of the certificate-holder's device or establishment by the U.S. Food and Drug Administration. SUNGO Technical Service Inc. assumes no liability to any person or entity in connection with the foregoing.*

*Pursuant to 21 CFR 807.39, "Registration of a device establishment or assignment of a registration number does not in any way denote approval of the establishment or its products. Any representation that creates an impression of official approval because of registration or possession of a registration number is misleading and constitutes misbranding." The U.S. Food and Drug Administration does not issue a certificate of registration, nor does the U.S. Food and Drug Administration recognize a certificate of registration, SUNGO Technical Service Inc. is not affiliated with the U.S. Food and Drug Administration.*



**Executive Director**  
**Issued: Dec. 23 2019**  
**Cert. No.: 2006US464538**  
**Expiration Date: Dec. 31 2020**



## Fiscal Year 2020 CERTIFICATION OF REGISTRATION

Annex to Cert. No.: 2006US464538

<b>Listing No</b>	<b>Code</b>	<b>Device Name</b>
D306060	FXP	COVER, SHOE, OPERATING-ROOM (Boot Cover; Disposable Slipper; Shoe Cover; Sleeve Cover)
D306061	FOG	HOOD, OXYGEN, INFANT (Disposable Hood)
D306062	FYF	CAP, SURGICAL (Cap/Disposable Cap/Surgical Cap/Non Woven Cap; Nurse Cap; Worker Cap; Doctor Cap)
D306063	KME	BEDDING, DISPOSABLE, MEDICAL (Disposable Bed Sheet; Disposable Pillow Cover)
D306064	LYU	ACCESSORY, SURGICAL APPAREL (Beard Cover)
D306065	FME	GOWN, EXAMINATION (Patient Gown)
D306066	IWO	APRON, PROTECTIVE (Apron)
D306067	FXO	SUIT, SURGICAL (Coverall)
D306068	OEA	NON-SURGICAL ISOLATION GOWN (Cpe Gown; Isolation Gown; Pe Rain Coat)
D306069	FYE	DRESS, SURGICAL (DRESSING/Visitor Gown/Clothing; Visitor Gown; Surgical Gown)
D306070	BWP	SHOE AND SHOE COVER, CONDUCTIVE (Shoe Cover)
D307072	KHA	MASK, SCAVENGING (Face Mask)

END OF THE ANNEX



# 检测报告

## (Test Report)

No. BOYBXQSR817757L1

样品名称 (Sample Description)	一次性医用口罩 Disposable Medical Face Mask
委托单位 (Applicant)	仙桃市通达无纺布制品有限公司 Xiantao Tongda Non-woven Products Co.,Ltd.

PONY 谱尼测试  
Pony Testing International Group  
[www.ponytest.com](http://www.ponytest.com)



声明  
Statement

1. 本报告无专用章和批准人签章无效。  
This report is invalid without the approver's signatures and special seal of inspection.
2. 本报告页面所使用“PONY”、“谱尼”字样为本单位的注册商标，其受《中华人民共和国商标法》保护，任何未经本单位授权的擅自使用和仿冒、伪造、变造“PONY”、“谱尼”商标均为违法侵权行为，本单位将依法追究其法律责任。  
The pattern and characters of "PONY" and "谱尼" used in this report are protected by the trademark law of the People's Republic of China. Any unauthorized usage, counterfeit, forgery and alteration of trademarks of "PONY" and "谱尼" are the violations of the law. The PONY has the right to pursue all legal liabilities of the subject of the delict.
3. 委托单位对报告数据如有异议，请于报告完成之日起十五日内（初级农产品报告请于报告收到之日起五日内）向本单位书面提出复测申请，同时附上报告原件并预付复测费。  
If the applicant has any questions about the results, shall provide a written retest application with the original report, and prepay the retest fees to PONY within fifteen days since the approval date (as an exception, it shall be within five days since the date received for the primary agriculture products report).
4. 委托单位办理完毕以上手续后，本单位会尽快安排复测。如果复测结果与异议内容相符，本单位将退还委托单位的复测费。  
After the applicant finishes the procedure mentioned above, PONY shall arrange the retest as soon as possible. If the retest result accords with the applicant dissent, PONY shall refund the retest fees.
5. 不可重复性或不能进行复测的实验，不进行复测，委托单位放弃异议权利。  
Tests that can not be repeated and tested shall not be carried out again.
6. 委托单位对样品的代表性和资料的真实性负责，否则本单位不承担任何相关责任。  
The applicant should undertake the responsibility for the provided samples' representativeness and document authenticity. Otherwise, PONY has not any relevant responsibilities.
7. 本报告仅对所测样品负责，报告数据仅反映对所测样品的评价，对于报告及所载内容的使用、使用所产生的直接或间接损失及一切法律后果，本单位不承担任何经济和法律后果。  
This report is only responsible for the provided sample. The test results only represent the evaluation of the tested sample. PONY will not be responsible for any economical or legal liability generated from direct or indirect usage of the test report.
8. 本单位有权在完成报告后处理所测样品。  
PONY has the right to dispose the tested sample after approval of the test report.
9. 本单位保证工作的客观公正性，对委托单位的商业信息、技术文件等商业秘密履行保密义务。  
PONY assures objectivity and impartiality of the test, and fulfills the obligation of confidentiality for applicant's commercial information, and technique document.
10. 本报告私自转让、盗用、冒用、涂改、未经本单位批准的复制（全文复制除外）或以其它任何形式的篡改均属无效，本单位将对上述行为严究其相应的法律责任。  
The report is invalid in case of illegal transfer, embezzlement, imposture, modification or any altering, reproducing except in full, without approval of PONY. PONY shall investigate and affix the applicant's legal liability accordingly.

\*\*\*\*\*  
**▲ 防伪说明 (Anti-counterfeiting Description):**

- (1) 报告编号是唯一的;  
The test report has exclusive report code.
  - (2) 报告采用特制防伪纸张印制, 纸张表面带有 "PONY" 防伪纹路, 该防伪纹路不支持复印, 即复制件不会带有 "PONY" 防伪纹路。  
The test report is printed by anti-copying paper whose surface shows "PONY" security print with specific anticounterfeiting technique. Security print will disappear after copying. Duplicates are not expected to give "PONY" security print under any circumstances.
- \*\*\*\*\*



全国服务热线

400-819-5688

WWW.PONYTEST.COM

扫描二维码

关注谱尼测试微信公众号

PONY4008195688



北京实验室: (010) 83055000 武汉实验室: (027) 83997127 哈尔滨实验室: (0451) 58627755  
上海实验室: (021) 64851999 长春实验室: (0431) 85150908 石家庄实验室: (0311) 85376660 温州实验室: (0577) 88271060  
青岛实验室: (0532) 88706866 大连实验室: (0411) 87336618 乌鲁木齐实验室: (0991) 6684186 合肥实验室: (0551) 63843474  
深圳实验室: (0755) 26050909 郑州实验室: (0371) 69350670 呼和浩特实验室: (0471) 3450025 广州实验室: (020) 89224310  
天津实验室: (022) 23607888 西安实验室: (029) 89608785 杭州实验室: (0571) 87219096 厦门实验室: (0592) 5568048  
苏州实验室: (0512) 62997900 太原实验室: (0351) 7555762 宁波实验室: (0574) 87736499 成都实验室: (028) 87702708

**检测结果**  
(Test Results)

No. BOYBXQSR817757L1

第 1 页, 共 2 页 (page 1 of 2)

样品名称 (Sample Description)	一次性医用口罩 Disposable Medical Face Mask	样品规格 (Sample Specification)	---
委托单位 (Applicant)	仙桃市通达无纺布制品有限公司 Xiantao Tongda Non-woven Products Co.,Ltd.	商标 (Trade Mark)	---
到样日期 (Received Date)	2020-03-25	生产日期或批号 (Manufacturing Date or Lot No.)	---
检测日期 (Test Date)	2020-03-25~2020-04-08	样品等级 (Sample Grade)	---
样品数量 (Sample Quantity)	50 只 50Pcs	检测类别 (Test Type)	委托检测 (Commissioning Test)
样品状态 (Sample Status)	全新 50 只, 包装完好, 3 层无纺布, 熔喷布过滤, 耳挂 New 50 pcs, packed in good condition, 3 layers of non-woven cloth, fusible spray cloth filter, ear hanging	检测环境 (Test Environment)	符合要求 (To meet the requirements)
检测项目 (Test Items)	见下页 See the next page		
检测方法 (Test Methods)	见下页 See the next page		
所用主要仪器 (Main Instruments)	恒温培养箱 Constant Temperature Incubator、细菌过滤效率检测仪 Bacterial filtration efficiency detector 等 etc		
备注 (Note)	限值标准 Limit on: EN 14683: 2019		
	编制人 (Edited by)	[Signature]	
	审核人 (Checked by)	郭惠	
	批准人 (Approved by)	戴晴	
	签发日期 (Issued Date)	2020-04-08	


## 检测结果 (Test Results)

No. BOYBXQSR817757L1

第 2 页, 共 2 页 (page 2 of 2)

样品名称和编号 (Sample Description and Number)	检测项目 (Test Items)	II R 型 Type II R 限值 (Limit)	检测结果 (Test Results)		单项结论 (Evaluation)	检测方法 (Test Methods)
			No.1	No.2		
R817757L1 一次性医用口罩 Disposable Medical Face Mask	细菌过滤效率, % Bacterial filtration efficiency(BFE)	≥98	No.1 99.2	No.2 99.2	符合 Qualified	EN 14683: 2019
			No.3 99.0	No.4 98.9		
			No.5 98.7			
	压力差, Pa/cm <sup>2</sup> Differential pressure	<60	23.3			
抗溅压力, kPa Splash resistance pressure	≥16.0	>16.0				
	微生物清洁度, cfu/g Microbial cleanliness	≤30	<1		符合 Qualified	EN 14683: 2019

样品编号和照片 (Sample Number and Photo):



仅对报告照片中的样品负责

Pony authenticate the photo on original report only

——以下空白——

(End of Report)



