

## IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

## CB TEST CERTIFICATE

Product

Ultrasound Diagnostic Systems

Name and address of the applicant

VINNO Technology (Suzhou) Co., Ltd.  
5F Building A, 4F Building C, No.27 Xinfu Rd., Suzhou Industrial  
Park, Suzhou, Jiangsu 215123, China

Name and address of the manufacturer

Same as applicant

Name and address of the factory

Note: When more than one factory, please report on page 2

Same as applicant

Ratings and principal characteristics

100-240V~, 50/60Hz, 400VA, Class I

Trademark (if any)



Customer's Testing Facility (CTF) Stage used

-

Model / Type Ref.

VINNO E20, VINNO E10, VINNO E10E, VINNO E10P,  
VINNO X3, VINNO X2, VINNO X2E, VINNO X2P, VINNO X1,  
VINNO X1E, VINNO X1P

Additional information (if necessary may also be reported on page 2)

See page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005+A1+A2  
IEC 60601-1-6:2010+A1+A2  
IEC 60601-2-37:2007+A1

As shown in the Test Report Ref. No. which forms part of this Certificate

211100746SHA-001, 211100746SHA-002,  
211100746SHA-003

This CB Test Certificate is issued by the National Certification Body

Intertek Semko AB  
Torshamnsgatan 43  
Box 1103  
SE-164 22 Kista, Sweden

intertek

Signature:

  
Hyden Li

Date: 20 January, 2022

**Additional information**

The following applicable requirements have not been evaluated by Intertek:

- Biocompatibility according to ISO 10993
- EMC according to IEC 60601-1-2

**Date:** 20 January, 2022

**Signature:**

A handwritten signature in blue ink, appearing to be 'unfllce', is written over the signature line.