



CONFIRMATION FOR INSTRUMENTS

Hereby we, GPC Medical Ltd. confirm that the following instruments

1. Drill Sleeve for Locking Screw 5.0mm, REF: ILBS716.50
2. Drill Sleeve for Locking Screw 3.5mm, REF: ILBS716.35
3. Drill Bit for Locking 4.3 mm, REF: S-ILBS736.43
4. fixLOCK Instrument Set for Large Fragment with Aluminium Graphic Box, REF GOS1439L

are:

- manufactured by GPC Medical Ltd;
- The instruments have increased resistance to mechanical wear;
- Instruments have heat resistance to pupil sterilization up to a temperature of 200 ° C, as well as resistance to vapor sterilization up to a temperature of 134 ° C;
- The instruments have resistance to processing with disinfectant solutions, used in the medical institution;
- The surface of the instruments is matte (so as not to reflect light in the eyes of medical staff);
- The following information is laser-engraved on each instrument:
 - Manufacturer or manufacturer's logo
 - Instrument code
 - CE marking
- Instruments are made of stainless steel, having the following composition (minimum - maximum):
 - C (carbon): 0.30 - 0.45%
 - Cr (chromium): 15.00 - 18.00%
 - Mo (molybdenum): 0.90 - 1.30%
 - Ni (nickel): ≤ 1.00%
 - Si (silicon): ≤ 1.00%
 - Mn (manganese): ≤ 1.00%
 - P (phosphorus): ≤ 0.05%

For GPC Medical Ltd.

Vikas Narang
Director



GPC Square, M Block, DDA LSC,
Vikas Puri, New Delhi - 110018, India



+91-11-43222600 (100 lines)
+91-11-45545172



info@gpcmedical.com
globalproducts@bol.net.in



www.gpcmedical.com
www.indianorthopaedic.com

CIN: U33119DL2004PLC130921

Serving to Grow



Growing to Serve