

To: Centrul Pentru Achizitii Publice Centralizate In Sanatate

Reference: ocds-b3wdp1-MD-1744877607033

Statement

Amstelveen, June 3rd, 2025

We, CANON MEDICAL SYSTEMS EUROPE BV, Bovenkerkerweg 59, 1185XB, The Netherlands, as authorized European Representative of Canon Medical Systems Corporation, 1385, Shimoishigami, Otawara-Shi, 324-8550, Tochigi, Japan, herewith state and confirm the following about our MRI system Vantage Orian 1,5T:

- The maximum sound pressure level (SPL) at peak gradient level and slew rate is not more than 110,5 dB (A)
- The recommended LHe level to take the system into commission is 72%. Out of experience we know that our partner Medexcom hands over the system with a much higher LHe level

CANON MEDICAL SYSTEMS EUROPE BV
Bovenkerkerweg 59, 1185 XB Amstelveen, THE NETHERLANDS


P. Bishop
Power of Attorney

Canon Medical Systems Europe B.V.
Bovenkerkerweg 59
1185 XB Amstelveen
The Netherlands
Phone +31(0)88 2285000

04 June 2025 | 15:18 CEST