

Technical Specification No The main stretcher

• Manufacturer: DAS Ambulance • Country of Origin: Turkey,

• Model: DAS DS250-1

• Certification ISO, CE, EN 1789

Technical Specification	Yes	No	Remarks
Configuration The main stretcher, Length 1960 mm, Width 560 mm Wheel diameter 200 mm with patient fastening system — the patient's weight up to 200 kg. Backrest and horizontal support surface for legs must to be easy removable. Wheight max 53 kg	X X X X X		Automatic release of the legs of the trolley when unloading from the ambulance. Height adjustable, 3.positions. Position Trendelenburg and anti- Trendelenburg when the trolley is on its own wheels. Adult fastening system, including over the patient's shoulders. Fastening system for new-borns and children. Folding support for infusions. Folding lateral handholds. Telescopic handholds for the transportation of the stretcher. Wheel brakes. Folding legs support of the stretcher, both anterior and posterior Manufactured from easy-to-maintain materials. The stretcher and the trolley will support a weight of up to 220 kg separately or combined, including when the equipment is on the wheels. Reusable mattress, made from resistant material which allows an easy wash and disinfection

Certif	ication:				
>	ISO 9001	X			ertificate attached
>	ISO 13485	X			ertificate attached
>	CE	X			ertificate attached
>	CE Declaration of Conformity 93/42/EEC	X			ertificate attached
>	EN 1865-1:2010+A1;2015, EN 1789	X		> c	ertificate attached
	Technical Specification	Yes	No		Remarks
Other	services included in offer				
>	transport to the final customer	X		> b	idder declaration attached
>	instalation and put in function	X		≻ b	idder declaration attached
>	training for the technical staff for final customer	X		> b	idder declaration attached
>	training for the medical staff for final customer	X		> b	idder declaration attached
>	warranty for 24 months	X		≻ b	idder declaration attached
GUAI	RANTEE				
>	24 months guaranteed from signing the receiving	X		> b	idder declaration attached
	minutes				
SERV	TICE AND MAINTENANCE				
>	the bidder has technical facilities necessary for the	X		> b	idder declaration attached
	service both for ambulances as well as for medical				
	equipment according to the general guarantee				
	conditions and the user guide of the manufacturer				
>	Maximum time for technical intervention – 48	X		> b	idder declaration attached
	hours.				
>	Maximum time for remedial measures, in total 72	X		> b	idder declaration attached
	hours.				
>	The technical service and current repair will be	X		≻ b	idder declaration attached
	performed without waiting in a queue.				
>	Temporary replacement of the equipment has to be	X		> b	idder declaration attached
	ensured according to the periods mentioned above.				
>	During the guarantee period, at the reasonable				
	request of the user, repair, adjustments and				
	maintenance of medical equipment ,according to				
	the specifications in the manufacturers' guides,				
	will be done free of charge.				
>	The spare parts and workmanship are for free,	v		∖ 1.	iddon do alonotion otto -11
	and space parts and workinghout for free,	X		> b	idder declaration attached

except from the consumables for vehicles established by the manufacturer			
Technical Specification	Yes	No	Remarks
AVAILABILITY OF SPARE PARTS			
 the bider will ensure the availability of spare parts, accessories and consumables for all offered positions on the market of the Republic of Moldova for at least 10 years from putting into operation, free of charge or against payment, as follows: spare parts free of charge, including the 	X		bidder declaration attached
workmanship for the 36 months of guarantee period o for the rest of the period – against payment for	X		bidder declaration attached
at least 10 years period is fulfilled.	X		bidder declaration attached
➤ The bidder will submit the pricelist for the spare parts and the maintenance program to the Ministry of Health, Labour and Social Protection as beneficiary	X		bidder declaration attached
GUIDES			
 Technical Guide and User Guide. All guides will be available in Romanian /Russian. 	X X		bidder declaration attachedbidder declaration attached