

LYNCMED MEDICAL TECHNOLOGY (BEIJING) CO.,LTD.
NO.1111, SOUTH HUIHE ROAD, CHAOYANG DISTRICT, BEIJING, CHINA 100020

The following samples were submitted and identified by/on behalf of the client as:

Name of Product/Item : PVC EXAMINATION GLOVES
Country of Origin : CHINA
Sample Receiving Date : JUN.21,2018
Testing Period : JUN.21,2018 TO JUL.09,2018
Test Performed : SELECTED TEST(S) AS REQUESTED BY APPLICANT
Test Requested : EN 455-2-2015 MEDICAL GLOVES FOR SINGLE USE – PART 2:
REQUIREMENTS AND TESTING FOR PHYSICAL PROPERTIES
(EXCEPT FOR CLAUSE 7 LABELLING)
Test Result(s) : FOR FURTHER DETAILS, PLEASE REFER TO THE FOLLOWING
PAGE(S)
Conclusion : THE SUBMITTED SAMPLE MET THE TEST REQUIREMENT.

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.
Approved by:



Zhou Xinkuan, SK
Lab Manager

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755)83071443, or email: CN.Doccheck@sgs.com



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

QDHL

001252

Test Report

No.: QDHL1806013404MD

Date: JUL.09,2018

Page: 2 of 4

Test Conducted:

EN 455-2-2015 Medical gloves for single use – part 2: Requirements and testing for physical properties

Number of test sample	:	39 Pieces
The type of gloves	:	examination/procedure gloves c)
Manufacturing batch code	:	/
Size	:	Examination/procedure gloves: M
Defects observed before testing	:	No defects
Test Result	:	Pass

Clause	Test Items	Result	Note
4	Dimensions	Pass	# 1
5	Strength	Pass	# 2
7	Labeling	N/C	# 3

Notes : #1 See result 1
 #2 See result 2
 #3 N/C means not conducted as per client's request.

Test Result:

1. Dimensions

Sample Quantity: 13 Pieces

Size	M												
Length(mm)	245	245	248	244	245	249	244	245	248	244	250	244	244
Width(mm)	96	98	97	97	96	98	96	98	97	96	96	98	97

Median value:

Length (mm): 245
 Width (mm): 97



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Requirements: see table 1&2

Table 1 Dimensions for surgical gloves

Size	Median length in mm	Median width in mm
5	≥250	67±4
5.5	≥250	72±4
6	≥260	77±5
6.5	≥260	83±5
7	≥270	89±5
7.5	≥270	95±5
8	≥270	102±6
8.5	≥280	108±6
9	≥280	114±6
9.5	≥280	121±6

**Table 2 Dimensions for
examination/procedure
gloves**

Size	Median length in mm	Median width in mm
Extra small	≥240	≤80
Small		80±10
Medium		95±10
Large		110±10
Extra Large		≥110

2. Strength

Sample Quantity: 26 Pieces

Size	M												
Force at break(N)	3.6	4.0	4.6	4.3	4.2	3.7	4.4	4.5	4.6	4.7	4.4	3.8	4.4
Force at break after challenge testing(N)	4.7	4.5	4.3	4.7	4.8	5.0	4.3	4.7	3.4	3.6	4.2	5.4	4.8

Median value:

Force at break (N): 4.4

Force at break after challenge testing (N): 4.7

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

QDHL 001254




Requirements: see table 3

Table 3 — Median values of force at break

	Force at break in Newton		
	1. Surgical gloves a)	Examination/procedure gloves b) c)	
Throughout shelf life tested according to 5.2 and within 12 months of manufacture tested according to 5.3	≥ 9,0	≥ 6,0	≥ 3,6
a) Requirements for all surgical gloves. b) Requirements for all examination gloves, except gloves made from thermoplastic materials (e.g. polyvinylchloride, polyethylene).. c) Requirements for gloves made from thermoplastic materials (e.g. polyvinylchloride, polyethylene).			

Sample Photo:

Received sample



SGS authenticate the photo on original report only

End of Report

