

ACON BIOTECH (HANGZHOU) CO.,LTD

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Urinalysis Reagent Strips

PURCHASE NO.: 37740066C

BATCH NO.: URS0030255

EXPIRATION DATE: 2022-06-27

QUANTITY: 100x100T

ANALYSIS		SPECIFICATIONS (Color card)	TEST RESULT
Glucose	Neg. urine	Negative	Complies
	250mg/dL	>100mg/dL <500mg/dL	Complies
Ketone	Neg. urine	Negative	Complies
	Yihua Urine controls	>40mg/dL <160mg/dL	Complies
Protein	Neg. urine37740053	Negative	Complies
	Yihua Urine controls	>+, ≤+++	Complies

Quality Department: Antony. Hu

Date: July 02, 2022

