

HORIBA ABX SAS  
Parc Euromédecine  
Rue du Caducée – BP7290  
34184 MONTPELLIER Cedex 4 – France  
Web : [www.horiba.com](http://www.horiba.com)  
Tél. : 33 (0) 4 67 14 15 16

## Déclaration de Conformité *Declaration of Conformity*

(No.dc 40153b)

Selon  
According to ISO 17050-1

### NOUS LE FABRICANT *WE THE MANUFACTURER*

Nom <i>Name</i>	<b>HORIBA ABX SAS</b>
Adresse <i>Address</i>	Parc Euromédecine Rue du Caducée – BP7290 34184 MONTPELLIER Cedex 4 - FRANCE

### DECLARONS QUE LE PRODUIT *DECLARE THAT THE PRODUCT*

Nom du produit <i>Product name</i>	<b>ABX Pentra CREATININE 120 CP</b> For application on : ABX PENTRA 400, PENTRA C200, PENTRA C400
Modèles <i>Models</i>	<b>A11A01933</b>
Pays d'origine <i>Country of origin</i>	France


### EST CONFORME AUX DIRECTIVES ET NORMES *CONFORMS TO DIRECTIVES AND STANDARDS*

Directives <i>Directives</i>	98/79/EC – IVD Medical Devices Classification : Outside Annex II and not for self-testing Conformity Assessment Procedure : Annex III
Normes <i>Standards</i>	N/A

Montpellier, France  
27 janvier 2014  
*January 27th, 2014*

**Christian DUBUC**

Directeur  
*Senior Director*



TEMP-4018 Rev.9