EU Declaration of Conformity

as per Annex IV of the Regulation EU 2017/746 on in-vitro diagnostic medical devices

Manufacturer: Roche Diagnostics GmbH
Address: Sandhofer Strasse 116
68305 Mannheim
Germany

Single Registration Number: DE-MF-000006260

Roche Diagnostics GmbH declares, under the sole responsibility, that the product/the product line

Product Name	Cat. No.	Basic UDI-DI
Elecsys Vitamin D total III	09038078190	761333601067AC
Elecsys Vitamin D total III	09038086190	761333601068AE

Intended Use:
Binding assay for the in vitro quantitative determination of total 25-hydroxyvitamin D in human serum and plasma. This assay is to be used as an aid in the assessment of vitamin D sufficiency.
The electrochemiluminescence binding assay is intended for use on cobas e immunoassay analyzers.

Product Name	Cat. No.	Basic UDI-DI
PreciControl Vitamin D total III	09038124190	7613336010709Z
PreciControl Vitamin D total III	09038124922	761333601071A3

Intended Use:
PreciControl Vitamin D total III is used for quality control of the Elecsys Vitamin D total III immunoassay on cobas e immunoassay analyzers.

Product Name	Cat. No.	Basic UDI-DI
CalSet Vitamin D total III	09038116190	761333601069AG

Intended Use:
CalSet Vitamin D total III is used for calibrating the quantitative Elecsys Vitamin D total III assay on cobas e immunoassay analyzers.

Risk Class: ☐ *A* ☒ *B* ☐ *C* ☐ *D*

Conformity Route:
☐ *Self-Declaration of Conformity (Class A)*
☐ *Self-Declaration of Conformity after Notified Body involvement for sterile manufacturing conditions acc. Art. 48 (10) (Class A sterile)*
☒ *Technical Documentation Assessment Class B/C – Annex IX*
☐ *Technical Documentation Assessment Class D – Annex IX*
☐ *Technical Documentation Assessment Class B/C/D for Self-Testing – Annex IX*
☐ *Technical Documentation Assessment Class B/C/D for Near-Patient Testing – Annex IX*

Roche Diagnostics GmbH; Sandhofer Straße 116; D-68305 Mannheim; Telefon +49-621-759-0; Telefax +49-621-759-2890

Certificates:	☐ *Technical Documentation Assessment Class C/D for Companion Diagnostics – Annex IX* ☒ *EU QM Certificate No.: V12 010283 0639* ☐ *EU Technical Documentation Assessment Certificate No. (Class D, Near-Patient Testing, Self-Testing and Companion Diagnostics):*
Other:	☐ *Common Specifications:*
Notified Body (NB) Name: *NB Address:* *NB Ident. No.:*	*TÜV Süd Product Service GmbH* *Ridlerstraße 65* *80339 Munich* *Germany* *0123*

to which this declaration relates fulfils the requirements of Regulation EU 2017/746 on in-vitro diagnostic medical devices.

Mannheim, 10 March 2023

Roche Diagnostics GmbH

i.V./*on behalf of the company*

DocuSigned by:
Christina Schmid
E3965E80F3E840E...

Dr. Christina Schmid
Head of Pre-Market Quality Core Lab

ppa./*on behalf of the company*

DocuSigned by:
Stefan Scheib
FC5EDEC1054B44C...

Dr. Stefan Scheib
Global Head of Regulatory Affairs, Core Lab

Contact address:

Roche Diagnostics GmbH
Abt./*Dept.* Global Regulatory Affairs
Sandhofer Strasse 116
D-68305 Mannheim