## Certificate



Certificate no.

CU 72225887 01

## **License Holder:**

DLAB SCIENTIFIC CO., LTD. Yu An Road 31, Airport Economic core zone, Shun yi District, Beijing, 101318 P.R. China

## **Manufacturing Plant:**

DLAB SCIENTIFIC CO., LTD. Yu An Road 31, Airport Economic core zone, Shun yi District, Beijing, 101318 P.R. China

Test report no.: USA-SOS CN22IIWZ 001

Client Reference: SHI HONGYUN

Tested to:

UL 61010-1:2012 R7.19

CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) + A1

Certified Product: PCR Thermocycler

License Fee - Units

Model Designation: TC1000-S, SCI1000-S, 50X4101300,

8

TC1000-G, SCI1000-G, 50X4102300 (X=1-9, X stands for different client)

Rating:

AC 100-120V, 50/60Hz, 600W, Class I

Appendix: 1

**Licensed Test mark:** 



**Date of Issue** (day/mo/yr) 11/05/2022



Date : 2022/05/11

DLAB SCIENTIFIC CO.,LTD. Yu An Road 31, Shun yi District, Beijing, 101318 P.R. China

Attn: Shi Hongyun

## Re. : CU US + Canada Certificate

Type of Equipment: PCR Thermocycler Model Designation: See Certificate Certificate No.: CU 72225887 0001 File No.: CN22IIWZ 001

Engineer/Contact : Sven Olaf Steinke

Standard(s) : UL 61010-1:2012 R7.19

CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) +

Α1

Dear Madame or Sir,

The above referenced technical equipment has been tested and was found to be in compliance with the listed test requirement(s). Enclosed, please find the TUV Rheinland approval document No. CU 72225887 0001. It authorizes you to label the listed product(s) with the TUV Rheinland Mark identified in the approval document. For compliance, the Test Mark must be on the approved unit.

Your product is subject to regular factory follow-up inspections as well as annual certificate and factory registration fees.

In using the TUV Rheinland Mark you are obligated to comply with the TUV Rheinland of North America Service Agreement.

If we can be of any further assistance to you, please do not hesitate to contact us.

Sincerely yours, Certification Body

Paddy Qiu QA Certification Officer

Enclosure