



浙江东方基因生物制品股份有限公司
Zhejiang Orient Gene Biotech Co.,LTD

STATEMENT

We, Zhejiang Orient Gene Biotech Co., Ltd , having a registered office at 3787#, East Yangguang Avenue, Dipu Street Anji 313300, Huzhou, Zhejiang, China assign SRL SANMEDICO having a registered office at A. Corobceanu street 7A, apt. 9, Chişinău MD-2012, Moldova, as exclusive authorized representative for Orient Gene Brand product in correspondence with the conditions of directive 98/79/EEC in Moldova only. The detailed product list is in the Annex 1 in the following pages.

We declare that the company mentioned above is authorized to register, notify, renew or modify the registration of medical devices on the territory of the Republic of Moldova.

This Statement letter will be valid from Mar.10th,2025to Mar.09th, 2027.

Zhejiang Orient Gene Biotech Co., Ltd

General Manager:

Date:2025/3/10



地址：浙江省湖州市安吉县递铺镇阳光大道东段 3787 号
Add: 3787#, East Yangguang Avenue, Dipu Street Anji 313300, Huzhou, Zhejiang, China
电话 Tel:+86-572-5226111 传真 Fax: +86-572-5226222 邮编 P.C.:313300



浙江东方基因生物制品股份有限公司
Zhejiang Orient Gene Biotech Co.,LTD

Annex 1

Numărul de catalog (referință)*	Denumire generică (denumirea dispozitivului)	Denumire comercială (brand)*	Modelul
GCCOV-702a-H1	TEST RAPID	Orient Gene	Rapid COVID-19 Antigen Oral Fluid Self-Test
GCCOV-702a-H5	TEST RAPID	Orient Gene	Rapid COVID-19 Antigen Oral Fluid Self-Test
GCCOV-702a-H20	TEST RAPID	Orient Gene	Rapid COVID-19 Antigen Oral Fluid Self-Test
GCCOV-502a-H1OGE	TEST RAPID	Orient Gene	Rapid COVID-19 (Antigen) Self-test
GCCOV-502a-H5OGE	TEST RAPID	Orient Gene	Rapid COVID-19 (Antigen) Self-test
GCCOV-502a-H20OGE	TEST RAPID	Orient Gene	Rapid COVID-19 (Antigen) Self-test

GIHSA-102a	TEST RAPID	Orient Gene	One Step Microalbumin Test Cassette
GIHSA-101a	TEST RAPID	Orient Gene	One Step Microalbumin Test Strip (Urine)
GCROA-602a	TEST RAPID	Orient Gene	Rotavirus rapid test cassette (feces)
GCMAL(pf/pv)-402a	TEST RAPID	Orient Gene	Malaria P.f./P.v. Ag Rapid Test Cassette (Whole Blood)
GCROA/ADE-602a	TEST RAPID	Orient Gene	Rotavirus and Adenovirus Combo Rapid Test Cassette (Feces)
GASPE-902a	TEST RAPID	Orient Gene	Male Fertility Rapid Test Cassette (Semen)
GAFSH-101a	TEST RAPID	Orient Gene	One Step Menopause Test Strip (Urine)
GAFSH-102a	TEST RAPID	Orient Gene	One Step Menopause Test Cassette (Urine)
GAIGF1-502a	TEST RAPID	Orient Gene	iGFBP-1 Rapid test Cassette (Swab)

地址：浙江省湖州市安吉县递铺镇阳光大道东段 3787 号

Add: 3787#, East Yangguang Avenue, Dipu Street Anji 313300, Huzhou, Zhejiang, China

电话 Tel:+86-572-5226111 传真 Fax: +86-572-5226222 邮编 P.C.:313300



浙江东方基因生物制品股份有限公司
Zhejiang Orient Gene Biotech Co.,LTD

GAIGF1-501a	TEST RAPID	Orient Gene	iGFBP-1 Rapid Test Strip (Cervical Secretion)
GALH-101a	TEST RAPID	Orient Gene	One Step Ovulation Test Strip (Urine) (25mlU)
GALH-101b	TEST RAPID	Orient Gene	One Step Ovulation Test Strip (Urine) (40mlU)
GALH-102a	TEST RAPID	Orient Gene	One Step Ovulation Test Cassette (Urine) (25mlU)
GALH-102b	TEST RAPID	Orient Gene	One Step Ovulation Test Cassette (Urine) (40mlU)
GCTYP-302a	TEST RAPID	Orient Gene	Typhoid IgG/IgM Rapid Test Cassette (serum/plasma)
GCMAL(pf/pan)-402a	TEST RAPID	Orient Gene	Malaria P.f./Pan Ag Rapid Test Cassette (Whole Blood)
GCDEN-425a	TEST RAPID	Orient Gene	Dengue NS1+IgM/IgG Combo Rapid Test Cassette (Whole blood/serum/plasma)
GCDEN(NS)-402c	TEST RAPID	Orient Gene	Dengue NS1 Antigen Rapid Test Cassette (Whole blood/serum/plasma)
GCDEN(ab)-402c	TEST RAPID	Orient Gene	Dengue IgM/IgG Rapid Test Cassette (Whole blood/serum/plasma)
GCVCH(O1/O9)-602a	TEST RAPID	Orient Gene	V.cholerae O1/O139 Ag Combo Rapid Test Cassette (Feces)
GCMAL(pf)-402a	TEST RAPID	Orient Gene	Malaria Pf Ag Rapid Test Cassette (Whole blood)
GCSAL(ST)-602a	TEST RAPID	Orient Gene	S. typhi Ag Rapid Test Cassette (Serum/plasma/Feces)
GCCHK(IgM)-402a	TEST RAPID	Orient Gene	Chikungunya IgM Rapid Test Cassette (Whole blood/Serum/Plasma)
GCCOV-502a-NN	TEST RAPID	Orient Gene	Coronavirus Ag Rapid Test Cassette (Swab)
GCCOV (Nab)-402b	TEST RAPID	Orient Gene	SARS-CoV-2 Neutralizing Antibody Rapid Test Cassette (Whole blood/Serum/Plasma)
GCCOV-502a-NA	TEST RAPID	Orient Gene	Coronavirus Ag Rapid Tests Cassette (Swab)

地址：浙江省湖州市安吉县递铺镇阳光大道东段 3787 号

Add: 3787#, East Yangguang Avenue, Dipu Street Anji 313300, Huzhou, Zhejiang, China

电话 Tel: +86-572-5226111 传真 Fax: +86-572-5226222 邮编 P.C.: 313300



浙江东方基因生物制品股份有限公司
Zhejiang Orient Gene Biotech Co.,LTD

GCCOV (Ag)-PN10	TEST RAPID	Orient Gene	COVID-19 Antigen Control Kit
GCCOV (Ag)-PN20	TEST RAPID	Orient Gene	COVID-19 Antigen Control Kit
GCFERA-545a	TEST RAPID	Orient Gene	Flu, COVID-19,RSV&Adeno Ag Combo Tests Cassette (Swab)
GCTV-502a	TEST RAPID	Orient Gene	Trichomonas Ag Rapid Test Cassette (Swab)
GCVCH(O1)-602a	TEST RAPID	Orient Gene	V.cholerae O1 Ag Rapid Test Cassette (Feces)
GCCHA-402a	TEST RAPID	Orient Gene	Chagas Ab Rapid Test Cassette (Whole blood/serum/plasma)
GCMAL(pf/pv Ab)-302a	TEST RAPID	Orient Gene	Malaria Pf/Pv Ab Rapid Test Cassette (Serum/plasma)
GCMAL(pf/pv Ab)-402a	TEST RAPID	Orient Gene	Malaria Pf/Pv Ab Rapid Test Cassette (Whole blood/Serum/plasma)
GCMKP-502b	TEST RAPID	Orient Gene	Monkeypox Ag Rapid Test Cassette (Swab)
GCCOV(Del)-T502a	TEST RAPID	Orient Gene	SARS-CoV-2 Delta-series Mutant Strain Ag Rapid Test cassette (Swab)
GCCOV-PN10	TEST RAPID	Orient Gene	COVID-19 IgG/IgM Control kit
GCCOV-PN20	TEST RAPID	Orient Gene	COVID-19 IgG/IgM Control kit
GCFER-T525a	TEST RAPID	Orient Gene	COVID-19/Flu A&B/ RSV Ag Combo Rapid Test Cassette (Swab)
GCCOV(B117)-525a	TEST RAPID	Orient Gene	COVID-19 Ag&B.1.1.7 Mutant Strain Combo Test Cassette (Swab)
GCFERA-T525a	TEST RAPID	Orient Gene	COVID-19/Flu A&B/ RSV/Adeno Ag Combo Rapid Test Cassette (Swab)
GCCOV-702a	TEST RAPID	Orient Gene	COVID-19 Ag Rapid Test Cassette (Oral fluid)
GCFE-T502a	TEST RAPID	Orient Gene	COVID-19/Flu A&B Ag Combo Rapid Test Cassette (Swab)

地址：浙江省湖州市安吉县递铺镇阳光大道东段 3787 号

Add: 3787#, East Yangguang Avenue, Dipu Street Anji 313300, Huzhou, Zhejiang, China

电话 Tel:+86-572-5226111 传真 Fax:+86-572-5226222 邮编 P.C.:313300



浙江东方基因生物制品股份有限公司
Zhejiang Orient Gene Biotech Co.,LTD

GCMKP-402a	TEST RAPID	Orient Gene	Monkeypox IgG/IgM Rapid Test Cassette (Whole blood/serum/plasma)
GCKa1-401a	TEST RAPID	Orient Gene	Leishmania Ab Rapid Test strip (Whole blood/serum/plasma)
GCKa1-T402a	TEST RAPID	Orient Gene	Leishmania IgG/IgM Rapid test cassette (whole blood/serum/plasma)
GCCOV-503a	TEST RAPID	Orient Gene	Smart Rapid COVID-19 Ag Test Device
GCBRU-402a	TEST RAPID	Orient Gene	Brucella Antibody Rapid Test Cassette (Whole blood/serum/plasma)
GCCHA-302a	TEST RAPID	Orient Gene	Chagas Ab Rapid Test Cassette (Serum/plasma)
GCCHK(IgM)-302a	TEST RAPID	Orient Gene	Chikungunya IgM Rapid Test Cassette (Serum/Plasma)
GCCOV-501a	TEST RAPID	Orient Gene	Rapid COVID-19 Antigen Test Strip
GCMON-352a	TEST RAPID	Orient Gene	Mononucleosis IgG/IgM Rapid Test Cassette (Serum/plasma)
GCMON-402a	TEST RAPID	Orient Gene	Mononucleosis Rapid Test Cassette (Whole blood/Serum/plasma)
GCMON-425a	TEST RAPID	Orient Gene	Mononucleosis IgG/IgM Rapid Test Cassette (Whole blood/Serum/plasma)
GCEV71 (IgM)-302a	TEST RAPID	Orient Gene	EV71 IgM Rapid Test Cassette (Serum/plasma)
GCEV71 (IgM)-402a	TEST RAPID	Orient Gene	EV71 IgM Rapid Test Cassette (Whole blood/Serum/plasma)
GENMP22-102a	TEST RAPID	Orient Gene	One Step Nuclear Matrix Protein 22 Test Cassette (Urine)
GEFOB/TF-602a	TEST RAPID	Orient Gene	Fecal Occult Blood and Transferrin Combo Rapid Test Cassette (Feces)
GCHEV-302a	TEST RAPID	Orient Gene	HEV IgM Rapid Test Cassette (Serum/Plasma)
GCMP (IgM)-302a	TEST RAPID	Orient Gene	M.pneumonia IgM Rapid Test Cassette (Serum/plasma)

地址：浙江省湖州市安吉县递铺镇阳光大道东段 3787 号

Add: 3787#, East Yangguang Avenue, Dipu Street Anji 313300, Huzhou, Zhejiang, China

电话 Tel:+86-572-5226111 传真 Fax: +86-572-5226222 邮编 P.C.:313300



浙江东方基因生物制品股份有限公司
Zhejiang Orient Gene Biotech Co.,LTD

FCCOV-502a	TEST RAPID	Orient Gene	SARS-CoV-2 Ag Fluorescence Rapid Test Cassette (Swab)
GCCOV-402Ba	TEST RAPID	Orient Gene	COVID-19 IgG/IgM Rapid test cassette (whole blood/serum/plasma)
GCCOV-402a	TEST RAPID	Orient Gene	COVID-19 IgG/IgM Rapid test cassette (whole blood/serum/plasma)
GCCOV-502a	TEST RAPID	Orient Gene	Coronavirus Ag Rapid Test Cassette (Swab)
GCFC-T503a	TEST RAPID	Orient Gene	Smart Rapid COVID-19 &Flu A/B Ag Test Device
GAHCG-101a	TEST RAPID	Orient Gene	One step pregnancy test strip (urine)
GAHCG-101d	TEST RAPID	Orient Gene	One step pregnancy test strip (urine)
GAHCG-101b	TEST RAPID	Orient Gene	One step pregnancy test strip (urine)
GAHCG-102a	TEST RAPID	Orient Gene	One step pregnancy test cassette (urine)
GAHCG-102d	TEST RAPID	Orient Gene	One step pregnancy test cassette (urine)
GAHCG-102b	TEST RAPID	Orient Gene	One step pregnancy test cassette (urine)

GEFOB-602c	TEST RAPID	Orient Gene	Fecal Occult Blood Rapid Test Cassette (Feces)
GEFOB-602b	TEST RAPID	Orient Gene	Fecal Occult Blood Rapid Test Cassette (Feces)
GEFOB-601c	TEST RAPID	Orient Gene	Fecal Occult Blood Rapid Test Strip (Feces)
GEFOB-601b	TEST RAPID	Orient Gene	Fecal Occult Blood Rapid Test Strip (Feces)
GECEA-402a	TEST RAPID	Orient Gene	Carcinoembryonic Antigen Rapid Test Cassette (Whole blood/serum/plasma)

地址：浙江省湖州市安吉县递铺镇阳光大道东段 3787 号
Add: 3787#, East Yangguang Avenue, Dipu Street Anji 313300, Huzhou, Zhejiang, China
电话 Tel: +86-572-5226111 传真 Fax: +86-572-5226222 邮编 P.C.: 313300



浙江东方基因生物制品股份有限公司
Zhejiang Orient Gene Biotech Co.,LTD

GECEA-401a	TEST RAPID	Orient Gene	Carcinoembryonic Antigen Rapid Test Strip (Whole blood/serum/plasma)
GETF-602a	TEST RAPID	Orient Gene	Transferrin Rapid Test Cassette (Feces)
GETF-601a	TEST RAPID	Orient Gene	Transferrin Rapid Test Strip (Feces)
GEAFP-401a	TEST RAPID	Orient Gene	Alpha-Fetoprotein Rapid Test Strip (Whole blood/serum/plasma)
GEAFP-402a	TEST RAPID	Orient Gene	Alpha-Fetoprotein Rapid Test Cassette (Whole blood/serum/plasma)
GIHSA-101a	TEST RAPID	Orient Gene	One step microalbumin test strip (urine)
GIHSA-102a	TEST RAPID	Orient Gene	One step microalbumin test cassette (urine)
GDCAR-335a	TEST RAPID	Orient Gene	Myoglobin/CK-MB/Troponin I Combo Test Cassette (Serum/plasma)
GDCKM-302a	TEST RAPID	Orient Gene	One step CK-MB Test Cassette (Serum/Plasma)
GDCKM-402a	TEST RAPID	Orient Gene	CK-MB Rapid Test Cassette (Whole blood/serum/plasma)
GDCRP-402a	TEST RAPID	Orient Gene	C-Reactive Protein Semi-Quantitative Rapid Test Cassette (Whole Blood/serum/plasma)
GDTRO-302a	TEST RAPID	Orient Gene	Troponin I Rapid Test Cassette (Serum/Plasma)
GDTRO-402a	TEST RAPID	Orient Gene	Troponin I Rapid Test Cassette (Whole blood/Serum/Plasma) (Except the tender No. ocds-b3wdp1-MD-1722410248839 din 05.09.2024, limited to the quantity 28060 pcs only, as per the tender)
GDTRO-402b	TEST RAPID	Orient Gene	Troponin I Rapid Test Cassette (Whole blood/Serum/Plasma)

地址：浙江省湖州市安吉县递铺镇阳光大道东段 3787 号

Add: 3787#, East Yangguang Avenue, Dipu Street Anji 313300, Huzhou, Zhejiang, China

电话 Tel:+86-572-5226111 传真 Fax: +86-572-5226222 邮编 P.C.:313300



浙江东方基因生物制品股份有限公司

Zhejiang Orient Gene Biotech Co.,LTD

GDMYO-402a	TEST RAPID	Orient Gene	Myoglobin Rapid Test Cassette (Whole blood/serum/plasma)
GDCAR-W435a	TEST RAPID	Orient Gene	Myoglobin/CK-MB/Troponin I Combo Rapid Test Cassette (whole blood/Serum/plasma)
GDPCT-402a	TEST RAPID	Orient Gene	Procalcitonin Rapid Test Cassette (Whole blood/serum/plasma)
GDPCT-T402a	TEST RAPID	Orient Gene	Procalcitonin Semi-Quantitative Rapid Test Cassette (Whole blood/serum/plasma)
GDPCT-T401a	TEST RAPID	Orient Gene	Procalcitonin Semi-Quantitative Rapid Test Strip (Whole blood/serum/plasma)
FDPCT -302a	TEST RAPID	Orient Gene	Procalcitonin Rapid Test Kit (serum/plasma)
GDDDI-402b	TEST RAPID	Orient Gene	D-Dimer Rapid Test Cassette (Whole blood/plasma)
FDCAR-T302a	TEST RAPID	Orient Gene	Troponin I/CK-MB/Myoglobin Fluorescence Combo Test Kit (Serum/plasma)
FDTRO-302a	TEST RAPID	Orient Gene	Troponin I Fluorescence Rapid Test Kit (Serum/plasma)
FDBNP-302a	TEST RAPID	Orient Gene	NT-ProBNP Fluorescence Rapid Test Kit (Serum/plasma)
FDCRP-402a	TEST RAPID	Orient Gene	C-Reactive Protein Rapid Test Kit (Whole blood/serum/plasma)
GDCKM-402a	TEST RAPID	Orient Gene	CK-MB Rapid Test Cassette (whole blood/serum/plasma)
GAHCG-201a	TEST RAPID	Orient Gene	One step pregnancy test strip (Urine/serum)
GAHCG-202a	TEST RAPID	Orient Gene	One step pregnancy test cassette (Urine/serum)
GAHCG-201b	TEST RAPID	Orient Gene	One step pregnancy test strip (Urine/serum)
GAHCG-202b	TEST RAPID	Orient Gene	One step pregnancy test cassette (Urine/serum)
GCHAV(IgM)-302Ba	TEST RAPID	Orient Gene	HAV IgM Rapid Test Cassette (Serum/plasma)
GCHAV-602a	TEST RAPID	Orient Gene	HAV Ag Rapid Test Cassette (Feces)

地址：浙江省湖州市安吉县递铺镇阳光大道东段 3787 号

Add: 3787#, East Yangguang Avenue, Dipu Street Anji 313300, Huzhou, Zhejiang, China

电话 Tel:+86-572-5226111 传真 Fax: +86-572-5226222 邮编 P.C.:313300



浙江东方基因生物制品股份有限公司

Zhejiang Orient Gene Biotech Co.,LTD

GCHAV(IgG/IgM) -302a	TEST RAPID	Orient Gene	HAV IgG/IgM Rapid Test Cassette (Serum/plasma)
GCHSV(IgG)-402a	TEST RAPID	Orient Gene	HSV IgG Rapid Test Cassette (Whole blood/serum/plasma)
GCHSV(IgM)-302a	TEST RAPID	Orient Gene	HSV IgM Rapid test Cassette (serum/plasma)
GCHP-601a	TEST RAPID	Orient Gene	H.pylori Ag Rapid Test Strip (feces)
GCHP-602a	TEST RAPID	Orient Gene	H.pylori Ag Rapid Test Cassette(feces)
GCTB-302a	TEST RAPID	Orient Gene	Tuberculosis IgG/IgM Rapid Test Cassette (serum/plasma)
GCTB-402a	TEST RAPID	Orient Gene	Tuberculosis IgG/IgM Rapid Test Cassette (whole blood/serum/plasma)
GCFLU(A/B)-501a	TEST RAPID	Orient Gene	Influenza A&B Ag Rapid Test Strip (Swab)
GCFLU(A/B)-502a	TEST RAPID	Orient Gene	Influenza A&B Ag Rapid Test Cassette (Swab)
GCFLU(A/B)-502Ca	TEST RAPID	Orient Gene	Influenza A&B Ag Rapid Test Cassette (Swab)
GCFLU(A)-501a	TEST RAPID	Orient Gene	Influenza A Ag Rapid Test Strip (Swab)
GCFLU(A)-502a	TEST RAPID	Orient Gene	Influenza A Ag Rapid Test Cassette (Swab)
GCHP-301a	TEST RAPID	Orient Gene	H.Pylori Ab Rapid Test Strip (serum/plasma)
GCHP-302a	TEST RAPID	Orient Gene	H.pylori Ab Rapid Test Cassette (serum/plasma)
GCHP-401a	TEST RAPID	Orient Gene	H.pylori Ab Rapid Test Strip (Whole blood/serum/plasma)
GCHP-402a	TEST RAPID	Orient Gene	H.pylori Ab Rapid Test Cassette (Whole blood/serum/plasma)
GCCA-502a	TEST RAPID	Orient Gene	Candida albicans Antigen rapid test cassette (swab)
GCGON-502b	TEST RAPID	Orient Gene	Gonorrhea Rapid Test Cassette (Swab)
GCGIA-602a	TEST RAPID	Orient Gene	Giardia lamblia Antigen Rapid tests cassette (feces)

地址：浙江省湖州市安吉县递铺镇阳光大道东段 3787 号

Add: 3787#, East Yangguang Avenue, Dipu Street Anji 313300, Huzhou, Zhejiang, China

电话 Tel:+86-572-5226111 传真 Fax: +86-572-5226222 邮编 P.C.:313300



浙江东方基因生物制品股份有限公司
Zhejiang Orient Gene Biotech Co.,LTD

GCSTR-501a	TEST RAPID	Orient Gene	Strep A Rapid Test Strip (Throat swab)
GCSTR-502a	TEST RAPID	Orient Gene	Strep A Rapid Test Cassette (Throat swab)
GCFC-525a	TEST RAPID	Orient Gene	Rapid COVID-19 + Influenza Antigen Test
GCRSV-502a	TEST RAPID	Orient Gene	RSV Antigen Rapid Test Cassette (swab)
GCADE-502a	TEST RAPID	Orient Gene	Adenovirus antigen rapid test cassette (swab)
GCADE-602a	TEST RAPID	Orient Gene	Adenovirus Rapid test cassette (feces)
GCCD(GDH)-602a	TEST RAPID	Orient Gene	Clostridium difficile Antigen GDH Rapid Test cassette (feces)
GCCD (Toxin A/B)-602a	TEST RAPID	Orient Gene	Clostridium difficile Toxin A&B rapid test cassette (feces)
GCCD-602a	TEST RAPID	Orient Gene	Clostridium difficile GDH & Toxin A/B Rapid Test Cassette (Feces)
GCHSV (IgM)-402a	TEST RAPID	Orient Gene	HSV IgM Rapid test Cassette (whole blood/serum/plasma)
GCHSV(IgG)-302a	TEST RAPID	Orient Gene	HSV IgG Rapid Test Cassette (serum/plasma)
GCSYP-301a	TEST RAPID	Orient Gene	Syphilis Ab Rapid Test Strip (serum/plasma)
GCSYP-302a	TEST RAPID	Orient Gene	Syphilis Ab Rapid Test Cassette (serum/plasma)
GCSYP-401a	TEST RAPID	Orient Gene	Syphilis Ab Rapid test strip (whole blood/serum/plasma)
GCSYP-402a	TEST RAPID	Orient Gene	Syphilis Ab Rapid test cassette (whole blood/serum/plasma)
GBBAR-101a	TEST RAPID	Orient Gene	One Step Barbiturates Test Strip (Urine)
GBBAR-102a	TEST RAPID	Orient Gene	One Step Barbiturates Test Cassette (Urine)
GBAMP-101a	TEST RAPID	Orient Gene	One Step Amphetamine Test Strip (Urine)
GBAMP-102a	TEST RAPID	Orient Gene	One Step Amphetamine Test Cassette (Urine)
GBAMP-105a	TEST RAPID	Orient Gene	One Step Amphetamine Dip Card (Urine)

地址：浙江省湖州市安吉县递铺镇阳光大道东段 3787 号

Add: 3787#, East Yangguang Avenue, Dipu Street Anji 313300, Huzhou, Zhejiang, China

电话 Tel:+86-572-5226111 传真 Fax: +86-572-5226222 邮编 P.C.:313300



浙江东方基因生物制品股份有限公司
Zhejiang Orient Gene Biotech Co.,LTD

GBPPX-101a	TEST RAPID	Orient Gene	One Step Propoxyphene Test Strip (Urine)
GBPPX-102a	TEST RAPID	Orient Gene	One Step Propoxyphene Test Cassette (Urine)
GBDSA-XXXXFX	TEST RAPID	Orient Gene	Oral Fluid Drug test Cube
GBDSA-XXXXEX	TEST RAPID	Orient Gene	Oral Fluid Drug test
GBDSA-XXXXFSI	TEST RAPID	Orient Gene	Oral Fluid Drug test Cube
GBDSA-XXXXCX	TEST RAPID	Orient Gene	Oral Fluid Drug test cylinder
GBOPI-102a	TEST RAPID	Orient Gene	One Step Opiate Test Cassette (Urine)
GBOPI-101a	TEST RAPID	Orient Gene	One Step Opiate Test Strip (Urine)
GBETG-101b	TEST RAPID	Orient Gene	One Step Ethyl Glucoronide Test Strip (urine)
GBETG-102b	TEST RAPID	Orient Gene	One Step Ethyl Glucoronide Test Cassette (urine)
GBMOP-101a	TEST RAPID	Orient Gene	One step Morphine Test strip (urine)
GBMOP-102a	TEST RAPID	Orient Gene	One step Morphine Test Cassette (urine)
GBMOP-105a	TEST RAPID	Orient Gene	One step Morphine Test dip card (urine)
GBTHC-101a	TEST RAPID	Orient Gene	One Step Marijuana Test Strip (Urine)
GBTHC-102a	TEST RAPID	Orient Gene	One Step Marijuana Test Cassette (Urine)
GBTHC-105a	TEST RAPID	Orient Gene	One Step Marijuana Test Dip Card (Urine)
GBMTD-101a	TEST RAPID	Orient Gene	One step Methadone Test strip (urine)
GBMTD-102a	TEST RAPID	Orient Gene	One step Methadone Test cassette (urine)
GBXXX-101	TEST RAPID	Orient Gene	One Step Drugs of Abuse Test Strip (Urine)
GBXXX-102	TEST RAPID	Orient Gene	One Step Drugs of Abuse Test Cassette (Urine)
GBXXX-105	TEST RAPID	Orient Gene	One Step Drugs of Abuse Test Dip Card (Urine)

地址：浙江省湖州市安吉县递铺镇阳光大道东段 3787 号
Add: 3787#, East Yangguang Avenue, Dipu Street Anji 313300, Huzhou, Zhejiang, China
电话 Tel: +86-572-5226111 传真 Fax: +86-572-5226222 邮编 P.C.: 313300



浙江东方基因生物制品股份有限公司
Zhejiang Orient Gene Biotech Co.,LTD

GBDSA-XXXXJSI	TEST RAPID	Orient Gene	Oral fluid drug test cylinder
GBDSA-XXXXJX	TEST RAPID	Orient Gene	Oral fluid drug test cylinder
GBDSA-XXXXKX	TEST RAPID	Orient Gene	Oral fluid drug test cylinder
GBDSA-XXXXMX	TEST RAPID	Orient Gene	Oral fluid drug test device
GBDSA-XXXXA/B/G/H/I	TEST RAPID	Orient Gene	Multi-drug rapid screen test cassette (oral fluid)
GBMTC-101a	TEST RAPID	Orient Gene	One Step Methcathinone Test Strip (Urine)
GBMTC-102a	TEST RAPID	Orient Gene	One Step Methcathinone Test Cassette (Urine)
GBKRA-101a	TEST RAPID	Orient Gene	One step kratom test strip (urine)
GBKRA-102a	TEST RAPID	Orient Gene	One step kratom test cassette (urine)
GBLSD-101a	TEST RAPID	Orient Gene	One Step Lysergic Acid Diethylamide Test Strip (Urine)
GBLSD-102a	TEST RAPID	Orient Gene	One Step Lysergic Acid Diethylamide Test Cassette (Urine)
FBXXX-1102	TEST RAPID	Orient Gene	Hair Multi-drug rapid test kit
GBETG-105a	TEST RAPID	Orient Gene	One step ethyl glucuronide test dip card (urine)
GBPGB-102b	TEST RAPID	Orient Gene	One step pregabalin test cassette (urine)
GBTRA-101a	TEST RAPID	Orient Gene	One step tramadol test strip (urine)
GBTRA-102a	TEST RAPID	Orient Gene	One step tramadol test cassette (urine)
GBOXY-101a	TEST RAPID	Orient Gene	One step oxycodone Test strip (urine)
GBOXY-102a	TEST RAPID	Orient Gene	One step oxycodone Test cassette (urine)
GBMDP-101a	TEST RAPID	Orient Gene	One step 3,4-Methylenedioxypyrovalerone Test strip (urine)

地址：浙江省湖州市安吉县递铺镇阳光大道东段 3787 号

Add: 3787#, East Yangguang Avenue, Dipu Street Anji 313300, Huzhou, Zhejiang, China

电话 Tel: +86-572-5226111 传真 Fax: +86-572-5226222 邮编 P.C.: 313300



浙江东方基因生物制品股份有限公司
Zhejiang Orient Gene Biotech Co.,LTD

GBMDP-101a	TEST RAPID	Orient Gene	One step 3,4-Methylenedioxypropylone Test cassette (urine)
GBMQL-102a	TEST RAPID	Orient Gene	One step Methaqualone Test cassette (urine)
GBMQL-101a	TEST RAPID	Orient Gene	One step Methaqualone Test strip (urine)
GBMPD-101a	TEST RAPID	Orient Gene	One step Methylphenidate Test strip (urine)
GBMPD-102a	TEST RAPID	Orient Gene	One step Methylphenidate Test cassette (urine)
GBUR-101a	TEST RAPID	Orient Gene	One step UR-144 test strip (urine)
GBUR-102a	TEST RAPID	Orient Gene	One step UR-144 test cassette (urine)
GBBUP-101a	TEST RAPID	Orient Gene	One step buprenorphine test strip (urine)
GBBUP-102a	TEST RAPID	Orient Gene	One step buprenorphine test cassette (urine)
GBPCP-101a	TEST RAPID	Orient Gene	One step Phencyclidine Test strip (urine)
GBPCP-102a	TEST RAPID	Orient Gene	One step Phencyclidine Test cassette (urine)
GBTCA-101a	TEST RAPID	Orient Gene	One step Tricyclic Antidepressants test strip (urine)
GBTCA-102a	TEST RAPID	Orient Gene	One step Tricyclic Antidepressants test cassette (urine)
GBEDD-101a	TEST RAPID	Orient Gene	One step EDDP test strip (urine)
GBEDD-102a	TEST RAPID	Orient Gene	One step EDDP test cassette (urine)
GBFEN-101b	TEST RAPID	Orient Gene	One step Fentanyl Test strip (urine)
GBFEN-102b	TEST RAPID	Orient Gene	One step Fentanyl Test cassette (urine)
GBALC-101a	TEST RAPID	Orient Gene	Urine Alcohol Test Strip
GBMAM-S102	TEST RAPID	Orient Gene	One step 6-MAM Test cassette (urine)

地址：浙江省湖州市安吉县递铺镇阳光大道东段 3787 号

Add: 3787#, East Yangguang Avenue, Dipu Street Anji 313300, Huzhou, Zhejiang, China

电话 Tel:+86-572-5226111 传真 Fax: +86-572-5226222 邮编 P.C.:313300



浙江东方基因生物制品股份有限公司
Zhejiang Orient Gene Biotech Co.,LTD

GBMAM-S101	TEST RAPID	Orient Gene	One step 6-MAM Test cassette (urine)
GBHCD-101a	TEST RAPID	Orient Gene	One step Hydrocodone test strip (urine)
GBHCD-102a	TEST RAPID	Orient Gene	One step Hydrocodone test cassette (urine)
GBNFT-101c	TEST RAPID	Orient Gene	One step Norfentanyl test strip (urine)
GBNFT-102c	TEST RAPID	Orient Gene	One step Norfentanyl test cassette (urine)
GBXXX-1102	TEST RAPID	Orient Gene	Hair Multi-Drug Rapid Test Kit (ICA)
GBDSA-XXXXLX	TEST RAPID	Orient Gene	Oral Fluid Drug Test Mini Cube
GBDUA-1X4	TEST RAPID	Orient Gene	One Step Multi-Drug Screen Test Dip Card (urine)
GBDOA-1X5	TEST RAPID	Orient Gene	One Step Multi-Drug Screen Test Cassette (urine)
GBDUA-1X6	TEST RAPID	Orient Gene	One Step Multi-Drug Screen Test Cup (urine)
GBCOT-102a	TEST RAPID	Orient Gene	One step cotinine test cassette (urine)
GBK2-101a	TEST RAPID	Orient Gene	One step K2 Test strip (urine)
GBK2-102a	TEST RAPID	Orient Gene	One step K2 Test cassette (urine)
GBKET-101a	TEST RAPID	Orient Gene	One step Ketamine Test strip (urine)
GBKET-102a	TEST RAPID	Orient Gene	One step Ketamine Test cassette (urine)
GBBZO-101a	TEST RAPID	Orient Gene	One step Benzodiazepines Test Strip (urine)
GBBZO-102a	TEST RAPID	Orient Gene	One step Benzodiazepines Test Cassette (urine)
GBCOC-101a	TEST RAPID	Orient Gene	One step Cocaine Test strip (urine)
GBCOC-102a	TEST RAPID	Orient Gene	One step Cocaine Test cassette (urine)
GBCOC-105a	TEST RAPID	Orient Gene	One step Cocaine Test dip card (urine)
GBMDM-101a	TEST RAPID	Orient Gene	One step ecstasy Test strip (urine)

地址：浙江省湖州市安吉县递铺镇阳光大道东段 3787 号

Add: 3787#, East Yangguang Avenue, Dipu Street Anji 313300, Huzhou, Zhejiang, China

电话 Tel:+86-572-5226111 传真 Fax: +86-572-5226222 邮编 P.C.:313300



浙江东方基因生物制品股份有限公司
Zhejiang Orient Gene Biotech Co.,LTD

GBMDM-102a	TEST RAPID	Orient Gene	One step ecstasy Test cassette (urine)
GBMET-101a	TEST RAPID	Orient Gene	One step Methamphetamine test strip (urine)
GBMET-102a	TEST RAPID	Orient Gene	One step Methamphetamine test cassette (urine)
GBMET-105a	TEST RAPID	Orient Gene	One step Methamphetamine test dip card (urine)
GCTOXI(IgG/IgM)-302a	TEST RAPID	Orient Gene	Toxoplasma gondii test cassette (serum/plasma)
GCTOXI(IgG)-302a	TEST RAPID	Orient Gene	Toxoplasma gondii IgG test cassette (serum/plasma)
GCTOXI(IgM)-302a	TEST RAPID	Orient Gene	Toxoplasma gondii IgM test cassette (serum/plasma)
GCCHL-502a	TEST RAPID	Orient Gene	Chlamydia Trachomatis Antigen test cassette (swab/urine)
GEPSA-402a	TEST RAPID	Orient Gene	Prostate specific antigen test cassette (whole blood/serum/plasma)
GEPSA-401a	TEST RAPID	Orient Gene	Prostate specific antigen test strip (whole blood/serum/plasma)
GEPSA-302a	TEST RAPID	Orient Gene	Prostate specific antigen test cassette (serum/plasma)
GEPSA-301a	TEST RAPID	Orient Gene	Prostate specific antigen test strip (serum/plasma)
GALH-101a-1T	TEST RAPID	Orient Gene	LH Ovulation Test Strip
GALH-101a-5T	TEST RAPID	Orient Gene	LH Ovulation Test Strip
GALH-101a-7T	TEST RAPID	Orient Gene	LH Ovulation Test Strip
GALH-101b-1T	TEST RAPID	Orient Gene	LH Ovulation Test Strip
GALH-101b-5T	TEST RAPID	Orient Gene	LH Ovulation Test Strip
GALH-101b-7T	TEST RAPID	Orient Gene	LH Ovulation Test Strip
GALH-102a-1T	TEST RAPID	Orient Gene	LH Ovulation Test Cassette
GALH-102a-5T	TEST RAPID	Orient Gene	LH Ovulation Test Cassette
GALH-102a-7T	TEST RAPID	Orient Gene	LH Ovulation Test Cassette
GALH-102b-5T	TEST RAPID	Orient Gene	LH Ovulation Test Cassette
GALH-102b-1T	TEST RAPID	Orient Gene	LH Ovulation Test Cassette

地址：浙江省湖州市安吉县递铺镇阳光大道东段 3787 号

Add: 3787#, East Yangguang Avenue, Dipu Street Anji 313300, Huzhou, Zhejiang, China

电话 Tel:+86-572-5226111 传真 Fax: +86-572-5226222 邮编 P.C.:313300



浙江东方基因生物制品股份有限公司
Zhejiang Orient Gene Biotech Co.,LTD

GALH-102b-7T	TEST RAPID	Orient Gene	LH Ovulation Test Cassette
--------------	------------	-------------	----------------------------

VPH-502a	TEST RAPID	Orient Gene	Vaginal pH test cassette (Vaginal secretions)
----------	------------	-------------	---

URS-1T to 14T with various combination	STRIPURI DE URINA	Orient Gene	LEU/NIT/URO/MA/PRO/PH/BLO/S G/ASC/CRE/KET/BIL/GLU/CA
GCHCV-302a	TEST RAPID	Orient Gene	HCV Hepatitis C Virus Rapid Test (serum/plasma) cassette
GCHCV-402a	TEST RAPID	Orient Gene	HCV Hepatitis C Virus Rapid Test (whole blood/serum/plasma) cassette
GCHIV-302a	TEST RAPID	Orient gene	HIV 1/2 Human Immunodeficiency Virus (Serum/Plasma) cassette
GCHIV-402a	TEST RAPID	Orient gene	HIV 1/2 Human Immunodeficiency Virus (Whole blood/serum/plasma)cassette
GCHBsg-302a	TEST RAPID	Orient gene	HBsAg Hepatitis B Surface Antigen Rapid Test (Serum/Plasma)
GCHBsg-402a	TEST RAPID	Orient gene	HBsAg Hepatitis B Surface Antigen Rapid Test(Whole Blood/Serum/Plasma)

The end.



地址：浙江省湖州市安吉县递铺镇阳光大道东段 3787 号
Add: 3787#, East Yangguang Avenue, Dipu Street Anji 313300, Huzhou, Zhejiang, China
电话 Tel:+86-572-5226111 传真 Fax: +86-572-5226222 邮编 P.C.:313300



Certificate

No. Q5 092305 0001 Rev. 02

Holder of Certificate: **Zhejiang Orient Gene Biotech Co., Ltd.**
3787#, East Yangguang Avenue, Dipu Street Anji
313300 Huzhou, Zhejiang
PEOPLE'S REPUBLIC OF CHINA

Certification Mark:



Scope of Certificate: **Design and Development, Production and Distribution of In Vitro Diagnostic Reagent and Instrument for the Detection of Drugs of Abuse, Fertility, Infectious Diseases, Oncology, Biochemistry, Cardiac Diseases, Allergic Disease based on Rapid Test, PCR and Liquid Biochip Method.**

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned above has established and is maintaining a quality management system, which meets the requirements of the listed standard(s). All applicable requirements of the testing and certification regulation of TÜV SÜD Group have to be complied with. For details and certificate validity see: www.tuvsud.com/ps-cert?q=cert:Q5 092305 0001 Rev. 02

Report No.: SH2398804

Valid from: 2024-03-17

Valid until: 2027-03-16

Date, 2024-03-01

Christoph Dicks
Head of Certification/Notified Body

Certificate

No. Q5 092305 0001 Rev. 02

Applied Standard(s): ISO 13485:2016
(EN ISO 13485:2016/AC:2018, EN ISO 13485:2016/A11:2021)
Medical devices - Quality management systems -
Requirements for regulatory purposes

Facility(ies): **Zhejiang Orient Gene Biotech Co., Ltd.**
3787#, East Yangguang Avenue, Dipu Street Anji, 313300
Huzhou, Zhejiang, PEOPLE'S REPUBLIC OF CHINA

See Scope of Certificate



浙江东方基因生物制品股份有限公司
Zhejiang Orient Gene Biotech Co., LTD



CE-DOC-OG038
Version 2.0

EC Declaration of Conformity

In accordance with Directive 98/79/EC

Legal Manufacturer: *Zhejiang Orient Gene Biotech Co., Ltd*

Legal Manufacturer Address: *3787#, East Yangguang Avenue, Dipu Street,
Anji 313300, Huzhou, Zhejiang, China*

Declares, that the products
Product Name and Model(s)

Troponin I Rapid Test Cassette (Whole Blood/Serum/Plasma)	GDTRO-402a
Troponin I Rapid Test Cassette (Whole Blood/Serum/Plasma)	GDTRO-402b

Classification: *Other*

Conformity assessment route: *Annex III (EC DECLARATION OF CONFORMITY)*

We, the Manufacturer, herewith declare with sole responsibility that our product/s mentioned above meet/s the provisions of the Directive 98/79/EC of the European Parliament and of the Council on In-Vitro Diagnostic Medical Devices.

We hereby explicitly appoint

EC Representative's Name: *Shanghai International Holding Corp. GmbH (Europe)*

EC Representative's Address: *Eiffestrasse 80, 20537 Hamburg, Germany*

to act as our European Authorized Representative as defined in the aforementioned Directive.

I, the undersigned, hereby declare that the medical devices specified above conform with the directive 98/79/EC on in vitro diagnostic medical devices and pertinent essential requirements

Date Signed: August 11, 2020

Name of authorized signatory: *Joyce Pang*
Position held in the company: *Vice-President*

CERTIFICATE OF ANALYSIS

Product Name: Troponin I Rapid Test (Whole blood/Serum/Plasma) (Cassette)

Catalog NO.: GDTRO-402b

Purchase NO.: 2025-IEU168#

Lot NO.: 2509269

Quantity: 500pcs

Expiration Date: 2027.08

CONTROLS		SPECIFICATION	TEST RESULT	CONCLUSION
Negative Specimens		Negative	Negative	☒ Pass ☐ Fail
Positive Specimens	0.5ng/ml	Positive	Positive	☒ Pass ☐ Fail
	1ng/ml	Positive	Positive	☒ Pass ☐ Fail
	10ng/ml	Positive	Positive	☒ Pass ☐ Fail



Conclusion: ☒Pass:All results meet QC standard.

☐Fail

Test by :

查妍

QC Supervisor:

雷伙恩

Date: 2025.09.25

Date:2025.09.25

Troponin I

Troponin I Rapid Test Cassette (Whole Blood/Serum/Plasma)

Package Insert

A rapid visual immunoassay for the qualitative presumptive detection of cardiac Troponin I in human whole blood, serum, or plasma specimens. For professional *in vitro* diagnostic use only.

INTENDED USE

The Troponin I Rapid Test Cassette (Whole Blood/Serum/Plasma) is a rapid visual immunoassay for the qualitative presumptive detection of cardiac Troponin I in human whole blood, serum, or plasma specimens. This kit is intended to be used as an aid in the diagnosis of myocardial infarction (MI).

SUMMARY

Cardiac Troponin I (cTnI) is a protein found in cardiac muscle with a molecular weight of 22.5 kDa.¹ Troponin I is part of a three subunit complex comprising of Troponin T and Troponin C. Along with tropomyosin, this structural complex forms the main component that regulates the calcium sensitive ATPase activity of actomyosin in striated skeletal and cardiac muscle.² After cardiac injury occurs, Troponin I is released into the blood 4-6 hours after the onset of pain. The release pattern of cTnI is similar to CK-MB, but while CK-MB levels return to normal after 72 hours, Troponin I remains elevated for 6-10 days, thus providing for a longer window of detection for cardiac injury. The high specificity of cTnI measurements for the identification of myocardial damage has been demonstrated in conditions such as the perioperative period, after marathon runs, and blunt chest trauma.³ cTnI release has also been documented in cardiac conditions other than acute myocardial infarction (AMI) such as unstable angina, congestive heart failure, and ischemic damage due to coronary artery bypass surgery.⁴ Because of its high specificity and sensitivity in the myocardial tissue, Troponin I has recently become the most preferred biomarker for myocardial infarction.⁵

PRINCIPLE

The Troponin I Rapid Test Cassette (Whole Blood/Serum/Plasma) has been designed to detect cardiac Troponin I through visual interpretation of color development in the strip. The membrane was immobilized with anti-cTnI antibodies on the test region. During the test, the specimen is allowed to react with colored anti-cTnI antibodies colloidal gold conjugates, which were precoated on the sample pad of the test. The mixture then moves on the membrane by a capillary action, and interact with reagents on the membrane. If there were enough cTnI in specimens, a colored band will form at the test region of the membrane.

Presence of this colored band indicates a positive result, while its absence indicates a negative result. Appearance of a colored band at the control region serves as a procedural control. This indicates that proper volume of specimen has been added and membrane wicking has occurred.

PRECAUTIONS

1. For professional *in vitro* diagnostic use only.
2. Warning: the reagents in this kit contain sodium azide which may react with lead or copper plumbing to form potentially explosive metal azides. When disposing of such reagents, always flush with large volumes of water to prevent azide build-up.
3. Do not use it if the tube/pouch is damaged or broken.
4. Test is for single use only. Do not re- use under any circumstances.
5. Handle all specimens as if they contain infectious agents. Observe established standard procedure for proper disposal of specimens
6. Wear protective clothing such as laboratory coats, disposable gloves and eye protection when specimens are assay.
7. Humidity and temperature can adversely affect results.

STORAGE AND STABILITY

All reagents are ready to use as supplied. Store unused test cassette unopened at 2°C-30°C. If stored at 2°C-8°C, ensure that the test cassette is brought to room temperature before opening. The test is not stable out of the expiration date printed on the sealed pouch. Do not freeze the kit or expose the kit over 30°C.

SPECIMEN COLLECTION AND PREPARATION

- The Troponin I Rapid Test Cassette (Whole Blood/Serum/Plasma) is intended only for use with human whole blood, serum, or plasma specimens.
- Only clear, non-hemolyzed specimens are recommended for use with this test.
- Serum or plasma should be separated with soonest possible opportunity to avoid hemolysis.
- Perform the testing immediately after the specimen collection. Do not leave the specimens at room temperature for prolonged periods. Specimens may be stored at 2-8°C for up to 3 days. For long term storage, specimens should be kept below -20°C.
- Bring specimens to room temperature prior to testing. Frozen specimens must be completely thawed and mixed well prior to testing. Avoid repeated freezing and thawing of specimens.
- Pack the specimens in compliance with applicable regulations for transportation of etiological agents, in case they need to be shipped.
- Icteric, lipemic, hemolysed, heat treated and contaminated sera may cause erroneous results.
- There is a slight possibility that some whole blood specimens with very high viscosity or which have been stored for more than 2 days may not run properly on the test cassette. Repeat the test with a serum or plasma specimen from the same patient using a new test cassette.

Materials Provided

1. Test cassettes
2. Droppers
3. Package insert

Materials Required But Not Provided

1. Specimen collection containers
2. Centrifuge (for plasma only)
3. Clock or Timer

DIRECTIONS FOR USE

Allow test cassette, specimen, buffer and/or controls to equilibrate to room temperature (15-30°C) prior to testing.


1. Remove the test from its sealed pouch, and place it on a clean, level surface. Label the cassette with patient or control identification. To obtain a best result, the assay should be performed within one hour.

2. Transfer **2-3 drops of serum or plasma** to the specimen well(S) of the cassette with a disposable pipette provided in the kit, and then start the timer.

OR

Transfer **4 drops of whole blood** specimen to the specimen well(S) of the cassette with a disposable pipette provided in the kit, and then start the timer.

2. Wait for the colored band(s) to appear. The result should be read at 15 minutes. Do not interpret the result after 20 minutes.



INTERPRETATION OF RESULTS

(Please refer to the illustration above)

POSITIVE: Two colored bands appear on the membrane. One band appears in the control region (C) and another band appears in the test region (T).

NEGATIVE: Only one colored band appears in the control region (C). No apparent colored band appears in the test region (T).

INVALID: Control band fails to appear. Results from any test which has not produced a control band at the specified reading time must be discarded. Please review the procedure and repeat with a new test. If the problem persists, discontinue using the kit immediately and contact your local distributor.

NOTE: Insufficient specimen volume, incorrect operation procedure, or performing expired tests are the most likely reasons for control band failure.

QUALITY CONTROL

Internal procedural controls are included in the test. A colored band appearing in the control region (C) is considered an internal positive procedural control. It confirms sufficient specimen volume and correct procedural technique.

External controls are not supplied with this kit. It is recommended that positive and negative controls be tested as a good laboratory practice to confirm the test procedure and to verify proper test performance.

LIMITATIONS

1. The Troponin I Rapid Test Cassette (Whole Blood/Serum/Plasma) is for professional *in vitro* diagnostic use, and should be used for the qualitative detection of cardiac Troponin I only. There is no meaning attributed to line color intensity or width.
2. The Troponin I Rapid Test Cassette (Whole Blood/Serum/Plasma) will only indicate the presence of Troponin I in the specimen and should not be used as the sole criteria for the diagnosis.
3. If the test result is negative and clinical symptoms persist, additional testing using other clinical methods is recommended. The test cannot detect less than 0.25 ng/mL of cTnI in specimens. Thus, a negative result does not at any time rule out the existence of Troponin I in blood, because the antibodies may be absent or below the minimum detection level of the test.
4. Like with all diagnostic tests, a confirmed diagnosis should only be made by a physician after all clinical and laboratory findings have been evaluated.
5. Some specimens containing unusually high titers of heterophile antibodies or rheumatoid factor (RF) may affect expected results. Even if the test results are positive, further clinical evaluation should be considered with other clinical information available to the physician.

PERFORMANCE CHARACTERISTICS

Table: Troponin I Rapid Test vs. EIA

Method		Troponin I Rapid Test Cassette		Total Results
		Positive	Negative	
EIA	Positive	138	2	140
	Negative	1	315	316
Total Results		139	317	456

Relative Sensitivity: 98.6% (94.9%-99.8%)*

Overall Agreement: 99.3% (98.1%-99.9%)*

Relative Specificity: 99.7% (98.3%-99.9%)*

*95% Confidence Interval

BIBLIOGRAPHY

1. Adams, et al. Biochemical markers of myocardial injury, Immunoassay Circulation 88:750-763, 1993.
2. Mehegan JP, Tobacman LS. Cooperative interaction between troponin molecules bound to the cardiac thin filament. J.Biol.Chem. 266:966, 1991.
3. Adams, et al. Diagnosis of Perioperative myocardial infarction with measurements of cardiac troponin I. N.Eng.J.Med 330:670, 1994.
4. Hossein-Nia M, et al. Cardiac troponin I release in heart transplantation. Ann. Thorac.Surg. 61: 227, 1996.
5. Alpert JS, et al. Myocardial Infarction Redefined, Joint European Society of Cardiology American College of Cardiology: J. Am. Coll. Cardio., 36(3):959, 2000.

REF GDTRO-402b



EC Declaration of Conformity

In accordance with Directive 98/79/EC

Legal Manufacturer: Healgen Scientific Limited Liability Company

Legal Manufacturer Address: 3818 Fuqua Street, Houston, TX 77047, USA.

Declares, that the products
Product Name and Model(s)

Healgen HCV Hepatitis C Virus Rapid Test (Serum/Plasma)(Cassette)	GCHCV-302a
Healgen HCV Hepatitis C Virus Rapid Test (Whole blood/Serum/Plasma)(Cassette)	GCHCV-402a

EDMA Code: 15 70 02 02

Classification: Annex II List A
Conformity assessment route: Annex IV (Full Quality Assurance)

Compliance of the designated product with the Directive 98/79/EC has been assessed and certified by the Notified Body

Notified Body: TÜV SÜD Product Service GmbH

Notified Body Address: Munich Branch Ridlerstraße 65 80339 München Germany

EC Certificate No.: V1 092378 0004 Rev. 02 Valid until: 2025-05-26

EC Design-Examination Certificate No.: V7 092378 0009 Rev. 00 Valid until: 2025-05-26

It bears the mark

CE 0123

We, the Manufacturer, herewith declare with sole responsibility that our product/s mentioned above meet/s the provisions of the Directive 98/79/EC of the European Parliament and of the Council on In-Vitro Diagnostic Medical Devices.


We hereby explicitly appoint

EC Representative Name: QARAD b.v.b.a.

EC Representative Address: Ciplastraat 3, B-2440 Geel, Belgium

to act as our European Authorized Representative as defined in the aforementioned Directive.

I, the undersigned, hereby declare that the medical devices specified above conform with the directive 98/79/EC on in vitro diagnostic medical devices and pertinent essential requirements

Signature: 
Name of authorized signatory: Joyce Pang
Position held in the company: Vice-President
Date: 2022.4.22



文件复审批准日期
2020 年 09 月 14 日

文件复审批准日期
2022 年 08 月 25 日

文件编号/Doc. No. :QC-M0111F-01

版本号/Version:2.0

生效日期/Effective Date:2018 年 10 月 25 日

Page 1 of 1

HCV Hepatitis C Virus Rapid Test (whole blood/serum/plasma)(Cassette) (CE)
Final Products inspection Report (Certificate of Analysis)

Item	HCV Hepatitis C Virus Rapid Test(whole blood/serum/plasma) (Cassette) Catalog No:GCHCV-402a	Specification	Cassette
Lot No.	2407232	Quantity	100000 tests
Source From	Workshop	Inspection Basis	HCV Hepatitis C Virus Rapid Test(whole blood/serum/plasma)(Cassette)Finished Product Quality Standards and Inspection SOP(CE)
Validity Date	2026.06	Date of Sampling	2024.07.26

Inspection Item		Acceptance Standard	Results
Functional Requirement : S-1 (AQL:2.5)	P1(20-22)	T line should be $\geq G7$ at 15 min.	Conformity ☒ / Non-Conformity ☐
	P2(12-14)	T line should be $\geq G6$ at 15 min.	Conformity ☒ / Non-Conformity ☐
	P3(6-8)	T line should be $\geq G5$ at 15 min.	Conformity ☒ / Non-Conformity ☐
	P4(1-3)	T line signal should be $G4 \leq T \leq G6$ at 15 min.	Conformity ☒ / Non-Conformity ☐
Functional Requirement : / 100 negative serum		T line should be $\leq G2$ at 15 min. C line should be visible $\geq G3$ within 3 min,and should be $\geq G7$ at 15 min. Note: when negative Serum/Plasma moving over T line, T line should be $< G5$ and fade within 3 min. If doing this, it is qualified. If not, it is unqualified	Conformity ☒ / Non-Conformity ☐
Functional Requirement: S-2 AQL:1.0 clinical whole blood specimens		Membrane background is clean at 15min, not impacting reading results.	Conformity ☒ / Non-Conformity ☐
Pouch Leakage tightness S-2 (AQL=0.65) Remarks:N/A		Leakage tightness is good	Conformity ☒ / Non-Conformity ☐
Conclusion: Conformity ☒ / Non-Conformity ☐			
Tester/Date: 查妍 2024.07.26		Reviewer/Date: 雷伙愚 2024.07.26	

HCV Hepatitis C Virus Rapid Test (Whole blood/Serum/Plasma)(Cassette)

CE 0123

REF GCHCV-402a

INTENDED USE

The HCV Hepatitis C Virus Rapid Test (Whole blood/Serum/Plasma)(Cassette) is a sandwich lateral flow chromatographic immunoassay for the qualitative detection of antibodies (IgG, IgM, and IgA) to Hepatitis C virus (HCV) in human whole blood, serum or plasma. It is intended to be used as a screening test and as an aid in the diagnosis of infection with HCV. Any reactive specimen with the HCV Hepatitis C Virus Rapid Test Cassette must be confirmed with alternative testing method(s) and clinical findings.

INTRODUCTION

Hepatitis C Virus (HCV) is a small, enveloped, positive-sense, single-stranded RNA virus. Antibody to HCV is found in over 80% of patients with well-documented non-A, non-B hepatitis. Conventional methods fail to isolate the virus in cell culture or visualize it by electron microscope. Cloning the viral genome has made it possible to develop serologic assays that use recombinant antigens (1, 2). Compared to the first generation HCV EIAs using single recombinant antigen, multiple antigens using recombinant protein and/or synthetic peptides have been added in new serologic tests to avoid nonspecific cross-reactivity and to increase the sensitivity of the HCV antibody tests (3, 4). HCV Hepatitis C Virus Rapid Test (Whole blood/Serum/Plasma)(Cassette) is a rapid test to qualitatively detect the presence of antibody to HCV in a whole blood, serum or plasma specimen. The test utilizes a combination of recombinant antigen to selectively detect elevated levels of HCV antibodies in whole blood, serum or plasma.

PRINCIPLE

The HCV Hepatitis C Virus Rapid Test Cassette is a lateral flow chromatographic immunoassay based on the principle of the double antigen–sandwich technique. The test Cassette consists of: 1) a burgundy colored conjugate pad containing HCV antigens conjugated with colloidal gold (HCV Ag conjugates) and rabbit IgG-gold conjugates, 2) a nitrocellulose membrane Cassette containing a test band (T band) and a control band (C band). The T band is pre-coated with non-conjugated HCV antigens, and the C band is pre-coated with goat anti-rabbit IgG. When an adequate volume of test specimen is dispensed into the sample well of the Cassette, the specimen migrates by capillary action across the Cassette. The antibodies: either the IgG, the IgM, or the IgA, to HCV if present in the specimen will bind to the HCV Ag conjugates. The immunocomplex is then captured on the membrane by the precoated HCV antigens, forming a burgundy colored T band, indicating a HCV Ab positive test result. Absence of the T band suggests a negative result. The test contains an internal control (C band) which should exhibit a burgundy colored band of the immunocomplex of goat anti-rabbit IgG and rabbit IgG-gold conjugate regardless the presence of any antibodies to HCV. Otherwise, the test result is invalid and the specimen must be retested with another Cassette.

PRODUCT CONTENTS

HCV Hepatitis C Virus Rapid Test (Whole blood/Serum/Plasma)(Cassette) containing HCV antigen (HCV antigen includes core, NS3, NS4 and NS5 segment) coated particles and HCV antigen (HCV recombinant antigen includes core, NS3, NS4 and NS5 segment) coated on the membrane.

MATERIALS SUPPLIED

- 25 sealed pouches each containing a test cassette, a pipette dropper and a desiccant (Test Cassette T band is pre-coated with non-conjugated HCV antigens, and the C band is pre-coated with goat anti-rabbit IgG on the nitrocellulose and coupled to colloidal gold on label pad)
- 1 Package insert

- 1 Buffer (4 mL) (Casein-salt: 1%, NaCl: 0.9%, Na₂HPO₄: 0.286%, NaN₃: 0.5%)



Warning

Warning: 0.5% NaN₃
Harmful if swallowed; Harmful to aquatic life with long lasting effects
Prevention
Wash face, hands and any exposed skin thoroughly after handling Wear protective gloves/protective clothing/eye protection/face protection
Do not breathe dust/fume/gas/mist/vapors/spray
Do not eat, drink or smoke when using this product
Avoid release to the environment.
Response
IF SWALLOWED: rinse mouth. Do NOT induce vomiting.
Get medical attention/advice if you feel unwell

MATERIAL REQUIRED BUT NOT PROVIDED

- Specimen collection containers
- Sterile lancets (for fingerstick whole blood only)
- Centrifuge (for plasma only)
- Timer
- Heparinized capillary tubes and dispensing bulb (for fingerstick whole blood only)

STORAGE AND STABILITY

The kit can be stored at room temperature or refrigerated (2-30°C). The test Cassette is stable through the expiration date printed on the sealed pouch. The test Cassette must remain in the sealed pouch until use. DO NOT FREEZE. Do not use beyond the expiration date.

WARNINGS AND PRECAUTIONS

- For professional *in vitro* diagnostic use only. Do not use after expiration date.
- Warning: the reagents in this kit contain sodium azide which may react with lead or copper plumbing to form potentially explosive metal azides. When disposing of such reagents, always flush with large volumes of water to prevent azide build-up.
- Do not use it if the tube/pouch is damaged or broken.
- Test is for single use only. Do not re-use under any circumstances.
- Handle all specimens as if they contain infectious agents. Observe established precautions against microbiological hazards throughout testing and follow the standard procedures for proper disposal of specimens.
- Wear protective clothing such as laboratory coats, disposable gloves and eye protection when specimens are assayed.
- Humidity and temperature can adversely affect results.
- Do not perform the test in a room with strong air flow, i.e. an electric fan or strong airconditioning.


SPECIMEN COLLECTION

- The HCV Hepatitis C Virus Rapid Test (Whole Blood/Serum/Plasma) (Cassette) can be performed using whole blood (from venipuncture and fingerstick), serum or plasma.
- For venipuncture whole blood and plasma: K-EDTA, Sodium Heparin, Sodium citrate Sterile, and Lithium heparin should be used as the anticoagulant. Other anticoagulants have not been tested and may give incorrect results.
- To collect Fingerstick Whole Blood specimens:
 - Wash the patient's hand with soap and warm water or clean with an alcohol wipe . Allow to dry.
 - Massage the hand without touching the puncture site by rubbing down the hand towards the fingertip of the middle or ring finger.
 - Puncture the skin with a new sterile lancet for each person. Wipe away the first sign of blood.
 - Gently rub the hand from wrist to palm to finger to form a rounded drop of blood over the puncture site.
 - Add the Fingerstick Whole Blood specimen to the test device by using a capillary tube:
 - Touch the end of the capillary tube to the blood until filled to approximately 60 µL. Avoid air bubbles.
 - Place the bulb onto the top end of the capillary tube, then squeeze the bulb to dispense the whole blood into the specimen well (S) of the test device.
 - Separate serum or plasma from blood as soon as possible to avoid hemolysis. Use only clear, nonhemolyzed specimens.
- Testing should be performed immediately after specimen collection. Do not leave the specimens at room temperature for prolonged periods. Serum and plasma specimens may be stored at 2-8°C for up to 3 days and may be stored at -20°C for 6 months. Whole blood collected by venipuncture should be stored at 2-8°C if the test is to be run within 2 days of collection. Do not freeze whole blood specimens. Whole blood collected by fingerstick should be tested immediately.
- Bring specimens to room temperature prior to testing. Frozen specimens must be completely thawed and mixed well prior to testing. Specimens should not be frozen and thawed repeatedly.
- If specimens are to be shipped, they should be packed in compliance with usual regulations for transportation of etiological agents.

TEST PROCEDURE

Allow test cassette, specimen, buffer and/or controls to equilibrate to room temperature (15-30°C) prior to testing.

- Remove the test Cassette from the foil pouch and use it as soon as possible. Best results will be obtained if the assay is performed within one hour.
- Place the test Cassette on a clean and level surface.
For Serum or Plasma Specimens: Hold the dropper vertically and transfer 1 drop of serum or plasma (approximately 30 µL) to the specimen well (S) of the test Cassette, then add 1 drop of buffer (approximately 40 µL) and start the timer. See illustration below.
For Venipuncture Whole Blood Specimens: Hold the dropper vertically and transfer 2 drops of venipuncture whole blood (approximately 60 µL) to the specimen well (S) of the test Cassette, then add 1 drop of buffer (approximately 40 µL) and start the timer. See illustration below.
For Fingerstick Whole Blood specimens: To use a capillary tube: Fill the capillary tube and transfer approximately 60 µL of fingerstick whole blood specimen to the specimen well (S) of the test device, then add 1 drops of buffer (approximately 40 µL) and start the timer. See illustration below.
- Wait for the red line(s) to appear. The result should be read at 15 minutes. Do not interpret the result after 30 minutes.



INTERPRETATION OF RESULTS

(Please refer to the illustration above)

Positive: Two lines appear. One colored line should be in the control line region (C) and another apparent colored line should be in the test line region (T).

Negative: One colored line appears in the control line region (C). No line appears in the test line region (T).

Invalid: Control line fails to appear. Insufficient specimen volume or incorrect procedural techniques are the most likely reasons for control line failure. Review the procedure and repeat the test with a new test Cassette. If the problem persists, discontinue using the test kit immediately and contact your local distributor.

QUALITY CONTROL

A procedural control is included in the test. A red line appearing in the control region (C) is the internal procedural control. It confirms sufficient specimen volume and correct procedural technique. Control standards are not supplied with this test. However, it is recommended that positive and negative controls are sourced from a local competent authority and tested as a good laboratory practice, to confirm the test procedure and verify the test performance.

LIMITATIONS

1. The HCV Hepatitis C Virus Rapid Test (Whole blood/Serum/Plasma)(Cassette) is for *in vitro* diagnostic use only. This test should be used for the detection of antibodies to HCV in whole blood, serum or plasma specimen.
2. The HCV Hepatitis C Virus Rapid Test (Whole blood/Serum/Plasma)(Cassette) will only indicate the presence of antibodies to HCV in the specimen and should not be used as the sole criteria for the diagnosis of Hepatitis C viral infection.
3. As with all diagnostic tests, all results must be considered with other clinical information available to the physician.
4. If the test result is negative and clinical symptoms persist, additional follow-up testing using other clinical methods is recommended. A negative result at any time does not preclude the possibility of Hepatitis C Virus infection.
5. A negative result can occur if the quantity of the antibodies to HCV present in the specimen is below the detection limits of the assay, or the antibodies that are detected are not present during the stage of disease in which a sample is collected.
6. Some specimens containing unusually high titer of heterophile antibodies or rheumatoid factor may affect expected results.
7. As with all diagnostic tests, a definitive clinical diagnosis should not be based on the results of a single test, but should only be made by the physician after all clinical and laboratory findings have been evaluated.
8. Results should not be used to determine the genotype of HCV infections.
9. Due to possible cross reactivity, the appearance of lines in T line does not necessarily indicate co-infection from IgG, IgM or IgA, nor can it identify the serotype.
10. The recommended anticoagulants are K₂EDTA, Sodium Heparin, Sodium citrate Sterile and Lithium heparin for venous whole blood. Other anticoagulants have not been evaluated with this test.

PERFORMANCE CHARACTERISTICS

Relative Sensitivity

A total of 506 HCV positive specimens were tested using the HCV Hepatitis C Virus Rapid Test (Whole blood/Serum/Plasma)(Cassette) and a commercially available test (Table 1). The relative sensitivity of the test is >99.9% (95% confidence interval: 99.27% – 100%).

Table 1: Sensitivity of HCV Hepatitis C Virus Rapid Test (Whole blood/Serum/Plasma)(Cassette)

Population	Specimen Type	Number of Specimens Tested	Positive by HCV Hepatitis C Virus Rapid Test	Positive by Commercially Available Test
Anti-HCV (any genotype)	plasma	329	329/329 (100%)	329/329 (100%)
Anti-HCV (any genotype)	Serum	26	26/26 (100%)	26/26 (100%)
Anti-HCV (genotype 1, 2, 3, 4 (non-subtype A), 4, 5, 6)	Serum/Plasma	151	151/151 (100%)	151/151 (100%)
Total		506	506/506 (100%)	506/506 (100%)

30 Seroconversion panels have been done and details of the 30 seroconversion are in the table below.

No.	Panel	Specimens No.	Results
1	PHV907	7	Positive from 0 days since first bleed
2	PHV908	13	Positive from 3 days since first bleed
3	PHV206(M)	25	/
4	PHV911(M)	5	Positive from 3 days since first bleed
5	PHV919	7	Positive from 28 days since first bleed
6	PHV920	10, No. 2 can't be got because of out of stock from the vendor	Positive from 16 days since first bleed
7	HCV9047	10	Positive from 28 days since first bleed

8	HCV9046	5	Positive from 69 days since first bleed
9	HCV6229	8	Positive from 17 days since first bleed
10	HCV10041	3	Positive from 6 days since first bleed
11	HCV9041	8	Positive from 62 days since first bleed
12	HCV9045	8	Positive from 37 days since first bleed
13	HCV6222	3	Positive from 40 days since first bleed
14	HCV6224	8	Positive from 19 days since first bleed
15	HCV6227	7	Positive from 75 days since first bleed
16	HCV6228	12	Positive from 31 days since first bleed
17	HCV10071	7	Positive from 84 days since first bleed
18	HCV6220	6	Positive from 18 days since first bleed
19	HCV10185	5	Positive from 130 days since first bleed
20	HCV10235	5	Positive from 96 days since first bleed
21	HCV6215	4	Positive from 20 days since first bleed
22	HCV9042	6	Positive from 8 days since first bleed
23	HCV9058	5	Positive from 10 days since first bleed
24	HCV9094	5	Positive from 9 days since first bleed
25	HCV9095	5	Positive from 10 days since first bleed
26	HCV9055	11	Positive from 65 days since first bleed
27	HCV9054	10	Positive from 72 days since first bleed
28	HCV9044	6	Positive from 21 days since first bleed
29	HCV10165	9	Positive from 19 days since first bleed
30	HCV6226	12	Positive from 39 days since first bleed

Relative Specificity

A total of HCV 1259 negative specimens were tested using the HCV Hepatitis C Virus Rapid Test (Whole blood/Serum/Plasma)(Cassette) and a commercially available test (Table 2). The relative specificity of the test is >99.9% (95% confidence interval: 99.71% – 100%).

Table 2: Specificity of the HCV Hepatitis C Virus Rapid Test (Whole blood/Serum/Plasma)(Cassette)

Population	Specimens Tested	Number of Specimens Tested	Negative by HCV Hepatitis C Virus Rapid Test	Negative by Commercially Available Test
Clinical Negative	Serum/plasma	202	202/202 (100%)	202/202 (100%)
Potentially cross-reacting	Serum/Plasma	30	30/30 (100%)	30/30 (100%)
Unselected Donors	Serum	1000	1000/1000 (100%)	1000/1000 (100%)
Inhibition Panel	Serum	27	27/27 (100%)	27/27 (100%)
Total		1259	1259/1259 (100%)	1259/1259 (100%)

Whole Blood vs. Serum vs. Plasma

Total 25 clinical negative samples (whole blood, serum, plasma) have been collected from patients in local hospital. The whole blood collected and separated into three tubes. One was stored as whole blood. One was collected into tube for plasma, one was collected into tube for serum (Table 3). There is a very good correlation of results between whole blood, serum, and plasma with HCV negative samples.

Table 3: A Comparison of HCV Hepatitis C Virus Rapid Test (Whole blood/Serum/Plasma)(Cassette) Specificity in negative Whole Blood and Paired Serum and Plasma Specimens

Specimen Type	Number of Specimens Tested	Negative by HCV Ab
Serum	25	25/25 (100%)
Plasma	25	25/25 (100%)
Whole blood	25	25/25 (100%)

A total of 25 positive specimens (whole blood, serum, plasma) were tested using the HCV Hepatitis C Virus Rapid Test (Whole blood/Serum/Plasma)(Cassette) (Table 4). There is a very good correlation of results between whole blood and paired plasma with HCV positive samples.

Table 4: A Comparison of HCV Hepatitis C Virus Rapid Test (Whole blood/Serum/Plasma)(Cassette) Specificity in positive Whole Blood and Paired Serum and Plasma Specimens.

Specimen Type	Number of Specimens Tested	Positive by HCV Ab
Serum	25	25/25 (100%)
Plasma	25	25/25 (100%)
Whole blood	25	25/25 (100%)

Precision

Intra Assay

Within-run precision has been determined by using 20 replicates of four specimens: a negative, a low positive, medium positive and a high positive. The negative, low positive, medium positive and high positive values were correctly identified >99% of the time.

Inter-Assay

Between-run precision has been determined by 5 independent assays on the same four specimens: a negative, a low positive, medium positive and a high positive. Three different lots of the HCV Hepatitis C Virus Rapid Test (Whole blood/Serum/Plasma)(Cassette) have been tested using negative, low positive, medium positive and high positive specimens. The specimens were correctly identified >99% of the time.

Cross Reactivity

No cross-reactivity was observed when samples positive for other diseases such as HIV, Syphilis, Infectious Mononucleosis, HBV, Rheumatoid Factor, HAMA, Hyper IgG, Hyper IgM, anti-HAV, anti-HSV2, anti-HEV, anti-EBV and anti-CMV were tested.

Interfering Substances

No interference was observed in samples with high concentrations of Uric acid, Ascorbic Acid, Hemoglobin, Gentistic Acid, Acetaminophen, Oxalic Acid, Albumin, Caffeine, Bilirubin, EDTA, Aspirin and Methanol.

Analytes	Conc	Analytes	Conc
Control	0	Control	0
Uric acid	0.15 mg/mL	Albumin	20 mg/mL
Ascorbic Acid	0.2 mg/mL	Caffeine	0.2 mg/mL
Hemoglobin	5.0 mg/mL	Bilirubin	0.3 mg/mL
Gentistic Acid	0.2 mg/mL	EDTA	0.2 mg/mL
Acetaminophen	1.0 mg/mL	Aspirin	0.2 mg/mL
Oxalic Acid	0.2 mg/mL	Methanol	1.0%

REFERENCE

1. Choo, Q.L., G.Kuo, A.J. Weiner, L.R. Overby, D.W. Bradley, and M. Houghton. Isolation of a cDNA clone derived from a blood-borne non-A, non-B viral hepatitis genome. Science 189; 244: 359
2. Kuo, G., Q.L. Choo, H.J. Alter, and M. Houghton. An assay for circulating antibodies to a major etiologic Virus of human non-A, non-B hepatitis. Science 1989; 244: 362
3. Van der Poel, C.L., H.T.M. Cuypers, H.W. Reesink, and P.N. Lelie. Confirmation of hepatitis C Virus infection by new four-antigen recombinant immunoblot assay. Lancet 1991; 337: 317
4. Wilber, J.C. Development and use of laboratory tests for hepatitis C infection: a review. J. Clin. Immunoassay 1993; 16: 204

INDEX OF SYMBOLS

	Consult instructions for use		Tests per kit		Authorized Representative
	For <i>in vitro</i> diagnostic use only		Use by		Do not reuse
	Store between 2-30°C		Lot Number		Catalog #
	Manufacturer		Warning		

 Healgen Scientific Limited Liability Company
Address: 3818 Fuqua Street, Houston, TX 77047, USA.
Tel: +1 713-733-8088 Fax: +1 713-733-8848
Website: www.healgen.com

 QARAD b.v.b.a.
Cipalstraat 3, B-2440 Geel, Belgium



Zhejiang Orient Gene Biotech Co., LTD

CERTIFICATE OF ANALYSIS

Product Name: HBsAg Rapid Test (Whole blood/Serum/Plasma) (Cassette)

Catalog NO.: GCHBsg-402a

Purchase NO.: 2025-IEU148#

Lot NO.: 2508344

Quantity:4255 pcs

Expiration Date: 2027.07

CONTROLS		SPECIFICATION	TEST RESULT	CONCLUSION
Negative Specimens		Negative	Negative	☒ Pass ☐ Fail
Positive Specimens	1ng/ml	Positive	Positive	☒ Pass ☐ Fail
	2ng/ml	Positive	Positive	☒ Pass ☐ Fail
	3ng/ml	Positive	Positive	☒ Pass ☐ Fail
	5ng/ml	Positive	Positive	☒ Pass ☐ Fail

Conclusion: ☒Pass: All results meet QC standard.
☐Fail



Test by: 查妍

Date: 2025.08.19

QC Supervisor: 雷似愚

Date: 2025.08.19

Hepatitis B Surface Antigen Rapid Test Cassette (Whole blood/Serum/Plasma)

INTENDED USE

The Hepatitis B Surface Antigen Rapid Test Cassette (Whole Blood/Serum/Plasma) is a lateral flow chromatographic immunoassay for the qualitative detection of Hepatitis B surface antigen (HBsAg) in human whole blood, serum or plasma. It is intended to be used as a screening test and as an aid in the diagnosis of infection with Hepatitis B virus (HBV). Any reactive specimen with the HBsAg Rapid Test Cassette must be confirmed with alternative testing method(s) and clinical findings.

INTRODUCTION

Viral hepatitis is a systemic disease primarily involving the liver. Most cases of acute viral hepatitis are caused by Hepatitis A virus, Hepatitis B virus (HBV) or Hepatitis C virus. The complex antigen found on the surface of HBV is called HBsAg. The presence of HBsAg in serum or plasma is an indication of an active Hepatitis B infection, either acute or chronic. In a typical Hepatitis B infection, HBsAg will be detected 2 to 4 weeks before the ALT level becomes abnormal and 3 to 5 weeks before symptoms or jaundice develop. HBsAg has four principal subtypes: adw, ayw, adr and ayr. Because of antigenic heterogeneity of the determinant, there are 10 major serotypes of Hepatitis B virus. The HBsAg Test Cassette (Whole Blood/Serum/Plasma) is a rapid test to qualitatively detect the presence of HBsAg in whole blood, serum or plasma specimens. The test utilizes a combination of double monoclonal antibodies to selectively detect elevated levels of HBsAg in whole blood, serum or plasma.

PRINCIPLE

The Hepatitis B Surface Antigen Rapid Test Cassette (Whole Blood/Serum/Plasma) is a lateral flow chromatographic immunoassay based on the principle of the double antibody-sandwich technique. The membrane is pre-coated with anti-HBsAg antibodies on the test line region of the test. During testing, Hepatitis B Surface Antigen in the whole blood, serum or plasma specimen reacts with the particle coated with anti-HBsAg antibody. The mixture migrates upward on the membrane, chromatographically by capillary action to react with anti-HBsAg antibodies on the membrane and generate a colored line. The presence of this colored line in the test region indicates a positive result, while its absence indicates a negative result. To serve as a procedural control, a colored line will always appear in the control line region indicating that the proper volume of specimen has been added and membrane wicking has occurred.

PRODUCT CONTENTS

The Hepatitis B Surface Antigen Rapid Test Cassette (Whole Blood/Serum/Plasma) containing anti-HBsAg antibodies particles and anti-HBsAg antibodies coated on the membrane.

MATERIALS SUPPLIED

Test cassette	Dropper	Buffer	Package insert
MATERIAL REQUIRED BUT NOT PROVIDED			
1. Specimen collection containers			
2. Lancets (for fingerstick whole blood only)			
3. Centrifuge (for plasma only)			
4. Timer			
5. Heparinized capillary tubes and dispensing bulb (for fingerstick whole blood only)			
STORAGE AND STABILITY			
The kit can be stored at room temperature or refrigerated (2-30°C). The test device is stable through the expiration date printed on the sealed pouch. The test cassette must remain in the sealed pouch until use. DO NOT FREEZE. Do not use beyond the expiration date.			
WARNINGS AND PRECAUTIONS			
1. For professional In Vitro diagnostic use only. Do not use after expiration date.			
2. Warning: the reagents in this kit contain sodium azide which may react with lead or copper plumbing to form potentially explosive metal azides. When disposing of such reagents, always flush with large volumes of water to prevent azide build-up.			

MATERIAL REQUIRED BUT NOT PROVIDED

STORAGE AND STABILITY

WARNINGS AND PRECAUTIONS

SPECIMEN COLLECTION AND PREPARATION

TEST PROCEDURE

Allow test cassette, specimen, buffer and/or controls to equilibrate to room temperature (15-30°C) prior to testing.


1. Remove the test cassette from the foil pouch and use it as soon as possible. Best results will be obtained if the assay is performed within one hour.
2. Place the test cassette on a clean and level surface.

For Serum or Plasma Specimens: Hold the dropper vertically and transfer 2-3 drops of serum or plasma (approximately 60-90µL) to the specimen well (S) of the test cassette. See illustration below.

For Venipuncture Whole Blood Specimens: Hold the dropper vertically and transfer 2 drops of venipuncture whole blood (approximately 50µL) to the specimen well (S) of the test cassette, then add 1 drop of buffer (approximately 40 µL) and start the timer. See illustration below.

For Fingerstick Whole Blood Specimens: Allow 2 hanging drops of fingerstick whole blood (approximately 50 µL) to fall into the center of the specimen well (S) on the test cassette, then add 1 drop of buffer (approximately 40 µL) and start the timer. See illustration below.

3. Wait for the red line(s) to appear. The result should be read in 15 minutes. Do not interpret the result after 15 minutes



INTERPRETATION OF RESULTS



POSITIVE: Two distinct red lines appear. One line should be in the control region (C) and another line should be in the test region (T).

NEGATIVE: One red line appears in the control region (C). No apparent red or pink line appears in the test region (T).

INVALID: Control line fails to appear. Insufficient specimen volume or incorrect procedural techniques are the most likely reasons for control line failure. Review the procedure and repeat the test with a new test Cassette. If the problem persists discontinue using the test kit immediately and contact your local distributor

QUALITY CONTROL

A procedural control is included in the test. A red line appearing in the control region (C) is the internal procedural control. It confirms sufficient specimen volume and correct procedural technique. Control standards are not supplied with this kit; however, it is recommended that positive and negative controls be tested as a good laboratory practice to confirm the test procedure and to verify proper test performance.

LIMITATIONS

1. Though the Hepatitis B Surface Antigen Rapid Test Cassette (Whole Blood/Serum/Plasma) is a reliable screening assay, it should not be used as a sole criterion for diagnosis of HBV infection.
2. The HBsAg Rapid Test Cassette is limited to the qualitative detection of HBsAg in human whole blood, serum or plasma. The intensity of the test band does not have linear correlation with HBsAg titer in the specimen.
3. A negative test result does not preclude the possibility of exposure to or infection with HBV. Infection through recent exposure (seroconversion) to HBV may not be detectable.
4. A negative result can occur if the quantity of HBsAg present in the specimen is below the detection limits of the assay (lower than 1 ng/mL), or the HBsAg that are detected are not present during the stage of disease in which a sample is collected.
5. Interference due to heterophile antibodies, Rheumatoid Factors and other nonanalyte substances in patient's serum, capable of binding antibodies multivalently and providing erroneous analyte detection in immunoassays, has been reported in various studies. Both laboratory professionals and clinicians must be vigilant to this possibility of antibody interference. Results that appear to be internally inconsistent or incompatible with the clinical presentation should invoke suspicion of the presence of an endogenous artifact and lead to appropriate in vitro investigative action.
6. This kit is intended ONLY for testing of individual samples. Do not use it for testing of cadaver samples, saliva, urine or other body fluids, or pooled (mixed) blood.
7. As with all diagnostic tests, a definitive clinical diagnosis should not be based on the result of a single test, but should only be made by the physician after all clinical and laboratory findings have been evaluated.

PERFORMANCE CHARACTERISTICS

Sensitivity:

The Hepatitis B Surface Antigen Rapid Test Cassette (Whole Blood/Serum/Plasma) has been tested with a sensitivity panel ranging from 0 to 300 ng/mL. All 10 HBsAg subtypes produced positive results on the Hepatitis B Surface Antigen Rapid Test Cassette (Whole Blood/Serum/Plasma). The test can detect 5ng/mL of HBsAg in 10 minutes, and 1 ng/mL of HBsAg in 15 minutes.

Specificity:

Antibodies used for the Hepatitis B Surface Antigen Rapid Test Cassette (Whole Blood/Serum/Plasma) were developed against whole Hepatitis B antigen isolated from Hepatitis B virus. Specificity of the Hepatitis B Surface Antigen Rapid Test Cassette (Whole Blood/Serum/Plasma) was also tested with laboratory strains of Hepatitis A and Hepatitis C. They all yielded negative results.

Hepatitis B Surface Antigen Rapid Test Cassette (Whole Blood/Serum/Plasma) vs. EIA test

Method	EIA		Total Results
	Results	Positive	Negative
Hepatitis B Surface Antigen Rapid Test Cassette (Whole Blood/Serum/Plasma)	Positive	345	5
	Negative	2	980
Total Results		347	985

Relative sensitivity: 99.4%

Relative specificity 99.5%

Accuracy: 99.5%

REFERENCE

1. Blumberg, B. S. The Discovery of Australian Antigen and its relation to viral hepatitis. Vitro. 1971; 7: 223

Catalogue No:GCHBsg-402a

Effective Date: 2023-08-22

B20137-05



Zhejiang Orient Gene Biotech Co., LTD

CERTIFICATE OF ANALYSIS

Product Name: HBcAb Rapid Test (Whole blood Serum Plasma) (Cassette)

Catalog NO.: GCHBcB-402a

Purchase NO.: 2025-SI268#

Lot NO.: S2503209

Quantity: 200 pcs

Expiration Date: 2027-02

CONTROLS	SPECIFICATION	TEST RESULT	CONCLUSION
Negative Specimens	Negative	Negative	☒ Pass ☐ Fail
Positive Specimens	Positive	Positive	☒ Pass ☐ Fail

Conclusion: ☒Pass: All results meet QC standard.

☐Fail

Test by:

查妍

QC Supervisor:

雷似愚

Date: 2025.04.01

Hepatitis B Core Antibody Rapid Test Cassette (Whole Blood/Serum/Plasma)

A rapid, one step test for the qualitative detection of Hepatitis B Core Antibody (HBcAb) in whole blood, serum or plasma. It is for professional in vitro diagnostic use only.

INTENDED USE

The Hepatitis B Core Antibody Rapid Test Cassette (Whole Blood/Serum/Plasma) is a rapid, one step test for the qualitative detection of Hepatitis B Core Antibody (HBcAb) in whole blood, serum or plasma.

SUMMARY

Chronic hepatitis B is a serious, debilitating illness that can cause cirrhosis of the liver, liver cancer and death. Chronic hepatitis B is the main cause of liver cancer and the tenth leading cause of death worldwide, with 400,000,000 people infected with the virus. Every year, one million people worldwide are expected to die from this infection.

Most people fight off the infection themselves, but approximately 5-10 percent of those infected with the virus become carriers, and an additional 5-10 percent of those infected each year will progress to chronic liver disease, cirrhosis and possibly liver cancer.

The Hepatitis B Core Antibody Rapid Test Cassette (Whole Blood/Serum/Plasma) is a rapid test to qualitatively detect the presence of HBcAb in whole blood, serum or plasma without the use of an instrument.

PRINCIPLE

Hepatitis B Core Antibody (HBcAb) is also known as anti-Hepatitis B Core Antigen (anti-HBc). These tests are immunoassays based on the principle of competitive binding.

During testing, the mixture migrates upward on the membrane chromatographically by capillary action. The membrane is pre-coated with HBcAg on the test line region of the strip. During testing, anti-HBc antibody, if present in the specimen, will compete with particle coated anti-HBc antibody for limited amount of HBcAg on the membrane, and no line will form in the test line region, indicating a positive result. A visible colored line will form in the test line region if there is no anti-HBc antibody in the specimen because all the antibody coated particles will be captured by the antigen coated in the test line region. To serve as a procedural control, a colored line will always appear in the control line region indicating that proper volume of specimen has been added and membrane wicking has occurred.

MATERIALS SUPPLIED

- | | | | |
|------------------|------------|-----------|-------------------|
| 1. Test cassette | 2. Dropper | 3. Buffer | 4. Package insert |
|------------------|------------|-----------|-------------------|

MATERIAL REQUIRED BUT NOT PROVIDED

- | | | |
|----------------------------------|---------------------------------|----------|
| 1. Specimen collection container | 2. Centrifuge (for plasma only) | 3. Timer |
|----------------------------------|---------------------------------|----------|

STORAGE AND STABILITY

Store as packaged in the sealed pouch either at room temperature or refrigerated (2-30°C). The test is stable through the expiration date printed on the sealed pouch. The test must remain in the sealed pouch until use. Do not freeze. Do not use beyond the expiration date.

WARNINGS AND PRECAUTIONS

1. For professional *in vitro* diagnostic use only. Do not use after expiration date.
2. Do not eat, drink or smoke in the area where the specimens or kits are handled.
3. Handle all specimens as if they contain infectious agents. Observe established precautions against microbiological hazards throughout testing and follow the standard procedures for proper disposal of specimens.
4. Wear protective clothing such as laboratory coats, disposable gloves and eye protection when specimens are being tested.
5. Humidity and temperature can adversely affect results.

SPECIMEN COLLECTION AND PREPARATION

1. The Hepatitis B Core Antibody Rapid Test Cassette (Whole Blood/Serum/Plasma) can be performed using whole blood, serum or plasma.
2. To collect venipuncture whole blood specimens: Collect anti-coagulated blood sample (EDTA-K2, heparin, and sodium citrate) following standard laboratory procedures.
3. Separate the serum or plasma from blood as soon as possible to avoid hemolysis. Only clear, no hemolyzed specimens can be used.
4. Testing should be performed immediately after the specimens have been collected. Do not leave the specimens at room temperature for prolonged periods. Specimens may be stored at 2-8°C for up to 3 days. For long term storage, specimens should be kept below -20°C.
5. Bring specimens to room temperature prior to testing. Frozen specimens must be completely thawed and mixed well prior to testing. Specimens should not be frozen and thawed repeatedly.
6. If specimens are to be shipped, they should be packed in compliance with local regulations for the transportation of etiologic agents.

TEST PROCEDURE

Allow test cassette, specimen, and/or controls to equilibrate to room temperature (15-30°C) prior to testing.

1. Bring the pouch to room temperature before opening it. Remove the test cassette from the sealed pouch and use it as soon as possible. Best results will be obtained if the assay is performed within one hour.
2. Place the test cassette on a clean and level surface.


For Serum or Plasma:

Hold the dropper vertically and transfer 3 full drops of serum or plasma (approx. 75 µL) to each specimen well (S) of the test cassette respectively, avoid trapping air bubbles in the specimen well (S). See illustration below.


For Venipuncture Whole Blood specimens:

Hold the dropper vertically and transfer 2 drops of whole blood (approx. 50 µL) to the specimen well (S) of the test cassette, then add 1 drop of buffer (approximately 40 µL) and starts the time. See illustration below.

3. Wait for the red line(s) to appear. The results should be read at 15 minutes. Do not interpret the results after 20 minutes.



INTERPRETATION OF RESULTS



- POSITIVE:** One red line appears in the control line region (C), No apparent red or pink line appears in the test region (T).
- NEGATIVE:** Two red lines appear. One line should be in the control line region (C) and another line should be in the test line region (T).
- INVALID:** Control line fails to appear. Insufficient specimen volume or incorrect procedural techniques are the most likely reasons for control line failure. Review the procedure and repeat the test with a new test cassette. If the problem persists, discontinue using the test kit immediately and contact your local distributor.

QUALITY CONTROL

A procedural control is included in the test. A red line appearing in the control line region (C) is the internal procedural control. It confirms sufficient specimen volume and correct procedural technique.

Control standards are not supplied with this kit; however, it is recommended that positive and negative controls be tested as good laboratory practice to confirm the test procedure and to verify proper test performance.

LIMITATIONS

- The Hepatitis B Core Antibody Rapid Test Cassette (Whole Blood/Serum/Plasma) is for *in vitro* diagnostic use only. This test should be used for the detection of HBcAb in whole blood, serum or plasma specimen. Neither the quantitative value nor the rate of increase in the concentration of HBcAb can be determined by this qualitative test.
- The Hepatitis B Core Antibody Rapid Test Cassette (Whole Blood/Serum/Plasma) will only indicate the presence of HBcAb in the specimen and should not be used as the sole criteria for the diagnosis of Hepatitis B viral infection.

- As with all diagnostic tests, all results must be considered with other clinical information.
- If the test result is negative and clinical symptoms persist, additional follow-up testing using other clinical methods is suggested. A negative result at any time does not preclude the possibility of Hepatitis B Virus infection.

PERFORMANCE CHARACTERISTICS

Sensitivity and Specificity

The Hepatitis B Core Antibody Rapid Test Cassette (Whole Blood/Serum/Plasma) was compared with leading commercial EIA HBcAb test, the results show that the Hepatitis B Core Antibody Rapid Test Cassette (Whole Blood/Serum/Plasma) has a high sensitivity and specificity.

Method		ELISA		Total Results
Hepatitis B Core Antibody Rapid Test Cassette	Results	Positive	Negative	
	Positive	443	4	
	Negative	17	120	
Total Results		460	124	584

Relative Sensitivity: 96.3%

Relative Specificity: 96.8%

Accuracy: 96.4%

REFERENCE

- Chizzali-Bonfadin C., Addlassnig K.P., Kreihsl M., Hatvan A., Horak W., Knowledge-based interpretation of serologic tests for hepatitis on the World Wide Web. Clin Perform Qual Health Care 1997 Apr-Jun 5: 61-3.
- ter Bog F., ten Kate F.J., Cuypers H.T., Leentvaar-Kuipers A., Oosting J., Wertheim-van Dillen P.M., Honkoop P, Rasch M.C., de Man R.A., van Hattum J., Chamelueau R.A., Reesink H.W., Jones E.A., Relation between laboratory results and histological hepatitis activity in individuals positive for hepatitis B surface antigen and antibodies to hepatitis B e antigen, Lancet 1998 June 351: 1914-8ng Infect Dis. 1997; 3: 213±221. <https://doi.org/10.3201/eid0302.970219> PMID: 9204307.