

Certificate of Analysis

Certificat d'Analyse

applied biosystems
by Thermo Fisher Scientific

VetMAX™ African Swine Fever Virus Detection Kit



100 Tests

LOT

ASFV-077



A28809



2024-11-20



2025-11-07

KIT CONFIGURATION CONFIGURATION DU KIT

COMPONENT DESCRIPTION <i>Description du composant</i>	REF	UNIT	LOT	
3 - Mix ASFV	MPEASFV	2 x 1000 µL	MPEASFV-077	2025-11-18
4a - EPC ASFV	EPCASFV	2 x 90 µL	EPCASFV-074	2025-11-07
5 - IPC ASFV	IPCASFV	500 µL	IPCASFV-061	2025-11-07

INSTRUCTIONS FOR USE

Notices d'utilisation



English MAN0010783 RevC.0
Spanish MAN0018204 RevB.0
French MAN0019541 RevA.0

QUALITY CONTROL RESULTS RESULTATS DE CONTROLE QUALITE

EPC <i>EPC</i>	Already extracted EPC <i>EPC déjà extrait</i>	Thermocycler used for QC <i>Thermocycleur utilisé pour le CQ</i>
4a - EPC ASFV	C _{t QC} ASFV = 25,4	ABI 7500 (Applied Biosystems)

IPC <i>IPC</i>	Extraction - <i>Extraction</i> : MagMAX™ CORE Nucleic Acid Purification Kit	Thermocycler used for QC <i>Thermocycleur utilisé pour le CQ</i>
5 - IPC ASFV	C _{t QC} IPC = 27,2	ABI 7500 (Applied Biosystems)

C_{t QC}: Indicative Ct of the Quality Control *C_t indicatif du Contrôle Qualité*

SUPPLEMENTAL INSTRUCTIONS - Available upon request to Eurotech@thermofisher.com

Notices d'utilisation complémentaires - Disponibles sur demande auprès de Eurotech@thermofisher.com

Amplification Protocol *Protocole d'amplification*

Nucleic acid purification Protocol

Protocole de purification d'acides nucléiques

Polish MAN0018346 RevB.0

N/A

Laboratoire Service International
6 allée des Ecureuils
Parc Tertiaire du Bois-Dieu
69380 Lissieu - France
www.thermofisher.com/support

Quality Assurance

Date: 12 December 2024