

EC Declaration of Conformity



Device Classification: Laboratory Equipment

Centurion Scientific Ltd.

Centrifuge ranges and accessories

PrO-Analytical.

C1015/R. C115/R. C106/R. C113. C104SO. C2000/R. C204/R. C206/R. C4000/R.
C404/R. C406/R. C5000/R. C504/R. C7000/R. C704/R. CFS4/R. CFS4A/R

Conform to the following Product specifications:

Directives

2006/42/EC

2014/35/EU

2014/30/EU

2014/65/EU

2014/65/EU

2011/65/EU

Standards

EN ISO 12100

EN61010-1. IEC 61010-2-020

EN 61326-1, EN55011. CFR 47 FCC Part 15 class a

EN 378-1

EN 378-2 For Refrigerated versions

EN 50581

A handwritten signature in black ink, appearing to read "K Cooper".

Certified by:

Company Representative
K Cooper QA manager

ISO 9001:2015

bsi.

ISO 13485:2016

CENTURION
SCIENTIFIC LIMITED

The Old Stables, Church Farm, Stoughton, Chichester, West Sussex. PO18 9JL. United Kingdom.