

A N A L Y S E N Z E R T I F I K A T
CERTIFICATE OF ANALYSIS

TE Puffer (100x)	Kat.Nr./Cat.No. : 42554
TE buffer (100x)	Lot/Contr.No. : P250242

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen appearance	klare farblose Lösung clear colorless solution	entspricht corresponds
pH	7.8 – 8.2	8.1
Zusammensetzung composition		
TRIS (# 37180)	1.0 M	entspricht corresponds
EDTA-Na ₂ ·2H ₂ O (# 11280)	0.1 M	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		28.07.2027
Lagerung (°C) storage		+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 29.07.2025

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.