



# DECLARATION OF CONFORMITY

ACCORDING TO In Vitro Diagnostic Medical Devices Regulation (EU) 2017/746

## EU Representative

SUNGO Cert Gmbh

Harffstr. 47, 40591 Düsseldorf, Germany

SRN: DE-AR-000010869

## Device Classification

**Classification:** Class A.

**Rule:** According to Rule 5, Annex VIII, of In Vitro Diagnostic Medical Devices Regulation (EU) 2017/746.

## Applicable Standards

EN ISO 20417: 2021,

EN ISO 15223-1:2016,

EN ISO 18113-1:2011,

EN ISO 14971:2019

## Remark

*The declaration of conformity is valid in connection with the release technical document CE/IVDR-UIP-05.*

*All the supporting documentation is retained at the premises of the manufacturer.*

*The Declaration of Conformity is exclusively under the sole responsibility of the manufacturer.*

## Manufacturer

**Name:** Shaoxing Shangyu Mingji Plastic Co., Ltd.

**Address:** Shengqiao, Xiaoyue, Shangyu, Shaoxing, Zhejiang, China 312367

## Product Information

**Name:** SEROLOGICAL PIPETTE

**Model:** MSP0101, MSP0201, MSP0501, MSP1001, MSP2501, MSP5001

**GMDN:** 43375

**Basic UDI-DI:**

**Classification:** Class A

## Conformity Assessment

Compliance of the designated product with the In Vitro Diagnostic Medical Devices Regulation (EU) 2017/746 has been assessed by issuing the EU declaration of conformity referred to in Article 17 after drawing up the technical documentation set out in Annexes II and III.

## Declaration

We herewith declare that the above-mentioned products meet the requirements of In Vitro Diagnostic Medical Devices Regulation (EU) 2017/746 and the applicable standards above.

Signature: *Song yiping* Date: 2021.8.19

Position: GM

Place: Shaoxing/China





