

EU Type Examination Certificate

This is to certify that:

WEINI TECHNOLOGY DEVELOPMENT CO., LTD
No2 Xingye Road, Xicheng Industrial Zone
Renhe Town, Baiyun District
Guangzhou
Guangdong
510470
China

Holds Certificate Number:

CE 709415

In respect of:

**Filtering half masks to EN 149:2001+A1:2009.
Various models - see continuation sheet for details.**


on the basis that BSI carried out the relevant Type Examination procedures under the requirements with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex V (Module B) and meets the relevant health and safety requirements specified in Annex II

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 2797):

Previous Notified Body: BSI 0086

First Issued: 2019-09-10

Latest Issue: 2019-10-24



Drs. Dave Hagenaaers, Managing Director

Effective Date: 2019-10-24

Expiry Date: 2024-09-10

Page: 1 of 12



...making excellence a habit.™

EU Type Examination Certificate

No. CE 709415

Product Specification – Flat Fold Particulate Half Masks

Technical specification: EN 149:2001+A1:2009 Respiratory protective devices – Filtering half masks to protect against particles
Classification – Single shift NR

Description: The particulate respirators are designed to protect against solid and non-volatile liquid particles. Half masks are vertical flat fold design, comprising of polypropylene nose bridge, a sponge nose pad and are held on the face by a pair of polyester elastic headbands. Some models incorporate a single, silicone exhalation valve, see details on models denoted by letter V.

FFP1 models comprise of the exact same filter media build up as FFP2 models. FFP3 filter media make up is different to those of FFP1/FFP2.

All models are single shift devices (denoted by the classification symbol NR) and have the Dolomite Clogging option (denoted by the classification symbol D). Model can also be supplied without the clogging option D.

FFP Class **FFP1** (without clogging option D) **or FFP1D** (with clogging option D)

Models: 951
951V (mask with exhalation valve)
951C (mask with activated charcoal)
951VC (mask with exhalation valve and activated charcoal)

FFP Class **FFP2** (without clogging option D) **or FFP2D** (with clogging option D)

Models: 952
952V (mask with exhalation valve fitted)
952C (mask with activated charcoal)
952VC (mask with exhalation valve and activated charcoal)

FFP Class **FFP3** (without clogging option D) **or FFP3D** (with clogging option D)

Models: 992V (mask with exhalation valve fitted)
992VC (mask with exhalation valve and activated charcoal)

First Issued: 2019-09-10

Latest Issue: 2019-10-24

Effective Date: 2019-10-24

Expiry Date: 2024-09-10

Page: 11 of 12

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

产品图片

Product picture



型号 : 952
model

执行标准 : CE/EN149:2001+A1:2009
Certificate

产品说明 : FFP2/Filter 0.3pm non-oil particular above 94%
specification
