

## CERTIFICATE OF ANALYSIS

### **B52 10X TBE Electrophoresis Buffer**

**Packaging Lot:** 3173443

**Expiry Date:** 29.02.2028 (DD.MM.YYYY)

**Storage:** at room temperature

**Note:** 10X TBE Electrophoresis buffer is manufactured in D class clean room facilities.

### **Filling lots for components in package:**

Lot	Quantity	Description
3172040	1 L	10X TBE Electrophoresis Buffer

### **QUALITY CONTROL**

Parameter	Method	Requirement	Result
pH	Determined according to Ph. Eur. 2.2.3	8.3±0.2	Conforms
Conductivity	Determined according to Ph. Eur. at 25 °C	3.63±0.2	Conforms
Density	Measured using a densitometer	1.057±0.002	Conforms
Refractive index	Measured using a refractometer at 20 °C	1.3567±0.0002	Conforms

### ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by QC: **J. Žilinskienė**

