

CERTIFICATE of Conformity



Certificate No.: MK 69265805 0001
Test Report No.: HU216S28 001
Certificate Holder: Snaigé AB
Pramones str. 6
62175 Alytus Lithuania

Product: Refrigerators for drugs

Identification: Model/Type Ref.: MD40DD

Rated voltage: 220-240 V;
Rated frequency: 50 Hz;
Rated current: 1,1 A;
Rated power: 180 W;
Refrigerant number: R600a;
Refrigerant charge: 65 g;
Protection against electric shock: Class I;
Climate class: 7

Tested according to: DIN 13277:2021-02

This certificate refers to the above mentioned product. This is to certify that the test sample is in conformity with the requirements stated above. This certificate does not imply assessment of the series-production of the product and does not permit the use of a TÜV Rheinland mark of conformity.

Certification Body

Date of Issue:
Budapest, 2021-11-30


Ferencne Schmidt

TÜV Rheinland InterCert Kft. – Product Certification Body — H-1143 Budapest, Gizella út 51-57. — www.tuv.com

Validity of certificate can be checked on www.CERTIPEDIA.com