

Manufacturer Declaration Letter

Date: 20-Mar-2025

To Whom It May Concern:

We hereby confirm following tech specifications of electric hospital bed model YA-D5-5 can be customized according to tender (code 140391) request.

- 1.Backboard X-ray transparency
- 2.Height adjustment range: Minimum range: 45 - 80 cm (Height range NOT include mattress height)
- 3.Wheels: Diameter: 15 cm
- 4.Back section frame for radiographic cassette/detector: 35 x 43 cm
- 5.Maximum load capacity: Labeled sticker on the headboard: 250 kg
- 6.Mattress: Minimum thickness: 14 cm
7. Trendelenburg: 0-16°

Signed & Stamp



Name: Eric CC

Title: Sales Manager

Dated: 20-Mar-2025