

TO WHOM IT MAY CONCERN,

We Kare Medical And Analytical Devices Limited Company (Kare Medikal ve Analitik Cihazlar Sanayi ve Ticaret Limited Şirketi) who are established and reputable manufacturers of medical devices located at Ziya Gokalp Cad. Kizilay No:36/23 Çankaya Ankara/TURKEY hereby declare that Hikoneb Oxybreath Mini 5 Oxygen Concentrator operating noise level is stated below. It is present:Audible alarm for Pressure below the norm and indication on the Monitor “ The Service Requirement” .

**Operating Noise Level: 48 dBA**

Yours Faithfully,

**Cağlar Yavas**

**Int. Trade Manager**