

Product No.	AM1005
Lot No.	2409241
Date of Manufacture	25SEP2024
Expiration Date	20SEP2025

TEST	SPECIFICATION	RESULT
Nuclease Tests - Enzyme		
RNase activity		Pass
Exonuclease activity		Pass
Nonspecific endonuclease activity		Pass
Functional Tests - Enzyme		
0 copies/reaction AIV input		Pass
40 copies/reaction AIV input	30.0 - 34.0 CT	Pass
80 copies/reaction AIV input	29.0 - 33.0 CT	Pass
160 copies/reaction AIV input	28.0 - 32.0 CT	Pass
320 copies/reaction AIV input	27.0 - 31.0 CT	Pass
1280 copies/reaction AIV input	25.5 - 29.5 CT	Pass
10240 copies/reaction AIV input	22.5 - 26.5 CT	Pass

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: Austin, TX, USA.

Corporate Legal Entity Life Technologies 5781 Van Allen Way Carlsbad, CA, USA 92008 www.thermofisher.com For inquiries, contact us at cofarequests@thermofisher.com



Product No.	AM1005		
Lot No.	2409241		
Date of Manufacture	25SEP2024		
Expiration Date	20SEP2025		
TEST		SPECIFICATION	RESULT
Xeno RNA		29.0 - 33.0 CT	Pass
Nuclease Tests - Buffer			
RNase activity			Pass
Exonuclease activity			Pass
Nonspecific endonuclease act	ivity		Pass
Functional Tests - Buffer			
0 copies/reaction AIV input			Pass
40 copies/reaction AIV input		30.0 - 34.0 CT	Pass
80 copies/reaction AIV input		29.0 - 33.0 CT	Pass
160 copies/reaction AIV inpu	t	28.0 - 32.0 CT	Pass
320 copies/reaction AIV inpu	t	27.0 - 31.0 CT	Pass
1280 copies/reaction AIV inp	out	25.5 - 29.5 CT	Pass

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: Austin, TX, USA.

Corporate Legal Entity Life Technologies 5781 Van Allen Way Carlsbad, CA, USA 92008 www.thermofisher.com For inquiries, contact us at cofarequests@thermofisher.com



Product No.	AM1005		
Lot No.	2409241		
Date of Manufacture	25SEP2024		
Expiration Date	20SEP2025		
TEST		SPECIFICATION	RESULT
10240 copies/reaction AIV inp	out	22.5 - 26.5 CT	Pass
Xeno RNA		29.0 - 33.0 CT	Pass
Nuclease Tests - Water			
RNase activity			Pass
Exonuclease activity			Pass
Nonspecific endonuclease activ	vity		Pass
Functional Tests - Water			
0 copies/reaction AIV input			Pass
40 copies/reaction AIV input		30 - 34 CT	Pass
80 copies/reaction AIV input		29 - 33 CT	Pass
160 copies/reaction AIV input		28 - 32 CT	Pass
320 copies/reaction AIV input		27 - 31 CT	Pass

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: Austin, TX, USA.

Corporate Legal Entity Life Technologies 5781 Van Allen Way Carlsbad, CA, USA 92008 www.thermofisher.com For inquiries, contact us at cofarequests@thermofisher.com



Product No.	AM1005		
Lot No.	2409241		
Date of Manufacture	25SEP2024		
Expiration Date	20SEP2025		
TEST		SPECIFICATION	RESULT
1280 copies/reaction AIV input	t	25.5 - 29.5 CT	Pass
10240 copies/reaction AIV inp	out	22.5 - 26.5 CT	Pass
Xeno RNA		29 - 33 CT	Pass
Functional Tests - Enhanced Testing			
500 copies/reaction of BVD R	NA	31.2 - 34.2 CT	Pass
50 copies/reaction of BVD RN	A	34.5 - 37.5 CT	Pass
No template control (NTC)		> = 40.0 CT	Pass

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: Austin, TX, USA.

Corporate Legal Entity Life Technologies 5781 Van Allen Way Carlsbad, CA, USA 92008 www.thermofisher.com For inquiries, contact us at cofarequests@thermofisher.com