

## CERTIFICATE OF ANALYSIS

**4408399 GeneScan™ - 600 LIZ™ Size Standard v2.0**

**Packaging Lot: 2954858**

**Expiry Date:** 19.09.2025 (DD.MM.YYYY)

**Storage:** at 5±3°C in the dark

**Note:** **For Research Use Only. Not for use in diagnostic procedures.**

To verify the conformity of the manufacturing process, kits were visually inspected for correct assembly and labeling.

### Filling lots for components in package:

Lot	Quantity	Description
2946132	2 x 200 µL	GeneScan™ - 600 LIZ™ Size Std. v2.0

### QUALITY CONTROL

Parameter	Method	Requirement	Result
Use test	Product is tested for the correct number of peaks and the absence of extra peaks on a multi-capillary Applied Biosystems Genetic Analyzer.	Correct peak number and absence of extra peaks	Pass
Functional test	Product is tested for overall average peak height on a multi-capillary Applied Biosystems Genetic Analyzer.	2500-14000 RFU	Pass

### ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by QC: **J. Žilinskienė**

