



Technical specification

Width	49.7 mm
Length	142 mm
Thickness	26 mm
Weight	65 g

Turbine



Reusable turbine
(code 910002)



Disposable turbine
(code 910004)

Supply voltage 5 V d.c. USB connection

Rated electrical power 0.25 W

Rated input current 50 mA max

Backup battery voltage none

Connectivity USB 2.0

Display none

Mouthpieces Ø 30 mm (1.18 inch)

IP protection level IPX1

Type of electrical protection Class II device

Safety level for shock hazard Type BF Apparatus

Conditions of use Apparatus for continuous use

Storage and Transportation conditions

Temperature: MIN -40 °C
MAX +70 °C
Humidity: MIN 10% RH;
MAX 95%RH

Operating conditions Temperature: MIN +10 °C
MAX +40 °C
Humidity: MIN 10% RH
MAX 95%RH

Memory capacity database PC software
PC software MIR Spiro

Spirometry

Flow sensor	bi-directional digital turbine
Volume range	10 L
Flow range	±16L/s
Volume accuracy (ATS 2019)	±2.5% or 50 mL
Flow accuracy	±5% or 200 mL/s
Dynamic resistance	< 0.5 cm H ₂ O/L/s
Temperature sensor	semiconductor (0-45°C)
Test available	FVC, VC, IVC, MVV, PRE-POST
Measured parameters	FVC, FEV1, FEV1/FVC%, FEV3, FEV3/FVC%, FEV6, FEV1/FEV6%, PEF, TPEF, FEF25, FEF50, FEF75, FEF2575, FEF7585, FET, FEV05, FEV05%, FEV075, FEV075%, FEV2, FEV2%, ELA, BEV, FIVC, FIV1, PIF, FIV1/FIVC%, PIF, IRV, VC, IVC, EVC, IC, EI, RFEV, ERV, FIVC, FIV1/FIVC, FEV1/VC%, FIF25, FIF50, FIF75, R50, VT, VE, RR, t _i , t _E , t _i /t _{TOT} , VT/t _i , MVV, MVV cal, MV

Certificates & Registrations

CE 0476	MDR 2017/745
FDA 510 (k)	K 122384
Health Canada	71191 (class II)
EMDN liv.4	Z121501
CND code	Z12150102
GMDN code	13680
Ministry of Health	2629712 /R (911006) 2621401 /R (911006T)
Applicable standards	IEC 60601-1:2005 + A1:2012 + A2:2020 IEC 60601-1-2:2014 + A1:2020 IEC 60601-1-6:2010 + A1:2013 + A2:2020 IEC 60601-1-9:2007 + A1:2013 IEC 62304:2006 + A1:2015 IEC 62366-1:2015 + A1:2020 ISO 15223-1:2021 ISO 14971:2019 ISO 10993-1:2018 EN ISO 10993-5:2009 ISO 10993-10:2021 ISO 10993-23:2021 ISO 20417:2021 ISO 2248:1985 ATS/ERS 2005, 2019 update ISO 23747:2015 EN ISO 26782:2009



Minispir



USB Spirometer
Real time tests directly on PC

Key features

Plug & Play

Device powered via integrated USB cable, without screen, without internal memory and with direct data recording on **MIR Spiro** software



Real-time tests

Real-time tests displayed on the PC screen via **MIR Spiro** software
Spirometry: FVC, VC, MVV, PRE/POST bronchodilator comparison

Pediatric incentive

Real-time animation for improved patient collaboration during the test

Integrated temperature sensor

Automatic BTPS Conversion

Predicted values

Wide selection of predicted values including GLI, ERS and others in PC mode

EMR/EHR connectivity

Integration via **MIR Spiro** software with EMR/EHR (in HL7, GDT, FHIR, EXCHANGE PROTOCOL)



Compatible turbines

		Mouthpiece	Turbine disinfection	Turbine calibration	Packaging	Antiviral filter
FlowMIR® disposable turbine		Disposable included	Not required	Not required	Individually packaged: packs of 60 pieces	Optional
Reusable turbine		Required, not included	Required	Required	Pack of 1 unit	Recommended by ATS

How to use

Minispir works when connected
to **PC via integrated USB cable**

MIR Spiro software

- \\ Comprehensive software for spirometry and oximetry
- \\ Designed to be integrated with EMR/EHR
- \\ Complies with the latest ATS/ERS guidelines
- \\ Available for desktop and laptop use
- \\ MacOS and Windows

All MIR professional devices
work with **MIR Spiro** software,
the latest generation software
for spirometry and oximetry.



Platinum Card

To subscribe to a Platinum
subscription plan
it is necessary to **have**
the MIR Spiro Platinum Card.

Measured parameters

	From MIR Spiro software via connection to the device
Spirometry	FVC, FEV1, PEF, FEF75, FEF25-75, FET, FEV1/FVC, FEV6, FEV1/FEV6, FEF25, FEF50, FIVC, FEV1/VC, ELA, MVV(cal), Time to PEF, FEV0.5, FEV0.5/FVC, FEV0.75, FEV0.75/FVC, FEF75-85, Extr. Vol, VC, EVC, IVC, IC, VC, ERV FEV3, FIV1, FIV1/FIVC, PIF, FEV3/FVC, PIF, FEV2, FEV2/FVC, FIF25, FIF50, FIF75, R50, FEV1/PEF (EI), FEV1/FEV0.5 (RFEV), TV, VE, RR, tI

Datasheet

Size	49.7 x 142 x 26 mm
Weight	65 g
Turbines	· Reusable Turbine (code 910002) · Disposable turbine (code 910004)
Supply voltage	5V USB connection
Nominal electrical power	0.25W
Nominal input current	50mA max
Backup battery voltage	absent
Connectivity	USB 2.0
Display	absent
Mouthpiece	Ø 30 mm (1.18 inches)
IP protection level	IPX1
Electrical protection	Class II device
Electrical safety level	Type BF device
Terms of use	Device for continuous use
Storage conditions	Temp: MIN -40°C, MAX+70°C Humidity: MIN 10% RH; MAX 95%RH
Terms of use	Temp: MIN +10°C, MAX +40°C Humidity: MIN 10% RH, MAX 95%RH
PC Software	MIR Spiro
Memory capacity	PC software database

Spirometry	
Flow sensor	two-way digital turbine
Flow Range	± 16 L/s
Volume accuracy (ATS 2019)	± 2.5% or 50 mL
Flow accuracy	± 5% or 200 mL/s
Dynamic resistance	< 0.5 cm H2O/L/s
Temperature sensor	semiconductor(0-45°C)
Available tests	FVC, VC, IVC, MVV, PRE-POST
Measured parameters	FVC, FEV1, FEV1/FVC%, FEV3, FEV3/FVC%, FEV6, FEV1/FEV6%, PEF, TPEF, FEF25, FEF50, FEF75, FEF2575, FEF7585, FET, FEV05, FEV05%, FEV075, FEV075%, FEV2, FEV2%, ELA, BEV, FIVC, FIV1, PIF, FIV1/FIVC%, PIF, IRV, VC, IVC, EVC, IC, EI, RFEV, ERV, FIVC, FIV1/FIVC, FEV1/VC%, FIF25, FIF50,FIF75, R50, VT, VE, RR, tI, tE, ti/tTOT, VT/tI, MVV, MVV cal, MV
Certificates and registrations	
CE 0476	MDR 2017/745
FDA 510 (k)	K 122384
Health Canada	71191 (Class II)
EMDN liv.4	Z121501
CND Code	Z12150102
GMDN Code	46906 (spiral)
Ministry of Health	2494682/R (911006I0) 2494688/R (911006I1)
Applicable regulations	Electrical Safety IEC 60601-1 Electro Magnetic Compatibility EN 60601-1-2 ISO 26782: 2009 ISO 23747: 2015 ATS/ERS:2005, 2019(update) IEC 60601-1-6:2010 IEC 60601-1-9:2007+AMD1:2013 IEC 62304:2006 + A1:2015 ISO 10993-1:2018

Compliance with guidelines and standards

Spirometry: ATS/ERS 2005 + update to 2019;
ISO 23747: 2015; ISO 26782: 2009

ITALY

MIR Medical
International Research
S.p.A.

Viale Luigi Schiavonetti,
270 00173, Rome

Tel. +39 06 22 754 777

Fax +39 06 22 754 785

mir@spirometry.com

spirometry.com

USA

MIR USA, Inc.
5462 S. Westridge Drive
New Berlin, WI 53151

Tel. +1 (262) 565-6797

Fax +1 (262) 364-2030

mirusa@spirometry.com

FRANCE

MIR Local Branch
Jardin des Entreprises, 290,
Chemin de Saint Dionisy
30980 LANGLADE

Tel. +33 (0)4 66 37 20 68

Fax +33 (0)4 84 25 14 32

mirfrance@spirometry.com

BRAZIL

MIR Local Branch
Rua Pinheiro Machado, 2659,
Sl.303, Caxias do Sul RS

Tel +55 5430253070

mirbrazil@spirometry.com

[in](#) [f](#) [@](#) [▶](#)