

<b>Declaration of Conformity</b>	
<b>Manufacturer</b>	<b>Endoaccess GmbH</b> Feldriethe 1 30826 Garbsen Germany
<b>Product</b>	<b>Reusable Biopsy Forceps</b>
<b>Product class &amp; rule according to annex IX of the MDD</b>	CE Class I, according to annex IX Class I, Rule 6 bullet 2, Reusable biopsy forceps is a surgically invasive device as well as reusable device for transitional application.
<b>Conformity Assessment route</b>	Annex VII MDD 93/42/EEC 2007/47/EC
<b>MDS/ MD/ GMDN/ UMDNS</b>	UMDNS Code: 11-775
<b>Notified Body</b>	<b>TÜV SÜD Product Service GmbH</b> Ridlerstraße 65 80339 München Germany CE
<b>EC Certificates</b>	-
<b>Start of CE-Marking</b>	-
<p><b>We herewith declare that all in Annex 1 Model List mentioned products meet the provisions of the council directive 93/42/EEC for medical devices. All supporting documentation is retained under the premisses of the manufacturer. We are exclusively responsible for the Declaration of Conformity.</b></p>	

Annex 1 Model List		Size			
EA-R18C10	Oval cups Biopsy Forceps	Ø 1.8*1000mm			
EA-R18C12	Oval cups Biopsy Forceps	Ø 1.8*1200mm			
EA-R18C15	Oval cups Biopsy Forceps	Ø 1.8*1500mm			
EA-R24C16	Oval cups Biopsy Forceps	Ø 2.4*1600mm			
EA-R24C17	Oval cups Biopsy Forceps	Ø 2.4*1700mm			
EA-R24C18	Oval cups Biopsy Forceps	Ø 2.4*1800mm			
EA-R24C21	Oval cups Biopsy Forceps	Ø 2.4*2100mm			
EA-R24C23	Oval cups Biopsy Forceps	Ø 2.4*2300mm			
EA-R18S10	Standard Biopsy Forceps autoclavable	Ø 1.8*1000mm			
EA-R18S12	Standard Biopsy Forceps autoclavable	Ø 1.8*1200mm			
EA-R18S15	Standard Biopsy Forceps autoclavable	Ø 1.8*1500mm			
EA-R18S16	Standard Biopsy Forceps autoclavable	Ø 1.8*1600mm			
EA-R24S16	Standard Biopsy Forceps autoclavable	Ø 2.4*1600mm			
EA-R24S17	Standard Biopsy Forceps autoclavable	Ø 2.4*1700mm			
EA-R24S18	Standard Biopsy Forceps autoclavable	Ø 2.4*1800mm			
EA-R24S21	Standard Biopsy Forceps autoclavable	Ø 2.4*2100mm			
EA-R24S23	Standard Biopsy Forceps autoclavable	Ø 2.4*2300mm			
EA-R18OW10	Oval cups with spike Biopsy Forceps	Ø 1.8*1000mm			
EA-R18OW12	Oval cups with spike Biopsy Forceps	Ø 1.8*1200mm			
EA-R18OW15	Oval cups with spike Biopsy Forceps	Ø 1.8*1500mm			
EA-R24OW16	Oval cups with spike Biopsy Forceps	Ø 2.4*1600mm			
EA-R24OW17	Oval cups with spike Biopsy Forceps	Ø 2.4*1700mm			
EA-R24OW18	Oval cups with spike Biopsy Forceps	Ø 2.4*1800mm			
EA-R24OW21	Oval cups with spike Biopsy Forceps	Ø 2.4*2100mm			
EA-R24OW23	Oval cups with spike Biopsy Forceps	Ø 2.4*2300mm			
EA-R18A10	Alligator cups Biopsy Forceps	Ø 1.8*1000mm			
EA-R18A12	Alligator cups Biopsy Forceps	Ø 1.8*1200mm			
EA-R18A15	Alligator cups Biopsy Forceps	Ø 1.8*1500mm			
EA-R24A16	Alligator cups Biopsy Forceps	Ø 2.4*1600mm			
EA-R24A17	Alligator cups Biopsy Forceps	Ø 2.4*1700mm			
EA-R24A18	Alligator cups Biopsy Forceps	Ø 2.4*1800mm			
EA-R24A21	Alligator cups Biopsy Forceps	Ø 2.4*2100mm			
EA-R24A23	Alligator cups Biopsy Forceps	Ø 2.4*2300mm			
EA-R18OP10	Alligator single moving cup Biopsy Forceps	Ø 1.8*1000mm			
EA-R18OP12	Alligator single moving cup Biopsy Forceps	Ø 1.8*1200mm			
EA-R18OP15	Alligator single moving cup Biopsy Forceps	Ø 1.8*1500mm			
EA-R24OP16	Alligator single moving cup Biopsy Forceps	Ø 2.4*1600mm			
EA-R24OP17	Alligator single moving cup Biopsy Forceps	Ø 2.4*1700mm			
EA-R24OP18	Alligator single moving cup Biopsy Forceps	Ø 2.4*1800mm			
EA-R24OP21	Alligator single moving cup Biopsy Forceps	Ø 2.4*2100mm			
EA-R24OP23	Alligator single moving cup Biopsy Forceps	Ø 2.4*2300mm			
EA-R18AW10	Alligator Jaws with spike Biopsy Forceps	Ø 1.8*1000mm			
EA-R18AW12	Alligator Jaws with spike Biopsy Forceps	Ø 1.8*1200mm			

[illegible]

Place, Date of Issue	Garbsen, 05.04.2018
Signature	 N. Kellermann QMR