

ANNEX I: Common list of rapid antigen tests¹⁰

As agreed by Member States on 23 July 2021

Disclaimer: This list was agreed by the HSC based on a proposal by the Technical Working Group on COVID-19 Diagnostic Tests. Experts participating in the Technical Working Group strongly recommend that use of rapid antigen tests is primarily intended for preliminary testing for SARS-CoV-2 infection in symptomatic patients, and note that rapid antigen tests should in particular be used in the specific contexts and circumstances referred to by the Commission Recommendation (EU) 2020/1743 of 18 November 2020 and the technical guidance by ECDC on 19 November 2020. The content of the common list is based on the clinical performance data and information that is available at this moment in time. The common list of rapid antigen tests does not include rapid antigen self-tests nor rapid antigen tests that are based on samples other than those collected from nasal, oropharyngeal or nasopharyngeal specimens. Updates to the common list are based on the criteria as described in Council Recommendation 2021/C 24/01 as well as the additional criteria and definitions agreed by the Technical Working Group on 29 June 2021. Discussions on criteria and definitions will continue during summer 2021, also taking into consideration the work carried out by the In Vitro Diagnostics Working Group of the Medical Device Coordination Group⁹ on guidance on the performance of COVID-19 tests in the context of CE-marking and common specifications under Article 9 of Regulation (EU) 2017/746.

| Manufacturer | RAT commercial name | CE marking | Clinical performance Data by manufacturer | Clinical performance Data used in MS | FINI evaluation studies | EU Member States using in practice | Other countries using in practice | Completed validation studies | MS currently validating | Device ID # in JRC database ¹¹ | Included in Common list of RATs as of: |
|--------------------------|---|------------|---|--|---|---|-----------------------------------|--|----------------------------|---|--|
| AAZ-LMB | COVID-VIRO® Rapid antigen test COVID-19 | Yes | 96.6% sensitivity 100% specificity | BE: 96.6% sensitivity, 100% specificity, NP swab FR: >95% sensitivity, 100% specificity SI: 96.6% sensitivity, 100% specificity, NP swab | | BE, FR, SI | CH | FR CH | | 1833 | 10 May 2021 |
| Abbott Rapid Diagnostics | Panbio™ COVID-19 Ag Rapid Test | Yes | 91.4% sensitivity 99.8% specificity NP swab (Ct ≤ 33) 98.1% sensitivity 99.8% specificity Nasal swab (Ct ≤ 33) | BE ¹⁰ : Small-scale head-to-head comparison of 5 RATs in Belgian hospital lab, Panbio overall sensitivity (Ct range 14,6 – 35,5): 45/57 samples (79%), Sensitivity for Ct≤25: 17/18 samples. Overall specificity 100%. DE: 91.4% sensitivity 99.8% specificity, NP swab; 98.1% sensitivity, 99.8 specificity, Nasal swab | DE (10 Dec 2020) 1108 samples, NP swab Clinical sensitivities: - Days ≤ 7: 90.8%; - Ct ≤ 33: 88.3%; - Ct ≤ 25: 95.8%; Clinical specificity: 99.9% | AT, BE, BG, CY, CZ, DE ¹² , DK, EE, EL, ES, FR ¹³ , HR, IT, LT, LV, MT, NL ¹⁵ , PL, PT, RO, SE, SK | CH, ME, MK, NO, UK, UA | DE ¹² , ES, FI, NL ¹⁵ , PT, CH, NO | CY, ES, HR, HU, IE, LU, SE | 1232 | 17 February 2021 |

¹⁰ This is the list of rapid antigen tests as referred to in Article 3 of the Regulation (EU) 2021/953 of the European Parliament and of the Council of 14 June 2021 on a framework for the issuance, verification and acceptance of interoperable COVID-19 vaccination, test and recovery certificates (EU Digital COVID Certificate) to facilitate free movement during the COVID-19 pandemic, OJ L 211, 15.6.2021, p. 1–22.

¹¹ See: <https://covid-19-diagnostics.jrc.ec.europa.eu/>.



[Handwritten signature]

| Manufacturer | RAT commercial name | CE marking | Clinical performance Data by manufacturer | Clinical performance Data used in MS | FIND evaluation studies | EU Member States using in practice | Other countries using in practice | Completed validation studies | MS currently validating | Device ID # in JRC database ¹¹ | Included in Common list of RATs as of: |
|--|---|------------|---|---|-------------------------|------------------------------------|-----------------------------------|------------------------------|-------------------------|---|--|
| BIOTEKE CORPORATION (WUXI) CO., LTD | SARS-CoV-2 Antigen Test Kit (colloidal gold method) | Yes | 96.49 % sensitivity 99.28 % specificity OP/NP swab | DE: Positive evaluation by Paul-Ehrlich-Institut (sensitivity of 95% at <Ct25) + Manufacturer specificity: 99.289% | | DE ⁽²⁾ | | DE ⁽²⁾ | | 2067 | 14 July 2021 |
| Biological Health S.L.U./BIOTICAL HEALTH S.L.U | biological SARS-CoV-2 Ag Card | Yes | Sensitivity: 96%, Specificity: 99% NP swab | BE: Validation study 1: sensitivity 91.7% for Ct<25; Validation study 2: 94% for Ct<25. Manufacturer specificity: 99% | | BE | | BE | | Yes (2013) | 23 July 2021 |
| Boditech Med Inc | AFIAS COVID-19 Ag | Yes | Sensitivity: 91.7%, Specificity: 98.7% NP swab | NL: Independent field study in mild symptomatic (n= 427): overall sensitivity: 81.1% (106 PCR+), Ct <30: 96.4% (85 PCR+), PCR on NP+OP, Target antigen = nucleoprotein | | FR, NL | | NL | | Yes (1989) | 23 July 2021 |
| BTNX Inc | Rapid Response COVID-19 Antigen Rapid Test | Yes | 90.2% sensitivity 100% specificity NP swab, NP swab, OP swab | DE: 94.55% sensitivity, 100% specificity | | AT, DE ⁽²⁾ , ES, SI | | DE ⁽²⁾ | | 1236 | 10 May 2021 |
| CerTest Biotec | CerTest SARS-CoV-2 Card test | Yes | 92.9% sensitivity 99.6% specificity NP swab | ES: Ct ≤ 25, sensitivity: 94.0%; sensitivity for samples within the first 5 days after symptom onset: 84.8% | | ES, PT, SI | | DE ⁽²⁾ , ES | | 1173 | 17 February 2021 |
| Core Technology Co., Ltd | Coretests COVID-19 Ag Test | Yes | 98.1% sensitivity 99.6% specificity NP swab | DE: 98.1% sensitivity, 99.6% specificity | | AT, DE ⁽²⁾ , RO | | DE ⁽²⁾ | | 1919 | 10 May 2021 |
| CTK Biotech, Inc | OnSite COVID-19 Ag Rapid Test | Yes | Clinical Sensitivity: 92.3 % Clinical Specificity: 100 % Nasal, NP swab | ES: 219 samples: Nasal swab - Clinical sensitivity 86% (90%: Ct <30) Specificity: 100% (Method B) DK: 107 samples: Nasal swab - clinical sensitivity 86%; (from asymptomatic and mild symptomatic individuals), Clinical specificity: 100% | To start | DK | | DK, ES | | 1581 | 7 July 2021 |
| DDS DIAGNOSTIC | Test Rapid Covid-19 Antigen (tampon nazofaringian) | Yes | 98.77% sensitivity 99.03% specificity Nasal swab | RO: Meets the minimum performance requirements. | | RO | | RO China | RO | 1225 | 10 May 2021 |
| DIALAB GmbH | DIAQUICK COVID-19 Ag Cassette | Yes | | BE: ZZ04Q1CE: 93.2% sensitivity, 100% specificity, NP swab | | AT, BE, DE ⁽²⁾ | | DE ⁽²⁾ | | 1375 | 10 May 2021 |



Handwritten signature in blue ink.