

EC DECLARATION OF CONFORMITY

We, **Air Liquide Medical Systems S.A.**
 Parc de haute technologie
 6, rue Georges Besse
 92182 ANTONY Cedex
 FRANCE

declare under our sole responsibility that the products range:

MONNAL T75,

blower lung ventilators for intensive care, and their accessories, comply with the requirements of the European Directive 93/42/EEC concerning medical devices, with all amendments and transposing legislation, and all others Union legislation applicable to the device.

Notified Body: **GMED (0459)**
 1, rue Gaston Boissier - 75724 PARIS Cedex 15 – France

Annex II conformity certificate, reference n° GMED-7405.

Classification: **class IIb**, under rule 9 as specified in annex IX of the Directive 93/42/EEC.

Reference of technical file: **DT_2008_MONNAL_T75**

Designation	Reference	Date of first issue
MONNAL T75	KB022600	28/07/2008
MONNAL T75 CO2	KB033600	23/08/2013

Associated detachable parts:

Designation	Reference
SINGLE USE EXPIRATORY VALVE FOR MONNAL EVA	KY694600
SINGLE USE EXPIRATORY VALVE FOR MONNAL EVA (x 5)	KY694800
SINGLE USE EXPIRATORY VALVE FOR MONNAL EVA (x 20)	KY694900
AUTOCLAVABLE EXPIRATORY VALVE FOR MONNAL EVA	KY694500
SINGLE USE EXPIRATORY FLOW SENSOR FOR MONNAL EVA	KY632500
SINGLE USE EXPIRATORY FLOW SENSOR FOR MONNAL EVA (x 5)	KY664500
SINGLE USE EXPIRATORY FLOW SENSOR FOR MONNAL EVA (x 20)	KY664600
AUTOCLAVABLE EXPIRATORY FLOW SENSOR FOR MONNAL EVA	KY632200

This EC Declaration of Conformity is valid for the placing on the market or the putting into service in the countries within the European Union and the European Economic Area where the following languages are authorized by the national competent authority:

CS čeština	DE deutsch	EN english	ES español
FR français	IT italiano	PL polski	PT português
RO română			

Antony, *February 16, 2022*

Olivier VAROMME
 Quality and Regulatory Affairs Director

