



Akkreditierte
Prüfstelle nach
EN ISO 17025



HygCen Austria GmbH | Werksgelände 28 | 5500 Bischofshofen

Certificate

FEDERAL STATE UNITARY ENTERPRISE

“MOSCOW ENDOCRINE PLANT“

23Novokhokhlovskaya Str.

Moscow

Russia, 109052

Endopharm Single use medical three-layer mask, type IIR

had been **tested successfully** according to

EN 14683:2019+AC:2019

„Medical face masks - Requirements and test methods “

inclusive appendix B, appendix C and ISO 22609

The conformity with these requirements

is confirmed by the test report numbers: B 25843, B 26044

Identification Number: 2025843/160

Bischofshofen, 2020-11-30

Prof. Dr. med. H.-P. Werner
Managing Director
Medical Specialist for Hygiene & Microbiology