



# TEST REPORT

3F, Building Block 2, No. 3400 Gonghexin Road,  
 Jing'an District - Shanghai 200436, P.R. CHINA  
 上海市静安区共和新路3400号2幢3层  
 Tél. : +86 21 68 55 50 32  
 Fax : +86 21 68 55 50 33  
 E-mail : [ctcshanghai@ctcgroupe.com](mailto:ctcshanghai@ctcgroupe.com)

Report No.: S200909070\_2

28 October 2020

naomizhou@lynmed.com

**APPLICANT:** LYNCMED MEDICAL TECHNOLOGY (BEIJING) CO., LTD (C41517)

ROOM1601, BUILDING NO. 2, ZHUBANG 2000  
 BUSINESS BUILDING, BALIZHUANG XILI99  
 CHAOYANG DISTRICT  
 100022 BEIJING  
 CHINA

Date of receipt : 07 Sept. 2020  
 Testing period : 10 Sept. 2020  
 : 11 Sept. 2020

**Buyer:** ---

**Sample description:** Disposable latex gloves

Test(s) requested : ---  
 Service : REGULAR  
 Brand / Section : ---  
 Season : ---  
 End use : ---  
 Factory name : ---  
 Factory code : ---

**For CE Marking : Yes**  
 Previous report : ---  
 Product category : ---  
 Product type : ---  
 Test stage : FIRST TEST  
 Supplier name : ---  
 Exported to : ---

## 1. Conclusion:

	Tests description	Conformity
	<b>EN ISO 374-1</b>	
1	Air Leak Test	Pass
2	Water Leak Test	Pass

Pass: requirements met    Fail: requirements not met    None: no requirement for this test    N/A: not applicable

Approved by

Henry YAN  
 Laboratory Manager

This test report in version S200909070\_2 supersedes the report S200909070\_1

The report is issued by CTC Shanghai under its General Conditions printed overleaf. The results shown in this report refer only to the sample(s) tested. Except by special arrangement, the test items will not be retained by CTC Shanghai for more than 3 months. The test report shall not be reproduced, except in full, without the written approval of the testing laboratory.

To declare the conformity to the requirement, the uncertainty of measurement, associated to the test results, has not been taken into account.



Report No.: S200909070\_2

28 October 2020

APPLICANT: LYNCMED MEDICAL TECHNOLOGY (BEIJING)  
CO., LTD (C41517)

**2. Sample(s) description assigned by laboratory:**

<u>Size</u>	<u>Analyzed product</u>	<u>Description</u>	<u>Sample information</u>
	GLOVE	Whole glove	



200909070

This test report in version S200909070\_2 supersedes the report S200909070\_1

The report is issued by CTC Shanghai under its General Conditions printed overleaf. The results shown in this report refer only to the sample(s) tested. Except by special arrangement, the test items will not be retained by CTC Shanghai for more than 3 months. The test report shall not be reproduced, except in full, without the written approval of the testing laboratory.

To declare the conformity to the requirement, the uncertainty of measurement, associated to the test results, has not been taken into account.





Report No.: S200909070\_2

28 October 2020

APPLICANT: LYNCMED MEDICAL TECHNOLOGY (BEIJING) CO., LTD (C41517)

### 3. GLOVE/

Whole glove

	Method	Client Requirement	Unit	Result	Conformity
<b>5.2. Air Leak Test</b>  Glove thickness Air pressure used to test Result	EN 374-2:2019	No air bubbles	mm kPa	0.10 3.5 No air bubbles	Pass
<b>5.2. Water Leak Test</b>  Result	EN 374-2:2019	No water leak		No water leak	Pass

### END OF TEST REPORT

: The test was carried out by external accredited laboratory, not within their accreditation scope.

This test report in version S200909070\_2 supersedes the report S200909070\_1

The report is issued by CTC Shanghai under its General Conditions printed overleaf. The results shown in this report refer only to the sample(s) tested. Except by special arrangement, the test items will not be retained by CTC Shanghai for more than 3 months. The test report shall not be reproduced, except in full, without the written approval of the testing laboratory.

To declare the conformity to the requirement, the uncertainty of measurement, associated to the test results, has not been taken into account.

