



Bioevokepeak Co., Ltd.

RM.2014,BLDG.3,LIGAOGUOJIHUAUYUAN,NO.1222,WEST AOTI
ROAD,LIXIA DISTRICT,JINAN,SHANDONG. P.C.: 250100
TEL:+86-531-88982330 FAX:+86-531-88983691
EMAIL: info@bioevokepeak.com WEBSITE: www.bioevokepeak.com

DECLARATION OF CONFORMITY

Technical file of the company mentioned below has been inspected and audit has been completed successfully 2014/35/EU Low Voltage Directive and 2014/30/EU Electromagnetic Compatibility Directive have been taken as reference for these processes

Company Name: Bioevokepeak Co., Ltd.

Brand: BIOEVOPEAK

Related Directives and Annex: 2014/35/EU Low Voltage Directive (LVD)
2014/30/EU Electromagnetic Compatibility (EMC)

Related Standards: EN 61326-1:2013; EN 61010-1:2010

Product(s): Autoclaves & Sterilizers

Type(s)/Model(s): STB-B23Z;STB-B45Z;STV-APV75;STB-B29-3A

Parameters: 220V,50 Hz

Classification: Laboratory Equipment

Examination Period: May 24, 2023

Date of Expiry: May 23, 2028

Review Result: We, Bioevokepeak Co., Ltd.declare that during the self-testing and performance evaluation, no Non-compliance according to the requirements of the Low Voltage Directive 2014/35/EC and Electromagnetic Compatibility Directive 2014/30/EU was detected.

Year of DOC marking: 2023
Signed for and on behalf of

Company: Bioevokepeak Co., Ltd.

General Manager: 

Document No: BEPSD-230524002

