



# LumiQuickDiagnostics, Inc

2946ScottBlvd.

SantaClara,CA95054

Tel.408-855-0061,Fax408-855-0063

## CERTIFICATE OF ANALYSIS

### Product Information

PartNumber: 71110B  
ProductName: COVID-19 Antigen Test Strip  
LotNumber: 20091489  
  
Quantity: 1430E.A  
Exp.Date: 2022-03  
IntegrityTest: Pass


### Buffer Bottle:

PartNumber: 71110B-XB  
ProductName: COVID-19 Extraction Buffer  
LotNumber: 20082160  
Quantity: 140E.A  
Exp.Date: 2022-02  
IntegrityTest: Pass

### Assay Results

Controls	No. Tested	No. Passed	No. Failed
Negative	5	5	0
Low Positive	5	5	0
High Positive	5	5	0

The lot (Check one)  passed  fails QC test

Authorized signature:   
Stan Tse, QA/QC Manager

Date: 2020-09-16