

Technical Data Sheets

Product Name: : Biohazard bag

Product Name:		Biohazard bag
Catalogue no		BHB-25190
Specification		700x1000mm \pm 10mm, 50 μ m
Material		PE
Appearance Quality		The surface is basically flat, no wrinkles, no stains, no impurities, no scratches, no bubbles, no pinholes
Tensile strength	Vertical/Mpa	≥ 20
	Horizontal/Mpa	≥ 20
Elongation at break	Vertical/%	≥ 250
	Horizontal/%	≥ 250
Leakage		No leakage
Heat seal strength/N/15mm		≥ 10
Package		The identification on the packaging box should meet the requirements, whether the printing is clear, whether the shipping mark is accurate, and whether the product quantity is correct.

浙江理能医疗器械有限公司
ZHEJIANG RENON MEDICAL INSTRUMENT CO.,LTD



EC DECLARATION OF CONFORMITY

Name of the manufacturer: **Zhejiang Renon Medical Instrument Co.,Ltd.**
Address: **No.777 East HuanZhan Road, Xizi Building B-614, Hangzhou
City, Zhejiang Province, China**

We declare under our sole responsibility that the medical device:
Biohazard bag
of class: **I, rule 1. according to annex IX of directive 93/42/EEC**

meets the provisions of the directive 93/42/EEC Annex VII and its
transpositions in national laws which apply to it. The declaration is valid
in connection with the "final inspection report" of the device.

Conformity assessment procedure: 93/42/EEC Annex VII

Place and date:
Hangzhou 2023-01-09

浙江理能医疗器械有限公司
ZHEJIANG RENON MEDICAL INSTRUMENT CO.,LTD
General manager:
Ye Yun Qin

