

**Detalii specifice și standarde tehnice**

**LOT: „Tub radiologic cu carcasa ACJB.38639.017.000.000-07”:**

<b>I.</b>	<b>Specificatii tehnice solicitate</b>		<b>Specificatii tehnice oferite</b>
	<b>Tub radiologic (X-ray Tube)</b>	Tub radiologic cu carcasa ACJB.38639.017.000.000-07 compatibil cu sistemul radiographic mobil К Ф П-Ц-ПП, inclusiv dispozitiv de stocare de energie IKE-57/420	Tub radiologic cu carcasa ACJB.38639.017.000.000-07 compatibil cu sistemul radiographic mobil К Ф П-Ц-ПП, inclusiv dispozitiv de stocare de energie IKE-57/420
	<b>Model</b>		ACJB.38639.017.000.000-07
	<b>Producator</b>		АО "ПЕНТГЕНПРОМ"
	<b>Cod produs</b>		8532100000 9022300000
<b>II.</b>	<b>Main specifications</b>		
	Pentru focalizare mare, mm	1,2	1,2
	Tensiune tub Ua, kV	40÷150	40÷150
	Tensiunea maximă a filamentului Uh, V	20	20
	Curentul tubului Ia, mA	280	280
	Pentru focalizare mică, mm	0.3	0.3
	Tensiune tub Ua, kV	40÷150	40÷150
	Curentul tubului Ia, mA	40	40
	Tip carcasă	RID	RID
	Ochelari cu carcasă	IEC	IEC
	Filtru emițător de raze X propriu, Al mm	1.5	1.5
	Compatibilitate cu dispozitivul KFPT-urile Promatrix	se conformează	Compatibil cu КФПЦ-Р-АМИКО (Proscan)

<b>III.</b>	<b>Documentation</b>		
	Manufacturer's certificate	<i>ISO 9001:2015</i> <i>ISO 13485:2016</i>	<i>ISO</i>
	Quality and safety standards	<i>EU Declaration of Conformity (DoC): Declaration that each product complies with EU requirements.</i>	<i>EC Certificate</i>
<b>IV.</b>	<b>Operation, maintenance and installation</b>	<i>At least one set of operation, maintenance and service manuals for each equipment.</i>  <i>The manuals to include instructions for:</i> <ul style="list-style-type: none"> <li>• <i>setting up the equipment</i></li> <li>• <i>routine cleaning and maintenance</i></li> <li>• <i>planning for periodic maintenance.</i></li> </ul>	<b>Operation, maintenance and installation included</b>
	Language of documentation	<i>Written in United Nations languages: at least in English.</i>	Language of documentation: English
	Installation and maintenance	The bidder must arrange for the equipment to be installed by certified or qualified personnel.  Any prerequisites for installation will be communicated to the beneficiary in advance, described in detail in written form.	The Installation and maintenance will be provided by the supplier
	Training	The bidder will provide user training (including how to use and maintain the equipment).	The Training will be provided by the supplier
	Standard maintenance tools	<i>All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication material, to be included in the offer. Bidders to specify the quantity of every item</i>	<i>All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication material, to be included in the offer</i>

		<i>included in their offer (including items not specified above).</i>	
	Warranty:	<i>At least 2 years.</i>	<i>6 moths</i>
<b>V.</b>	<b>Conditii de mediu</b>		
	TEMPERATURA DE OPERARE	= +10°C până la +35°C	= +10°C până la +35°C
	TEMPERATURA DE DEPOZITARE	= -10°C până la +55°C	= -10°C până la +55°C
	UMIDITATE RELATIVA Depozitare	= 10% până la 95%; Funcționare 30% până la 90%	= 10% până la 95%; Funcționare 30% până la 90%
	PRESIUNEA ATMOSFERICĂ	= Stocare 700 hPa până la 1060 hPa; Funcționare de la 800 hPa până la 1060 hPa	= Stocare 700 hPa până la 1060 hPa; Funcționare de la 800 hPa până la 1060 hPa

**Toate bunurile vor fi testate de beneficiar. Acceptarea bunurilor și semnarea actelor de recepție va avea loc după testarea pozitivă efectuată de beneficiar.**

În calitate de ofertant, confirmăm prin semnătura noastră de mai jos, că ne asumăm responsabilitatea pentru informațiile prezentate în cadrul acestei achiziții.

NUMELE FURNIZORULUI VIPROMED SERVICE SRL

Semnătura și ștampila autorizate \_\_\_\_\_

Locul: mun.Chisinau

Data:11.02.2024