

SONIC SHOT 7 Sharpened Functionality

Each performance specification was honed to perfection, for safe and reliable contrast delivery.



Ultrasonic Motor

Completely non-magnetic design. Exclusive for MR, long-lifetime and trusting performance.

Powerhead Lamps

Quickly & Easily recognize injector operation.

Plunger Detection

Insert the syringe, and the injector will automatically detect the plunger position. No need to manipulate manually.



One-Touch Syringe Adapter

Exclusive adapters designed for pre-filled syringes. Easily inserted and removed with One-Touch design.



Operational Buttons


Powerhead buttons are designed for easy identification and ease of operation.



Powerhead


Completely non-magnetic structure, new features, and enhanced ease of use.

Body Weight Input Mode




Injector will automatically calculate the protocol based only on patient weight.

Flow Rate Mode




Manually set desired injection parameters.

Drip Infusion



Simplified Drip Mode, easy access and easy to setup and use.

Injection



Real-time display of injection progress.

Scan Timing



Real-time information, Elapsed Time bargraph and audible sound indicate when to scan.



Console

Design for improved reliability and safety with Nemoto's advanced Body Weight Protocol.

REGISTRUL DE STAT AL DISPOZITIVELOR MEDICALE

Введите текст для поиска...										
Nr	Denumire	Den.comerc.	Model	Nr. catalog	Tara	Producatorul	Reprezentant	Ordin	Data	Cod v
						NEMOTO				
DM000422607	SISTEM DE INJECTARE A MEDIULUI DE CONTRAST ANGIOGRAFIC		CT CONTRAST, SMART SHOT ALPHA		Japonia	NEMOTO KYORINDO CO., LTD.	INTERMED S.R.L.	Rg04-000012	18-01-2023	
DM000422608	SISTEM DE INJECTARE A MEDIULUI DE CONTRAST ANGIOGRAFIC		PRESS DUO ELITE		Japonia	NEMOTO KYORINDO CO., LTD.	INTERMED S.R.L.	Rg04-000012	18-01-2023	
DM000422600	SISTEM DE INJECTARE A MEDIULUI DE CONTRAST ANGIOGRAFIC		CT CONTRAST, DUAL SHOT ALPHA7		Japonia	NEMOTO KYORINDO CO., LTD.	INTERMED S.R.L.	Rg04-000012	18-01-2023	
DM000422609	SISTEM DE INJECTARE A MEDIULUI DE CONTRAST ANGIOGRAFIC		MR CONTRAST, SONIC SHOT 7		Japonia	NEMOTO KYORINDO CO., LTD.	INTERMED S.R.L.	Rg04-000012	18-01-2023	

EC Certificate
Directive 93/42/EEC Annex II, excluding Section 4
Full Quality Assurance System
Medical Devices

Registration No.: HD 60134044 0001

Report No.: 12022660 007

Manufacturer: Nemoto Kyorindo Co., Ltd.
2-27-20 Hongo, Bunkyo-ku
Tokyo 113-0033
Japan

Products: CT Contrast Delivery Systems, MR Contrast Delivery Systems,
Angiography Contrast Delivery Systems, Extravasation Sensors

(see attachment for sites included)

Replaces Approval, Registration No.: HD 60103647 0001

Expiry Date: 2023-11-07

The Notified Body hereby declares that the requirements of Annex II, excluding section 4 of the directive 93/42/EEC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex II, section 5 of the aforementioned directive. For placing on the market of class III devices covered by this certificate an EC design-examination certificate according to Annex II, section 4 is required.

Effective Date: 2018-11-08

Date: 2018-11-07



Notified Body

M. Aihara
M.Sc. M. Aihara

TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg
TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC
concerning medical devices with the identification number 0197.

TÜV Rheinland
LGA Products GmbH
Tillystraße 2, 90431 Nürnberg

**Attachment to
Certificate**

Registration No.: HD 60134044 0001
Report No.: 12022660 007

Manufacturer: Nemoto Kyorindo Co., Ltd.
2-27-20 Hongo, Bunkyo-ku
Tokyo 113-0033
Japan

Sites included :

Nemoto Kyorindo Co., Ltd.

Test site

1-4-18 Asahi, Kawaguchi-shi, Saitama 332-0001, Japan

Manufacture

Nemoto Kyorindo Co., Ltd.

Kawaguchi Plant

2-12-23 Aoki, Kawaguchi-shi, Saitama 332-0031, Japan

Manufacture

Nemoto Kyorindo Co., Ltd.

Technical Center

2-12-4 Aoki, Kawaguchi-shi, Saitama 332-0031, Japan

Design/ development

Nemoto Kyorindo Co., Ltd.

Kawaguchi Warehouse

1-7-5 Asahi, Kawaguchi-shi, Saitama 332-0001, Japan

Manufacture

Date: 2018-11-07



Notified Body

M. Aihara
M.Sc. M. Aihara

Nemoto Kyorindo Co., Ltd.
2-27-20 Hongo,
Bunkyo-ku,
Tokyo
113-0033
Japan
2023-08-25

Notified Body Confirmation Letter
Reference: EU2023-607/642255

To whom it may concern,

Confirmation of the status of a formal application, written agreement, and appropriate surveillance in the framework of Regulation (EU) 2023/607 amending Regulations (EU) 2017/745 and (EU) 2017/746 as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices

This letter confirms that, **BSI Group The Netherlands B.V.**, a Notified Body (NB) designated against Regulation (EU) 2017/745 (MDR) and identified by the number **2797** on NANDO, has received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of MDR and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of MDR with the following manufacturer:

Nemoto Kyorindo Co., Ltd.
2-27-20 Hongo,
Bunkyo-ku,
Tokyo
113-0033
Japan

SRN Number: JP-MF-000004827

The devices covered by the formal application and the written agreement mentioned above are identified in the Tables below. Table 1 identifies the devices for which an MDR application has been received, written agreement concluded and for which the NB is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive. Table 2 identifies the devices for which an MDR application has been received and a written agreement concluded, but the NB has not yet taken the responsibility for appropriate surveillance of the corresponding devices under the applicable Directive.

In the case of devices covered by certificates issued under Directive 90/385/EEC (AIMDD) or Directive 93/42/EEC (MDD) that expired after 26 May 2021 and before 20 March 2023, without having been withdrawn, this letter also confirms that the manufacturer signed the written agreement under MDR by the date of MDD/AIMDD certificate expiry; or provided evidence that a competent authority of a Member State had granted a derogation or exemption from the applicable conformity assessment procedure in accordance with Article 59(1) of MDR or Article 97(1) of the MDR respectively, by the 20 Mar 2023 for the relevant devices.

The transition timelines that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 120.3c of MDR (as amended by (EU) 2023/607), are shown below:

- 26 May 2026 for Class III custom-made implantable devices
- 31 December 2027 for Class III devices and Class IIb implantable devices excluding Well-established technologies (WET - sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors)
- 31 December 2028 for other Class IIb devices, Class IIa, Class I devices placed on the market in sterile condition or have a measuring function
- 31 December 2028 for devices not requiring the involvement of a notified body under MDD but requiring it under MDR (e.g., class I devices that qualify as re-usable surgical instruments)

On behalf of BSI Group The Netherlands B.V.,

Takato Akimoto
BSI Scheme Manager

Table 1: Devices covered by this letter and for which the NB is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive:

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Not applicable			

Table 2: Devices covered by this letter and for which the NB is NOT responsible for appropriate surveillance of the corresponding devices under the applicable Directive:

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
CT Contrast Delivery Systems DUAL SHOT alpha7 456013848170AM	Class IIb excluding Class IIb implantable non-WET	N/A	MDD Certificate number: HD 60134044 0001 NB Number: 0197
CT Contrast Delivery Systems SmartShot alpha 456013848200A5	Class IIb excluding Class IIb implantable non-WET	N/A	MDD Certificate number: HD 60134044 0001 NB Number: 0197
MR Contrast Delivery Systems SONIC SHOT 7 456013848210A8	Class IIb excluding Class IIb implantable non-WET	N/A	MDD Certificate number: HD 60134044 0001 NB Number: 0197
Angiography Contrast Delivery Systems PRESS DUO elite 456013848260AP	Class IIb excluding Class IIb implantable non-WET	N/A	MDD Certificate number: HD 60134044 0001 NB Number: 0197
Extravasation Sensors LD 456013848130A9	Class IIb excluding Class IIb implantable non-WET	N/A	MDD Certificate number: HD 60134044 0001 NB Number: 0197

Confirmation Letter Revision History

Date	Action
2023/08/25	Initial issue

MB2791

Declaration of Conformity

Manufacturer: **Nemoto Kyorindo Co., Ltd.**

2-27-20 Hongo, Bunkyo-ku, Tokyo 113-0033, Japan

Authorized European Representative:

Medicor International NV

Wingepark 5B-101 3110 Rotselaar Belgium

Product: **MR Contrast Delivery Systems SONIC SHOT 7**

Valid from: Starting from serial number EMB30078G, and in conjunction with the release documents for the product.

We herewith declare that the above mentioned device meets all applicable provisions of the EC Directive 93/42/EEC.

The following Standards were applied:

EN ISO 13485:2016	Medical devices- Quality management systems - Requirements for regulatory purposes
EN ISO 14971:2012	Medical devices-Application of risk management to medical devices
EN 60601-1:2006+A1:2013	Medical electrical equipment-Part 1: General requirements for basic safety and essential performance
EN 60601-1-2:2015	Medical electrical equipment-Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic disturbances - Requirements and tests
EN 60601-1-6:2010	Medical electrical equipment-Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability
EN 62304:2006	Medical device software – Software life-cycle processes
EN 62366:2008	Medical devices – Application of usability engineering to medical devices
EN 1041:2008	Information supplied by the manufacturer with medical devices
EN ISO 15223-1:2016	Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements

Conformity Assessment Procedure:

Annex II (exclude Sec. 4) of the EC Directive 93/42/EEC.

Notified Body: **TÜV Rheinland LGA Products GmbH CE 0197**

Registration No.: **HD 60134044 0001**

Classification: **IIb according to Annex IX of MDD, Rule 11, Sub-clause 1, Indent 1**



Place: **Tokyo, Japan**

Date: **February 18, 2019**

Makoto Yasuda

Makoto Yasuda
General Manager