

JOB REF NO.: 2021-05-21-009 **DATE RECEIVED:** MAY 21, 2021 DATE REPORTED: JUNE 04, 2021

PAGE: 1 of 1

: CPSA/210663485-CA64950 **Test Report No** : ASAP International Sdn. Bhd. **Company**

No.1, Jalan Sitar 33/6, Seksyen 33,

40400 Shah Alam, Selangor,

Malavsia

TEST REPORT

Sample Description Powder Free Nitrile Examination Gloves,

Device Basic UDI-DI: 955589090ASAPPFN4L

Brand Name ASAP Country of Origin Malaysia All countries Country of Destination

Size

Quantity Tested 5 pieces

Test Conducted Powder Content Test Method EN455 Part 3:2015

21 May 2021 – 04 June 2021 **Testing Period**

On testing the samples, the following results were obtained:-

SIZE Average Powder Mass per Glove

0.08 mg M

SIGNED FOR AND ON BEHALF OF SGS (MALAYSIA) SDN BHD

**** End of Report ****

CHEE TUCK CHOON SECTION HEAD

IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at https://www.sgs.com/en/terms-and-conditions and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at https://www.sgs.com/en/terms-and-conditions/terms-e-document. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.