

BOEN HEALTHCARE CO., LTD.
CERTIFICATE OF ANALYSIS

Report No.: GM01-BN24

Product Name	Microscope Slide	Specification	7105
Order Quantity	4000BXS	Test Quantity	13bxs
Lot. No.	20240301	Mfg. Date	20240301
Exp. Date	20270228	Report Date	2024.03.26
Inspection Standard	JB/T 8230.3-1997		
No.	Technical Requirement	Conclusion	
1	The size of microscope slide is 25.4 x 76.2mm (1"x 3").	Meet the requirement	
2	The Thickness of microscope slide is 1.1mm±0.05mm.	Meet the requirement	
3	The microscope slide should be colorless and transparent.	Meet the requirement	
4	The surface of microscope slide should be smooth and thickness should be uniformity.	Meet the requirement	
5	The microscope slide should be clean, no mist, fingerprint or water stains.	Meet the requirement	
6	The edges of microscope slide should be smooth and should be polished complete.	Meet the requirement	
7	After boiled test, the microscope slide should be no blur, bubble or crack.	Meet the requirement	
Conclusion	According to the requirements of JB/T 8230.3-1997, concluded that this batch pass.		

Inspector: Leo Sun
Date: 2024.03.26



Checker: Irene Qiao
Date: 2024.03.26