



Benannt durch Designated by  
 Zentralstelle der Länder  
 für Gesundheitsschutz  
 bei Arzneimitteln und  
 Medizinprodukten  
 www.zl.de  
 BS-MDR-099



Product Service

## EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III  
 (Class IIa and Class IIb Devices)

**No. G10 010578 0039 Rev. 09**

**Manufacturer:** **Drägerwerk AG & Co. KGaA**

Moisinger Allee 53-55  
 23542 Lübeck  
 GERMANY

**SRN Manufacturer:** DE-MF-000005329

The Certification Body of TÜV SÜD Product Service GmbH certifies that the manufacturer has established, documented and implemented a quality management system as described in Article 10 (9) of the Regulation (EU) 2017/745 on medical devices. Details on device categories covered by the quality management system are described on the following page(s). The Report referenced below summarises the result of the assessment and includes reference to relevant CS, harmonized standards and test reports. The conformity assessment has been carried out according to Annex IX Chapter I and III of this regulation with a positive result. The quality management system assessment was accompanied by the assessment of technical documentation for devices selected on a representative basis. The certified quality management system is subject to periodical surveillance by TÜV SÜD Product Service GmbH. The surveillance assessment shall also include an assessment of the technical documentation for the device or devices concerned on the basis of further representative samples. All applicable requirements of the testing and certification regulation of TÜV SÜD Group have to be complied with. For details and certificate validity see: [www.tuvsud.com/ps-cert?q=cert:G10\\_010578\\_0039 Rev. 09](http://www.tuvsud.com/ps-cert?q=cert:G10_010578_0039_Rev_09)

**Report No.:** 713253108\_CN

**Preceding Certificate No.:** G10 010578 0039 Rev. 08

**Valid from:** 2023-03-14

**Valid until:** 2025-03-17

**Date of Initial Issuance:** 2020-03-18

**Issue date:** 2023-03-14

Christoph Dicks

Head of Certification/Notified  
 Body





Benannt durch Designated by  
 Zentralsstelle der Länder  
 für Gesundheitsschutz  
 bei Arzneimitteln und  
 Medizinprodukten  
 www.zdM.de  
 BS-MDR-099



Product Service

## EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III  
 (Class IIa and Class IIb Devices)

**No. G10 010578 0039 Rev. 09**

<b>Classification:</b>	Class IIa
<b>Device Group:</b>	R02 - BREATHING CIRCUITS AND CATHETER MOUNTS R0301 - RESPIRATORY MASKS R030201 - VENTILATION BALLOONS R0401 - VENTILATION FILTERS R0402 - NATURAL BREATHING FILTERS Z120301 - ANAESTHESIA AND PULMONARY VENTILATION SUPPORT INSTRUMENTS Z120309 - MEDICAL/MEDICINAL GAS PIPELINE SYSTEMS AND RELATED ACCESSORIES
<b>Intended Purpose:</b>	-
<b>Classification:</b>	Class IIa
<b>Device Group:</b>	Z12040192 - GENERAL MEDICINE DIAGNOSIS AND MONITORING INSTRUMENTS - MEDICAL DEVICE SOFTWARE
<b>Intended Purpose:</b>	-
<b>Classification:</b>	Class IIa
<b>Device Group:</b>	Z120302 - VITAL SIGNS MONITORING INSTRUMENTS
<b>Intended Purpose:</b>	-
<b>Classification:</b>	Class IIa
<b>Device Group:</b>	A060304 - INTRA-OPERATION FLUID COLLECTION DEVICES
<b>Intended Purpose:</b>	-
<b>Classification:</b>	Class IIb
<b>Device Group:</b>	Z12040192 - GENERAL MEDICINE DIAGNOSIS AND MONITORING INSTRUMENTS - MEDICAL DEVICE SOFTWARE
<b>Intended Purpose:</b>	Software intended to provide clinical information for the purpose of supporting patient management and the decision making process
<b>Classification:</b>	Class IIb
<b>Device Group:</b>	Z120804 - NEONATOLOGY INSTRUMENTS





Benannt durch Designated by  
 Zentralstelle der Länder  
 für Gesundheitsschutz  
 bei Arzneimitteln und  
 Medizinprodukten  
 www.zfl.de  
 BS-MDR-099



Product Service

## EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III (Class IIa and Class IIb Devices)

**No. G10 010578 0039 Rev. 09**

**Intended Purpose:** Warming therapy devices intended to provide controlled ambient conditions for premature babies and neonates in closed and open care therapy

**Classification:** Class IIb

**Device Group:** Z120301 - ANAESTHESIA AND PULMONARY VENTILATION SUPPORT INSTRUMENTS

**Intended Purpose:** Devices for the purpose of ventilation and/or anesthesia

**Classification:** Class IIb

**Device Group:** Z120309 - MEDICAL/MEDICINAL GAS PIPELINE SYSTEMS AND RELATED ACCESSORIES

**Intended Purpose:** Devices intended to distribute or supply gases, vacuum, electricity or data to equipment in diagnostic, therapy or surgery

**Classification:** Class IIb

**Device Group:** R020107 - THERMOREGULATED BREATHING CIRCUITS

**Intended Purpose:** Inspiratory (and expiratory) heated disposable breathing circuit for conducting humidified breathing gas from humidifier to patient

**Classification:** Class IIb

**Device Group:** R020101 - STANDARD BREATHING CIRCUITS

**Intended Purpose:** Devices intended to administer gases for the purpose of ventilation

The validity of this certificate depends on conditions and/or is limited to the following: /.

### Revision History:

Rev.	Dated	Report	Description
00	2020-03-18	713169482	-
01	2021-07-02	713184148	-
02	2021-09-30	713215188	-
03	2021-10-01	713215832	-
04	2021-10-04	713215842	-
05	2021-10-04	713219421	-
06	2021-11-22	713229134	-





Benannt durch Designated by  
 Zentralstelle der Länder  
 für Gesundheitsschutz  
 bei Arzneimitteln und  
 Medizinprodukten  
 www.zl.de  
 BS-MDR-099



Product Service

## EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III  
 (Class IIa and Class IIb Devices)

**No. G10 010578 0039 Rev. 09**

07	2022-02-21	713213004	-
08	2022-10-06	713225304_CN	-
09	2023-03-14	713253108_CN	Supplemented: Device(s)/group of device(s) added



## CATHETER MOUNTS AND AIRWAY CONNECTORS

### 4. CATHETER MOUNTS AND AIRWAY CONNECTORS

- Smoothbore products provide low resistance to flow while external coil reduces kinks and blockages
- Extendable product provides a lightweight and flexible solution
- Double swivel connector increases ease of use
- Double cap to maintain PEEP during suctioning or bronchoscopy
- Each catheter mount and airway connector is equipped with a red safety cap so that the system is completely closed and protected until directly prior to use on the patient
- 22F/15M connector on device side as a two-in-one solution
- Standardized connectors for safe connection to other components
- Luer Lock port for gas sampling
- Disposable convenience enhances workflow
- PVC-free design of gas leading components eliminates risk of softeners such as DEHP

<b>A</b>	<b>Catheter mount ErgoStar® CM 40, disposable, 10 cm, double swivel elbow, 50 pcs.</b>	<b>MP01840</b>
<b>B</b>	<b>Catheter mount ErgoStar® CM 45, disposable, 15 cm, double swivel elbow, 50 pcs.</b>	<b>MP01845</b>
<b>C</b>	<b>Catheter mount ErgoStar® CM 50, disposable, 10 – 16 cm, extendable, 50 pcs.</b>	<b>MP01850</b>
<b>D</b>	<b>Catheter mount ErgoStar® CM 55, disposable, 15 cm, straight, 50 pcs.</b>	<b>MP01855</b>
<b>E</b>	<b>Catheter mount ErgoStar® CM 60, disposable, 10 cm, small, double swivel elbow, 50 pcs.</b>	<b>MP01860</b>
<b>F</b>	<b>Airway connector ErgoStar® AC 90, disposable, double-swivel elbow w/ double cap, 100 pcs.</b>	<b>MP01890</b>
<b>G</b>	<b>Airway connector ErgoStar® AC 95, disposable, w/ Luer Lock sampling port, 100 pcs.</b>	<b>MP01895</b>



Not all articles are available worldwide.

## BREATHING SYSTEM FILTERS AND HMEs

### 2. BREATHING SYSTEM FILTERS AND HMEs

#### ELECTROSTATIC FILTER AND HME

The TwinStar® Plus filter/HME from Dräger is a combination of an efficient HME and a high-performance breathing system filter, thereby significantly contributing to infection prophylaxis in ventilation treatment.

- Combination of efficient HME and high-performance breathing system filter
- High retention rates for bacteria and viruses
- High return of moisture
- Minimal workload and low cost
- Sampling connector with tethered cap
- Transparent housing for visual control
- Standardized connectors for safe connection to other components
- Clear labeling and blue color coding for quick identification

<b>A</b>	<b>Filter + HME TwinStar® 90 Plus, disposable, ~90 ml deadspace, 100 pcs.</b>	<b>MP05800</b>
<b>B</b>	<b>Filter + HME TwinStar® 60A Plus, disposable, ~60 ml deadspace, angled, 100 pcs.</b>	<b>MP05810</b>
<b>C</b>	<b>Filter + HME TwinStar® 55 Plus, disposable, ~55 ml deadspace, 100 pcs.</b>	<b>MP05805</b>
<b>D</b>	<b>Filter + HME TwinStar® 25 Plus, disposable, ~25 ml deadspace, 100 pcs.</b>	<b>MP05815</b>
<b>E</b>	<b>Filter + HME TwinStar® 9 Plus, disposable, ~9 ml deadspace, 100 pcs.</b>	<b>MP05820</b>

#### MECHANICAL FILTER AND HME

<b>F</b>	<b>Mechanical filter + HME TwinStar® HEPA Plus, disposable, ~90 ml deadspace, 100 pcs.</b>	<b>MP05801</b>
----------	--	----------------

