



# Declaration of Conformity



**We: ELITechGroup B.V.**  
**Van Rensselaerweg 4**  
**6956 AV Spankeren**  
**The Netherlands**

declare under sole responsibility that the product indicated below (including all accessories) and to which this declaration relates, conforms to the provisions of:

- Directive 98/79/EC of the European Parliament and of the Council of 27 October 1998 on *in vitro* diagnostic medical devices (“IVD Directive”)
- Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment (“RoHS2 Directive”)

It is certified that this product is registered in accordance with the requirements of above mentioned EU Directives and carries the CE marking.

<b>Product</b>	<b>Clinical chemistry analyzer, automated</b>
<b>Model</b>	<b>Selectra ProS</b>
<b>Reference numbers</b>	<b>6003-500</b>
<b>GMDN code</b>	<b>56678</b>
<b>Accessories</b>	<b>See separate document ‘Regulatory status of parts &amp; accessories’</b>

### Product classification

As per Article 9, section 1 the products are categorized as other devices (“self declaration”).

### Conformity assessment procedure

In accordance with:

- Annex III of the IVD Directive
- Article 4 of the RoHS2 Directive

The product (including all accessories) may be marketed without any restrictions within the following countries and regions:

- The Netherlands (NL);
- All other member states of the European Union (EU);
- All member states of the European Free Trade Association (EFTA) and Switzerland.

Spankeren, July 2016

A. Altink

Managing Director



# Declaration of Conformity



## List of applied (harmonized) standards

	Standard version	Description	Tested / certified by
<b>Safety</b>	IEC 61010-1:2001	Safety requirements for electrical equipment for measurement, control, and laboratory use. Part 1: General requirements	<b>UL</b>
	IEC 61010-2-010:2003	Safety requirements for electrical equipment for measurement, control and laboratory used – Part 2-010: particular requirements for laboratory equipment for the heating of material	
	IEC 61010-2-081:2001	Safety requirements for electrical equipment for measurement, control, and laboratory use. Part 2-081: Particular requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes	
	IEC 61010-2-101:2002	Safety requirements for electrical equipment for measurement, control, and laboratory use -- Part 2-101: Particular requirements for in vitro diagnostic (IVD) medical equipment	
<b>EMC</b>	IEC 61326-1:2005	Electrical equipment for measurement, control and laboratory use - EMC requirements – Part 1: General requirements	<b>PHILIPS</b>
	IEC 61326-2-6:2005	Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 2-6: Particular requirements – In Vitro diagnostic (IVD) medical equipment	
<b>Quality systems</b>	EN ISO 13485:2012	Medical devices—Quality management systems—Requirements for regulatory purposes.	<b>LRQA</b>
	CAN/CSA ISO 13485:2003	Medical devices—Quality management systems—Requirements for regulatory purposes.	