

Anode Buffer Container 3500 Series

Product No. 4393927
Lot No. 2503021
Date of Manufacture 16MAY2025
Expiration Date 19JUL2026

TEST	SPECIFICATION	RESULT
PHYSICAL INSPECTION Material is clear, colorless and free of particulates		
Appearance - Conformance		Pass
pH TEST Measured at 25.0°C ± 0.3°C		
pH Value	7.6 - 8.2	7.899
CONDUCTIVITY TEST Measured at 25.0°C ± 0.1°C		
Conductivity	1500 - 2000 µS/cm	1773 µS/cm

*For Research Use Only.
Not for use in diagnostic procedures.*

**Manufactured in compliance with our ISO 13485 certified quality management system.
Site: Bedford, MA, USA**

Corporate Legal Entity
Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at
cofarequests@thermofisher.com

Elsa Qirollari
Quality Assurance
Issued 19MAY2025