

Resona I9 series

Diagnostic Ultrasound System

Datasheet

Release 1.2.0 (Rev23361)

Innovation, in every facet



Powered by **ZST+**

1. System Overview

Powered by the most revolutionary ZONE Sonography® Technology, i9's new ZST⁺ platform brings the ultrasound image quality to a higher level by zone acquisition and channel data processing.

1.1 Advantages of ZST⁺ platform

- Advanced Acoustic Acquisition
- Dynamic Pixel Focusing (DPF)
- Sound Speed Compensation (SSC)
- Total Recall Imaging (TRI)
- Powerful Processing Architecture
- Enhanced Channel Data Processing
- ZONE Sonography® Technology
- Up to 8,257,536 channels

1.2 Application

- Abdomen
- Obstetrics
- Gynecology
- Cardiology
- Small parts

- Urology
- Vascular
- Pediatrics
- Nerve
- Emergency & Critical
- Pelvic Floor
- Others

1.3 Transducer types

- Curved array transducer
- Linear array transducer
- Phased array transducer
- 4D Volume transducer
- Pencil transducer
- Bi-Plane transducer

1.4 Advanced Imaging technics

- THI (Tissue Harmonic Imaging) and PSH™ (Phase Shift Harmonic Imaging)
- iBeam™ (Spatial Compound Imaging)
- iClear™ (Speckle Suppression Imaging)

- iClear⁺
 - iTouchTM (Auto Image Optimization)
 - Echo BoostTM
 - Zoom/iZoom (Full Screen Zoom)
 - FCI (Frequency Compound Imaging)
 - B steer
 - ExFOV (Extended Field of View)
 - HD Scope
 - Smart Track
 - HR FlowTM (High Resolution Flow)
 - Glazing Flow
- 1.5 Imaging modes
- B-Mode
 - M-Mode/Color M-mode
 - Color Doppler Imaging
 - Power Doppler Imaging/Directional PDI
 - Pulsed Wave Doppler
 - Continuous Wave Doppler
 - Free Xros MTM (Anatomical M-mode)
- 1.6 Function Modes
- Free Xros CMTM (Curved Anatomical M-mode)
 - iScapeTM View (Panoramic Imaging)
 - TDI (Include TVI, TVD, TVM, TEI)
 - TDI QA (TDI Quantitative Analysis)
 - UWN+ (Ultra Wideband Non-linear Plus) Contrast ImagingTM
 - Contrast Imaging QA (Quantitative Analysis)
 - LVO (Left Ventricular Opacification)
 - Low MI Contrast
 - Volume CEUS
 - Natural Touch Elastography Imaging
 - STE Imaging (Sound Touch Elastography)
 - STQ Imaging (Sound Touch Quantification)
 - Endocavity STE
 - High frame rate STE
 - Stress Echo
 - TTQA (Tissue Tracking Quantitative

- Analysis)
- FH TTQA
- Smart 3D™ (Freehand 3D)
- Real-time 4D
- iPage⁺ (Multi-Slice Imaging)
- STIC (Spatial-Temporal Image Correlation)
- SCV⁺ (Slice Contrast View)
- Smart-V™ (Smart Volume)
- Smart V Trace
- Color 3D
- Niche
- iLive
- Smart Planes CNS
- Smart Planes FH Analysis on Vessel Stiffness)
- Smart Face
- Smart ICV
- Smart Scene 3D
- Ultrasound Fusion Imaging
- Endocavity Fusion Imaging
- Fusion RESP
- V Flow (Vector Flow)
- DICOM
- Clinical Measurement Package
- Smart Pelvic Floor
- Smart OB™ (Auto OB measurement)
- Smart NT™ (Auto NT measurement)
- Smart Fetal HR (Fetal Heart Rate)
- Smart FLC
- IVF
- IMT
- RIMT (RF-Data based IMT)
- Auto EF
- R-VQS (RF-Data based Quantitative
- Smart Thyroid
- Smart Breast
- Smart Hip
- Smart HRI
- Smart Caliper
- Smart Trace
- CPP (Color Pixel Percentage)

	<ul style="list-style-type: none"> • V-Mapping • iNeedle™ (Needle Visualization Enhancement) • ECG function 		<ul style="list-style-type: none"> • Touch gestures • Anti-virus software: McAfee, Windows Defender • iVocal
1.7	Other features	1.8	<p>Language support</p> <ul style="list-style-type: none"> • Software: English, Chinese, German, Spanish, French, Italian, Portuguese, Russian, Czech, Polish, Turkish, Finnish, Danish, Icelandic, Norwegian, Swedish, Hungarian, Serbian, Dutch, Lithuanian, Greek, Thai • Keyboard input: English, Chinese, German, Spanish, French, Italian, Portuguese, Russian, Czech, Polish, Icelandic, Norwegian, Swedish, Finnish, Turkish, Danish, Hungarian, Serbian • User manual: English, Chinese, German, Spanish, French, Italian, Portuguese, Russian, Polish, Turkish, Serbian
	<ul style="list-style-type: none"> • Ultrasound gel • Ultrasound gel warmer • Built-in wireless adapter • Replaceable battery assembly • Central brake • Probe adapter • DVD assembly • Wipes box bracket • DVR • Ambient light • iStorage • iWorks™ (Auto Workflow Protocol) • MedSight • MedTouch • UltraAssist (Off-line software) • UltraView™ (Off-line analysis software) 		

- Small keyboard: English, Germany, Russian, Spanish, French
- From front to back: 300 ± 20 mm
- From bottom to top: 150 ± 20 mm

2 Physical Specification

2.1 Dimension and weight

- The control panel and the monitor are in the lowest position.
- Configured with floating support arm and 23.8 inch monitor
 - Depth: 1020 ± 20 mm;
 - Width: 550 ± 10 mm;
 - Height: 1000 ± 20 mm
- Weight: $105kg\pm4kg$ (net weight, standard configuration but not including the probe)

2.2 Audio speakers

Stereo audio speakers

2.3 Dual-wing floating support arm

- Rotate angle: 90 ± 5 degrees (to the left); 150 ± 5 degrees (to the right)
- Tilt angle (when positioned vertically): 20 ± 5 degrees (backward); 85 ± 5 degrees (forward)

2.4 Wheels

- Diameter: 125mm
- When the central brake is configured: Central brake for total lock and break
- When the central brake is not configured: 4 Castors for total lock and break

2.5 Transducer port and holder

- Transducer ports with dust prevention: 5 active ports and 1 pencil probe port
- Support hot plug with active indicator lights
- Transducer holder: 5, plus 1 dedicated endocavity transducer holder and 1 dedicated pencil transducer holder

2.6 Electrical power

- Voltage: $100-240V\sim$

- Frequency: 50/60 Hz
 - Power consumption: 650 VA
- 2.7 Operating Environment
- Ambient temperature: 0-40°C
 - Relative humidity: 20%-85% (no condensation)
 - Atmospheric pressure: 700hPa-1060hPa
- 2.8 Storage & Transportation Environment
- Ambient temperature: -20-55 °C
 - Relative humidity: 20%-95% (no condensation)
 - Atmospheric pressure: 700hPa-1060hPa
- 2.9 System Noise
- ≤26dB @25°C
- iConsole: intelligent control panel for clinical-exam specific layout and adaptive adjustment: 6 programmable E-ink keys for dynamic display of user-defined functions
 - Automatic light indication for residual battery power
 - Full-space floating control panel adjustment and can be fixed at any position (when centered in the trackball):
 - Left/right rotation:180±5 degrees (90 degree for both left and right)
 - Left/right adjustment: 1100mm±50mm (550mm for both left and right)
 - Down/up adjustment: 300±20mm
 - Front/back adjustment: 350±20mm
- 3 User Interface**
- 3.1 Floating control panel
- Brightness adjustable for the backlight of the whole control panel
 - Full-sized, backlit QWERTY keyboard
- 3.2 Monitor
- 23.8-inch bezel-less LED monitor

- with high resolution
 - Resolution: 1920x1080
 - Viewing angle: 178 degrees
 - Digital on screen display of brightness and contrast controls
 - Automatic adjustment of monitor light with the changing environment
 - Automatic LED brightness
 - Tilt/Rotate independent adjustment
 - Tilt angle range: 105 degrees
 - Rotate angle range: 240 degrees
- 3.3 Touch screen
- 15.6-inch high sensitivity anti-glare color touch screen
 - Resolution: 1920*1080
 - Digital brightness and contrast adjustment through preset
 - Angle adjustable range: 40 degrees
 - Viewing angle: 170 degrees
 - Support touch screen gestures
 - Support either hand writing or with gloves on
- Editable buttons: long press to add, delete or move buttons
 - Moveable 3D/4D tabs
 - Clinical scenario-based 3D/4D user interface
 - Digital TGC
 - Short-cut switch of latest used probe & exams
- 3.4 Touch gestures
- Swipe down/up: display/remove projected image on touch screen
 - Swipe horizontally: page up/down or review images/cine loops one by one
 - Swipe from left edge to right: display hidden menu on projected image.
 - Image parameter adjustment.
 - Measurement on projected image on touch screen
 - Zoom in/out the projected image on touch screen

- Rotate or erase on projected 3D/4D image on touch screen
 - 8 user defined gestures using two fingers for more functions, such as freeze, print, activate specific imaging modes, measurements, and some other special functions.
- 3.5 System boot-up
- Boot-up from shut-down: <30 sec
 - Boot-up from stand-by: <5 sec
 - Shut-down: <30 sec
- 3.6 Comments
- Supports text input and arrow
 - Voice annotation: record voice as annotation for images and cine
 - Support freehand marking on touch screen
 - Adjustable text size and arrow size
 - Supports home position
 - Covers various application
 - User customizable
- 3.7 Bodymark
- More than 232 bodymarks for versatile application
 - User customizable
- ## 4 Imaging Parameters
- 4.1 B-mode
- Display formats:
 - Single(B)
 - Dual (B+B), support by B/ M/ Color/ Power/ PW/ CW/ Color M mode
 - Quad (4B), support by B/ Color/ Power
 - iClear™/ iClear+: Off, 7 steps
 - iBeam™: Off, 3 steps
 - iTouch™: on/off
 - FCI (Frequency Compound Imaging)
 - Dual Live: On/off
 - Image quality: Pen/Gen/Res/HPen/HGen/HRes/HG en-FFR/HRes-FFR (dependent on transducer)
 - B steer: 5 levels, available on linear

- transducers
- ExFOV: off, 1-2 (dependent on transducer)
 - Depth: 30 levels, 1.5-40cm
 - Frame rate (max): 1701f/s
 - Acoustic output power: dependent on transducer
 - TGC/LGC: 8 segments on touch screen
 - Dynamic range: 30-260 (dependent on transducer)
 - Gain: 0-100, 1/step
 - FOV: continuously adjustable
 - Line density: L/M/H/UH
 - Persistence: 0-7 levels
 - Horizontal Scale: on/off
 - L/R flip and U/D flip: on/off
 - Rotation: 0°, 90°, 180°, 270°
 - TSI: general/muscle/fluid/fat
 - Gray Map: 8 types
 - Tint map: off, 8 types
 - Echo Boost: On/Off
- Smooth: 0-6 levels
- HD Scope: off, 1-3 levels
- SSC (Sound Speed Compensation): On/Off
- Free view: -45°~45°, 5°/step
- Dehaze: 0-30 levels
- Ref Lines: on/off (under GYN and Pelvic Floor exam mode and using intracavity probe)
- V 1:1: on/off (available on linear probe under dual-split mode)
- ExtImage: On/Off
- Auto Merge: On/Off
- ZoneVue
- Edge enhance: 0-6, 1/step
- #### 4.2 THI and PSH™
- Patent PSH™ technology, obtains purer harmonic, better contrast resolution, higher SNR, exceptional high frequency harmonic
 - iClear™ available
 - Image quality: HPen/HGen/HRes or

HPen/ HPen-FFR/HGen/HRes/HRes-	1 level (HR Flow)
FFR (depends on transducers)	<ul style="list-style-type: none"> Max velocity: 146.5cm/s
• Echo Boost™: On/Off	<ul style="list-style-type: none"> Steer: available on linear transducer
4.3 M-mode	<ul style="list-style-type: none"> Max frame rate: 529f/s
• Display formats: V2:3, V3:2, V3:1,	<ul style="list-style-type: none"> Acoustic output power: same as B
H2:3, Full (V: vertical, H: horizontal)	mode
• Color M-mode available	<ul style="list-style-type: none"> Gain: 0-100, 2/step
• Acoustic output power: same as B	<ul style="list-style-type: none"> ROI size/position: adjustable
• Depth: same as B	<ul style="list-style-type: none"> Scale: max. 30 steps
• Dynamic range: 30-180, 5/step	<ul style="list-style-type: none"> Baseline: -8 – 8, 1/step
• Gain: 0-100, 1/step	<ul style="list-style-type: none"> Wall filter: 8 steps, 5-433Hz
• M sweep speed: 6 steps	<ul style="list-style-type: none"> PRF: 0.1-12.6kHz
• M soften: 0-4, 1/ step	<ul style="list-style-type: none"> Packet size: 0-3, 1/ step
• Tint map: off, 8 types	<ul style="list-style-type: none"> Flow state: L/M/H
• Gray Map: 8 types	<ul style="list-style-type: none"> Smooth: 0-6, 1/ step
• Edge enhancement: 0-3, 1/ step	<ul style="list-style-type: none"> B/C align: On/Off
4.4 Color Doppler Imaging	<ul style="list-style-type: none"> Priority: 0%-100%, 1%/step
• Dual live: On/Off	<ul style="list-style-type: none"> Color map: V0-V10; VV0-VV9
• HR Flow™: High Resolution Flow	<ul style="list-style-type: none"> Invert: On/Off
provides better image quality and	<ul style="list-style-type: none"> Auto Invert: On/Off
flow sensitivity	<ul style="list-style-type: none"> Persistence: 0-6, 1/ step
• Image quality: Pen/Gen/Res (color),	<ul style="list-style-type: none"> Velocity tag: On/Off

- Line density: L/M/H/UH
- iTouch™: on/off
- Smart track: On/Off
- Glazing Flow: On/Off, L/M/H
- Flow state: L/M/H
- Smooth: 0-6, 1/ step
- B/C align: On/Off
- Priority: 0%-100%, 1%/step
- Color map: 4 types
- Directional color map: 4 types
- Persistence: 0-6, 1/step
- Line density: L/M/H/UH
- Invert: On/Off
- iTouch™: On/Off
- Smart track: On/Off
- Glazing Flow: On/Off, L/M/H

4.5 Power Doppler Imaging

- Dual live: On/Off
- HR Flow™: High Resolution Flow

provides better image quality and sensitivity
- Support directional power Doppler
- Image quality: Pen/Gen/Res (Power), 1 level (HR Flow)
- Acoustic output power: same as B
- Dynamic range: 10-70, 5/step
- Gain: 0-100, 2/step
- ROI size/position: adjustable
- Steer: available on linear transducers)
- Scale: max. 30 steps
- Wall filter: 8 steps
- PRF: max.12.6kHz
- Packet size: 0-3, 1/ step
- Flow state: L/M/H
- Smooth: 0-6, 1/ step
- B/C align: On/Off
- Priority: 0%-100%, 1%/step
- Color map: 4 types
- Directional color map: 4 types
- Persistence: 0-6, 1/step
- Line density: L/M/H/UH
- Invert: On/Off
- iTouch™: On/Off
- Smart track: On/Off
- Glazing Flow: On/Off, L/M/H

4.6 PW/CW Mode

- Display formats: V2:3, V3:2, V3:1, H2:3, Full, Duplex/Triplex (PW only)

(V: vertical, H: horizontal)
- Image quality: Pen/Gen/Res
- PW velocity: max. 8.681m/s
- min. 2.03cm/s
- CW velocity: max. 39m/s
- min. 5.34cm/s
- Sample volume size: 0.5-30mm

- (PW only)
 - Sample gate depth: continuously adjustable
 - Baseline: 9 steps
 - PW Steer: available on linear transducer
 - Volume: 0%-100%, 2%/step
 - PW PRF: 1.0-23.1kHz
 - CW PRF: 0.3-104.0kHz
 - Gain: 0-100, 2/step
 - Dynamic range: 24-72, 2/step
 - Sweep speed: 6 steps
 - Wall Filter: 0 ~ 9 steps
 - Invert: On/Off
 - Auto invert: On/Off
 - Angle correction: -89~89 degrees, 1/step
 - HPRF: On/Off
 - Auto calc: On/Off
 - Auto calc cycle: 1, 2, 3, 4, 5
 - Auto Calc Loop
 - Trace Sensitivity: 0-5, 1/step
 - Trace Smooth: Off, 1-4, 1/step
 - Trace area: Above, Below, All
- 4.7 Free Xros M™
- Display formats: V2:3, V3:2, V3:1, H2:3 (V: vertical, H: horizontal)
 - Color Free Xros M available
 - Up to 3 lines
 - Display all lines
 - Sweep speed: 6 steps
 - M Tint map: Off, 8 types
 - Gray Map: 8 types
- 4.8 Free Xros CM™
- Quick angle: 0, -60, 60 degrees
 - Gray map: 10 types
 - Tint map: Off; 8 types
 - Time/frequency resolution: 0-6, 1/step
 - Only available in TDI mode
 - Display formats: V2:3, V3:2, V3:1, H2:3 (V: vertical, H: horizontal)
 - Sweep speed: 6 steps
 - Tint map: Off; 8 types

- Gray Map: 8 types
 - Edit, Undo, Delete function for curved line
 - 4.9 iBeam™
 - Spatial compound imaging
 - Off, 1-3, 1/step
 - 4.10 iClear™
 - Speckle suppression imaging
 - Available on B, 3D, 4D mode
 - 4.11 iTouch™
 - Auto image optimization
 - B-mode: gain, TGC, dehaze
 - Color: gain, color box position
 - Power: gain
 - PW: gain, scale, PRF, WF, SV size, SV position, steering angle
 - Contrast imaging: gain
 - 4.12 Echo Boost™
 - Available in cardiac exam mode when using a phased array probe
 - Improve the homogeneity through the whole field of view
 - Better noise control in cardiac chambers and muscles
- 4.13 Zoom
- Zoom: Spot zoom (write zoom), Pan zoom (read zoom) 0.8x-10x
 - iZoom: convertible 3 steps; normal image, zoom standard area, zoom only image area
- 4.14 QSave
- Quickly save image parameter setting after image adjustment done
 - Support Save, Create, Restore
 - IP (Image Process) Preset Manager
 - Support save, create, restore or delete IPs
 - Quickly switch to different exam modes without returning to B mode
 - A default set of image parameters under each IP are provided for different exam modes

- Only image parameters are switched while measurements, comments, and bodymarks remain the same
- PRF: 0.4-14.9kHz
- Acoustic output power: same as B
- Gain: 0-100, 2/step
- Dynamic range: 10-70, 5/step (TEI only)
- 3D/4D Preset Manager
 - The scenarios and subpresets can be renamed, restored, deleted, added, set to active, or moved
 - Show scenario and subpreset parameters
 - Provide multiple groups of preset 3D/4D parameters based on different application scenarios to quickly obtain expected image effect
 - ROI size/position: adjustable
 - Scale: max. 30 steps
 - Baseline: -8 – 8, 1/step (TVI only)
 - Wall filter: 8 steps
 - Packet size: 0-3,1/step
 - Flow state: L/M/H
 - Smooth: 0-6, 1/ step
 - B/C align: On/Off
 - Priority: 0%-100%, 1%/step
 - Color map: 10 types (TVI), 8 types (TEI)
- 4.15 Tissue Velocity/Energy Imaging (included in TDI option)
 - Available on phased array, and SC6-1s, SC8-2s, SP5-1s, and P7-3Ts
 - Dual live: side by side displays B and B+TVI
 - Max frame rate: 3175 f/s
 - Invert: On/Off
 - Persistence: 0-6, 1/ step
 - Velocity tag: On/Off (TVI only)
 - Line density: L/M/H/UH
 - Image quality: 2 levels
- 4.16 Tissue Velocity Doppler (included in

- TDI option)
 - Display formats: V2:3, V3:2, V3:1, H2:3, Full, Duplex/Triplex (V: vertical, H: horizontal)
 - Sample volume size: same as PW
 - Sample gate depth: adjustable
 - Sample volume depth: continuously adjustable
 - Scale: 30 levels
 - Volume: 0%-100%, 2%/step
 - PRF: 0.7-23.1kHz
 - Gain: 0-100, 2/step
 - Baseline: 9 steps
 - Dynamic range: 24-72, 2/step
 - Sweep speed: 6 steps
 - Wall Filter: 10 steps
 - Invert: On/Off
 - Angle correction: -89-89 degrees, 1/step
 - Image quality: 2 levels
 - Time/frequency resolution: 0-6, 1/step
- 4.17 Tissue Velocity Motion (included in TDI option)
- Display formats: V2:3, V3:2, V 3:1, H2:3, Full (V: vertical, H: horizontal)
 - Acoustic output power: same as B
 - Gain: 0-100, 2/step
 - M sweep speed: 6 steps
 - Smooth: 0-6, 1/ step
 - Color Map: 10 types
 - Image quality: 2 levels
 - Persistence: 0-6, 1/ step
 - Packet size: 0-3, 1/ step
 - Priority: 0%-100%, 1%/step
 - Velocity tag: On/Off
 - Tissue state: L/M/H
- 4.18 Smart 3D™
- Smart 3D
 - Acquisition preparation:
 - 3D/4D Scenario setting: Routine

- Acquisition Methods: Rocked, Linear
- Brightness: 0-100%, 2%/step
- Reset VOI: On/Off
- Contrast: 0-100%, 2%/step
- Flip VOI: On/Off
- Smooth: 0-10, 1/ step
- Angle: 10-80°
- Depth VR: Off/Black/Cyan/Blue/Rose
- Distance: 10-200mm
- Tint: off; 8 types
- Acquiring Time: 1.0s-20.0s
- Degree: 10-80°
- VR:
- Distance: 10-200mm
- 3D/4D Scenario: Routine (Surf., iLive Gen., iLive Transp., Skeleton)
- MPR:
- Active quadrant: A, B, C
- Reset: All, Orientation, Curve
- Gray Map: 1-8
- VOI: On/Off/Fixed
- Brightness: 0-100%, 2%/step
- Active quadrant: A, B, C, VR
- Contrast: 0-100%, 2%/step
- VR orientation: 0°, 90°, 180°, 270°
- iClear: Off; 7 types
- Tint: Off; 8 types
- Thickness: 0-30mm
- Flip: flip VR
- Adv.:
- Sync: synchronize VR with selected plane
- Direction: Up/Down, Left/Right, Front/Back, Down/Up, Right/Left, Back/Front
- Orientation Assist: On/Off
- Threshold: 0-100%, 1%/step
- VR Refine: Off; 7 steps
- Opacity: 0-100%, 5%/step
- Surface enhance: 0-7, 1/step

- MagiClean:
Off/Low/Mid/High/Max
 - Inversion: On/Off
 - A3:1: On/Off
 - Move light: On/Off
 - Degree: 10-80°
 - Distance: 10-200mm
 - Main render: Surface, Max, Min,
X Ray, iLive
 - Sub render: Surface, Max, Min,
X Ray
 - Mix: Set the mix ratio of the two
render modes
 - Tool
 - Edit:
 - Rubber: On/Off

 - Eraser Diameter: 8-80, 1/step

 - Cut (area selection): Polygon,
Contour, Rectangle, Line

 - Undo: Undo, Undo All
 - 3D Layout:
 - Niche Views: Inner, Outer
- 4.19 4D
- 3Slice
 - Active Quadrant: A, B, C,
Niche/3Slice
 - Auto rotation:
 - Position: Set Start/Set End

 - Direction: Left/Right, Up/Down

 - Step: 1-15°

 - Quick Angle: 30-180°

 - Rotation control: play, single
loop, loop

 - Save AVI to USB
 - Available on all volume transducers
 - Static 3D and real time 4D
- Acquisition preparation:
- 4D frame rate: max. 80 vps
 - 3D/4D Scenario setting: Smart
 - Scene3D (Spine, Brain, Long
Bone, Face, Endometrium,
Pelvic), Routine, iLive Pro, Bone,
 - Tissue (not all scenarios are
listed)

- Refresh: On/Off
 - Angle: 10-120°
 - Quality: low1, low2, mid, high1, high2
 - Other parameters are the same as Smart 3D
- Adv.:
- The parameters are the same as Smart 3D
- Tool
- The parameters are the same as Smart 3D
- VR:
- 3D/4D Scenario: Smart Scene 3D, Routine, iLive Pro, Bone, Tissue, Routine (not all scenarios are listed)
 - 3D iClear: Off; 7 steps
 - Face⁺: Off, 3 steps
 - Auto Play: Stop, x1, x2, x3, x1/2, x1/3
 - Frame: Select a frame
 - Other parameters are the same as Smart 3D
- 3D Reference Point
 - Enable the operator to define one or more reference points on MPRs, which are then projected to VR image; helpful for the operator to better understand the corresponding spatial relations of VR image and MPRs
 - Display: Point only, H line, V line
 - Delete All
 - Hide All
 - Color 3D
 - Available on volume transducers
 - Supports Color and Power mode
 - Only Available in Static 3D mode
 - STIC
- MPR
- 3D iClear: Off; 7 types
 - Other parameters are the same as Smart 3D

- Color STIC available
- Move plane
- SCV+
 - SCV
 - Reset: All
 - Orientation Assist: On/Off
 - Active Quadrant: A, B, C
 - Main render modes: Surface, Max, Min, X-ray
 - Sub render modes: Surface, Max, Min, X-ray
 - Mix: Set the mix ratio of the two render modes
- iPage+
 - Slice display mode: Slice only, Slice with SCV
 - RefMPR: A Plane, B Plane, C Plane
 - Reset Ori: On/Off
 - Sync MPR: A/B/C
 - Quick Rotation: 90°, -90°, -180°
 - CMPR
 - Slice layout: 2*2, 3*3, 4*4, 5*5
 - Reset Curve
 - Slice Number: 3-25
 - Undo Last
 - Spacing: 0.5-10mm
 - Flip: Thickness, Direction
 - Line (Start)/Line (End)
 - Ref. Image (A B, C), 1, 2, 3
 - Single window
 - Trace Options: Line, Trace, Spline
 - Display Ref plane
 - Hide Ref plane
 - Quick Rotation: 90°, -90°, -180°
 - Vertically display ref plane
 - Hide Annotation
 - Horizontally display ref plane
 - Rotate RL

- Support labeled measurements
 - iLive
 - Shading: 0-10, 1/step
 - Grad View
 - Hyaline: On/Off, 0~100%, 5%/step
 - Move Light
 - Light Position: 6 default positions
selectable
 - Light 1/2/3: Parallel, Point, Torch
 - VL Saturation: 0~100%, 1%/step
 - VL Hue: 0~100%, 1%/step
 - VL Distance: 0.0 – 5.0
 - VL Bright: 0~100%, 1%/step
 - VL Angle: 0~100%, 1%/step
 - Tint: 0-4
 - Reset

Classic/IntPoint/ExtPoint/Parallel/Tor
ch/ 3-Light/ User 1/User 2

 - Copy to: Copy the current lighting mode to customized lighting mode "User 1" or "User 2"
 - Smart Planes CNS
- Detect automatically the standard sections of TCP, TTP, MSP and TVP
 - Rotation around X/Y/Z axes
 - Hide line: On/Off
 - Reset: All planes/current plane
 - Auto comment supported:
A(anterior), P(posterior), L(Left),
R(Right), U(Up), D(down), CSP, T,
CH, CV, CM, LV on TCP, TTP, MSP
and TVP
 - Auto measurement supported:
- TCD and Cist Maga (CM) on section TCP;
 - BPD, OFD and HC on section TTP;
 - LVW on section TVP
 - Support editing measurement results
 - Hide Measure: On/Off
 - MSP adjust: A/B/C
 - Support comment and bodymark on sectional plane
 - Font Size: Small, Medium, Large

- Smart ICV
 - Measure fetal cerebral volume automatically
- Smart Planes FH
 - Detect automatically left ventricular outflow tract view, right ventricular outflow tract view, LAV-DA view, 3VV-T view and stomach bubble view
 - Navigate: On/Off
 - Fetal heart: Stomach, 4 chamber, LVOT, RVOT, Ductal Arch, 3VT
 - Quick Adjust: -10-10
- Smart Face
 - Recognize fetal face automatically and then display the face in a recommended viewing angle
 - AutoDirect
 - MixRender: 0~3
 - Erase size: 0-2
 - Eraser: For Rubber/Re-Rubber
 - SubTint: 1-4
- FaceContact: -15~15
- VR Orientation: 0°, 90°, 180°, 270°
- Smart FLC (Smart Follicle)
 - Automatic follicle calculation
 - Edit ROI and detect follicle contour automatically
 - Undo: Undo, Redo, Undo All
 - Active Quadrant: A, B, C, Follicle, Left, Right
 - Calc: On/Off
 - Edit ROI: On /Off
 - Edit: Divide, Merge, Add/Del
 - Save to Report
- Smart-V™ (Smart Volume)
 - Auto 3D volume calculation
 - Smart-V ROI: Manual ROI on A, B, C planes separately
 - Smart-V Trace: Manual trace on A, B, C planes separately
 - Smart-V Vocal: Separately Trace contours on each slice that is generated by rotation

- Smart-V Parallel: Separately Trace contours on each slice that is generated by translation optimization, to planes acquisition, quantification and with automated workflow throughout whole procedure.
 - Display: 2D, 2D&C
 - Trace Mode: Smart Trace, Spline, Trace, Control Point
 - Window format: 1*1, 3*2, 3*3, 4*3
 - Shell: Inside, Outside, Sym.
 - Auto detect contour of target
 - Volume result shows in result window
 - Reset: Orientation
 - Active Quadrant: A, B, C, Smart-V
 - Support MRP measurement
 - Smart Scene 3D
 - As an intelligent scenario-oriented volume scan technique, it automatically enables the identification of tissue characteristics and provides an organ-specific diagnosis with full-stack intelligence from imaging
- 4.20 Smart Track
- Available on linear probes
 - Enable the function under Color/Power mode, the angle and the position of the ROI are adjusted automatically.
 - Enable the function under Color/Power+PW mode, the angle

and the position of the PW sampling line, SV size, SV angle and SV position are adjusted automatically.	4.23 iNeedle <ul style="list-style-type: none">• Needle visualization enhancement• Available on linear (except for the linear plane of ELC13-4s), SC6-1s, and C6-2Gs probes
4.21 RIMT (RF-Data based IMT) <ul style="list-style-type: none">• Available in single/dual B carotid exam mode• Side: left/right• Calculation of 6 RIMT values, RIMT average value, SD and ROI W• Report operation:<ul style="list-style-type: none">- Data deleting- RIMT trend graphic viewing- Preview	4.24 V Flow (Vector Flow) <ul style="list-style-type: none">• Needle direction: Auto, Left, Right• B/iNeedle: On/Off• In-plane biopsy and Out-plane biopsy• GPS real-time guiding biopsy
4.22 iScape™ View <ul style="list-style-type: none">• Acquisition method: B and Power• Supports speed indicator• Actual size: On/Off• Fit size: On/Off• Ruler: On/Off• Tint map: Off; 8 types• Rotation: 0~355 degrees, 5/step	4.24 V Flow (Vector Flow) <ul style="list-style-type: none">• Available on L9-3s in real-time B/Color carotid exam mode• Quality: 1-8• V Flow map: 4 levels• Gain: 0-8, 1/step• Persistence: 0-5, 1/step• Edge smooth: Low/Mid/High• Wall Filter: 8 steps• Brightness: 0-100%, 5%/step• Colorbar scale: 50-100cm/s, 10cm/s/step

- Arrow life time: 10-50, 1/step available
 - Arrow size: S/M/L/XL/XXL
 - Arrow density: 1%-10%,1%/step
 - Packet size: 0-3, 1/step
 - Acquire time: 0.5s, 0.6s, 0.8s, 1.0s, 1.2s, 1.5s
 - TGC: same as B
 - HR Range: Random, HR48-53, HR54-59, HR60-69, HR70-79, HR80-89, HR>89
- 4.25 UWN+ Contrast Imaging™
- UWN+ (Ultra Wideband Non-linear Plus) contrast imaging technology, which provides exceptional contrast agent detecting capability, not only extracts second harmonic, but also non-linear fundamental signals
 - Available on SC6-1s, SC8-2s, C11-3s, C6-2Gs, L9-3s, L14-3Ws, L20-5s, V11-3Hs, SP5-1s, SD8-1s, DE11-3Ws, and ELC13-4s probes
 - Micro Flow Enhancement (MFE)
- Timer1: On/Off
 - Timer2: On/Off
 - Pro capture: captures prospective image less than 480s preset table
 - Retro capture: captures retrospective image less than 120s preset table
 - Dual live: side by side displays tissue image and contrast image
 - MFE: On/Off
 - MFE period: 0.1s, 0.2s, 0.4s, 0.6s, 0.8s, 1.0s, MAX
 - Destruct: instantly destroy contrast bubbles
 - Destruct AP: -30~0dB, 0.3/step
 - Destruct time: 500-2000 ms
 - Contrast Agent: SonoVue and SonaZoid
 - iClear: Off; 7 steps
 - Mix: mix contrast image with tissue image

- Mix map: 7 types, available when Mix mode is active compatibility with commercially available ultrasound contrast agents.
 - Persistence: 8 steps Because the availability of these agents is subject to government regulation and approval, product features intended for use with these agents may not be commercially marketed nor made available before the contrast agent is cleared for use. Contrast related product features are enabled only on systems for delivery to an authorized country or region of use. Mindray medical systems makes no claims concerning the safety or effectiveness of contrast agents.
 - Dynamic range: same as B mode
 - Gray map: 8 types
 - Tint map: Off; 8 types
 - Supports U/D Flip and L/R Flip
 - Rotation: same as B
 - CEUS Position: On/Off
 - Line density: L/M/H/UH
 - FOV: On/Off
 - FOV size/position: continuously adjustable
 - ExFov: Off, 1-2, 1/step
 - Gain: 0-100, 1/step
 - iTouch: On/Off, -8~8, 2/step
 - Image quality: 3 levels
 - Smooth: 0-6, 1/step
 - Enhance: On/Off
 - Markline: On/Off
 - LGC: 8 points
- *The I9 series is designed for
- #### 4.26 Contrast Imaging QA
- Support Time-Intensity Curve analysis
 - Table display: display data in table
 - Up to 8 ROIs
 - Delete all
 - Delete current

	<ul style="list-style-type: none"> • Copy ROI • Fit curve • Raw curve • Motion tracking: Reduce the effect of tissue movement • X scale: 1-5, 1/step • Auto play: Stop, X1/10, X1/5, X1/4, X1/3, X1/2, X1, X2, X3 	4.29 Volume CEUS	<ul style="list-style-type: none"> • Available on SD8-1s and DE11-3Ws probes • Display: Contrast, Tissue • Capture 3D image • Other parameters are the same as those of 4D imaging
		4.30 STE Imaging (Sound Touch Elastography Imaging)	
4.27 LVO			<ul style="list-style-type: none"> • Available on SC6-1s, C6-2Gs, L9-3s, L20-5s, L14-3Ws, V11-3Hs, DE11-3Ws, and ELC13-4s probes • Display Format: V1:1, H1:1, Full • Invert: On/Off • HQ Elasto: On/Off • E Quality: Pen, Gen, Res • Elas.Metric: E, Cs, G • Scale: 30 levels • Opacity: 0-5, 1/step • Map: 3 types • ROI Width/Height: continuously random adjustable
4.28 Low MI Contrast Imaging			
	<ul style="list-style-type: none"> • Only available in cardiac exam mode • Dedicated left ventricle contrast imaging tool 		

- ROI Center Depth: continuously
adjustable
 - Depth: same as B mode
 - iLayering: On/Off
 - Filter: 0, 1
 - RLB View: On/Off
 - M-STB Index: On/Off
 - M-STB Sensibility: 0~4, 1/step
 - Smooth: 0~2, 1/step
 - Persistence: 0~2, 1/step
 - RLB Map: On/Off, RLB, RLB&E,
RLB&B&E
 - Map Position: 0%~100%, 5%/step
 - E bar: Mean, Max, Min, SD
 - E Avg: Off, 8 levels
 - Select/Bad: On/Off
 - Fixed ROI: On/Off
 - Save All: On/Off
 - Lesion: Off, 1-10
- 4s probes
- Support strain, strain ratio and strain histogram measurement
 - Unique shell analysis function
 - Stress compensation technology reduces deeper tissue artifacts, obtain more uniform stress throughout whole field
 - Stress indicator: supports frame by frame stress indication.
 - Map: 6 types
 - Smooth: 0-5, 1/ step
 - Opacity: 0-5, 1/step
 - ROI Width/height: continuously adjustable
 - Invert: On/Off
 - Display Format: V1:1, H1:1, Full
 - Strain mode: 0~1, 1/step
 - Dynamic Range: 0~5, 1/step
 - Map Position: 0%~100%, 5%/step
 - Strain Scale: 0-5, 1/step
 - E Sensitivity: 0-5, 1/step
- 4.31 Natural Touch Elastography
- Available on L9-3s, L14-3Ws, L20-5s, V11-3Hs, DE11-3Ws, and ELC13-

4.32 High frame rate STE

- Available on SC6-1s, C6-2Gs, L9-3s, L20-5s, L14-3Ws, V11-3Hs, DE11-3Ws, and ELC13-4s probes
- Improve the image refresh rate and provide a smoother STE image

4.33 STQ Imaging (Sound Touch Quantification Imaging)

- Available on SC6-1s, C6-2Gs, L9-3s, L20-5s, L14-3Ws, V11-3Hs, DE11-3Ws, ELC13-4s probes
- ROI Adjustment: adjust the ROI fixed size
- Elas.Metric: E, Cs, G
- E bar: Mean, Max, Min, SD
- M-STB Index: On/Off
- M-STB Sensibility: 0-4, 1/step
- Filter: 0, 1
- Smooth: 0-2, 1/step
- Persistence: 0-2, 1/step
- Map Position: 0~100%, 5%/step
- Lesion: Off, 1~10

- The square height of the elasto curve represents the average value of the elasto metric for current frame.

- Scale: 0-9, 1/step
- E Avg: off, 8 levels
- HQElasto: On/Off
- Select/Bad: On/Off
- Fixed ROI: On/Off
- Save All: On/Off
- Lesion: Off, 1-10

4.34 iFusion

- Available on SC6-1s, C6-2Gs, SP5-1s, SC8-2s, L14-3Ws, ELC13-4s, and V11-3Hs probes in B/ Color/ Power/ contrast imaging mode (non-cardiac)
- Single window display
- Fusing CT/MR/PET/freehand volume data with the ultrasonic image
- CT/MR/PET data reconstruction for

- 3D displaying
 - Tracking system: On/Off
 - Fusion ratio: -1~1, 0.1/step
 - Axis rotation: 0° ~ 360° based on X-axis, Y-axis or z-axis in increment of 2°.
 - ROI Offset X: -630~630
 - ROI Offset Y: -566~566
 - Window W/L: 1/step. Adjust the CT/MR/PET/freehand brightness and the contrast by changing the width and the level.
 - Reset Window W/L
 - Reset CT/MR
 - Display marks
 - Respiration curve: On/Off
 - Respiration Range: 0/1/2
 - View Type: Axial, Coronal, Sagittal
 - Quick Translation
 - Registration
 - Mark on Volume Data/Mark on Fusion Image
 - Support general measurement
 - Support adding comment and bodymark
- 4.35 Auto EF
- Output EDV/ESV/EF/SV/CO by Simpson method
 - Activated with or without ECG
 - Adjustment for the border of endocardium by single point or multi points
 - Adjust Frame
 - Layout: Dual/ Single
 - Diastole FR
 - Systole FR
 - Volume curve: On/Off
- 4.36 TDI QA
- Dedicated quantification tool for TDI velocity, strain, strain rate analysis
 - Ellipse ROI, Standard ROI
 - Up to 8 of ROI
 - Delete all

- Delete current
- ROI tracking: tracking ROI along with cardiac movement
- Smooth: 1-7, 1/step
- X scale: 1-5,1/step
- Std.Height: 1.5-50 mm
- Std.Width: 1.5-50 mm
- Std.Angle: -89-90 degrees
- Export: export current data as CSV format file
- Accept & compute: start tracking myocardium movement when user accept trace result
- Display effect: 0/1; at 1, tracking in velocity vector arrow; at 0, tracking in dots
- Trace method: 3 point or manual for ALAX, A4C, A2C; manual for PSAXB, PSAXM, PSAXAP
- Bull's eye: trace result in bull's eye model

4.37 TT QA

- Available on the SP5-1s probe under the cardiac mode.
- Tissue tracking quantitative analysis
- Mandatory ECG connection before TTQA cine acquisition
- Six views for analysis: ALAX, A4C, A2C, PSAXB, PSAXM, PSAXAP
- Reload: reload cine again for new study
- Edit: modify trace points
- Start tracking
- Valve's open and close time index: MVC, MVC', AVC, AVO, MVO
- Data export: export data in CSV file
- Cycle: ECG triggered cardiac cycle recognition for analysis; adjustable
- Auto play: stop, X1/10, X1/5, X1/4, X1/3, X1/2, X1, X2, X3
- Thickness: 1-30mm, 1mm/step; adjust trace thickness
- Track point: 20-40, 1/step
- Parameter: Volume, Speed,

Displacement, L Strain, L Strain R, T

- Available on the SP5-1s probe in cardiac mode

Strain, T Strain R, Area, R Strain, R

- 14 factory protocols

Strain R, C Strain, C Strain R, C

- User-defined protocols

Rotation, C Rotation R

- ECG triggered acquisition, display, selection, comparison, evaluation and archiving of multiple cardiac loops during various stages of a stress echo examination

- Smooth: 0-4, 1/step
- Trace method: 3 point, manual
- Tracking cycles: 1, 3
- Select Cycle: select among 3 cycles when the Tracking Cycle is set to 3

- Average Cycle: On/Off
- Velocity scale: 0%-200%
- Display style: All, Endo., Myo., Epi.,
- Curve Display

- Torsion & Torsion Rate Curve
- LGC adjustment
- Data Export: export current data as CSV format file

- View: standard views (PLAX, SAB,

PSAX, SAA, A4C, A2C, ALAX), and

customized views

- Image acquisition

- R-wave trigger

- Acquire mode: Manual ROI or full screen

- Ability to acquire frames or clips in B-mode, M-mode, Color, PW and

4.38 Fetal heart TT QA

- Available Under Fetal Echo related exam modes
- Trace method: Manual

4.39 Stress Echo

TDI

- Image selection
 - Attach the images with view annotation label PLAX, SAB, PSAX, SAA, A4C, A2C, ALAX, and customized views)
 - Review
 - Automatically adjust to the number of images user defined
 - Wall Motion Scoring
 - ASE 16 (with score 4-7), or ASE 17 (with score 4-7)
 - Graphical display of scoring (Normal, Hyperkinetic, Severely Hyperkinetic, Akinetic, Dyskinetic)
 - LV volume measurement
 - Measurement of LV Volume in all phases of cardiac cycle
 - Report
 - Reporting for both Wall Motion Scoring and LV volume measurement
- 4.40 R-VQS
- Track movements of the upper and lower vessel walls automatically
 - Displacement and Vessel diameter display in the result window.
 - Motion curve of vessel walls display under the image in real time.
 - Side: Left, Right
 - Speed: 6 levels
 - Position: 0-100%
 - Curve Disp: Adjust the height of the curve
- 4.41 Smart Pelvic
- Enter smart pelvic in 2D or 3D/4D scanning mode.
 - Set Rest, Valsalva, and Contraction frame
 - Measure automatically
- 4.42 Smart Fetal HR (Fetal Heart Rate)
- Measure the fetal heart rate automatically on B and M modes
 - Smart FHR cycle: 1-3
 - Alignment Out-Plane

- | | |
|---|---|
| <p>4.43 Smart Thyroid</p> <ul style="list-style-type: none"> • Recognize nodules, separate the nodule boundary, list the characteristics of the nodules, trace the nodule boundary, and perform measurements • Side: Left, Right, Isthmus • Select nodule • Add nodule/Delete nodule • Select view • Add/delete malignant signs view • Display/Hide measurement results • Analyze Nodule • Confirm Nodule • Review Nodule Analysis <p>4.44 Smart Breast</p> <ul style="list-style-type: none"> • Recognize lesions, list the characteristics of the lesions and perform measurements • Side: Left, Right • Select Mass • Add Mass/Delete Mass | <ul style="list-style-type: none"> • Select view • Add/delete malignant signs view • Display/Hide measurement results • Analyze Mass • Confirm Mass • Review Mass Analysis <p>4.45 iScanHelper</p> <ul style="list-style-type: none"> • Tutorial function as a guidance to show basic scanning skill with graphic of probe position, schematic of anatomy and example clinical image • Support ABD, SMP, URO, OB, GYN applications • Support broadcasting the scanning skill in multi languages <p>4.46 iCompare</p> <ul style="list-style-type: none"> • Allow to compare real-time ultrasound imaging to the past DICOM CT/MRI/Mammography/X-Ray/Ultrasound images without external workstation |
|---|---|

- Helpful to easily evaluate and follow up the progression of disease, treatment effect monitoring.
- 4.47 DVR
- Digital video recorder, a useful tool for education and memory
 - Max storage length each time: 60 min
- settable) and prospective storage (1-480s pre-settable)
 - Frame compare: displays one cine in dual format and allows frame by frame compare side by side
 - Cine compare: compare cines which are saved in same imaging mode
 - Jump to first and jump to last: one keystroke go to first or last frame in the cine

5 Cine Review and Raw Data

Processing

5.1 Cine review

- Available in all modes
- Frame by frame manual cine loop review or auto playback with variable speed
- Maximum cine memory up to 63575 frames (B storage server) or 210.65s (M storage server)
(depends on the mode)
- Maximum 4D cine memory: 5423 volumes (SD8-1s)
- Retrospective storage (1-120s pre-

5.2 Raw data processing

- B-mode:
TGC
Gain
Dynamic range
- Gray map
Tint map
iClear
- L/R Flip
U/D Flip
Rotation
- iTouch

LGC	
Dual live	Baseline
Auto Merge	Smooth
H Scale	Color map
Echo Boost	Priority
B/iNeedle	Dual Live
Smooth	Invert
Zoom	Velocity tag
Ref Lines	Glazing flow
Dehaze	• PW:
V1:1	Gain
ExtImage	Baseline
ZoneVue	Volume
Edge Enhance	Angle
• M-mode:	Speed
Gain	Dynamic range
Speed	Gray map
Dynamic Range	Tint Map
Gray Map	Display format
Tint Map	Invert
Display format	WF
• Color:	T/F Res

6 Measurement/Analysis and Report*

6.1 Generic measurements

• B-Mode	Color Vel
Depth	Strain Hist
Distance	Elas. Hist
Ellipse	Color Vel Profile
Trace	Elas.
Spline	Strain
Cross	TSM
Angle(2L)	Color Pixel Percent(Ellipse)
Angle(3P)	Color Pixel Percent(Trace)
Double Dist	Color Pixel Percent(Rectangle)
Trace Len	Color Pixel Percent(Recall)
Trace Len(Spline)	Smart Caliper
Parallel	Smart Trace
Distance P-L	-----
IMT	Volume
B-Profile	Volume(E+Dist.)
B-Hist(Ellipse)	Ratio(D)
B-Hist(Trace)	B Ratio
B-Hist(Spline)	-----
B-Hist(Rectangle)	Volume
	Volume

Volume(Ellipse)	● M-Mode
Volume(E+Dist.)	HR
Ratio(A)	HR(R-R)
Area1	Slope
Area2	Distance
Directional Ratio	Time
D1	Velocity
D2	
RAC	● D-Mode
Sag	PS
XS	ED
Volume Flow	PS/ED
Vas Area	Vel
TAMEAN	HR
TAMAX	HR(R-R)
Elas. Ratio	Time
A	Auto Trace
B	Manual Trace
Strain Ratio	Spline Trace
A	Acceleration
B	-----
	Ratio(Vel)

Ratio(VTI)	S/D
-----	D/S
Volume Flow	PI
Vas Area	RI
TAMEAN	PV
TAMAX	HR
6.2 AutoCalc	6.3 Clinical option measurement package
PS	● Abdomen
ED	B-Mode
MD	- Aorta Bif
PPG	- Aorta Aneurysm Status
TAMAX	- Shunt Diam
Vol Flow(TAMAX)	- Portal V Diam
TAMEAN	- M Portal V Diam
Vol Flow(TAMEAN)	- Splenic V Diam
Vas Diam	- PS Conflnc Diam
Vas Area	- Renal V Diam
DT	- SMV Diam
MPG	- IMV Diam
MMPG	- CHD
VTI	- GB L
AT	- GB H

- GB W
- Renal V Diam(Renal Transplant1)
- GB wall th
- Ureter Diam(Renal Transplant1)
- Cystic Duct
- Cortex(Renal Transplant2)
- CBD
- Renal V Diam(Renal Transplant2)
- Panc duct
- Panc head
- Panc neck
- Panc body
- Panc tail
- Appendix
- Appendix Wall
- Pylorus
- Pylorus Wall
- Renal L
- Renal H
- Renal W
- Cortex
- Adrenal L
- Adrenal H
- Adrenal W
- Ureter
- Cortex(Renal Transplant1)
- Pre-BL L
- Pre-BL H
- Pre-BL W
- Post-BL L
- Post-BL H
- Post-BL W
- Spleen L
- Spleen H
- Spleen W
- Spleen Area
- Skin-L.Capsule Dist.
- Hepatic Lesion1 Elas.
- Hepatic Lesion2 Elas.
- Hepatic Lesion3 Elas.
- LSM

- Rt DT(Insp) - SMA
- Rt DT(Expir) - Anterior-Posterior
- Lt DT(Insp) - Transverse
- Lt DT(Expir) - C Hepatic A
- Free Fluid - Anterior-Posterior
- Smart HRI - Transverse
- ----- - Proper Hepatic A
- Renal Vol - Anterior-Posterior
- Pre-BL Vol - Transverse
- Post-BL Vol - Hepatic A
- Mictur.Vol - Anterior-Posterior
- ----- - Transverse
- Aorta - Splenic A
- Anterior-Posterior - Anterior-Posterior
- Transverse - Transverse
- Outer Diameter - GDA
- Inner Diameter - Anterior-Posterior
- Outer Area - Transverse
- Inner Area - IMA
- Celiac Axis - Anterior-Posterior
- Anterior-Posterior - Transverse
- Transverse - Aorta Aneurysm

- Long Transverse
- Anterior-Posterior Splenic A Aneurysm
- Transverse Long
- Celiac A Aneurysm Anterior-Posterior
- Long Transverse
- Anterior-Posterior GDA Aneurysm
- Transverse Long
- SMA Aneurysm Anterior-Posterior
- Long Transverse
- Anterior-Posterior IMA Aneurysm
- Transverse Long
- C Hepatic A Aneurysm Anterior-Posterior
- Long Transverse
- Anterior-Posterior EVAR Residual Aneurysm
- Transverse Sac(2D)
- Proper Hepatic A Aneurysm Anterior-Posterior
- Long Transverse
- Anterior-Posterior EVAR Inflow(2D)
- Transverse Anterior-Posterior
- Hepatic A Aneurysm Transverse
- Long EVAR Graft Body(2D)
- Anterior-Posterior Anterior-Posterior

- Transverse
- Transverse
- EVAR Limb(2D)
- Outer Diameter
- Anterior-Posterior
- Inner Diameter
- Transverse
- Outer Area
- EVAR Outflow(2D)
- Inner Area
- Anterior-Posterior
- ABD Stenosis 3(2D)
- Transverse
- Anterior-Posterior
- Anterior-Posterior
- Transverse
- Transverse
- Aortic Bypass Graft Anast(2D)
- Transverse
- Anterior-Posterior
- Outer Diameter
- Transverse
- Inner Diameter
- Aortic Bypass Graft Graft(2D)
- Outer Area
- Anterior-Posterior
- Inner Area
- Transverse
- ABD Stenosis 4(2D)
- Anterior-Posterior
- Anterior-Posterior
- Transverse
- Transverse
- Outer Diameter
- Inner Diameter
- Inner Diameter
- Outer Area
- Outer Area
- Inner Area
- IVC
- ABD Stenosis 2(2D)
- Anterior-Posterior
- Anterior-Posterior
- Transverse

- Checklist - L
- Hepatic V(2D) - H
- Anterior-Posterior - W
- Transverse - Hepatic Lesion 1
- Lt Hepatic V(2D) - d1
- Anterior-Posterior - d2
- Transverse - d3
- M Hepatic V(2D) - Hepatic Lesion 2
- Anterior-Posterior - d1
- Transverse - d2
- Rt Hepatic V(2D) - d3
- Anterior-Posterior - Hepatic Lesion 3
- Transverse - d1
- Liver - d2
- L - d3
- H - Hepatic Cyst 1
- W - d1
- R Liver Lobe - d2
- L - d3
- H - Hepatic Cyst 2
- W - d1
- L Liver Lobe - d2

- d3 - GB Finding 4
- Hepatic Cyst 3 - d1
- d1 - d2
- d2 - d3
- d3 - GB Finding 5
- GB - d1
- GB L - d2
- GB H - d3
- GB W - Panc Finding 1
- GB wall th - d1
- GB Finding 1 - d2
- d1 - d3
- d2 - Panc Finding 2
- d3 - d1
- GB Finding 2 - d2
- d1 - d3
- d2 - Panc Finding 3
- d3 - d1
- GB Finding 3 - d2
- d1 - d3
- d2 - Panc Finding 4
- d3 - d1

- d2 - d3
- d3 - Renal Lesion 3
- Panc Finding 5 - d1
- d1 - d2
- d2 - d3
- d3 - Renal Cyst 1
- Kidney - d1
- Renal L - d2
- Renal H - d3
- Renal W - Renal Cyst 2
- Cortex - d1
- Adrenal - d2
- Adrenal L - d3
- Adrenal H - Renal Cyst 3
- Adrenal W - d1
- Renal Lesion 1 - d2
- d1 - d3
- d2 - Kidney(Superior)
- d3 - H
- Renal Lesion 2 - W
- d1 - Kidney(Mid)
- d2 - H

- W - H
- Kidney(Inferior) - W
- H - Finding 2(Renal Transplant1)
- W - L
- Renal A - H
- Long - W
- Anterior-Posterior - Finding 3(Renal Transplant1)
- Transverse - L
- Renal A Aneurysm - H
- Long - W
- Anterior-Posterior - Finding 4(Renal Transplant1)
- Transverse - L
- Kidney(Renal Transplant1) - H
- L - W
- H - Finding 5(Renal Transplant1)
- W - L
- Adrenal(Renal Transplant1) - H
- L - W
- H - Finding 6(Renal Transplant1)
- W - L
- Finding 1(Renal Transplant1) - H
- L - W

- Renal Transplant 1(2D) - H
- Cortex(Renal Transplant1) - W
- Renal V Diam(Renal Transplant1) - Finding 3(Renal Transplant1)
- Ureter Diam(Renal Transplant1) - L
- Kidney(Renal Transplant1) - W
- L - Finding 4(Renal Transplant1)
- H -
- W - L
- Adrenal(Renal Transplant1) - H
- L - W
- H - Finding 5(Renal Transplant1)
- W -
- Finding 1(Renal Transplant1) - L
- - H
- L - W
- H - Finding 6(Renal Transplant1)
- W -
- Finding 2(Renal Transplant1) - L
- - H
- L - W

- Kidney(Renal Transplant2) - H
- L - W
- H - Finding 5(Renal Transplant2)
- W - L
- Adrenal(Renal Transplant2) - H
- L - W
- H - Finding 6(Renal Transplant2)
- W - L
- Finding 1(Renal Transplant2) - H
- L - W
- H - Renal Transplant 2(2D)
- W - Cortex(Renal Transplant2)
- Finding 2(Renal Transplant2) - Renal V Diam(Renal
- L - Transplant2)
- H - Ureter Diam(Renal
- W - Transplant2)
- Finding 3(Renal Transplant2) - Kidney(Renal Transplant2)
- L - L
- H - H
- W - W
- Finding 4(Renal Transplant2) - Adrenal(Renal Transplant2)
- L - L

- H - Finding 5(Renal)
- W - Transplant2)
- Finding 1(Renal) - L
- Transplant2) - H
- L - W
- H - Finding 6(Renal)
- W - Transplant2)
- Finding 2(Renal) - L
- Transplant2) - H
- L - W
- H - Bladder
- W - Pre-BL L
- Finding 3(Renal) - Pre-BL H
- Transplant2) - Pre-BL W
- L - Post-BL L
- H - Post-BL H
- W - Post-BL W
- Finding 4(Renal) - Spleen
- Transplant2) - Spleen L
- L - Spleen H
- H - Spleen W
- W - Spleen Area

- Hepatic Lesion1 ElasRatio
- A
- B
- Hepatic Lesion2 ElasRatio
- A
- B
- Hepatic Lesion3 ElasRatio
- A
- B
-
- M-Mode
- Rt DT(Insp M)
- Rt DT(Expir M)
- Lt DT(Insp M)
- Lt DT(Expir M)
- RDE(QB)
- RDE(DB)
- LDE(QB)
- LDE(DB)
-
- D-Mode
- Aorta
- Celiac Axis
- SMA
- C Hepatic A
- Proper Hepatic A
- Hepatic A
- Splenic A
- GDA
- IMA
- Aorta(Post)
- Celiac Axis(Post)
- SMA(Post)
- C Hepatic A(Post)
- Proper Hepatic A(Post)
- Hepatic A(Post)
- Splenic Artery(Post)
- GDA(Post)
- IMA(Post)
- EVAR Residual Aneurysm Sac
- EVAR Inflow
- EVAR Graft Body
- EVAR Limb
- EVAR Outflow

- Aortic Bypass Graft Anast
- Aortic Bypass Graft Graft
- IVC Reflux
- IVC
- Hepatic V
- Lt Hepatic V
- M Hepatic V
- Rt Hepatic V
- Portal V
- M Portal V
- Splenic V
- Renal V
- SMV
- IMV
- Hepatic A Anast(Liver Transplant)
- Hepatic V Anast(Liver Transplant)
- Portal V Anast(Liver Transplant)
- IVC(Liver Transplant)
- Hep V Confl(Liver Transplant)
- Donor IVC(Liver Transplant)
- Renal A
- Ren A Org
- M Renal A
- Renal A1
- Renal A2
- Hilum
- Interlobar A
- Arcuate A
- Segment A
- Artery Anast(Renal Transplant1)
- Artery Anast 2(Renal Transplant1)
- Vein Anast(Renal Transplant1)
- Vein Anast 2(Renal Transplant1)
- Renal A(Renal Transplant1)
- Renal A1(Renal Transplant1)
- Renal A2(Renal Transplant1)
- Hilum(Renal Transplant1)
- Interlobar A(Renal Transplant1)
- Arcuate A(Renal Transplant1)
- Segmental A(Renal Transplant1)
- Renal Vein 1(Renal Transplant1)

- Renal Vein 2(Renal Transplant1) - Sten
- Artery Anast(Renal Transplant2) - Post Sten
- Artery Anast 2(Renal Transplant2) - ABD Stenosis 2
- Vein Anast(Renal Transplant2) - Pre Sten
- Vein Anast 2(Renal Transplant2) - Sten
- Renal A(Renal Transplant2) - Post Sten
- Renal A1(Renal Transplant2) - ABD Stenosis 3
- Renal A2(Renal Transplant2) - Pre Sten
- Hilum(Renal Transplant2) - Sten
- Interlobar A(Renal Transplant2) - Post Sten
- Arcuate A(Renal Transplant2) - ABD Stenosis 4
- Segmental A(Renal Transplant2) - Pre Sten
- Renal Vein 1(Renal Transplant2) - Sten
- Renal Vein 2(Renal Transplant2) - Renal Transplant 1(Doppler)
- TIPS - Artery Anast(Renal Transplant1)
- ----- - SMA/Ao
- CA/Ao - Artery Anast 2(Renal Transplant1)
- ----- - Vein Anast(Renal Transplant1)
- ABD Stenosis 1 - Vein Anast 2(Renal Transplant1)
- Pre Sten - Vein Anast 2(Renal Transplant1)

- Transplant1)
 - Renal A(Renal Transplant1)
 - Renal A1(Renal
- Transplant1)
 - Renal A2(Renal
- Transplant1)
 - Hilum(Renal Transplant1)
 - Interlobar A(Renal
- Transplant1)
 - Arcuate A(Renal
- Transplant1)
 - Segmental A(Renal
- Transplant1)
 - Renal Vein 1(Renal
- Transplant1)
 - Renal Vein 2(Renal
- Renal Transplant 2(Doppler)
 - Artery Anast(Renal
- Transplant2)
 - Artery Anast 2(Renal
- Vein Anast(Renal
- Transplant2)
 - Vein Anast 2(Renal
- Transplant2)
 - Renal A(Renal Transplant2)
 - Renal A1(Renal
- Transplant2)
 - Renal A2(Renal
- Transplant2)
 - Hilum(Renal Transplant2)
- Transplant2)
 - Interlobar A(Renal
- Transplant2)
 - Arcuate A(Renal
- Transplant2)
 - Segmental A(Renal
- Transplant2)
 - Renal Vein 1(Renal
- Transplant2)
 - Renal Vein 2(Renal
- Renal Vein 1(Renal
- Renal Vein 2(Renal
- Gynecology
- B-Mode

- UT L
- UT H
- UT W
- Endo
- Cervix L
- Cervix H
- Cervix W
- Ovary L
- Ovary H
- Ovary W
- Follicle1 L
- Follicle1 W
- Follicle1 H
- Follicle2 L
- Follicle2 W
- Follicle2 H
- Follicle3 L
- Follicle3 W
- Follicle3 H
- Follicle4 L
- Follicle4 W
- Follicle4 H
- Follicle5 L
- Follicle5 W
- Follicle5 H
- Follicle6 L
- Follicle6 W
- Follicle6 H
- Follicle7 L
- Follicle7 W
- Follicle7 H
- Follicle8 L
- Follicle8 W
- Follicle8 H
- Follicle9 L
- Follicle9 W
- Follicle9 H
- Follicle10 L
- Follicle10 W
- Follicle10 H
- Follicle11 L
- Follicle11 W
- Follicle11 H
- Follicle12 L

- Follicle12 W - PVA(R)
- Follicle12 H - PVA(Va)
- Follicle13 L - PUA(R)
- Follicle13 W - PUA(Va)
- Follicle13 H - BPW-SP Dist.(R)
- Follicle14 L - BPW-SP Dist.(Va)
- Follicle14 W - Cx-SP Dist.(R)
- Follicle14 H - Cx-SP Dist.(Va)
- Follicle15 L - RA-SP Dist.(R)
- Follicle15 W - RA-SP Dist.(Va)
- Follicle15 H - Shuttle(R)
- Follicle16 L - Shuttle(Va)
- Follicle16 W - Rectocele Depth
- Follicle16 H - Intus. Depth
- DWT - ARA(R)
- BSD(R) - ARA(Va)
- BSD(Va) - ARA(C)
- RVA(R) - LH AP Diam(R)
- RVA(Va) - LH AP Diam(Va)
- UTA(R) - LH AP Diam(C)
- UTA(Va) - LH Lateral Diam(R)
- URA - LH Lateral Diam(Va)

- LH Lateral Diam(C) - Fibroid1 Elas.
- LH Area(R) - Fibroid2 Elas.
- LH Area(Va) - Fibroid3 Elas.
- LH Area(C) - -----
- LA Angle(R) - UT Vol
- LA Angle(Va) - UT SUM
- LA Angle(C) - UT-L/CX-L
- LA Thickness(R) - Ovary Vol
- LA Thickness(Va) - Follicle1
- LA Thickness(C) - Follicle2
- LUG(R) - Follicle3
- LUG(Va) - Follicle4
- LUG(C) - Follicle5
- GYN Lesion1 Strain - Follicle6
- GYN Lesion2 Strain - Follicle7
- GYN Lesion3 Strain - Follicle8
- Lesion1 Elas. - Follicle9
- Lesion2 Elas. - Follicle10
- Lesion3 Elas. - Follicle11
- Fibroid1 Strain - Follicle12
- Fibroid2 Strain - Follicle13
- Fibroid3 Strain - Follicle14

- Follicle15 - d2
- Follicle16 - d3
- Mean DWT - Fibroid 3
- BND - d1
- IAS Damage - d2
- EAS Damage - d3
- ----- - Uterine Finding 1
- Uterus - d1
- UT L - d2
- UT H - d3
- UT W - Uterine Finding 2
- Endo - d1
- Uterine Cervix - d2
- Cervix L - d3
- Cervix H - Uterine Finding 3
- Cervix W - d1
- Fibroid 1 - d2
- d1 - d3
- d2 - Uterine Finding 4
- d3 - d1
- Fibroid 2 - d2
- d1 - d3

- Uterine Finding 5 - d2
- d1 - d3
- d2 - Ovarian Finding 1
- d3 - d1
- Uterine Finding 6 - d2
- d1 - d3
- d2 - Ovarian Finding 2
- d3 - d1
- Ovary - d2
- Ovary L - d3
- Ovary H - Ovarian Finding 3
- Ovary W - d1
- Ovarian Cyst 1 - d2
- d1 - d3
- d2 - Ovarian Finding 4
- d3 - d1
- Ovarian Cyst 2 - d2
- d1 - d3
- d2 - Ovarian Finding 5
- d3 - d1
- Ovarian Cyst 3 - d2
- d1 - d3

- Ovarian Finding 6
- Follicle5 W
- d1
- Follicle5 H
- d2
- Follicle6
- d3
- Follicle6 L
- Follicle1
- Follicle6 W
- Follicle1 L
- Follicle6 H
- Follicle1 W
- Follicle7
- Follicle1 H
- Follicle7 L
- Follicle2
- Follicle7 W
- Follicle2 L
- Follicle7 H
- Follicle2 W
- Follicle8
- Follicle2 H
- Follicle8 L
- Follicle3
- Follicle8 W
- Follicle3 L
- Follicle8 H
- Follicle3 W
- Follicle9
- Follicle3 H
- Follicle9 L
- Follicle4
- Follicle9 W
- Follicle4 L
- Follicle9 H
- Follicle4 W
- Follicle10
- Follicle4 H
- Follicle10 L
- Follicle5
- Follicle10 W
- Follicle5 L
- Follicle10 H

- Follicle11
- Follicle11 L
- Follicle11 W
- Follicle11 H
- Follicle12
- Follicle12 L
- Follicle12 W
- Follicle12 H
- Follicle13
- Follicle13 L
- Follicle13 W
- Follicle13 H
- Follicle14
- Follicle14 L
- Follicle14 W
- Follicle14 H
- Follicle15
- Follicle15 L
- Follicle15 W
- Follicle15 H
- Follicle16
- Follicle16 L
- Follicle16 W
- Follicle16 H
- GYN Lesion 1
- d1
- d2
- d3
- GYN Lesion 2
- d1
- d2
- d3
- GYN Lesion 3
- d1
- d2
- d3
- Residual Urine
- BL Height
- BL Depth
- GYN Lesion1 Strain Ratio
- A
- B
- GYN Lesion2 Strain Ratio
- A

- B - Fibroid1 Elas. Ratio
- GYN Lesion3 Strain Ratio - A
- A - B
- B - Fibroid2 Elas. Ratio
- Lesion1 Elas. Ratio - A
- A - B
- B - Fibroid3 Elas. Ratio
- Lesion2 Elas. Ratio - A
- A - B
- B -
- Lesion3 Elas. Ratio M-Mode
- A -
- B D-Mode
- Fibroid1 Strain Ratio • Obstetrics
- A B-Mode
- B GS
- Fibroid2 Strain Ratio - YS L
- A CRL
- B NT
- Fibroid3 Strain Ratio - BPD
- A OFD
- B HC

- AC
- APTD
- FL
- TTD
- TAD
- FTA
- APAD
- THD
- TCD
- HrtC
- CM
- TC
- IT
- Umb VD
- LVW
- F-kidney L
- HW
- Mat Kidney
- OOD
- Cervix L
- IOD
- AF
- HUM
- NF
- Ulna
- Orbit
- RAD
- PL Thickness
- Tibia
- Sac Diam1
- FIB
- Sac Diam2
- CLAV
- Sac Diam3
- Vertebrae
- AF1
- MP
- AF2
- Foot
- AF3
- NBL
- AF4
- Ear
- LVIDd

- LVIDs - Ao Asc Diam(Z-Score)
- LV Diam - Ao Desc Diam(Z-Score)
- LA Diam - Duct Art Diam(Z-Score)
- RVIDd - TV Diam(Z-Score)
- RVIDs - LPA Diam(Z-Score)
- RV Diam - RPA Diam(Z-Score)
- RA Diam - IVC Diam(Z-Score)
- IVSd - AV Diam(Z-Score)
- IVSs - MPA Diam(Z-Score)
- IVS - RV Diam(Z-Score)
- LV Area - LV Diam(Z-Score)
- LA Area - RV Area(Z-Score)
- RV Area - LV Area(Z-Score)
- RA Area - RVIDd(Z-Score)
- Ao Diam - LVIDd(Z-Score)
- MPA Diam - UT L
- LVOT Diam - UT H
- RVOT Diam - UT W
- Facial Angle - Endo
- HrtA - AH
- MV Diam(Z-Score) - PH
- PV Diam(Z-Score) - 3th Ventricle

- NT Above Cord
- NT Below Cord
- Mandible
- Prenasal th
- Heart AP
- Heart T
- LV Width
- LV Length
- RV Width
- RV Length
- LA Width
- RA Width
- LVWd
- LVWs
- RVWd
- RVWs
- AV Diam
- AV Area
- PV Area
- F-kidney H
- F-kidney W
- Lung
- Stomach
- YS H
- YS W
- Amniotic Sac L
- Amniotic Sac H
- Amniotic Sac W
- Ovary Cyst L
- Ovary Cyst H
- Ovary Cyst W
- UT AW
- UT PW
- CSP
- FMF
- MMF
- Lung CCAM L
- Lung CCAM H
- Lung CCAM W
- AD
- Lliac Wing Angle
- FAGL
- FAG
- Intestinum Crassum

- Liver Length - AoD/MPAD
- Rib Length - LAD/AoD
- Shoulder Blade - UT Vol
- ----- - UT SUM
- MAD - UT-L/CX-L
- Mean Sac Diam - -----
- AFI - AFI
- EFW - AF1
- EFW2 - AF2
- HC/AC(Campbell) - AF3
- FL/AC - AF4
- FL/BPD - Uterus
- AXT - UT L
- CI - UT H
- FL/HC(Hadlock) - UT W
- AC(c) - Endo
- HC(c) -
- HrtC/TC - M-Mode
- TCD/AC - FHR (M)
- LVW/HW - LVIDd
- LVD/RVD - LVIDs
- LAD/RAD - RVIDd

- RVIDs
- Fetal Ao
- IVSd
- Desc Aorta
- IVSs
- Ut A
- RVIDd(Z-Score)
- Ovarian A
- LVIDd(Z-Score)
- FHR (Doppler)
- MVE
- Asc Aorta
- TVE
- RVOT
- AVE
- LVOT
- MAPSE
- MV E
- TAPSE
- MV A
- LV ICT
- TV E
- LV IRT
- TV A
- LV ET
- MV E'
- RV ICT
- MV A'
- RV IRT
- MV S'
- RV ET
- TV E'
-
- TV A'
- D-Mode
- TV S'
- Umb A
- AV PV
- Duct Veno
- AV VTI
- Placenta A
- PV PV
- MCA
- PV VTI

- Duct Art PV B-Mode
- Duct Art VTI - RVAWd(2D)
- AV TPV - RVAWs(2D)
- PV TPV - RVDd(2D)
- Duct Art TPV - RVDs(2D)
- Thoracic Aorta - IVSd(2D)
- Hepatic Vein - IVSs(2D)
- IVC - LVIDd(2D)
- Umb V - LVIDs(2D)
- Ovary - LVPWd(2D)
- Endometrium - LVPWs(2D)
- Cervical Cancer - Diastole(2D)
- Fibroid - Systole(2D)
- Duct Art - LVd apical
- ICA - LVLs apical
- Celiac A - LVAd apical
- ----- - LVAs apical
- MV E/A - LVAd sax MV
- TV E/A - LVAs sax MV
- MV E/E' - LVAd sax Endo
- TV E/E' - LVAd sax Epi
- Cardiology - LV Major

- LV Minor - Ao Sinus Diam(2D)
- LV Area(d) - Ao st junct(2D)
- LV Area(s) - AVA
- HR(2D) - Ao Arch Diam(2D)
- RA Major - Ao Asc Diam(2D)
- RA Minor - Ao Desc Diam(2D)
- RA Area - Duct Art Diam
- RA Vol(A4C) - Post Ductal
- RAP - Pre Ductal
- RV Area(d) - MCS(2D)
- RV Area(s) - MV Diam
- RV Major - MV EPSS(2D)
- RV Minor - MVA
- LA Diam(2D) - TV Diam
- LA Major - TVA
- LA Minor - PV Diam
- LA Area - RVOT Diam
- LVOT Diam - MPA Diam(2D)
- Ao Diam(2D) - RPA Diam(2D)
- ACS(2D) - LPA Diam(2D)
- AV Diam - IVC Diam(Expir)
- Ao Isthmus(2D) - IVC Diam(Insp)

- SVC Diam(Expir) - LVIDd(2D)
- SVC Diam(Insp) - LVPWd(2D)
- LCA Diam - IVSs(2D)
- RCA Diam - LVIDs(2D)
- PEd(2D) - LVPWs(2D)
- PEs(2D) - HR(2D)
- VSD Diam - Simpson
- ASD Diam - A4Cd
- PDA Diam - A4Cs
- PFO Diam - A2Cd
- AutoEF - A2Cs
- Rt DT(Insp) - HR(2D)
- Rt DT(Expir) - Mod.Simpson
- Lt DT(Insp) - LVLD apical
- Lt DT(Expir) - LVLs apical
- ----- - LVAd sax MV
- LA/Ao(2D) - LVAs sax MV
- ----- - LVAd sax PM
- LV(2D) - LVAs sax PM
- Diastole(2D) - HR(2D)
- Systole(2D) - S-P Ellipse
- IVSd(2D) - LVLD apical

- LVAd apical - LV Mass(A-L)
- LVLs apical - LVld apical
- LVAs apical - LVAd sax Epi
- HR(2D) - LVAd sax Endo
- B-P Ellipse - LV Mass(T-E)
- LVIDd(2D) - LVAd sax Epi
- LVAd sax MV - LVAd sax Endo
- LVIDs(2D) - a
- LVAs sax MV - d
- LVAd apical - LA Vol(Simp)
- LVAs apical - LA Vol(A2C)
- HR(2D) - LA Vol(A4C)
- Bullet - LA Vol(A-L)
- LVld apical - LA apical
- LVLs apical - LAA(A2C)
- LVAd sax MV - LAA(A4C)
- LVAs sax MV - MVA(VTI)
- HR(2D) - LVOT Diam
- LV Mass(Cube-2D) - LVOT VTI
- IVSd(2D) - MV VTI
- LVIDd(2D) - AVA(VTI)
- LVPWd(2D) - LVOT Diam

- LVOT VTI - PISA AR
- AV VTI - AR Rad
- CO(LVOT) - AR Als Vel
- LVOT Diam - AR VTI
- LVOT VTI - PISA TR
- AV HR - TR Rad
- CO(RVOT) - TR Als Vel
- RVOT Diam - TR VTI
- RVOT VTI - PISA PR
- PV HR - PR Rad
- CO(MV) - PR Als Vel
- MV Diam - PR VTI
- MV VTI - Qp/Qs
- MV HR - LVOT Diam
- CO(TV) - LVOT VTI
- TV Diam - RVOT Diam
- TV VTI - RVOT VTI
- TV HR - Z-Scores (\leq 3Y) (2D)
- PISA MR - AV Diam
- MR Rad - Ao Sinus Diam
- MR Als Vel - Ao st junct
- MR VTI - PV Diam

-	Ao Arch IA-LCA	-	RV Area(d) A4C
-	Ao Arch LCA-LSA	-	RV Area(s) A4C
-	Ao Arch after LSA	-	RVd Major A4C
-	Ao Isthmus	-	RVs Major A4C
-	Thoracic Ao Diam	-	RVd Minor (basal) A4C
-	IVC Diam	-	RVd Minor (midcavity) A4C
-	MV Diam	-	LV Area(d) A2C
-	TV Diam	-	LV Area(s) A2C
-	MPA Diam	-	LVIDd A2C(2D)
-	RPA Diam	-	LVIDs A2C(2D)
-	LPA Diam	-	
-	Z-Scores (<18Y) (2D)		M-Mode
-	LV Area(d) A4C	-	RVAWd(M)
-	LV Area(s) A4C	-	RVAWs(M)
-	LVIDd A4C(2D)	-	RVDD(M)
-	LVIDs A4C(2D)	-	RVDs(M)
-	LA AP Diam A4C	-	Ao Arch Diam(M)
-	LA LL Diam A4C	-	Ao Asc Diam(M)
-	LA Area A4C	-	Ao Desc Diam(M)
-	RA AP Diam A4C	-	Ao Diam(M)
-	RA LL Diam A4C	-	Ao Isthmus(M)
-	RA Area A4C	-	Ao Sinus Diam(M)

- Ao st junct(M) - MV E-F Slope
- ACS(M) - MV EPSS(M)
- HR(M) - PEd(M)
- IVSd(M) - PEs(M)
- IVSs(M) - RPA Diam(M)
- LA Diam(M) - RVET(M)
- LPA Diam(M) - RVOT Diam
- Diastole(M) - RVPEP(M)
- Systole(M) - MAPSE
- LVET(M) - TAPSE
- LVIDd(M) - MV ALL
- LVIDs(M) - IVC Diam(Insp)(M)
- LVOT Diam - IVC Diam(Expir)(M)
- LVPEP(M) - SVC Diam(Insp)(M)
- LVPWd(M) - SVC Diam(Expir)(M)
- LVPWs(M) - Rt DT(Expir M)
- MCS(M) - Rt DT(Expir M)
- MPA Diam(M) - Lt DT(Expir M)
- MV A Amp - Lt DT(Expir M)
- MV E Amp - RDE(QB)
- MV D-E Slope - RDE(DB)
- MV D-E Amp - LDE(QB)

- LDE(DB) - IVSd(M)
- ----- - LVPWd(M)
- LA/Ao(M) - Z-Scores (<18Y) (M)
- ----- - LVIDd(M)
- LV(M) - LVIDs(M)
- Diastole(M) -
- Systole(M) D-Mode
- IVSd(M) - MV Aa(lateral)
- LVIDd(M) - MV Aa(medial)
- LVPWd(M) - AAo Vmax
- IVSs(M) - AV VTI
- LVIDs(M) - AV HR
- LVPWs(M) - AV Vmax
- HR(M) - AR DecT
- LV Mass(Cube-M) - AR Time
- IVSd(M) - AR PHT
- LVIDd(M) - AR Ved
- LVPWd(M) - AR Vmax
- LV Tei Index(M) - AR VTI
- MV C-O dur(M) - MV ARa(lateral)
- LVET(M) - MV ARa(medial)
- Z-Scores ($\leq 3Y$) (M) - ASD Vmax

- AV AccT
- AV DecT
- Coarc Post-Duct
- Coarc Pre-Duct
- DAO Vmax
- MV DRa(lateral)
- MV DRa(medial)
- MV Ea(lateral)
- MV Ea(medial)
- IVC Vel(Expir)
- IVC Vel(Insp)
- IVCT
- LPA Vmax
- LVET(Doppler)
- LVOT AccT
- LVOT VTI
- LVOT Vmax
- LVPEP(Doppler)
- MPA Vmax
- dP/dt
- Tau(BAI)
- MR VTI
- MR Vmax
- MS Vmax
- MV A Dur
- MV A Vel
- MV A VTI
- MV AccT
- MV DecT
- MV E Dur
- MV E Vel
- MV E VTI
- IVRT
- MV VTI
- MV HR
- MV Vmax
- PVein A Dur
- PVein A Vel
- PVein D Vel
- PVein D VTI
- PVein DecT
- PVein S Vel
- PVein S VTI
- PDA Vel(d)

- PDA Vel(s) - TV A Dur
- PR PHT - TV A Vel
- PR VTI - TV AccT
- PR Ved - TV DecT
- PR Vmax - TV E Vel
- PR DecT - TV VTI
- PV AccT - TV HR
- PV VTI - TV Vmax
- PV HR - VSD Vmax
- PV Vmax - Hepatic V S Vel
- RAP - Hepatic V D Vel
- RPA Vmax - -----
- RVET(Doppler) - MV E/A
- RVOT Vmax - MVA(PHT)
- RVOT VTI - TV E/A
- RVPEP(Doppler) - TVA(PHT)
- MV Sa(lateral) - -----
- MV Sa(medial) - LV Tei Index(Doppler)
- SVC Vel(Expir) - MV C-O dur(Doppler)
- SVC Vel(Insp) - LVET(Doppler)
- TR VTI - RVSP
- TR Vmax - TR Vmax

- RAP - MV VTI
- PAEDP - MV HR
- PR Ved - CO(TV)
- RAP - TV Diam
- MVA(VTI) - TV VTI
- LVOT Diam - TV HR
- LVOT VTI - RV Tei Index
- MV VTI - TV C-O dur
- AVA(VTI) - RVET(Doppler)
- LVOT Diam - PISA MR
- LVOT VTI - MR Rad
- AV VTI - MR Als Vel
- CO(LVOT) - MR VTI
- LVOT Diam - PISA AR
- LVOT VTI - AR Rad
- AV HR - AR Als Vel
- CO(RVOT) - AR VTI
- RVOT Diam - PISA TR
- RVOT VTI - TR Rad
- PV HR - TR Als Vel
- CO(MV) - TR VTI
- MV Diam - PISA PR

- PR Rad - Cortex(Renal Transplant2)
- PR Als Vel - Renal V Diam(Renal
- PR VTI - Transplant2)
- Qp/Qs - Ureter Diam(Renal Transplant2)
- LVOT Diam - Prostate L
- LVOT VTI - Prostate H
- RVOT Diam - Prostate W
- RVOT VTI - Seminal L
- Urology - Seminal H
- B-Mode - Seminal W
- Renal L - Urethra
- Renal H - Pre-BL L
- Renal W - Pre-BL H
- Cortex - Pre-BL W
- Adrenal L - Post-BL L
- Adrenal H - Post-BL H
- Adrenal W - Post-BL W
- Ureter - Testicular L
- Cortex(Renal Transplant1) - Testicular H
- Renal V Diam(Renal - Testicular W
- Transplant1) - Epididymis L
- Ureter Diam(Renal Transplant1) - Epididymis H

- Epididymis W - Cortex
- Scrotal Wall - Adrenal
- Testis V(2D) - Adrenal L
- Testis V(Valsalva 2D) - Adrenal H
- Prostate Mass1 Strain - Adrenal W
- Prostate Mass2 Strain - Renal Lesion 1
- Prostate Mass3 Strain - d1
- Prostate Mass1 Elas. - d2
- Prostate Mass2 Elas. - d3
- Prostate Mass3 Elas. - Renal Lesion 2
- ----- - d1
- Renal Vol - d2
- Prostate Vol - d3
- Pre-BL Vol - Renal Lesion 3
- Post-BL Vol - d1
- Mictur.Vol - d2
- Testicular Vol - d3
- ----- - Renal Cyst 1
- Kidney - d1
- Renal L - d2
- Renal H - d3
- Renal W - Renal Cyst 2

- d1 - Anterior-Posterior
- d2 - Transverse
- d3 - Kidney(Renal Transplant1)
- Renal Cyst 3 - L
- d1 - H
- d2 - W
- d3 - Adrenal(Renal Transplant1)
- Kidney(Superior) - L
- H - H
- W - W
- Kidney(Mid) - Finding 1(Renal Transplant1)
- H - L
- W - H
- Kidney(Inferior) - W
- H - Finding 2(Renal Transplant1)
- W - L
- Renal A - H
- Long - W
- Anterior-Posterior - Finding 3(Renal Transplant1)
- Transverse - L
- Renal A Aneurysm - H
- Long - W

- Finding 4(Renal Transplant1) - Adrenal(Renal Transplant1)
- L - L
- H - H
- W - W
- Finding 5(Renal Transplant1) - Finding 1(Renal Transplant1)
- L - L
- H - L
- W - H
- Finding 6(Renal Transplant1) - W
- L - Finding 2(Renal Transplant1)
- H - L
- W - L
- Renal Transplant 1(2D) - H
- Cortex(Renal Transplant1) - W
- Renal V Diam(Renal Transplant1) - Finding 3(Renal Transplant1)
- Ureter Diam(Renal Transplant1) - L
- Kidney(Renal Transplant1) - H
- L - W
- Finding 4(Renal Transplant1) - Finding 4(Renal Transplant1)
- H - L
- W - L

- H
- W
- Finding 5(Renal Transplant1)
- L
- H
- W
- Finding 3(Renal Transplant2)
- Finding 6(Renal Transplant1)
- L
- H
- L
- W
- Finding 4(Renal Transplant2)
- W
- L
- H
- W
- Finding 5(Renal Transplant2)
- L
- W
- L
- H
- W
- Finding 6(Renal Transplant2)
- W
- L
- H
- W

- Renal Transplant 2(2D) - H
- Cortex(Renal Transplant2) - W
- Renal V Diam(Renal Transplant2) - Finding 3(Renal Transplant2)
- Ureter Diam(Renal Transplant2) - L
- Kidney(Renal Transplant2) - W
- L - Finding 4(Renal Transplant2)
- H -
- W - L
- Adrenal(Renal Transplant2) - H
- L - W
- H - Finding 5(Renal Transplant2)
- W -
- Finding 1(Renal Transplant2) - L
- - H
- L - W
- H - Finding 6(Renal Transplant2)
- W -
- Finding 2(Renal Transplant2) - L
- - H
- L - W

- Prostate - d2
- Prostate L - d3
- Prostate H - Bladder
- Prostate W - Pre-BL L
- Prostate2 - Pre-BL H
- Long - Pre-BL W
- Anterior-Posterior - Post-BL L
- Coronal - Post-BL H
- Seminal Vesicle - Post-BL W
- Seminal L - Testis
- Seminal H - Testicular L
- Seminal W - Testicular H
- Prostate Mass 1 - Testicular W
- d1 - Testis Mass 1
- d2 - d1
- d3 - d2
- Prostate Mass 2 - d3
- d1 - Testis Mass 2
- d2 - d1
- d3 - d2
- Prostate Mass 3 - d3
- d1 - Testis Mass 3

-	d1	-	H
-	d2	-	W
-	d3	-	Epididymal Tail
-	Epididymis	-	L
-	Epididymis L	-	H
-	Epididymis H	-	W
-	Epididymis W	-	Prostate Mass1 Strain Ratio
-	Testicle(Superior)	-	A
-	H	-	B
-	W	-	Prostate Mass2 Strain Ratio
-	Testicle(Mid)	-	A
-	H	-	B
-	W	-	Prostate Mass3 Strain Ratio
-	Testicle(Inferior)	-	A
-	H	-	B
-	W	-	Prostate Mass1 Elas. Ratio
-	Epididymal Head	-	A
-	L	-	B
-	H	-	Prostate Mass2 Elas. Ratio
-	W	-	A
-	Epididymal Body	-	B
-	L	-	Prostate Mass3 Elas. Ratio

- A - Renal A2(Renal Transplant1)
- B - Hilum(Renal Transplant1)
- - Interlobar A(Renal Transplant1)
- M-Mode - Arcuate A(Renal Transplant1)
- Segmental A(Renal Transplant1)
- D-Mode - Renal Vein 1(Renal Transplant1)
- Renal A - Renal Vein 2(Renal Transplant1)
- Ren A Org - Artery Anast(Renal Transplant2)
- M Renal A - Artery Anast 2(Renal
- Renal A1 - Transplant2)
- Renal A2 - Vein Anast(Renal Transplant2)
- Hilum - Vein Anast 2(Renal Transplant2)
- Interlobar A - Renal A(Renal Transplant2)
- Arcuate A - Renal A1(Renal Transplant2)
- Segment A - Renal A2(Renal Transplant2)
- Artery Anast(Renal Transplant1) - Hilum(Renal Transplant2)
- Artery Anast 2(Renal - Interlobar A(Renal Transplant2)
- Transplant1) - Arcuate A(Renal Transplant2)
- Vein Anast(Renal Transplant1) - Segmental A(Renal Transplant2)
- Vein Anast 2(Renal Transplant1) - Renal Vein 1(Renal Transplant2)
- Renal A(Renal Transplant1) - Renal Vein 2(Renal Transplant2)
- Renal A1(Renal Transplant1) - Testis A

- Testis V - Arcuate A(Renal)
- Testis V(Valsalva) - Transplant1)
- Epididymis A - Segmental A(Renal)
- Epididymis V - Transplant1)
- ----- - Renal Vein 1(Renal)
- Renal Transplant 1(Doppler) - Transplant1)
- Artery Anast(Renal) - Renal Vein 2(Renal)
- Transplant1) - Transplant1)
- Artery Anast 2(Renal) - Renal Transplant 2(Doppler)
- Transplant1) - Artery Anast(Renal)
- Vein Anast(Renal) - Transplant2)
- Transplant1) - Artery Anast 2(Renal)
- Vein Anast 2(Renal) - Transplant2)
- Transplant1) - Vein Anast(Renal)
- Renal A(Renal Transplant1) - Transplant2)
- Renal A1(Renal) - Vein Anast 2(Renal)
- Transplant1) - Transplant2)
- Renal A2(Renal) - Renal A(Renal Transplant2)
- Transplant1) - Renal A1(Renal)
- Hilum(Renal Transplant1) - Transplant2)
- Interlobar A(Renal) - Renal A2(Renal)
- Transplant1) - Transplant2)

- Hilum(Renal Transplant2) - ECA IMT
- Interlobar A(Renal Transplant2) - CCA
- Arcuate A(Renal Transplant2) - Anterior-Posterior
- Segmental A(Renal Transplant2) - Transverse
- Renal Vein 1(Renal Transplant2) - Outer Diameter
- Renal Vein 2(Renal Transplant2) - Inner Diameter
- Renal Vein 1(Renal Transplant2) - Outer Area
- Renal Vein 2(Renal Transplant2) - Inner Area
- B-Mode - Bulb
- Vascular
 - B-Mode - Outer Diameter
 - CCA IMT - Inner Diameter
 - Bulb IMT - Outer Area
 - ICA IMT - Inner Area
 - ECA IMT - Carotid Bifurcation
 - ----- - Anterior-Posterior
 - ----- - Transverse
 - IMT - Outer Diameter
 - CCA IMT - Inner Diameter
 - Bulb IMT - Outer Area
 - ICA IMT - Inner Area

- ICA
 - Anterior-Posterior
 - Transverse
 - Outer Diameter
 - Inner Diameter
 - Outer Area
 - Inner Area
 - Innom A
- ECA
 - Anterior-Posterior
 - Transverse
 - Outer Diameter
 - Inner Diameter
 - Outer Area
 - Inner Area
 - Mammary A
- Vert A
 - Anterior-Posterior
 - Transverse
 - Outer Diameter
 - Inner Diameter
 - Outer Area
 - Inner Area
- CCA Aneurysm
- Subclav A
 - Long

- Anterior-Posterior
- Subclav A Aneurysm
- Transverse
- Long
- Bulb Aneurysm
- Anterior-Posterior
- Anterior-Posterior
- Transverse
- Long
- Anterior-Posterior
- Innom A Aneurysm
- Transverse
- Long
- Carotid Bifurcation Aneurysm
- Anterior-Posterior
- Anterior-Posterior
- Long
- Transverse
- Anterior-Posterior
- Mammary A Aneurysm
- Transverse
- Long
- ICA Aneurysm
- Anterior-Posterior
- Anterior-Posterior
- Long
- Transverse
- Anterior-Posterior
- Carotid Graft 1 Anast
- Transverse
- Long
- ECA Aneurysm
- Anterior-Posterior
- Anterior-Posterior
- Long
- Transverse
- Carotid Graft 1 Graft
- Transverse
- Long
- Vert A Aneurysm
- Anterior-Posterior
- Anterior-Posterior
- Long
- Transverse
- Carotid Graft 2 Anast
- Transverse
- Long

- Anterior-Posterior
- Carotid Stent 3
- Transverse
- Long
- Carotid Graft 2 Graft
- Anterior-Posterior
- Long
- Transverse
- Anterior-Posterior
- Carotid Stenosis 1
- Transverse
- Anterior-Posterior
- Carotid Graft 3 Anast
- Transverse
- Long
- Outer Diameter
- Anterior-Posterior
- Inner Diameter
- Transverse
- Outer Area
- Carotid Graft 3 Graft
- Inner Area
- Long
- Carotid Stenosis 2
- Anterior-Posterior
- Transverse
- Carotid Stent 1
- Outer Diameter
- Long
- Inner Diameter
- Anterior-Posterior
- Outer Area
- Transverse
- Inner Area
- Carotid Stent 2
- Carotid Stenosis 3
- Long
- Anterior-Posterior
- Anterior-Posterior
- Transverse
- Transverse
- Outer Diameter

- Inner Diameter - Outer Area
- Outer Area - Inner Area
- Inner Area - Radial A
- Carotid Stenosis 4 - Anterior-Posterior
- Anterior-Posterior - Transverse
- Transverse - Outer Diameter
- Outer Diameter - Inner Diameter
- Inner Diameter - Outer Area
- Outer Area - Inner Area
- Inner Area - Ulnar A
- Axill A - Anterior-Posterior
- Anterior-Posterior - Transverse
- Transverse - Outer Diameter
- Outer Diameter - Inner Diameter
- Inner Diameter - Outer Area
- Outer Area - Inner Area
- Inner Area - Axill A Aneurysm
- Brachial A - Long
- Anterior-Posterior - Anterior-Posterior
- Transverse - Transverse
- Outer Diameter - Brachial A Aneurysm
- Inner Diameter - Long

- Anterior-Posterior
- UE A Graft 2 Native Inflow
- Transverse
- Anterior-Posterior
- Radial A Aneurysm
- Transverse
- Long
- UE A Graft 2 Anast
- Anterior-Posterior
- Anterior-Posterior
- Transverse
- Transverse
- Ulnar A Aneurysm
- UE A Graft 2 Graft
- Long
- Anterior-Posterior
- Anterior-Posterior
- Transverse
- Transverse
- UE A Graft 2 Native Outflow
- Anterior-Posterior
- UE A Graft 1 Native Inflow
- Anterior-Posterior
- Transverse
- Transverse
- UE A Graft 1 Anast
- Anterior-Posterior
- Transverse
- Transverse
- UE A Graft 3 Native Inflow
- Anterior-Posterior
- UE A Graft 1 Antast
- Anterior-Posterior
- Transverse
- Transverse
- Anterior-Posterior
- Transverse
- UE A Graft 3 Anast
- Anterior-Posterior
- UE A Graft 1 Graft
- Anterior-Posterior
- Transverse
- Transverse
- UE A Graft 3 Anast
- Anterior-Posterior
- Anterior-Posterior
- Transverse
- Transverse
- UE A Graft 1 Native Outflow
- Anterior-Posterior
- Anterior-Posterior
- Transverse
- Transverse
- UE A Graft 3 Graft
- Anterior-Posterior
- UE A Graft 1 Native Outflow
- Anterior-Posterior
- Anterior-Posterior
- Transverse
- Transverse
- UE A Graft 3 Graft
- Anterior-Posterior
- UE A Graft 3 Native Outflow

- Anterior-Posterior
- Transverse
- UE A Stent 1
- Long
- Anterior-Posterior
- Transverse
- UE A Stent 2
- Long
- Anterior-Posterior
- Transverse
- UE A Stent 3
- Long
- Anterior-Posterior
- Transverse
- UE A Stenosis 1(2D)
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- C.Iliac A
- UE A Stenosis 2(2D)
- Anterior-Posterior

- Transverse - Outer Diameter
- Outer Diameter - Inner Diameter
- Inner Diameter - Outer Area
- Outer Area - Inner Area
- Inner Area - DFA
- Ex.Iliac A - Anterior-Posterior
- Anterior-Posterior - Transverse
- Transverse - Outer Diameter
- Outer Diameter - Inner Diameter
- Inner Diameter - Outer Area
- Outer Area - Inner Area
- Inner Area - SFA
- IIA - Anterior-Posterior
- Anterior-Posterior - Transverse
- Transverse - Outer Diameter
- Outer Diameter - Inner Diameter
- Inner Diameter - Outer Area
- Outer Area - Inner Area
- Inner Area - Pop A
- CFA - Anterior-Posterior
- Anterior-Posterior - Transverse
- Transverse - Outer Diameter

- Inner Diameter - Outer Area
- Outer Area - Inner Area
- Inner Area - PTib A
- TP Trunk A - Anterior-Posterior
- Anterior-Posterior - Transverse
- Transverse - Outer Diameter
- Outer Diameter - Inner Diameter
- Inner Diameter - Outer Area
- Outer Area - Inner Area
- Inner Area - Dors.Ped. A
- A.Tib A - Anterior-Posterior
- Anterior-Posterior - Transverse
- Transverse - Outer Diameter
- Outer Diameter - Inner Diameter
- Inner Diameter - Outer Area
- Outer Area - Inner Area
- Inner Area - C.Iliac A Aneurysm
- Peroneal A - Long
- Anterior-Posterior - Anterior-Posterior
- Transverse - Transverse
- Outer Diameter - Ex.Iliac A Aneurysm
- Inner Diameter - Long

- Anterior-Posterior
- TP Trunk A Aneurysm
- Transverse
- Long
- IIA Aneurysm
- Anterior-Posterior
- Long
- Transverse
- Anterior-Posterior
- A.Tib A Aneurysm
- Transverse
- Long
- CFA Aneurysm
- Anterior-Posterior
- Long
- Transverse
- Anterior-Posterior
- Peroneal A Aneurysm
- Transverse
- Long
- DFA Aneurysm
- Anterior-Posterior
- Long
- Transverse
- Anterior-Posterior
- PTib A Aneurysm
- Transverse
- Long
- SFA Aneurysm
- Anterior-Posterior
- Long
- Transverse
- Anterior-Posterior
- Dors.Ped. A Aneurysm
- Transverse
- Long
- Pop A Aneurysm
- Anterior-Posterior
- Long
- Transverse
- Anterior-Posterior
- LE A Graft 1 Native Inflow
- Transverse
- Anterior-Posterior

- Transverse - LE A Graft 3 Native Inflow
- LE A Graft 1 Anast - Anterior-Posterior
- Anterior-Posterior - Transverse
- Transverse - LE A Graft 3 Anast
- LE A Graft 1 Graft - Anterior-Posterior
- Anterior-Posterior - Transverse
- Transverse - LE A Graft 3 Graft
- LE A Graft 1 Native Outflow - Anterior-Posterior
- Anterior-Posterior - Transverse
- Transverse - LE A Graft 3 Native Outflow
- LE A Graft 2 Native Inflow - Anterior-Posterior
- Anterior-Posterior - Transverse
- Transverse - LE A Stent 1
- LE A Graft 2 Anast - Long
- Anterior-Posterior - Anterior-Posterior
- Transverse - Transverse
- LE A Graft 2 Graft - LE A Stent 2
- Anterior-Posterior - Long
- Transverse - Anterior-Posterior
- LE A Graft 2 Native Outflow - Transverse
- Anterior-Posterior - LE A Stent 3
- Transverse - Long

- Anterior-Posterior - Inner Area
- Transverse - LE A Stenosis 4(2D)
- LE A Stenosis 1(2D) - Anterior-Posterior
- Anterior-Posterior - Transverse
- Transverse - Outer Diameter
- Outer Diameter - Inner Diameter
- Inner Diameter - Outer Area
- Outer Area - Inner Area
- Inner Area - LE A Finding 1
- LE A Stenosis 2(2D) - Long
- Anterior-Posterior - Anterior-Posterior
- Transverse - Transverse
- Outer Diameter - LE A Finding 2
- Inner Diameter - Long
- Outer Area - Anterior-Posterior
- Inner Area - Transverse
- LE A Stenosis 3(2D) - LE A Finding 3
- Anterior-Posterior - Long
- Transverse - Anterior-Posterior
- Outer Diameter - Transverse
- Inner Diameter - LE A Finding 4
- Outer Area - Long

- Anterior-Posterior
- Ax V
- Transverse
- Anterior-Posterior
- LE A Finding 5
- Transverse
- Long
- Checklist
- Anterior-Posterior
- Brachial V
- Transverse
- Anterior-Posterior
- LE A Finding 6
- Transverse
- Long
- Checklist
- Anterior-Posterior
- Radial V
- Transverse
- Anterior-Posterior
- Int Jug V
- Transverse
- Anterior-Posterior
- Checklist
- Transverse
- Ulnar V
- Checklist
- Anterior-Posterior
- Innom V
- Transverse
- Anterior-Posterior
- Checklist
- Transverse
- Volar V
- Checklist
- Anterior-Posterior
- Subclav V
- Transverse
- Anterior-Posterior
- Checklist
- Transverse
- Cephalic V
- Checklist
- Anterior-Posterior

- Transverse
- BA Junction
- Checklist
- Anterior-Posterior
- Basilic V
- Transverse
- Checklist
- Anterior-Posterior
- Transverse
- Upper Arm Basilic V
- Checklist
- Anterior-Posterior
- CA Junction
- Transverse
- Checklist
- Anterior-Posterior
- Transverse
- Checklist
- Transverse
- Checklist
- Transverse
- Checklist
- Anterior-Posterior
- Upper Arm Cephalic V
- Transverse
- Checklist
- Anterior-Posterior
- Transverse
- Checklist
- Transverse
- Checklist
- Transverse
- Checklist
- Forearm Basilic V
- Anterior-Posterior
- Checklist
- Cephalic-Antecubital V
- Transverse
- Checklist
- Anterior-Posterior
- Transverse
- Checklist
- Transverse
- Digital V
- Anterior-Posterior
- Checklist
- Forearm Cephalic V
- Transverse
- Checklist
- Anterior-Posterior
- Transverse
- Checklist
- Median Cubital V
- Anterior-Posterior
- Checklist

- Transverse
- Transverse
- Checklist
- AVF-Outflow Vein Level 6
- AVF-Inflow Artery
- Anterior-Posterior
- Transverse
- Transverse
- AVF-Stenosis 1
- AVF-Anast
- Anterior-Posterior
- Transverse
- AVF-Stenosis 2
- AVF-Outflow Vein Level 1
- Anterior-Posterior
- Transverse
- Transverse
- AVF-Stenosis 3
- AVF-Outflow Vein Level 2
- Anterior-Posterior
- Transverse
- Transverse
- AVF-Aneurysm 1
- AVF-Outflow Vein Level 3
- Anterior-Posterior
- Transverse
- AVF-Aneurysm 2
- AVF-Outflow Vein Level 4
- Anterior-Posterior
- Transverse
- Transverse
- AVF-Aneurysm 3
- AVF-Outflow Vein Level 5
- Anterior-Posterior
- Transverse

- AV Graft-Inflow Artery - Anterior-Posterior
- Anterior-Posterior - Transverse
- Transverse - AV Graft-Outflow Vein Level 5
- AV Graft-Arterial Anast - Anterior-Posterior
- Anterior-Posterior - Transverse
- Transverse - AV Graft-Outflow Vein Level 6
- AV Graft-Graft - Anterior-Posterior
- Anterior-Posterior - Transverse
- Transverse - C.Iliac V
- AV Graft-Venous Anast - Anterior-Posterior
- Anterior-Posterior - Transverse
- Transverse - Checklist
- AV Graft-Outflow Vein Level 1 - Ex.Iliac V
- Anterior-Posterior - Anterior-Posterior
- Transverse - Transverse
- AV Graft-Outflow Vein Level 2 - Checklist
- Anterior-Posterior - IIV
- Transverse - Anterior-Posterior
- AV Graft-Outflow Vein Level 3 - Transverse
- Anterior-Posterior - Checklist
- Transverse - CFV
- AV Graft-Outflow Vein Level 4 - Anterior-Posterior

- Transverse
- Sural V
- Checklist
- Anterior-Posterior
- Femoral V
- Transverse
- Anterior-Posterior
- Checklist
- Transverse
- Soleal V
- Checklist
- DFV
- Transverse
- Anterior-Posterior
- Checklist
- Transverse
- A.Tib V
- Checklist
- Checklist
- Pop V
- Transverse
- Anterior-Posterior
- Checklist
- Transverse
- TP Trunk V
- Checklist
- Checklist
- Transverse
- Anterior-Posterior
- PTib V
- Transverse
- Anterior-Posterior
- Checklist
- Transverse
- Saph V
- Checklist
- Anterior-Posterior
- Peroneal V
- Transverse
- Anterior-Posterior
- Checklist
- Transverse
- SSV
- Checklist

- Transverse
- SSV Thigh Extension
- Checklist
- Anterior-Posterior
- SF Junction
- Transverse
- Anterior-Posterior
- Checklist
- Transverse
- AASV
- Checklist
- GSV Thigh
- Transverse
- Anterior-Posterior
- Checklist
- Transverse
- PASV
- Checklist
- GSV Knee
- Transverse
- Anterior-Posterior
- Checklist
- Transverse
- Thigh Perf
- Checklist
- GSV Calf
- Transverse
- Anterior-Posterior
- Checklist
- Transverse
- Prox Calf Perf
- Checklist
- SP Junction
- Transverse
- Anterior-Posterior
- Checklist
- Transverse
- Mid Calf Perf
- Checklist

- Transverse - PComA
- Checklist - PCA
- Dist Calf Perf - P1 PCA
- Anterior-Posterior - P2 PCA
- Transverse - Ophthalmic A
- Checklist - ICA Siphon
- Pseudoaneurysm - Terminal Vert A
- Long - BA
- Anterior-Posterior - Ba V
- Transverse - CCA
- Neck - ICA
- - ECA
- M-Mode - Bulb
- - Carotid Bifurcation
- D-Mode - Vert A
- ACA - Subclav A
- A1 ACA - Innom A
- MCA - Mammary A
- M1 MCA - Subclav V
- M2 MCA - CCA Aneurysm
- AComA - ICA Aneurysm
- Terminal ICA - ECA Aneurysm

- Bulb Aneurysm
- Carotid Bifurcation Aneurysm
- Vert A Aneurysm
- Subclav A Aneurysm
- Innom A Aneurysm
- Mammary A Aneurysm
- Carotid Graft 1 Native Inflow
- Carotid Graft 1 Anast Pre
- Carotid Graft 1 Anast Max
- Carotid Graft 1 Anast Post
- Carotid Graft 1 Graft
- Carotid Graft 1 Native Outflow
- Carotid Graft 2 Native Inflow
- Carotid Graft 2 Anast Pre
- Carotid Graft 2 Anast Max
- Carotid Graft 2 Anast Post
- Carotid Graft 2 Graft
- Carotid Graft 2 Native Outflow
- Carotid Graft 3 Native Inflow
- Carotid Graft 3 Anast Pre
- Carotid Graft 3 Anast Max
- Carotid Graft 3 Anast Post
- Carotid Graft 3 Graft
- Carotid Graft 3 Native Outflow
- Carotid Stent 1
- Carotid Stent 2
- Carotid Stent 3
- Axill A
- Brachial A
- Ulnar A
- Radial A
- UE A Graft 1 Native Inflow
- UE A Graft 1 Anast
- UE A Graft 1 Graft
- UE A Graft 1 Native Outflow
- UE A Graft 2 Native Inflow
- UE A Graft 2 Anast
- UE A Graft 2 Graft
- UE A Graft 2 Native Outflow
- UE A Graft 3 Native Inflow
- UE A Graft 3 Anast
- UE A Graft 3 Graft
- UE A Graft 3 Native Outflow
- UE A Stent 1

- UE A Stent 2
- UE A Stent 3
- C.Iliac A
- Ex.Iliac A
- IIA
- CFA
- DFA
- SFA
- Pop A
- TP Trunk A
- A.Tib A
- Peroneal A
- PTib A
- Dors.Ped. A
- LE A Graft 1 Native Inflow
- LE A Graft 1 Anast Pre
- LE A Graft 1 Anast Max
- LE A Graft 1 Anast Post
- LE A Graft 1 Graft
- LE A Graft 1 Native Outflow
- LE A Graft 2 Native Inflow
- LE A Graft 2 Anast Pre
- LE A Graft 2 Anast Max
- LE A Graft 2 Anast Post
- LE A Graft 2 Graft
- LE A Graft 2 Native Outflow
- LE A Graft 3 Native Inflow
- LE A Graft 3 Anast Pre
- LE A Graft 3 Anast Max
- LE A Graft 3 Anast Post
- LE A Graft 3 Graft
- LE A Graft 3 Native Outflow
- LE A Stent 1
- LE A Stent 2
- LE A Stent 3
- Axill V
- Brachial V
- Radial V
- Ulnar V
- Cephalic V
- Basilic V
- AVF-Inflow Artery
- AVF-Anast
- AVF-Outflow Vein Level 1

- AVF-Outflow Vein Level 2 - Pre Sten
- AVF-Outflow Vein Level 3 - Sten
- AVF-Outflow Vein Level 4 - Post Sten
- AVF-Outflow Vein Level 5 - ICA(Sten)
- AVF-Outflow Vein Level 6 - Pre Sten
- AVF-Stenosis 1 - Sten
- AVF-Stenosis 2 - Post Sten
- AVF-Stenosis 3 - ECA(Sten)
- AV Graft-Inflow Artery - Pre Sten
- AV Graft-Arterial Anast - Sten
- AV Graft-Graft - Post Sten
- AV Graft-Venous Anast - Bulb(Sten)
- AV Graft-Outflow Vein Level 1 - Pre Sten
- AV Graft-Outflow Vein Level 2 - Sten
- AV Graft-Outflow Vein Level 3 - Post Sten
- AV Graft-Outflow Vein Level 4 - Carotid Bifurcation(Sten)
- AV Graft-Outflow Vein Level 5 - Pre Sten
- AV Graft-Outflow Vein Level 6 - Sten
- ASP - Post Sten
- BSP - Vert A(Sten)
- ----- - Pre Sten
- CCA(Sten) - Sten

- Post Sten
- Subclav A(Sten)
- Pre Sten
- Sten
- Post Sten
- Innom A(Sten)
- Pre Sten
- Sten
- Post Sten
- Mammary A(Sten)
- Pre Sten
- Sten
- Post Sten
- Carotid Stenosis 1
- Pre Sten
- Sten
- Post Sten
- Carotid Stenosis 2
- Pre Sten
- Sten
- Post Sten
- Carotid Stenosis 3
- Pre Sten
- Sten
- Post Sten
- Carotid Stenosis 4
- Pre Sten
- Sten
- Post Sten
- Axill A(Sten)
- Pre Sten
- Sten
- Post Sten
- Brachial A(Sten)
- Pre Sten
- Sten
- Post Sten
- Ulnar A(Sten)
- Pre Sten
- Sten
- Post Sten
- Radial A(Sten)
- Pre Sten
- Sten

- Post Sten
- Pre Sten
- UE A Stenosis 1
- Sten
- Pre Sten
- Post Sten
- II A(Sten)
- Post Sten
- UE A Stenosis 2
- Pre Sten
- Post Sten
- Sten
- Pre Sten
- Post Sten
- CFA(Sten)
- Post Sten
- UE A Stenosis 3
- Pre Sten
- Post Sten
- Sten
- Pre Sten
- Post Sten
- DFA(Sten)
- Post Sten
- UE A Stenosis 4
- Pre Sten
- Post Sten
- SFA(Sten)
- Post Sten
- C.Iliac A(Sten)
- Pre Sten
- Post Sten
- Sten
- Post Sten
- Sten
- Pop A(Sten)
- Post Sten
- Ex.Iliac A(Sten)
- Sten

- Post Sten
- Pre Sten
- TP Trunk A(Sten)
- Pre Sten
- Sten
- Post Sten
- A.Tib A(Sten)
- Pre Sten
- Sten
- Post Sten
- Peroneal A(Sten)
- Pre Sten
- Sten
- Post Sten
- P.Tib A(Sten)
- Pre Sten
- Sten
- Post Sten
- Dors.Ped. A(Sten)
- Pre Sten
- Sten
- Post Sten
- LE A Stenosis 1
- Post Sten
- Sten
- Pre Sten
- LE A Stenosis 2
- Pre Sten
- Sten
- Post Sten
- LE A Stenosis 3
- Pre Sten
- Sten
- Post Sten
- LE A Stenosis 4
- Pre Sten
- Sten
- Post Sten
- C.Iliac V
- PV
- Reflux
- Checklist
- Ex.Iliac V
- PV
- Reflux

- Checklist - PV
- IIV - Reflux
- PV - Checklist
- Reflux - Peroneal V
- Checklist - PV
- CFV - Reflux
- PV - Checklist
- Reflux - Sural V
- Checklist - PV
- Femoral V - Reflux
- PV - Checklist
- Reflux - Soleal V
- Checklist - PV
- DFV - Reflux
- PV - Checklist
- Reflux - A.Tib V
- Checklist - PV
- Pop V - Reflux
- PV - Checklist
- Reflux - TP Trunk V
- Checklist - PV
- PTib V - Reflux

- Checklist - PV
- Saph V - Reflux
- PV - Checklist
- Reflux - SP Junction
- Checklist - PV
- SSV - Reflux
- PV - Checklist
- Reflux - SSV Thigh Extension
- Checklist - PV
- SF Junction - Reflux
- PV - Checklist
- Reflux - AASV
- Checklist - PV
- GSV Thigh - Reflux
- PV - Checklist
- Reflux - PASV
- Checklist - PV
- GSV Knee - Reflux
- PV - Checklist
- Reflux - Thigh Perf
- Checklist - PV
- GSV Calf - Reflux

- Checklist
- THY Nodule1 Strain
- Prox Calf Perf
- THY Nodule2 Strain
- PV
- THY Nodule3 Strain
- Reflux
- THY Nodule1 Elas.
- Checklist
- THY Nodule2 Elas.
- Mid Calf Perf
- THY Nodule3 Elas.
- PV
- Breast Mass1 Strain
- Reflux
- Breast Mass1 Elas.
- Checklist
- Breast Mass2 Strain
- Dist Calf Perf
- Breast Mass2 Elas.
- PV
- Breast Mass3 Strain
- Reflux
- Breast Mass3 Elas.
- Checklist
- Breast Mass4 Strain
- ABI
- Breast Mass4 Elas.
- ASP
- Breast Mass5 Strain
- BSP
- Breast Mass5 Elas.
- Small Parts
- B-Mode
- Thyroid L
- Breast Mass6 Strain
- Thyroid H
- Breast Mass6 Elas.
- Thyroid W
- Breast Mass7 Strain
- Isthmus H
- Breast Mass7 Elas.
- Breast Mass8 Strain
- Breast Mass8 Elas.

- Breast Mass9 Strain - Transverse
- Breast Mass9 Elas. - Thyroid(Inferior)
- Breast Mass10 Strain - Anterior-Posterior
- Breast Mass10 Elas. - Transverse
- Testicular L - Parathyroid 1
- Testicular H - Long
- Testicular W - Anterior-Posterior
- Epididymis L - Transverse
- Epididymis H - Parathyroid 2
- Epididymis W - Long
- Scrotal Wall - Anterior-Posterior
- Testis V(2D) - Transverse
- Testis V(Valsalva 2D) - Parotid
- ----- - Long
- Thyroid Vol - Anterior-Posterior
- Testicular Vol - Transverse
- ----- - Lymph Node 1
- Thyroid(Superior) - Long
- Anterior-Posterior - Anterior-Posterior
- Transverse - Transverse
- Thyroid(Mid) - Lymph Node 2
- Anterior-Posterior - Long

- Anterior-Posterior
- Thyroid Nodule 1
- Transverse
- Long
- Lymph Node 3
- Anterior-Posterior
- Transverse
- Long
- Anterior-Posterior
- Thyroid Nodule 2
- Transverse
- Long
- Lymph Node 4
- Anterior-Posterior
- Transverse
- Long
- Anterior-Posterior
- Thyroid Nodule 3
- Transverse
- Long
- Lymph Node 5
- Anterior-Posterior
- Transverse
- Long
- Anterior-Posterior
- Thyroid Nodule 4
- Transverse
- Long
- Lymph Node 6
- Anterior-Posterior
- Transverse
- Long
- Anterior-Posterior
- Thyroid Nodule 5
- Transverse
- Long
- Thyroid
- Anterior-Posterior
- Transverse
- Thyroid L
- Transverse
- Thyroid H
- Thyroid Nodule 6
- Thyroid W
- Long

- Anterior-Posterior - A
- Transverse - B
- Thyroid Nodule 7 - THY Nodule3 Strain Ratio
- Long - A
- Anterior-Posterior - B
- Transverse - THY Nodule1 Elas. Ratio
- Thyroid Nodule 8 - A
- Long - B
- Anterior-Posterior - THY Nodule2 Elas. Ratio
- Transverse - A
- Thyroid Nodule 9 - B
- Long - THY Nodule3 Elas. Ratio
- Anterior-Posterior - A
- Transverse - B
- Thyroid Nodule 10 - Breast Mass 1
- Long - L
- Anterior-Posterior - H
- Transverse - W
- THY Nodule1 Strain Ratio - Nip. Dist.
- A - Skin Dist.
- B - Breast Mass 2
- THY Nodule2 Strain Ratio - L

- H - Breast Mass 6
- W - L
- Nip. Dist. - H
- Skin Dist. - W
- Breast Mass 3 - Nip. Dist.
- L - Skin Dist.
- H - Breast Mass 7
- W - L
- Nip. Dist. - H
- Skin Dist. - W
- Breast Mass 4 - Nip. Dist.
- L - Skin Dist.
- H - Breast Mass 8
- W - L
- Nip. Dist. - H
- Skin Dist. - W
- Breast Mass 5 - Nip. Dist.
- L - Skin Dist.
- H - Breast Mass 9
- W - L
- Nip. Dist. - H
- Skin Dist. - W

- Nip. Dist. - B
- Skin Dist. - Breast Mass3 Elas. Ratio
- Breast Mass 10 - A
- L - B
- H - Breast Mass4 Strain Ratio
- W - A
- Nip. Dist. - B
- Skin Dist. - Breast Mass4 Elas. Ratio
- Breast Mass1 Strain Ratio - A
- A - B
- B - Breast Mass5 Strain Ratio
- Breast Mass1 Elas. Ratio - A
- A - B
- B - Breast Mass5 Elas. Ratio
- Breast Mass2 Strain Ratio - A
- A - B
- B - Breast Mass6 Strain Ratio
- Breast Mass2 Elas. Ratio - A
- A - B
- B - Breast Mass6 Elas. Ratio
- Breast Mass3 Strain Ratio - A
- A - B

- Breast Mass7 Strain Ratio - A
- A - B
- B - Testis
- Breast Mass7 Elas. Ratio - Testicular L
- A - Testicular H
- B - Testicular W
- Breast Mass8 Strain Ratio - Testis Mass 1
- A - d1
- B - d2
- Breast Mass8 Elas. Ratio - d3
- A - Testis Mass 2
- B - d1
- Breast Mass9 Strain Ratio - d2
- A - d3
- B - Testis Mass 3
- Breast Mass9 Elas. Ratio - d1
- A - d2
- B - d3
- Breast Mass10 Strain Ratio - Epididymis
- A - Epididymis L
- B - Epididymis H
- Breast Mass10 Elas. Ratio - Epididymis W

- Testicle(Superior) M-Mode
- H -
- W D-Mode
- Testicle(Mid) - STA
- H - ITA
- W - Isthmus
- Testicle(Inferior) - Parathyroid 1
- H - Parathyroid 2
- W - Testis A
- Epididymal Head - Testis V
- L - Testis V(Valsalva)
- H - Epididymis A
- W - Epididymis V
- Epididymal Body • Emergency
- L B-Mode
- H - Renal L
- W - Renal H
- Epididymal Tail - Renal W
- L - CBD
- H - Portal V Diam
- W - CHD
- - GB wall th

- Aorta Bif
- Ureter
- Pre-BL L
- Pre-BL H
- Pre-BL W
- Post-BL L
- Post-BL H
- Post-BL W
- GS
- YS L
- CRL
- BPD
- UT L
- UT H
- UT W
- Endo
- Ovary L
- Ovary H
- Ovary W
- Endo
- Ovary L
- Ovary H
- Ovary W
- Renal Vol
- Pre-BL Vol
- Post-BL Vol
- Mictur.Vol
- Ovary Vol
- UT Vol
- UT SUM
- -----
- Uterus
- UT L
- UT H
- UT W
- Endo
- Ovary
- Ovary L
- Ovary H
- Ovary W
- Kidney
- Renal L
- Renal H
- Renal W
- Cortex
- Bladder
- Pre-BL L

- Pre-BL H
 - Intima-Media Thickness
 - Pre-BL W
 - Measurement
 - Post-BL L
 - Automatic detection of IMT when ROI is set
 - Post-BL H
 - Near wall and far wall detection
 - Post-BL W
 - Support CCA, ICA, ECA, Bulb IMT
 - M-Mode
 - Angle selectable
 - FHR (M)
 - 6.5 IVF
 - The uterus and follicle growth curve can be displayed in the IVF report.
 - D-Mode
 - Data of IVF history exams can be checked in the IVF report.
 - Orthopedic
 - B-Mode
 - 6.6 Smart OBTM
 - Auto measurement for OB, a special tool for easy OB scan, and greatly reduce time and increase productivity
 - Support BPD, HC, OFD, FL, AC
 - Better get GA before start auto AC
 - Measurement result can be modified by user
 - M-Mode
 - Measurement result can be modified by user
 - D-Mode
 - Measurement result can be modified by user
- 6.4 IMT
- 6.7 Smart NTTM

- NT auto measurement
 - Auto detection of NT inside ROI
- 6.8 Smart HIP
- HIP auto measurement
- application
 - Editable value in report
 - Images selectable
 - Anatomy information for vascular and OB report
- 6.9 Smart Trace
- Measures the lengths of major axis and minor axis, area and circumference of a closed region on the image semi-automatically
- Editing through iReport
 - User-defined report template
 - Selecting report modules
 - Adding/removing measurement items from the report
- 6.10 Smart Caliper
- Measures lengths of the major axis and the minor axis semi-automatically
- Changing report layout
 - Load/save comment
 - Viewing history reports
 - Preview and printing reports
- 6.11CPP
- Measures the pixel proportion of blood flow signal in the region of interest
- Able to Export as PDF/RTF file
 - V-Mapping
 - Add anatomical graphics for illustration
- 6.12 Smart HRI
- Auto measurement method for Hepatorenal Index (HRI)
- Mini report
 - Quickly displaying Mini report in the thumbnail area of the main screen
- 6.13 Report
- Specific report template by

- Including both general measurement and application measurement results
 - Support deleting measurement results
- 6.14 iWorks™
- Auto workflow protocol
 - Templates are user configurable
 - Functions: pause, stop, replace, repeat, skip, insert single step, return and continue, steps in thumbnail, iNSert™ another template
 - iWorks setup mode: B; B/B (Dual Live); Dual B/B; Color; B/Color (Dual Live); Power; B/Power (Dual Live); B + PW; Color + PW; Power + PW; B + CW; Color + CW; B+M; B+TVI; TVI+TVD; iScape.
 - iWorks setup annotation: support up to 2 annotations, location and font size are configurable.
- iWorks setup bodymark: select existing library, and transducer indicator is pre-settable
 - iWorks setup measurement: select existing measurement library
 - Template import and export are available
 - Support create user-defined iWorks protocol
- 6.15 UltraView™
- Components:
 - DICOM Basic
 - DICOM Query/Retrieve
 - DICOM OB/GYN SR
 - DICOM Cardiac SR
 - DICOM Vascular SR
 - DICOM Breast SR
 - DICOM Urology SR
 - DICOM Pediatric SR
 - DICOM Small Parts SR
 - DICOM Abdomen SR
 - TDI QA

- Contrast Imaging QA
- Tissue Tracking QA
- Stress Echo
- SCV (Slice Contrast View)
- Niche
- iLive
- iPage⁺
- IVF
- Ultrasound Fusion Imaging
- Auto EF
- Smart Planes CNS
- PC-based off-line software
- iStation™ workstation dedicated for patient exam management
- Patient exam query/retrieve
- Support review of current and past exam
- New exam, Activate exam, Continue exam functions, End exam are available
- Support measurements and calculations on archived exam and images
- Export images in BMP/JPG/TIFF/DCM/AVI/MP4 format
- Support backup/send to USB devices, DVD-RW, CD-R, DVD+R, DVD-R, DVD+RW media

*Not all measurements are listed in this part; For more detailed information please refer to User Manual

7 Exam Storage and Management

7.1 Exam storage

- 1TB hard drive
- 128G SSD (Solid State Drive)
- Direct digital storage of single frame and cine 2D, color and Doppler.

7.2 Exam management

8 Connectivity

8.1 Ethernet Network Connection

- Cable connection
- Wireless connection: built-in wireless adapter

8.2 DICOM 3.0

- DICOM basic
 - Verify (SCU, SCP)
 - Print
 - Store
 - Storage Commitment
 - Media Exchange
 - DICOM Worklist
 - DICOM Query/Retrieve
 - DICOM Modality Performed
- Procedure Step - MPPS
- DICOM OB/GYN structure report
 - DICOM Cardiac structure report
 - DICOM Vascular structure report
 - DICOM Breast Report
 - DICOM Abdominal structure report
 - DICOM Small Parts structure report
- An interactive app that lets you transfer clinical images straight from Mindray Ultrasound system to a smart device, such as mobile phone or tablet PC
- Needs to be installed on mobile terminal
 - Transfer images or clips from system to mobile terminal through Wi-Fi
 - Support both iOS (7.0 and above) and Android (4.0 and above) system
- For iOS powered smart device: DICOM is mandatory
 - For Android powered smart device: DICOM not necessary

8.3 iStorage (included in UltraAssist)

Direct network storage tool between ultrasound system and personal computer

8.4 MedSight

- patient information, and tutorial software iScanHelper study on smart devices
- Support Android and iOS powered smart devices
 - Android 4.0 and above
 - iOS 7.0 and above
 - DICOM not necessary
- B-mode Frequencies: 1.2~3.8, 1.7~5.2, 2.0~6.0MHz
- Harmonic Frequencies: 4.0, 5.0, 6.0 MHz
- Color Frequencies: 2.0, 2.5, 3.0, 3.3 (HR Flow) MHz
- PW Frequencies: 2.0, 2.5, 3.0 MHz
- Biopsy Guide: NGB-022, multi angle, reusable; LPUBKG60 (disposable)

9 Transducers

9.1 Curved array

- SC6-1s
 - Application: Obstetrics, Gynecology, Abdomen, Small Parts, Musculo-skeletal, Vascular, Urology, Nerve
 - Bandwidth: 1.2-6.0MHz
 - Number of Elements:192
 - FOV (max): 60°
 - Extended FOV: 72°
 - Convex Radius: 60mm
 - Depth: 4-40cm
 - Physical Footprint: 65.1mm × 16.4mm
 - Footprint: 64.9mm × 16.2mm
 - C11-3s
 - Application: Abdomen, Vascular, Cardiac, Small Parts, Pediatric, Cephalic
 - Bandwidth: 2.6-12.8MHz
 - Number of Elements:128
 - FOV (max): 101°
 - Extended FOV: 113°
 - Convex Radius: 15mm
 - Depth: 1.5-35cm
 - Physical Footprint: 32.8mm× 25mm
 - Footprint: 27.4mm × 8.4mm
 - B-mode Frequencies: 2.6~6.5,

3.2~7.9, 4.7~12.8 MHz	MHz	
- Harmonic Frequencies: 6.0, 7.0, 8.0		- Color Frequencies: 2.0, 2.5, 3.0, 3.5(HR
MHz		Flow) MHz
- Color Frequencies: 4.4, 5.0, 5.7, 5.0		- PW Frequencies: 2.0, 2.5, 3.0 MHz
(HR Flow) MHz		- Biopsy Guide: NGB-024, multi angle,
- PW Frequencies: 4.4, 5.0, 5.7 MHz		reusable; LPUBKG81 (disposable)
- Biopsy Guide: NGB-018, multi angle,		● V11-3Hs
reusable		- Application: Obstetrics, Gynecology,
● C6-2Gs		Urology
- Application: Obstetrics, Gynecology,		- Bandwidth: 3.0-11.0MHz
Abdomen, Urology		- Number of Elements:192
- Bandwidth: 1.2-6.0MHz		- FOV (max): 170°
- Number of Elements:128		- Extended FOV: 210°
- FOV (max): 94°		- Convex Radius: 11mm
- Extended FOV: 106°		- Depth: 1.5-28cm
- Convex Radius: 20mm		- Physical Footprint: 24.9mm× 21.8mm
- Depth: 4-40cm		- Footprint: 24mm × 9mm
- Physical Footprint: 37.6mm × 19mm		- B-mode Frequencies: 3.0~7.0,
- Footprint: 31.5mm × 11.2mm	4.0~9.0, 5.0~11.0 MHz	
- B-mode Frequencies: 1.2~3.8, 1.7~5.2,		- Harmonic Frequencies: 8.0, 9.0, 10.0
2.0~6.0MHz		MHz
- Harmonic Frequencies: 3.2, 4.0, 5.0		- Color Frequencies: 4.4, 5.0, 5.7, 5.5(HR

Flow) MHz	- Biopsy Guide: NGB-029, multi angle, reusable
- PW Frequencies: 4.5, 5.0, 5.5 MHz	
- Biopsy Guide: NGB-025, single angle, reusable	9.2 Volume curved array
● SC8-2s	● SD8-1s
- Application: Obstetrics, Gynecology, Abdomen, Urology, Vascular	- Application: Obstetrics, Gynecology, Abdomen
- Bandwidth: 1.8 - 8.2 MHz	- Bandwidth: 1.8-8.2 MHz
- Number of Elements:192	- Number of Elements: 192
- FOV (max): 76°	- FOV (max): 66°
- Extended FOV: 88°	- Extended FOV: 91°
- Convex Radius: 40mm	- Volume Sweep Angle (max): 85°
- Depth: 4-40cm	- Convex Radius: 45mm
- Physical Footprint: 26.3mm× 66.9mm	- Depth: 4-40cm
- Footprint: 15mm × 52mm	- Physical Footprint: 75.7mm × 52.6mm
- B-mode Frequencies: 1.8~5.4, 2.3~6.8, 2.8~8.2 MHz	- B-mode Frequencies: 1.8-5.4, 2.3-6.8, 2.8-8.2 MHz
- Harmonic Frequencies: 4.0, 5.5, 6.0 MHz	- Harmonic Frequencies: 4.0, 5.0, 5.5, 6.0 MHz
- Color Frequencies: 3.0, 3.5, 4.0, 3.5(HR Flow) MHz	- Color Frequencies: 3.0, 3.5, 4.0, 3.5(HR Flow) MHz
- PW Frequencies: 3.0, 3.5, 3.8 MHz	- PW Frequencies: 3.0, 3.5, 3.8 MHz
	- Biopsy Guide: NGB-039, multi angle,

- reusable
 - DE11-3Ws
 - Application: Obstetrics, Gynecology, Urology
 - Bandwidth: 2.0-9.0 MHz
 - Number of Elements: 192
 - FOV (max): 162°
 - Extended FOV: 187°
 - Volume Sweep Angle (max): 120°
 - Convex Radius: 10mm
 - Depth: 4-40cm
 - Physical Footprint: 24 mm x 24mm
 - B-mode Frequencies: 2.0~6.0, 2.8~8.2, 3.0~9.0 MHz
 - Harmonic Frequencies: 4.0, 5.0, 6.0 MHz
 - Color Frequencies: 4.4, 5.0, 5.7, 5.0(HR Flow) MHz
 - PW Frequencies: 4.4, 5.0, 5.7 MHz
 - Biopsy Guide: NGB-047, single angle, reusable
 - L9-3s
 - Application: Abdomen, Pediatric, Small Parts, Musculo-skeletal, Vascular, Nerve, Obstetrics
 - Bandwidth: 2.5-9.0MHz
 - Number of Elements: 192
 - Field of View (max): 43.7mm
 - Extended FOV: 60° (OB1, NT); 40°(Others)
 - Steered Angle:
 - B: 0°, +/−6°, +/−12°;
 - C/PW: −30°-30°
 - Depth:1.5-35cm
 - Physical Footprint: 62mm × 22mm
 - Footprint: 48mm × 11mm
 - B-mode Frequencies: 2.5~7.0, 3.4~8.2, 3.6~9.0 MHz
 - Harmonic Frequencies: 5.0, 6.0, 7.0 MHz
 - Color Frequencies: 3.0, 3.6, 5.0, 4.0(HR Flow) MHz
 - PW Frequencies: 3.0, 3.6, 5.0 MHz
- 9.3 Linear

- Biopsy Guide: NGB-034, multi angle, reusable
- L20-5s
- Application: Abdomen, Small Parts, Musculo-skeletal, Vascular, Nerve
- Bandwidth: 6-23MHz
- Number of Elements: 192
- Field of View (max): 28.5mm
- Extended FOV: 20°
- Steered Angle:
 - B: 0, +/−6°, +/−12°;
 - C/PW: −20°-20°
- Depth: 1.5-29cm
- Physical Footprint: 42.23mm × 22.10mm
- Footprint: 31.5mm × 4.5mm
- B-mode Frequencies: 6.0~13.0, 9.0~16.6, 12.5~23.0 MHz
- Harmonic Frequencies: 14.0, 16.0, 18.0 MHz
- Color Frequencies: 9.0, 11.0, 13.0, 13.0(HR Flow) MHz
- PW Frequencies: 8.3, 10.0, 12.5 MHz
- Biopsy Guide: not available
- L14-3Ws
- Application: Small Parts, Musculo-skeletal, Vascular, Abdomen, Pediatric, Thoracic/Pleural
- Bandwidth: 3.0-14.0MHz
- Number of Elements: 256
- Field of View (max): 50.8mm
- Extended FOV: 20°
- Steered Angle:
 - B: 0, +/−6°, +/−12°;
 - C/PW: −30°-30°
- Depth: 1.5-35cm
- Physical Footprint: 66.8mm × 25.5mm
- Footprint: 55.5mm × 8.2mm
- B-mode Frequencies: 3.0~9.0, 5.0~12.0, 6.0~14.0 MHz
- Harmonic Frequencies: 10.0, 11.0, 12.0, 13.0 MHz
- Color Frequencies: 5.0, 6.2, 7.2, 8.3(HR flow) MHz

- PW Frequencies: 5.0, 6.2, 7.2 MHz
- Application: Abdomen, Cardiac, Vascular, Cephalic, Thoracic/Pleural
- Biopsy Guide: NGB-054, multi angle/depth, reusable
- Bandwidth: 1.5 - 4.5MHz
- Number of Elements: 80
- Field of View (max): 90°
- Extended FOV: 90°
- Depth: 2-38cm
- Physical Footprint: 38.2mm × 30.5mm
- 9.4 Phased array
 - P7-3Ts
 - Application: Cardiac
 - Bandwidth: 2.3-7.2MHz
 - Number of Elements: 64
 - Field of View (max): 90°
 - Extended FOV: 90°
 - Depth: 2.0-38.0cm
 - Physical Footprint: 14mm × 12mm
 - Footprint: 12.2mm × 12.2mm
 - B-mode Frequencies: 2.3~5.4, 2.8~6.4, 3.3~7.2 MHz
 - Harmonic Frequencies: 5.0, 6.0, 7.0 MHz
 - Color Frequencies: 2.7, 3.3, 4.0, 4.0(HR Flow) MHz; TDI: 3.0, 3.8 MHz
 - PW Frequencies: 2.0, 2.3, 2.5 MHz; TDI: 2.5, 4.0 MHz
 - CW Frequencies: 2.0 MHz
 - Biopsy Guide: NGB-011, multi angle, reusable
- PW Frequencies: 2.7, 3.3, 4.0 MHz
- Biopsy Guide: not available
- 9.5 Bi-Plane
 - ELC13-4s (Convex & Linear)
- SP5-1s

- Application: Urology	Linear: 71.4mm × 9mm
- Bandwidth:	- B-mode Frequencies:
Convex: 3.5-9.5 MHz	Convex: 3.5~6.0, 4.5~7.5, 5.5~8.8,
Linear: 4.8-11.0MHz	6.5~9.5 MHz
- Number of Elements: 192	Linear: 4.8~7.8, 5.3~9.0, 5.8~10.0,
- Field of View (max):	7.0~11.0 MHz
Convex: 177°	- Harmonic Frequencies:
Linear: 64.8mm	Convex: 8.0, 9.0, 10.0 MHz
- Extended FOV: 217° (Convex); 40°(Linear)	Linear: 10.0, 11.0, 12.0 MHz
- Convex Radius: 10mm (Convex)	- Color Frequencies:
- Depth:	Convex: 4.4, 5.0, 5.2, 5.5(HR Flow) MHz
Convex: 1.5-28cm	Linear: 6.0, 7.2, 8.3, 9.0(HR Flow) MHz
Linear: 1.5-35cm	- PW Frequencies:
- Steered Angle (Linear):	Convex: 4.5, 5.0, 5.5 MHz
-B: 0, +/−6°, +/−12°	Linear: 5.0, 6.3, 7.1, 8.3 MHz
-C/PW: −15°-15°	- Biopsy Guide: NGB-051, multi depth, reusable
- Physical Footprint:	9.6 Pencil
Convex: 20mm × 20mm	• CW5s
Linear: 19.5mm × 19.5mm	- Application: Vascular, Cephalic,
- Footprint:	Pediatric
Convex: 20mm × 9mm	

- Number of Elements: 2
- Biopsy Guide: not available
- CW2s
- Application: Cardiac, Cephalic, Pediatric
- Number of Elements: 2
- Biopsy Guide: not available
- Temperature with 4 levels:
off/34°C/37°C/40°C, with deviation
of ±1°C
- Light indicator for temperature protecting
- Dimension: 82(D)*78(W)*119(H) mm
- Weight: approx. 240g (net)
- Continuous operation time: >12h

10 Peripheral Devices and Accessories

- | | |
|---|---|
| 10.1 Black/white video printer | 10.5 Footswitch |
| <ul style="list-style-type: none"> • Digital
MITSUBISHI P95DW-N • Analog
SONY UP-X898MD | <ul style="list-style-type: none"> • USB port: FS-81-SP-2(single pedal),
971-SWNOM (2/3-pedal) • Support User-definable functions
(Freeze, Save, Print) |
| 10.2 Color digital printer | 10.6 ECG |
| SONY UP-D25MD | <ul style="list-style-type: none"> • 6-pin, AHA/IEC, for 3-lead wires |
| 10.3 Graph/text printer | <ul style="list-style-type: none"> • ECG wave display: on/off |
| HP OFFICEJET PRO 8100 | <ul style="list-style-type: none"> • ECG source: Lead/External |
| 10.4 Gel warmer | <ul style="list-style-type: none"> • Position: 0~100%, 5%/step |
| <ul style="list-style-type: none"> • Support gel warming with 3 angle position: 15, 55, 90 degrees • Easily be disassembled off system for cleaning | <ul style="list-style-type: none"> • Trig mode: off/single/dual/timer • Gain: 0-30, 1/step • Sweep speed: 6 steps • Invert: on/off |

10.7	PCG (not for sale in EU countries)	<ul style="list-style-type: none"> PCG wave display: on/off Gain: 0-30, 1/step Speed: 6 steps Smooth: 1-4, 1/step 	<ul style="list-style-type: none"> Battery fully-recharged time: less than 6h Continuous scanning time: more than 120 mins for 2H battery or 240 mins for 4H battery.
10.8	Barcode reader	<ul style="list-style-type: none"> SYMBOL LS2208 (1D) SYMBOL DS4308 (2D) 	10.12 Probe adapter
10.9	Built-in Wi-Fi 5 Wireless adapter	<ul style="list-style-type: none"> Encryption: WPA, WPA2 Max transfer speed: 300Mbps Protocols: IEEE 802.11 ac/a/b/g/n Frequency: 2.4G/5G 	<ul style="list-style-type: none"> PCM-ES01: transforming E socket to S socket, only for SC6-1E, SD8-1E, P7-3TE, and SP5-1E PCM-US01: transforming U socket to S socket, only for P7-3TU and SP5-1U
10.10	iVocal Microphone	<ul style="list-style-type: none"> SAMSON XPD2 	11 System Inputs and Outputs
10.11	Built-in Battery	<ul style="list-style-type: none"> Replaceable and rechargeable lithium battery Full battery lasts for more than 24h in standby mode Battery capacity indicators without powering on the system 	11.1 Audio input/output <ul style="list-style-type: none"> Microphone/Audio: 1 port 11.2 Video output <ul style="list-style-type: none"> S-Video out: 1 port, PAL/NTSC HDMI: 1 Port VGA out: 1 port 11.3 Physio input <ul style="list-style-type: none"> Support ECG/PCG signal ECG: 1 port

- PCG: 1 port

11.4 Other input/output

- USB: 6 ports (5 USB 3.0 and 1 Type-C)
- Ethernet: 1 port

12 Safety and Conformance

12.1 Quality standards

- ISO 9001
- ISO 13485

12.2 Design standards

- EN 60601-1 and IEC 60601-1
- EN 60601-1-2 and IEC 60601-1-2
- EN 60601-1-6 and IEC 60601-1-6
- EN 60601-2-37 and IEC 60601-2-37
- EN 62304 and IEC 62304
- EN 62366 and IEC 62366
- EN ISO 17664 and ISO 17664

12.3 CE declaration

The ultrasound system is fully in conformance with the Council Directive 93/42/EEC Concerning Medical Devices. The number adjacent to the CE marking (0123) is the code of the EU-notified body that certified meeting the requirements of Annex II excluding

(4). of the Directive.

NOTICE:

Not all features or specifications described in this document may be available in all transducers and/or modes.

Mindray reserves the right to make changes in specifications and features shown herein, or discontinue the product at any time without notice or obligation.

Contact Mindray Representative for the most current information.