

**Immucor, Inc.**  
 3130 Gateway Drive  
 Norcross, GA 30091-5625  
 Tel: (770) 441-2051 Toll-free: (800) 510-5110  
 FAX: (770) 441-3807  
 Web site <http://www.immucor.com>

## Certificate of Analysis

This is to certify that the product listed hereunder has been manufactured and tested by Immucor, Inc., of Norcross, Georgia. The product listed has been tested in accordance with current requirements for the product in question, and has been found to meet all specifications.

**Product:** Checkcell  Checkcell (Weak)  Complement Control Cells   
**Catalog Nos.** 0002225 / 3 x 10 mL  
**Lot No.** 38288  
**Expiration date:** 2022-11-25 **Date of manufacture:** 2022-09-19

**Chemical Composition:** Checkcell: A 4-6% suspension of group O pooled human red blood cells sensitized with an IgG antibody.  
 Checkcell (Weak): A 2-3% suspension of group O pooled human red blood cells sensitized with an IgG antibody.  
 Complement Control Cells: A 2-4% suspension of pooled human red blood cells sensitized with human complement.  
 These red blood cells are suspended in buffered preservative solution containing Neomycin Sulfate 0.1 mg/mL, Chloramphenicol 0.25 mg/mL, and Gentamycin Sulfate 0.05 mg/mL.

Viral Marker Testing:	Requirement	Non-reactive
	Source material nonreactive for antibodies to human immunodeficiency virus types 1 and 2 (anti-HIV) and human hepatitis C virus (Anti-HCV), and hepatitis B surface antigen (HBsAg); and nonreactive for HIV-1 RNA and HCV RNA by NAT using FDA-licensed tests and nonreactive for serologic test for syphilis (STS).	HBsAg <input checked="" type="checkbox"/> Anti-HCV <input checked="" type="checkbox"/> Anti-HIV <input checked="" type="checkbox"/> HIV-1 RNA <input checked="" type="checkbox"/> HCV RNA <input checked="" type="checkbox"/> STS <input checked="" type="checkbox"/>

**Identity testing:** Product tested for identity as labeled.

**Approved for Release:** All specifications met; labeling checked and confirmed.

DocuSigned by Lisa Hugg  
 I am the author of this document  
07-Oct-2022 | 10:30:26 AM EDT  
DF21BD0A391E4587BC124B33D0B3337C

DocuSigned by Kimberly Nguyen  
 I approve this document  
12-Oct-2022 | 11:38:47 AM PDT  
23939EC988DF4795A3D5706DDC3D8397

QC Reviewed by/Date: \_\_\_\_\_ Approved by/Date: \_\_\_\_\_