

Endurity MRI™

Single-Chamber Pacemaker



Merlin@home™
Transmitter
Compatible

Product Highlights - Pacemaker

- Allows patients to undergo 1.5 T or 3 T MRI scans when used with MRI Ready leads from Abbott*
- Physician preferred size and physiologic shape minimize pocket size
- Outstanding longevity provides 14,4 years of service life,¹ which is supported by a 10-year warranty²
- AutoCapture™ pacing system offers the maximum in threshold adaptability and patient safety with ventricular Beat-by-Beat™ capture confirmation. The AutoCapture pacing system automatically delivers a 5,0 V backup safety pulse when noncapture is detected, and it may be programmed to either a bipolar or unipolar configuration
- A suite of state-of-the-art features—such as automaticity, Ventricular AutoCapture™ pacing system and SenseAbility™ technology—are designed to deliver optimal therapy for patients at implant and throughout their lives
- Real-time electrogram (EGM) waveform, as well as the associated event markers that precede and follow a specific triggering event, can be programmed to automatically record up to 14 minutes of stored EGMs when encountering one or more programmable trigger options
- 6-month ERI-EOL interval
- An optional, easy-to-use hand-held device (SJM MRI Activator™ device) can be used to program the device to pre-approved MRI settings pre- and post-MRI scan, decreasing the number of workflow steps and increasing clinic efficiency

*See MRI Conditional Parameters

Ordering Information

Contents: MRI Ready Pacing System

MODEL NUMBER	DESCRIPTION	DIMENSIONS (H X W X T, MM)	WEIGHT (G)	VOLUME (CC)	CONNECTOR
PM1172	Endurity MRI™ Pacemaker	41 x 50 x 6	19	9.7 (± 0,5)	IS-1

Indications: Implantation is indicated in one or more of the following permanent conditions: syncope, presyncope, fatigue, disorientation due to arrhythmia/bradycardia, or any combination of those symptoms. **Rate-Modulated Pacing** is indicated for patients with chronotropic incompetence, and for those who would benefit from increased stimulation rates concurrent with physical activity. **Atrial Pacing** is indicated for patients with sinus node dysfunction and normal AV and intraventricular conduction systems. **Ventricular Pacing** is indicated for patients with significant bradycardia and normal sinus rhythm with only rare episodes of A-V block or sinus arrest, chronic atrial fibrillation, severe physical disability.

Contraindications: **Single-chamber pulse generators** are contraindicated in patients with an implanted cardioverter-defibrillator. **Rate-Adaptive Pacing** may be inappropriate for patients who experience angina or other symptoms of myocardial dysfunction at higher sensor-driven rates. An appropriate Maximum Sensor Rate should be selected based on assessment of the highest stimulation rate tolerated by the patient. **Single-Chamber Ventricular Demand Pacing** is relatively contraindicated in patients who have demonstrated pacemaker syndrome, have retrograde VA conduction, or suffer a drop in arterial blood pressure with the onset of ventricular pacing. **Single-Chamber Atrial Pacing** is relatively contraindicated in patients who have demonstrated compromise of AV conduction. For specific contraindications associated with individual modes, refer to the programmer's on-screen help.

Potential Adverse Events: The following are potential complications associated with the use of any pacing system: arrhythmia, heart block, thrombosis, threshold elevation, valve damage, pneumothorax, myopotential sensing, vessel damage, air embolism, body rejection phenomena, cardiac tamponade or perforation, formation of fibrotic tissue/ local tissue reaction, inability to interrogate or program a device because of programmer malfunction, infection, interruption of desired device function due to electrical interference, loss of desired pacing and/or sensing due to lead displacement, body reaction at electrode interface, or lead malfunction (fracture or damage to insulation), loss of normal device function due to battery failure or component malfunction, device migration, pocket erosion, or hematoma, pectoral muscle stimulation, phrenic nerve or diaphragmatic stimulation. The following, in addition to the above, are potential complications associated with the use of rate-modulated pacing systems: inappropriate, rapid pacing rates due to sensor failure or to the detection of signals other than patient activity, loss of activity-response due to sensor failure, palpitations with high-rate pacing.

Refer to the User's Manual for detailed indications, contraindications, warnings, precautions and potential adverse events.

Product Specifications

PHYSICAL SPECIFICATIONS

Model	PM1172
Telemetry	Inductive
Dimensions (mm)	41 x 50 x 6
Weight (g)	19
Volume (cc)	9.7
Connector	IS-1

Remote Monitoring

Compatible with Merlin@home™ Transmitter

PARAMETER SETTINGS

Rate/Timing

(Fixed) (ms)	125; 160–400 in steps of 30; 440; 470 ¹
Base Rate (min ⁻¹)	30–130 in steps of 5; 140–170 in steps of 10
Mode	VOO(R); VVI(R); VVT(R); Pacing Off AOO(R); AAI(R); AAT(R)
Hysteresis Rate (min ⁻¹)	Off; 30 ⁴ –150 in steps of 5
Search Interval (min ⁻¹)	Off; 1; 5; 10; 15; 30
Cycle Count	1–16 in steps of 1
Intervention Rate (min ⁻¹)	Off; 80–120 in steps of 10; Intrinsic +0; Intrinsic +10; Intrinsic +20; Intrinsic +30; Same as Base Rate
Intervention Duration (min)	1–10 in 1 minute intervals
Recovery Time	Fast; Medium; Slow; Very Slow
Rest Rate (min ⁻¹)	Off; 30–150; in steps of 5
Rate Responsive VREF	Off; Low; Medium; High
Shortest VREF	125–475 in steps of 25

Output/Sensing

ACap™ Confirm ⁶	On; Off; Monitor
Primary Pulse Configuration	Bipolar
Backup Pulse Configuration	Bipolar
Backup Pulse Amplitude (V)	5.0 ⁴
Search Interval (hours)	8; 24
A or V Pulse Amplitude (V)	0.25–4.0 in steps of 0.25; 4.5–7.5 in steps of 0.5
A or V Pulse Width (ms)	0.05; 0.1–1.5 in steps of 0.1
A or V Pulse Configuration	Unipolar (tip-case); Bipolar (tip-ring)
A or V Sense Configuration	Unipolar Tip (tip-case); Bipolar (tip-ring); Unipolar Ring (ring-case)
Atrial Sensitivity (mV)	0.1–0.4 ⁶ in steps of 0.1; 0.5; 0.75–2.0 in steps of 0.25; 2.5–4.0 in steps of 0.5; 5.0 ⁷
V Sensitivity (mV)	0.5–5.0 in steps of 0.5; 6–10 in steps of 1.0; 12.5 ⁷
Ventricular AutoCapture™ Pacing System	On; Off
Primary Pulse Configuration	Unipolar; Bipolar
Backup Pulse Configuration	Unipolar; Bipolar
Backup Pulse Amplitude (V)	5.0 ⁸
Search Interval (hours)	8; 24
SenseAbility™ Technology	Off; On (Automatic Sensitivity Control adjustment for atrial or ventricular events)
A Max Sensitivity (mV)	0.2–1.0 in steps of 0.1
V Max Sensitivity (mV)	0.2–2.0 in steps of 0.1
Threshold Start	(Atrial and Ventricular Post-Sense) 50; 62.5; 75; 100% (Atrial Post-Pace) 0.2–3.0 in steps of 0.1 mV (Ventricular Post-Pace) Auto; 0.2–3.0 in steps of 0.1 mV (Atrial and Ventricular Post-Sense) 0; 30; 60; 95; 125; 160; 190; 220 (Atrial Post-Pace) 0; 30; 60; 95; 125; 160; 190; 220 (Ventricular Post-Pace) Auto; 0; 30; 60; 95; 125; 160; 190; 220
Decay Delay (ms)	

AF Management

AF Suppression™ Algorithm	Off; On
Lower Rate Overdrive (min ⁻¹)	10 ⁴
Upper Rate Overdrive (min ⁻¹)	5 ⁴
No. of Overdrive Pacing Cycles	15–40 in steps of 5
Rate Recovery (ms)	8; 12 ⁴
Maximum AF Suppression Rate (min ⁻¹)	80–150 in steps of 5; 160–180 in steps of 10
Atrial Tachycardia Detection Rate (min ⁻¹)	110–200 in steps of 10; 225–300 in steps of 25

Rate-Modulated Parameters

Maximum Sensor Rate (min ⁻¹)	80–150 in steps of 5; 160–180 in steps of 10
Reaction Time	Very Fast; Fast; Medium; Slow
Recovery Time	Fast; Medium; Slow; Very Slow
Sensor	On; Off; Passive
Slope	Auto (-1); Auto (+0); Auto (+1); Auto (+2); Auto (+3); 1–16 in steps of 1
Threshold	Auto (-0.5); Auto (+0.0); Auto (+0.5); Auto (+1.0); Auto (+1.5); Auto (+2.0); 1–7 in steps of 0.5

Stored Electrograms

Options	
Priority Options	Off; Low; High
Channel	1; 2; 3
Triggers	
Magnet Response	Off; Low; High
High Atrial Rate	Off; Low; High
Rate (min-1)	125–300 in steps of 25
No. of Consecutive Cycles	2; 3; 4; 5; 10; 15; 20
Advanced Hysteresis	Off; Low; High
Noise Reversion	Off; Low; High

High Ventricular Rate can alternately be High Atrial Rate; they use the same sub-parameters.

Other

V Lead Monitoring	Monitor; Auto Polarity Switch
V Low Impedance Limit (Ω)	100–500 in steps of 25
V High Impedance Limit (Ω)	750–2500 in steps of 250; 3000
Atrial limits apply when implanted in the atrium.	
Lead Type	Uncoded; Unipolar; Bipolar
Magnet Response	Off; Battery Test
NIPS Options	
Stimulation Chamber	Atrial; Ventricular
Coupling Interval (ms)	100–800 in steps of 10 ⁴
S1 Count	2–25 in steps of 1
S1 ⁴ ; S2; S3 and S4 Cycle (ms)	Off; 100–800 in steps of 10 (Fixed or Adaptive)
Diagnostic Trends	AT/AF Activity, Exercise; Lead Impedance; R (or P) Wave; V (or A) Threshold

MRI Settings

MRI Mode	AOO; VOO; DOO; Pacing Off
MRI Base Rate	30–120 min ⁻¹
MRI Pulse Configuration	Bipolar
MRI Pulse Amplitude	5.0 V; 7.5 V
MRI Pulse Width	1.0 ms
MRI Paced AV Delay	25–120 ms

MRI Scan Parameters**

MRI Ready Lead	Lead Lengths	Magnet (Tesla)	Scanner Mode	Scan Region
Tendril™ 2088TC Lead	46, 52, 58 cm	1.5 T, 3 T	Normal Operating Mode	Full Body
IsoFlex™ Optim™ 1944 Lead	46, 52 cm			
IsoFlex™ Optim™ 1948 Lead	52, 58 cm			

**Refer to the MRI Ready Systems Manual for more detailed information.

Customer Support: 46-8-474-4756

Brief Summary: Prior to using these devices, please review the Instructions for Use for a complete listing of indications, contraindications, warnings, precautions, potential adverse events and directions for use.

™ Indicates a trademark of the Abbott group of companies.

© 2018 Abbott. All Rights Reserved.

26020-SJM-END-0914-0004(7)
It is approved for international use only.

1. A.V = 2.5 V @ 0.4 ms; 500 ohms; 100% VVI pacing @ 60 bpm; AutoCapture™ Pacing System OFF; SEGMs ON.
2. Terms and conditions apply; refer to the warranty for details.
3. Programming options dependent on pacing mode.
4. The highest available setting for hysteresis rate will be 5 min-1 below the programmed base rate.
5. Atrial Implants Only.
6. Values 0.1–0.4 not available in a unipolar sense configuration.
7. Sensitivity is with respect to a 20 ms haversine test signal.
8. This parameter is not programmable.
9. S1 Burst Cycle is applied at the preprogrammed S1 cycle length.



Tendril™ STS

Pacing Lead

Product Highlights - Pacing Lead

- The Tendril STS lead allows patients to undergo MRI scans when used in conjunction with a MRI Ready pacemaker from St. Jude Medical
 - Allows MRI scans (See Parameter Settings for scan exclusion zone)
 - Permits a maximum whole body averaged specific absorption rate (SAR) of 2 watts per kilogram (W/kg)
- Soft silicone tip offers more compliance and less tip pressure at the lead tip-endocardium interface
- Small diameter lead offers improved ease of venous passage, reduced risk of venous thrombosis or rib-clavicle crush and ability to accommodate additional leads more easily
- Optim™ lead insulation—a chemical co-polymer that blends the best features of polyurethane and silicone for improved handling and increased durability
- Titanium nitride (TiN) fractal coating on the tip and ring electrodes is designed to promote precise sensing and to provide improved contact with the myocardium
- Lubricious Fast-Pass™ coating facilitates lead insertion through the introducer and veins to ease implantation
- Fits through a 6 F introducer



Ordering Information - MRI-Ready Pacing System

Model Number	Description	Insulation	Fixation	Min. Introducer (F)	Connector	Length (cm)
2088TC	Tendril™ STS Pacing Lead	Optim™	Ext/Ret helix	6	IS-1 bipolar	46*; 52*; 58*; 65; 100

* Indicates lead lengths that are MRI conditional with a scan exclusion zone.

Model Number	Description	Dimensions (H x W x T, mm)	Weight (g)	Volume (cc)	Connector
PM1140	Endurity™ Core Pacemaker	41 x 50 x 6	19	9,7 (± 0,5)	IS-1
PM2140	Endurity Core Pacemaker	46 x 50 x 6	19	10,4 (± 0,5)	IS-1
PM1152	Endurity Core Pacemaker	41 x 50 x 6	19	9,7 (± 0,5)	IS-1
PM2152	Endurity Core Pacemaker	46 x 50 x 6	19	10,4 (± 0,5)	IS-1
PM1162	Endurity Pacemaker	41 x 50 x 6	19	9,7 (±0,5)	IS-1
PM2162	Endurity Pacemaker	46 x 50 x 6	19	10,4 (± 0,5)	IS-1
PM1172	Endurity MRI™ Pacemaker	41 x 50 x 6	19	9,7 (± 0,5)	IS-1
PM2172	Endurity MRI Pacemaker	46 x 50 x 6	19	10,4 (± 0,5)	IS-1
PM1272	Assurity MRI™ Pacemaker	47 x 50 x 6	20	10,4 (±0,5)	IS-1
PM2272	Assurity MRI Pacemaker	47 x 50 x 6	20	10,4 (±0,5)	IS-1

Indications: Tendril™ STS lead is designed for permanent sensing and pacing in either the right atrium or the right ventricle, in combination with a compatible device. Active leads such as the Tendril STS lead may be indicated for patients where permanent fixation of passive leads is suspected to be unstable.

In atrial applications, the use of screw-in leads such as Tendril STS lead may be indicated in the presence of an abnormal, surgically altered or excised atrial appendage.

Contraindications: Tendril STS lead is contraindicated: in the presence of tricuspid atresia, for patients with mechanical tricuspid valves, in patients who are expected to be hypersensitive to a single dose of one milligram of dexamethasone sodium phosphate.

Adverse Events: Potential complications associated with the use of Tendril STS lead are the same as with the use of other active fixation leads and include: cardiac tamponade, diaphragmatic stimulation, embolism, excessive bleeding, induced ventricular ectopy, infection, loss of pacing and/or sensing due to dislodgment or mechanical malfunction of the pacing lead, phrenic nerve stimulation, thrombosis. Complications reported with direct subclavian venipuncture include pneumothorax, hemothorax, laceration of the subclavian artery, arteriovenous fistula, neural damage, thoracic duct injury, cannulation of other vessels, massive hemorrhage and, rarely, death.

Refer to the User's Manual for detailed indications, contraindications, warnings, precautions and potential adverse events.

Tendril™ STS

Pacing Lead

Product Specifications - Pacing Leads

PHYSICAL SPECIFICATIONS

Model	2088TC
Minimum Introducer Size	6 F
Type of Lead	Active-fixation, bipolar, steroid-eluting, endocardial, pacing lead
Lead Connector	IS-1 bipolar
Lead Lengths	46; 52; 58; 65; 100 cm
Fixation Mechanism	Extendable/Retractable helix
Typical Number of Rotations for Helix Extension	6-11 (straight stylet)
Lead Body Diameter	1,9 mm (max)
Tip-to-Ring Spacing	10 mm
Lead Tip Electrode (Cathode)	Active titanium-nitride-coated Pt/Ir helix (2,0 mm extension)
Tip Electrode Surface Area	6,9 mm ²
Ring Electrode (Anode)	Titanium-nitride-coated Pt/Ir
Ring Electrode Surface Area	16 mm ²
Mapping	Capable with titanium-nitride-coated Pt/Ir helix
Steroid	< 1 mg dexamethasone sodium phosphate
Inner Conductor/Outer Conductor	MP35N™* coil
Inner Insulation	Silicone rubber
Outer Insulation	Optim™ lead insulation
Lead Body Coating	Fast-Pass™ coating

In Pack

Straight stylets	1 x-soft in lead; 1 x-soft; 1 soft
J-curved stylets	2 soft
Helix extension/retraction clip-on tools	2 clip-on tools

Accessory Kits

Available Separately	Model Number	Compatible Lengths	Description
Stylet Kit	DS06002 with appropriate length designation	46; 52; 58; 65; 100 cm	1 fixation tool; 1 clip-on tool; 1 J-shaped soft; 1 x-soft; 1 soft; 1 firm; 1 x-firm
	DS06003 with appropriate length designation	46; 52; 58; 65; 100 cm	1 clip-on tool; 1 J-shaped soft; 1 x-soft; 1 soft; 1 firm; 1 x-firm
Locator™ Plus	1281 with appropriate length designation	46; 52; 58; 65 cm	Disposable implant tool to facilitate precise lead positioning and manipulation with one hand
Deflectable Stylet	1292 with appropriate length designation	46; 52; 58; 65 cm	

MRI Conditional Parameters

Magnet strength: 1.5 Tesla
 SAR: ≤ 2 W/kg
 Scan region: Isocenter must be inferior to L4 or 10 cm superior to C1



*MP35N is a trademark of SPS Technologies, Inc.

Customer Support: 46-8-474-4756

Brief Summary: Prior to using these devices, please review the Instructions for Use for a complete listing of indications, contraindications, warnings, precautions, potential adverse events and directions for use. Devices depicted may not be available in all countries. Check with your St. Jude Medical representative for product availability in your country.

Unless otherwise noted, ™ indicates that the name is a trademark of, or licensed to, St. Jude Medical or one of its subsidiaries. ST. JUDE MEDICAL and the nine-squares symbol are trademarks and service marks of St. Jude Medical, Inc. and its related companies. © 2015 St. Jude Medical, Inc. All Rights Reserved.
 SJM-TND-1214-0003 | This document is for International use only.



TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD
 ZERTIFIKAT ♦ CERTIFICATE ♦ 認證證書 ♦ CERTIFICADO ♦ CERTIFICAT



Benannt durch/Designated by:
 Zentralstelle der Länder
 für Gesundheitsschutz
 bei Arzneimitteln und
 Medizinprodukten
 www.zlg.de
 ZLG-BS-200.14.03



Product Service

EC Certificate

EC Design-Examination Certificate

Directive 90/385/EEC on Active Implantable Medical Devices (AIMDD), Annex 2 (4)
 (Other devices than custom made or intended for clinical investigation)

No. I7 014607 0234 Rev. 00

Model(s):

see below

Facility(ies):

St. Jude Medical Cardiac Rhythm Management Division
 15900 Valley View Court, Sylmar CA 91342, USA

St. Jude Medical Puerto Rico LLC
 Lot A Interior - #2 Rd Km. 67.5, Santana Industrial Park, Arecibo
 PR 00612, USA

St. Jude Medical Operations (M) Sdn.Bhd.
 Plot 102, Lebuhraya Kampung Jawa, Bayan Lepas Industrial Zone,
 11900 Penang, MALAYSIA

Parameters

./.

**Design
Facility(ies):**

St. Jude Medical Cardiac Rhythm Management Division
 15900 Valley View Court, Sylmar, CA 91342, USA

Product:

Implantable Pacemakers

Test Report No.: 70069297

Model:

Model No.:

Variant:

Microny™ II SR+

2525T

Test Report No.: 70110810

Model:

Model No.:

Variant:

Zephyr™ SR
 Zephyr™ DR
 Zephyr™ XL DR

5620
 5820
 5826

Page 2 of 4

TÜV SÜD Product Service GmbH is Notified Body with identification no. 0123

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany

TUV®

TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD
 ZERTIFIKAT ♦ CERTIFICATE ♦ 認證證書 ♦ СЕРТИФИКАТ ♦ CERTIFICADO ♦ CERTIFICAT



Benannt durch/Designated by
 Zentralstelle der Länder
 für Gesundheitsschutz
 bei Arzneimitteln und
 Medizinprodukten
 www.zlg.de
 ZLG-BS-200.14.03



Product Service

EC Certificate

EC Design-Examination Certificate
 Directive 90/385/EEC on Active Implantable Medical Devices (AIMDD), Annex 2 (4)
 (Other devices than custom made or intended for clinical investigation)

No. I7 014607 0234 Rev. 00

Test Report No.: 71321436

Model:	Model No.:	Variant:
Zephyr™ XL SR	5626	

Test Report No.: 713017309_1

Model:	Model No.:	Variant:
Assurity™	PM1240	
Assurity™	PM2240	
Endurity™	PM1160	
Endurity™	PM2160	
Allure™	PM3120	
Allure™ RF	PM3222	
Allure Quadra™ RF	PM3242	

Test Report No.: 713028360

Model:	Model No.:	Variant
Quadra Allure MP™RF	PM3262	

Test Report No.: 713043621

Model:	Model No.:	Variant:
Assurity MRI™	PM1272	MR Conditional
Assurity MRI™	PM2272	MR Conditional
Endurity MRI™	PM1172	MR Conditional
Endurity MRI™	PM2172	MR Conditional
Endurity™	PM1162	MR Conditional
Endurity™	PM2162	MR Conditional





ST. JUDE MEDICAL

90264376 Rev. G

Declaration of Conformity

SJM Declaration of Conformity
Implantable Pacemakers
ATTACHMENT TO DECLARATION OF CONFORMITY

St. Jude Medical (SJM) hereby declares that the following SJM facilities and products conform to the applicable provisions of Annex 2 the European Union's Active Implantable Medical Devices Directive, AIMDD, 90/385/EEC. All supporting documentation is retained under the premises of SJM. We declare no application has been lodged with any other notified body for the same products. This declaration is issued under the sole responsibility of the manufacturer. This declaration supersedes any declaration issued previously for the same product(s).

Manufacturer Address: St. Jude Medical Cardiac Rhythm Management Division
15900 Valley View Court
Sylmar, CA 91342

European Representative: St. Jude Medical Coordination Center BVBA
The Corporate Village
Da Vincilaan 11 Box F1
1935 Zaventem, Belgium

Product Type: Implantable Pacemakers

Product Name(s): See Attachment

Model Number(s): See Attachment

Classification: AIMD

GMDN Code(s): See Attachment

Original CE Mark Date: See Attachment

(FQA or EC as appropriate) Certificate No and expiration date: EC
Certification No: I7 014607 0234 Rev. 00
Expiration Date: 2024-05-26

FQA
Certificate No: I1 16 12 14607 211
Expiration Date: 2021-07-25

ISO13485
Certificate No: Q1N 17 09 14607 217
Expiration Date: 2020-10-31

Signature:



Kathy Berg
Manager Regulatory Affairs


Issue Date



SJM Declaration of Conformity
Implantable Pacemakers
ATTACHMENT TO DECLARATION OF CONFORMITY**Applicable Quality System Standards:**

Fulfills the requirements of Annex 2 of the European Union's Active Implantable Medical Devices Directive, AIMDD, 90/385/EEC and corresponding national legislation.

Fulfills applicable requirements including CE marking and the Essential Requirements of the AIMDD, 90/385/EEC and corresponding national legislation.

Notified Body:

TÜV SÜD Product Service GmbH Zertifizierstelle
Ridlerstraße 65, 80339, Munich, Germany

Notified Body Number:


0123

Manufacturing Facilities:

St. Jude Medical Cardiac Rhythm Management Division
15900 Valley View Court Sylmar, CA 91342 USA

St. Jude Medical Puerto Rico LLC
Lot A Interior - #2 Rd Km. 67.5, Santana Industrial Park,
Arecibo PR 00612, USA

St. Jude Medical Operations (M) Sdn. Bhd
Plot 102, Lebuhraya Kampung Jawa, Bayan Lepas
Industrial Zone, 11900 Penang, MALAYSIA

Signature:

Kathy Berg
Manager Regulatory Affairs



Issue Date



SJM Declaration of Conformity

Implantable Pacemakers

ATTACHMENT TO DECLARATION OF CONFORMITY

The following product(s) is/are approved under EC-certificate number **I7 014607 0230 Rev. 00:**

Product Name	Model No.	GMDN Codes	First Date of CE Marking
Microny™ II SR+	2525T	47267	1999-9-17
Zephyr™ XL DR	5826	47265	2006-5-9
Zephyr™ DR	5820	47265	2006-5-9
Zephyr™ SR	5620	47267	2006-5-9
Zephyr™ XL SR	5626	47267	2007-6-13
Assurity™	PM1240	47267	2013-3-7
Assurity™	PM2240	47265	2013-3-7
Endurity™	PM1160	47267	2013-3-7
Endurity™	PM2160	47265	2013-3-7
Allure™	PM3120	47263	2013-3-7
Allure™ RF	PM3222	47263	2013-3-7
Allure Quadra™ RF	PM3242	47263	2013-3-7
Quadra Allure MP™ RF	PM3262	47263	2014-7-31
Assurity MRI™	PM1272 (MR Conditional)	47267	2014-12-18
Assurity MRI™	PM2272 (MR Conditional)	47265	2014-12-18
Endurity MRI™	PM1172 (MR Conditional)	47267	2014-12-18
Endurity MRI™	PM2172 (MR Conditional)	47265	2014-12-18
Endurity™	PM1162 (MR Conditional)	47267	2014-12-18
Endurity™	PM2162 (MR Conditional)	47265	2014-12-18
Endurity™ Core	PM1140 (MR Conditional)	47267	2015-7-24
Endurity™ Core	PM2140 (MR Conditional)	47265	2015-7-24
Endurity™ Core	PM1152 (MR Conditional)	47267	2015-7-24
Endurity™ Core	PM2152 (MR Conditional)	47265	2015-7-24
Quadra Allure™	PM3542 (MR Conditional)	47263	2016-10-21
Quadra Allure MP™	PM3562 (MR Conditional)	47263	2016-10-21
Zenex™	PM1250	47267	2018-10-12
Zenex™	PM2250	47265	2018-10-12
Zenus™	PM1170	47267	2018-10-12
Zenus™	PM2170	47265	2018-10-12
Zenex MRI™	PM1282 (MR Conditional)	47267	2018-10-12

Signature:


Kathy Berg
Manager Regulatory Affairs

14 Jun 2019
Issue Date





ST. JUDE MEDICAL

90264376 Rev. G

Declaration of Conformity


SJM Declaration of Conformity
Implantable Pacemakers
ATTACHMENT TO DECLARATION OF CONFORMITY

Product Name	Model No.	GMDN Codes	First Date of CE Marking
Zenex MRI™	PM2282 (MR Conditional)	47265	2018-10-12
Zenus MRI™	PM1182 (MR Conditional)	47267	2018-10-12
Zenus MRI™	PM2182 (MR Conditional)	47265	2018-10-12

Signature:



Kathy Berg
Manager Regulatory Affairs


Issue Date

86480 SJM Declaration of Conformity Template Rev D

Page 4 of 4

This confidential document is the property of St. Jude Medical and shall not be reproduced, distributed, disclosed or used without the express written consent of St. Jude Medical.

