

EU DECLARATION OF CONFORMITY

V2.2/91/168-206-218-186 from 29.04.2021

Matrix Medical Disposable Face Mask Type II R

We, as Authorized Representative

MATRIX IT SOLUTIONS SRL

str. Bucuresti Nr.230, Biroul Nr.1, 077125 ors. MAGURELE, jud. ILFOV, ROMANIA
VAT-No.: RO30208314, Tel/Fax: 0216670711/0216670722, E-mail: info@itmatrix.ro

of the manufacturer

name and address of manufacturer

NEOMATRIX SRL

#121, 31 August 1989 Street, MD-2012 Chisinau, Republic of Moldova

declare under our sole responsibility,
that the following product:

Product Model[s]

MASCĂ FACIALĂ DE UZ MEDICAL TYPE II R DE UNICĂ FOLOSINȚĂ

Medical-Mask-TYPE II R_N1; Medical-Mask-TYPE II R_N5;

Medical-Mask-TYPE II R_N10; Medical-Mask-TYPE II R_N50.

Part Number[s]

91168; 91206; 91218; 91186.

Brand Name [trade mark & logo]

MATRIX



Product Type classification
according to harmonised European
Standard EN 14683:2019+AC:2019

Type II R Medical Face Mask

Product Type classification
according to Medical Devices
Directive 93/42/EEC and
Regulation (EU) 2017/745

Class I medical device: non-sterile, non-measuring

Intended Use

The purpose of this Type II R Medical Face Mask is to reduce the risk of infection transfer especially from the person wearing the mask to their surroundings. The product should cover the mouth and nose for providing a barrier to minimize direct transmission of infectious germs between staff and patients.

meets all essential technical requirements and is in conformity with the relevant applicable EU legislation

Applied Directive(s)

Council Directive 93/42/EEC concerning medical devices;
Regulation (EU) 2017/745 of the European Parliament and of the
Council of 5 April 2017 on medical devices, amending Directive 2001/83/EC,
Regulation (EC) No 178/2002 and Regulation (EC) No 1223/2009
and repealing Council Directives 90/385/EEC and 93/42/EEC

Applied Standard(s)

EN 14683:2019 + AC:2019 Medical Face Masks

- Requirements and test methods

(including: ISO 22609:2004 | ISO 11737-1:2018 | ISO 10993-5:2009)

Manufacturer's Quality Standard(s)

ISO 9001:2015 (QMS) Quality Management System
[Manufacturing of the surgical masks]

The declaration has been carried out in accordance with conformity assessment procedures for protective equipment.

Assessment has been carried out in accordance with the evaluation of:

Test Report(s) No

GZHT02407734-S1

Test conducted by

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch
on behalf of Intertek Deutschland GmbH

Test Lab address

3/F., Hengyun Building, 235 Kafa Ave., Guangzhou Economic & Technological
Development District, Guangzhou, P.R.C. (510730)

Issue Date

15-Apr-2021

Revision Date/No

-

* The undersigned herewith declares, that the above-mentioned product(s) meet the provisions of the relevant EC Council Directives, Regulations and harmonized standards. All supporting documentation is retained under the premises of the manufacturer.

ors. MAGURELE, jud. ILFOV, ROMANIA
29-Apr-2021

General Manager,
MIRCEA CIOBANU





NeoMATRIX SRL

#7/1 Nikolai Zelinski Str., MD-2015 Chisinau, Republic of Moldova
Legal address: #121, 31 August 1989 Str., MD-2012 Chisinau,
Republic of Moldova

Bureau Veritas Certification Holding SAS – UK Branch certifies that the Management System of the above organisation has been audited and found to be in accordance with the requirements of the management system standards detailed below

ISO 9001:2015

Scope of certification

Manufacturing of the surgical masks.

Original cycle start date: **29 April 2021**
Certification / Recertification Audit date: **24 February 2021**
Certification / Recertification cycle start date: **29 April 2021**
Subject to the continued satisfactory operation of the organization's Management System,
this certificate expires on: **28 April 2024**

Certificate No. UA230071 Version: 0 Revision date: 29 April 2021



Anatoliy Zvon
Signed on behalf of BVCH SAS UK Branch




0008

Certification body address: **5th Floor, 66 Prescott Street, London E1 8HG, United Kingdom**
Local office: **5th floor, 28, Simon Petlyura St., Kyiv, 01032, UKRAINE**

Further clarifications regarding the scope and validity of this certificate, and the applicability of the management system requirements, please call: **+380 44 354 16 00**



MEDICAL DISPOSABLE FACE MASK TYPE IIR



3-Folds | 3-Layers | with Elastic Ear-loops

GENERAL INFORMATION

MANUFACTURER

Name: **NEOMATRIX SRL**
Address: #121, 31 August 1989 Street,
MD-2012 Chisinau, Republic of MOLDOVA
ISO 9001:2015 Bureau Veritas Certified Company
Quality Management System:
Manufacturing of the surgical masks

Authorized Representative in Moldova:
NEOMATRIX SRL

Authorized Representative
in Romania & European Union:

MATRIX IT SOLUTIONS SRL
str. Bucuresti Nr.230, Biroul Nr.1,
077125 ors. MAGURELE, jud. ILFOV, ROMANIA
VAT-No.: RO30208314
Tel.: / Fax: 021 667 07 11/021 667 07 22
E-mail: info@itmatrix.ro

CONFORMITY ASSESSMENT PROCEDURE

According to Regulation (EU) 2017/745 of the
European Parliament and of the Council of 5 April
2017 on medical devices, amending Directive
2001/83/EC, Regulation (EC) No 178/2002 and
Regulation (EC) No 1223/2009 and repealing Council
Directives 90/385/EEC and 93/42/EEC

NOTIFIED BODY

No involvement of a Notified Body is needed for this
Non-Sterile class I device.



PRODUCT INFORMATION

INTENDED USE

The purpose of this Medical Face Mask is to reduce the risk of
infection transfer especially from the person wearing the mask
to their surroundings. The product should cover the mouth and
nose for providing a barrier to minimize direct transmission of
infectious germs between staff and patients.

DESCRIPTION

Rectangular medical face masks with a shapeable nose piece
and two ear-loops on each side to hold mask in place.

Trade Mark & logo: **MATRIX**  

Model[s]: **Medical-Mask-TYPE II R_N50 | Medical-Mask-TYPE II R_N1
Medical-Mask-TYPE II R_N10 | Medical-Mask-TYPE II R_N5**

Romanian: **Masca faciala de uz medical TYPE II R de unica folosinta**

Part Number[s]: **91168; 91206; 91218; 91186**

This product is **Type IIR Medical Face Mask** according to European
Standard: **EN 14683:2019 + AC:2019**
+ ISO 10993 Biocompatibility
in vitro cytotoxicity, Test method ISO 10993-5-2009


MATERIAL

Outside Layer: **Spunbond** non-woven 100% pure Polypropylene - [blue] 23g/m²
Middle Layer: **BFE99 Meltblown** non-woven Polypropylene - [white] 25g/m²
Inner Layer: **Spunbond** non-woven 100% pure Polypropylene - [white] 23g/m²
Nose piece: Plastic covered iron
Elastic Band: Polyester
Total mask density: 71g/m²
Latex Free & Fiber-glass Free

DIMENSIONS

Length: 175mm Width 97mm [+/- ~3%]
Length of ear loop: 180mm each
Length of nose piece: 100mm

MANUFACTURING COUNTRY
MADE IN MOLDOVA



REGULATION & TESTING INFORMATION

REGULATORY INFORMATION

Product CE marked as per 93/42/EEC Directive on Medical Devices & EU Regulation 2017/745.

Class I Medical Device - Non-Sterile & Non-Measurable
Type II R Medical Face Mask - Non-Sterile

TEST METHODS

Bacterial Filtration Efficiency (BFE)

The ability of the face mask to filter our bacteria so that they are not released into the user's surroundings (BFE), (%)
 When tested in accordance with Annex B of EN 14683, the BFE of the medical face mask shall conform to the minimum value given for the relevant type in Table 1 of EN 14683:2019 + AC:2019.

Differential Pressure (Breathability)

The lower this value, the easier it is for the user to breath normally (Pa/cm²)

The differential pressure for the medical face mask shall be <40 Pa/cm² for Type I / Type II masks and <60 Pa/cm² for medical face masks Type II R in accordance with Annex C of EN 14683:2019 + AC:2019.

Splash Resistance Pressure

Splash resistance pressure: The ability of the face mask to withstand the penetration of liquid splashes (kPa) [protection from droplets]. No penetration at 16.0 kPa should be attested in accordance to EN 14683:2019 + AC:2019 using ISO 22609:2004 Test method - for resistance against penetration by synthetic blood.

Microbial cleanliness (Bioburden)

The bioburden (microbial cleanliness) of the medical face masks shall be ≤ 30 cfu/g when tested in accordance to Annex D of EN 14683:2019 + AC:2019 and namely 5 test specimens selected randomly from top, bottom and center of the same box/batch/lot (ISO 11737-1:2018).

TABLE
Performance Requirements for Medical Face Masks
according to EN 14683:2019+AC:2019

TEST	TYPE I	TYPE II	TYPE II R	MATRIX Medical Mask TESTS RESULTS
Bacterial Filtration Efficiency [BFE], %	≥95	≥98	≥98	99.9 PASS
Differential Pressure (Pa/cm ²)	<40	<40	<60	47.84 PASS
Splash Resistance Pressure (kPa)	not required	not required	≥16,0	PASS
Microbial Cleanliness (cfu/g) [Bioburden]	≤30	≤30	≤30	7.63 PASS

ADDITIONAL TEST PERFORMED FOR MATRIX MASKS

Cytotoxicity
 In Vitro Cytotoxicity Test [MTT Method]
 according to ISO 10993-5:2009 to test the
 potential cytotoxicity of the test article

test article has no
 potential toxicity to
 L-929 cells in the
 MTT Method

STERILIZATION


This mask is non-sterile



TESTED BY:



THREE LAYER PROTECTION



STORAGE

Store in a dry and cool place, away from intense sources of heat. Keep the masks as much as practicably possible in their dispenser box. Keep dispenser boxes as much as practicably possible in their shipping case.



PACKING

Shipping case of 2500 Units (masks)
 50 Units are placed in 1 [one] box and 50 boxes are placed in 1 [one] Shipping Case

Box dimension: 190x100x87mm

Box weight: 170gr

Box material: 225gr-paper

Shipping case dimension: 510x390x435mm


Shipping case material: Carton

Shipping case weight: 9kg

The uninterrupted use duration of the device is usually less than 2 hours.

SHELF LIFE

shelf-life is 2 years from production, if stored properly.
 marked on package:



BARCODE



VISUALS

