

DECLARATION OF CONFORMITY

MANUFACTURER: Alcohol Countermeasure Systems Corp.
ADDRESS: 60 International Boulevard, Toronto, ON M9W 6J2, CANADA
PRODUCT: ALERT™ 5 breath alcohol tester
INTENDED USE: Law enforcement
POWER SUPPLY: 2 x 1.5V (AA batteries) alkaline
Classification: Group 1, Class B

The device noted above has been tested with the essential requirements and other relevant provisions of EMC Directive 2004/108/EC when used for its intended purpose:

Electromagnetic compatibility pursuant to EN 55011:2007 +A2

Applied standard(s): EN 55011:2007 + A2 (Emissions)
EN 61326-1: 2006 (ESD Susceptibility)

Alcohol Countermeasure Systems Corp hereby declares that the ALERT J5 product complies with the above Directives and that the following marking applies to this product :



Alcohol Countermeasure Systems Corp further declares that manufacture of the ALERT J5 product met the requirements of Quality Management System according to ISO 9001 and IATF 16949 quality standards and in accordance with the Good Manufacturing Practice of US FDA 21 CFR Part 820.

Dated at Toronto, this 30th day of June 2021,

For and on behalf of the company,
ALCOHOL COUNTERMEASURE
SYSTEMS CORP.



Felix JE Comeau
President