CERTIFICATE



of Conformity

Registration No.:

AK 50231450 0001

Report No.:

10020210 002

Holder:

Broadsound Corporation 5F, No. 31, Shintai Rd., Jupei City, Hsinchu 302 Taiwan

Product:

Medical Device

(Medical Diagnostic Ultrasound Transducer Assemblies)

Identification:

ME3C40B, ME3C60B, ME3C36B, ME7L38B, ME6E10B, ME5C50B,

AT8L125, PH3C52B, PH7L95B, PH6E94B, PH5C73B

Tested acc. to:

IEC 60601-1:1988+A1+A2

The certificate of conformity refers to the above mentioned product. This is to certify that the specimen is in conformity with the assessment requirement mentioned above. This certificate does not imply assessment of the production of the product and does not permit the use of a TÜV Rheinland mark of conformity.

Date

22.06.2012

Certification Body

Dipl.-Ing. W. Hsu

YÜVRheinland

TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg