

TÜV Rheinland LGA Products GmbH • 51105 Köln

SCW Medicath Ltd.
No.4, Baolong 6th Road,
Baolong Industrial Town,
Longgang District, Shenzhen,
518116 Guangdong,
P.R. China

Contact

Tel. +49 911 655-5225
Mail: medical_products@de.tuv.com

Date April 02, 2024

Notified Body Confirmation Letter

Reference. : 10924200

To whom it may concern,

Confirmation of the status of a formal application, written agreement, and appropriate surveillance in the framework of Regulation EU 2023/607 amending Regulations (EU) 2017/745 and (EU) 2017/746 as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices.

This letter confirms that **TÜV Rheinland LGA Products GmbH**, a Notified Body (NB) designated against Regulation (EU) 2017/745 (MDR) and identified by the number **0197** on NANDO, has received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of MDR and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of MDR with the following manufacturer:

SCW Medicath Ltd.
No.4, Baolong 6th Road,
Baolong Industrial Town,
Longgang District, Shenzhen,
518116 Guangdong,
P.R. China
SRN Number (if available): CN-MF-000019140

TÜV Rheinland
LGA Products GmbH

Am Grauen Stein
51105 Köln
Germany

Headquarter

Tillystraße 2
90431 Nuremberg

Phone. +49 911 655 5225
Fax +49 911 655 5226
service@de.tuv.com
www.tuv.com/safety

Board of Management

Dipl.-Ing. Thomas Weigand, Spokesman

Dipl.-Kfm. Dr. Jörg Schlosser

Nuremberg HRB 26013
VAT No.: DE 811835490

Chairman of the
Supervisory Board

Dr.-Ing. Michael Fübi

The transition timelines that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 120.3c of MDR (as amended by (EU) 2023/607), are shown below:

- May 26, 2026 for Class III custom-made implantable devices
- December 31, 2027 for Class III devices and Class IIb implantable devices excluding Well-established technologies (WET - sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors)
- December 31, 2028 for other Class IIb devices, Class IIa, Class I devices placed on the market in sterile condition or have a measuring function
- December 31, 2028 for devices not requiring the involvement of a notified body under MDD but requiring it under MDR (e.g., class I devices that qualify as re-usable surgical instruments)

On behalf of the Notified Body



Samuel QIN

Certification body

Table 1: Devices covered by this letter and for which the NB is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive:

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Postpartum Balloon Basic UDI-DI: 69388205000000000000006027	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Postpartum Balloon Basic UDI-DI: 6938820500000000000000362A	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Cervical Ripening Balloon Basic UDI-DI: 69388205000000000000006129	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Cervical Ripening Balloon Basic UDI-DI: 6938820500000000000000372C	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Guide Wire Basic UDI-DI: 693882050000000000000013ZT	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Hemostasis Valve Sets Basic UDI-DI: 693882050000000000000012ZR	Class I devices placed on the market in sterile condition	N/A	Certificate # HD 60144232 0001 NB #0197
Stopcock Basic UDI-DI: 693882050000000000000014ZV	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Balloon Inflation Device Basic UDI-DI: 6938820500000000000000002ZN	Class I devices placed on the market in sterile condition	N/A	Certificate # HD 60144232 0001 NB #0197
Connecting Tubing Basic UDI-DI: 693882050000000000000004ZS	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Manifold Basic UDI-DI: 69388205000000000000001828	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Pressure Bandage Basic UDI-DI: 693882050000000000000006ZW	Class I devices placed on the market in sterile condition	N/A	Certificate # HD 60144232 0001 NB #0197
Colored piston specialty Syringe Basic UDI-DI: 693882050000000000000022ZU	Class I devices placed on the market in sterile condition	N/A	Certificate # HD 60144232 0001 NB #0197
Dose-control Syringe Basic UDI-DI: 693882050000000000000020ZQ	Class I devices placed on the market in sterile condition	N/A	Certificate # HD 60144232 0001 NB #0197
Manifold Set Basic UDI-DI: 69388205000000000000002729	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Ureteral Stent Set Basic UDI-DI: 693882050000000000000031ZV	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Tracheostomy Tube Kits Basic UDI-DI: 6938820500000000000000292D	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Tracheostomy Tube Kits Basic UDI-DI: 6938820500000000000000622B	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Tracheostomy Tube Kits Basic UDI-DI: 6938820500000000000000632D	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Percutaneous Nephrostomy Sets Basic UDI-DI: 6938820500000000000000282B	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Locking Drainage Catheter Basic UDI-DI: 6938820500000000000000392G	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
ERCP Guidewire Basic UDI-DI: 693882050000000000000040ZW	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Percutaneous Access Set Basic UDI-DI: 6938820500000000000000382E	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Infusion Sets with needleless adapter Basic UDI-DI: 69388205000000000000002525	Class I devices placed on the market in sterile condition	N/A	Certificate # HD 60144232 0001 NB #0197
Drainage Catheter Sets Basic UDI-DI: 693882050000000000000023ZW	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Introducer Needles Basic UDI-DI: 69388205000000000000002627	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Introducer Sets Basic UDI-DI: 693882050000000000000011ZP	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Transradial Introducer Sets Basic UDI-DI: 693882050000000000000030ZT	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Hemodialysis Catheterization Kit Basic UDI-DI: 693882050000000000000021ZS	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Hemodialysis Catheterization Kit Basic UDI-DI: 6938820500000000000000662K	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Hemodialysis Catheterization Kit Basic UDI-DI: 6938820500000000000000672M	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Angiographic Syringes Basic UDI-DI: 69388205000000000000001ZL	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Disposable Infusion Pumps Basic UDI-DI: 6938820500000000000000652H	Class IIb excluding Class IIb implantable non-WET	N/A	Certificate # HD 60144232 0001 NB #0197
Patient-Controlled Analgesic Infusion Pumps Basic UDI-DI: 6938820500000000000000642F	Class IIb excluding Class IIb implantable non-WET	N/A	Certificate # HD 60144232 0001 NB #0197
Disposable Pressure Transducers Basic UDI-DI: 6938820500000000000000192A	Class IIb excluding Class IIb implantable non-WET	N/A	Certificate # HD 60144232 0001 NB #0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre- application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
I.V Cannulas Basic UDI-DI: 69388205000000000000002423	Class IIa	N/A	Certificate # HD 60144232 0001 NB #0197
Injection Cap Basic UDI-DI: 6938820500000000000000592N	Class I devices placed on the market in sterile condition	N/A	Certificate # HD 60144232 0001 NB #0197

**Table 2: Devices covered by this letter and for which the NB is NOT
responsible for appropriate surveillance of the corresponding devices under
the applicable Directive:**

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre- application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
N/A	N/A	N/A	N/A

Confirmation Letter Revision History

Date	NB internal reference traceable to each version of the letter	Action
2024/04/02	SCWME_CL607_2024- 04-02	Initial issue



EC Certificate
Directive 93/42/EEC Annex II, excluding Section 4
Full Quality Assurance System
Medical Devices

Registration No.: HD 60144232 0001

Report No.: 17047213 010

Manufacturer: SCW Medicath Ltd.
No. 4 Baolong 6th Road
Baolong Industrial Town
Longgang District, Shenzhen
518116 Guangdong
P.R. China

Products: Medical Devices

(see attachment for products included)

Replaces Approval, Registration No.: HD 60139711 0001

Expiry Date: 2024-05-26

The Notified Body hereby declares that the requirements of Annex II, excluding section 4 of the directive 93/42/EEC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex II, section 5 of the aforementioned directive. For placing on the market of class III devices covered by this certificate an EC design-examination certificate according to Annex II, section 4 is required.

Effective Date: 2020-05-26

Date: 2020-05-26

Notified Body



TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg
TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC
concerning medical devices with the identification number 0197.



TÜV Rheinland
LGA Products GmbH
Tillystraße 2, 90431 Nürnberg

Doc. 1/2 Rev. 0

**Attachment to
Certificate**

Registration No.: HD 60144232 0001
Report No.: 17047213 010

Manufacturer: SCW Medicath Ltd.
No. 4 Baolong 6th Road
Baolong Industrial Town
Longgang District, Shenzhen
518116 Guangdong
P.R. China

Products:

- Disposable Pressure Transducers
- Introducer Sets
- Guide Wires
- Angiographic Syringes
- Hemodialysis Catheterization Kits
- Patient-Controlled Analgesic Infusion Pumps
- Disposable Infusion Pumps
- Tracheostomy Tube Kits
- Percutaneous Nephrostomy Sets
- Ureteral Stent Sets
- Drainage Catheter Sets
- Transradial Introducer Sets
- Introducer Needles
- I.V Cannulas
- Cervical Ripening Balloon
- Postpartum Balloon

Notified Body

Date: 2020-05-26

Fuxiu Sheng





**TÜV Rheinland
LGA Products GmbH
Tillystraße 2, 90431 Nürnberg**

Doc. 2/2, Rev. 0

**Attachment to
Certificate**

Registration No.: HD 60144232 0001
Report No.: 17047213 010

Manufacturer: SCW Medicath Ltd.
No. 4 Baolong 6th Road
Baolong Industrial Town
Longgang District, Shenzhen
518116 Guangdong
P.R. China

Products:

- Locking Drainage Catheters
- Percutaneous Access Sets
- ERCP Guidewires
- Manifolds
- Stopcocks
- Manifold Sets
- Connecting Tubings

**Aspects of manufacture concerned with securing and
maintaining sterile conditions:**

- Dose-control Syringes
- Balloon Inflation Devices
- Colored Piston Specialty Syringes
- Infusion Sets with Needleless Adapters
- Pressure Bandages
- Hemostasis Valve Sets
- Injection Caps

Notified Body



Date: 2020-05-26

EC Declaration of Conformity

Manufacturer:

SCW MEDICATH LTD.

NO.4, Baolong 6th Road, Baolong Industrial
Town, Longgang District, Shenzhen, 518116,
Guangdong, P.R. China

whose single Authorized Representative:

OBELIS S.A.

Bd. Général Wahis 53
1030 Brussels, Belgium

We, the manufacturer, herewith declare that the products

Angiographic Syringes

Models: DSA-60-A1, DSA-150-A1, CT-100-X1, CT-200-A1, MRI-115-A1,
DSA-100-B1, DSA-150-B1, CT-200-B1, MRI-65-A1, CT-200-A2,
DSA-200-A1, DSA-130-A1, CT/DSA-150-B1, MRI-60-B1, CT-200-NE,
DSA-90-NE, DSA-120-NE, CT-100-NE, CT-190-A1, CT-190/190-A1,
CT-200/200-A2, CT-200/200-B1, CT-100/100-NE, CT-100/200-NE,
CT-200/200-NE, CT-200-EM, CT-200/200-EM, CT-200-MED,
CT-200/200-MED, MRI-65/115-A1, MRI-65/65-A1, MRI-115/115-A1,
MRI-60/60-B1, MRI-60/60-NE, MRI-65-MED, MRI-65/200-MED

GMDN Code: 15286

meet the provisions of Directive 93/42/EEC which apply to them.

The medical device has been assigned to class IIa according to Annex IX Rule 2 of the Directive 93/42/EEC. It bears the mark



The product concerned has been manufactured under a quality management system according to Annex II of Directive 93/42/EEC.

Compliance of the designated product with the Directive 93/42/EEC has been assured via assessment of the quality management system by the Notified Body

TÜV Rheinland LGA Products GmbH
Tillystraße 2
90431 Nürnberg
Deutschland

Certificate No.: HD 60144232 0001

Issue date: 26.05.2020

Expiry date: 26.05.2024

following the procedure relating to the EC Declaration of Conformity set out in Annex II of Directive 93/42/EEC.

This Declaration of conformity is valid in connection with the release document for the respective batch of produced devices.

The above mentioned declaration of conformity is exclusively under the responsibility of



SCW MEDICATH LTD.
NO.4, Baolong 6th Road, Baolong Industrial Town,
Longgang District, Shenzhen, 518116, Guangdong, P.R. China

Shenzhen, 2022/09/23

Place, date

Miniam Xie

Legally binding signature, Function