

Certificate of Analysis

Product name: BioPro RABIES ELISA Ab kit
Catalogue number: RAB 01-02
Kit batch number: R2501
Expiry date: 31-12-2026
Kit components:

Quantity	Component	Batch
2 x 1	Rabies coated microplates	1R240301
1 x 0,6 ml	Rabies Positive control serum	2R241212
1 x 0,6 ml	Rabies Negative control serum	3R241212
1 x 0,3 ml	Rabies control serum 1	12R241212
1 x 0,3 ml	Rabies control serum 2	13R241212
1 x 0,3 ml	Rabies control serum 3	14R241212
1 x 15 ml	Sample diluent	4R241212
1 x 30 ml	Diluent for biotinylated antibody	5R241212
1 x 0,3 ml	Biotinylated anti-Rabies antibody (100x)	6R241231
1 x 30 ml	Diluent for Streptavidin peroxidase conjugate	7R241212
1 x 0,3 ml	Streptavidin peroxidase conjugate (100x)	8R241231
1 x 30 ml	TMB Substrate	9R241212
1 x 15 ml	Stop solution	10R241212
1 x 250 ml	Washing solution (10x)	11R241212
6 x 1	Adhesive foils	-

Kit control

Negative control serum OD : 1,625
Positive control serum OD: 0,171
Rabies control serum 1 PB: 64,1%
Rabies control serum 2 PB: 43,2%
Rabies control serum 3 PB: 15,9%
Inter plate: 2,0%
Intra plate: 1,9%

<u>Rabies Panel</u>	<u>result</u>	<u>expected result</u>
1	positive	positive
2	positive	positive
3	positive	positive
4	negative	negative / positive
5	negative	negative

Publication date: 24.1.2025


-QM Signature: