

Certificate of Analysis

Cathode Buffer Container 3500/Flex Series

Product No. **4408256**
Lot No. **2604155**
Date of Manufacture **13MAY2026**
Expiration Date **05AUG2027**

TEST	SPECIFICATION	RESULT
Visual Inspection		
Clear, particulate free, viscous solution		Pass
pH Test		
Measured at 25.0°C ± 0.1°C		
pH Value	7.6 – 8.2	8.2
Conductivity		
Measured at 25.0°C ± 0.1°C		
Conductivity	1500 – 2000 µS/cm	1702 µS/cm

For Research Use Only. Not for use in diagnostic procedures.

Manufactured under a Quality Management System that meets the requirements of ISO 13485 and ISO 9001.
Site: Warrington, UK

Life Technologies
5781 Van Allen Way
www.thermofisher.com
For inquiries, contact us at
cofarequests@thermofisher.com



Quality Assurance
Issued 21MAY2026

CRT1000874_B