

CERTIFICATE OF CONFORMITY

Product: Pharmacy refrigerator

Tested by request of Zhongke Meiling Cryogenics Company Limited
No.1862 Zishi Road, Economic & Technological
Development Zone, 230601 Hefei, China

Manufactured at (name and place) Zhongke Meiling Cryogenics Company Limited
No.1862 Zishi Road, Economic & Technological
Development Zone, 230601 Hefei, China

Trade mark (if any): 

Model/Type Ref: YC-130EL, YC-130L, YC-315EL, YC-315L, YC-330EL,
YC-330L, YC-395EL, YC-395L, YC-525EL, YC-525L,
YC-55EL, YC-55L, YC-725EL, YC-725L, YC-75EL, YC-75L

Ratings and principal characteristics: 220-240 VAC, 50 Hz
YC-130EL, YC-130L : 0.98 A
YC-315EL, YC-315L : 1.35 A
YC-330EL, YC-330L : 1.31 A
YC-395EL, YC-395L : 1.8 A
YC-525EL, YC-525L : 2.49 A
YC-55EL, YC-55L : 0.9 A
YC-725EL, YC-725L : 3.90 A
YC-75EL, YC-75L : 0.92 A

Additional information (if any) : /

A sample of the product has been tested and found to be in conformity with :

IEC 61010-1:2010 + A1:2016, EN 61010-1:2010 + A1:2019

as shown in the test report No. SHES220300547301

This Certificate of Conformity is the result of testing a sample of the product submitted, in accordance with the provisions of the relevant specific standard. It does not imply an assessment of the whole production. This Certificate of Conformity has been issued in accordance with the SGS CEBEC certification program CPRStd003, which is available on request. Conformity of the produced products with the specimen tested remains on the full responsibility of the manufacturer.

Brussels, 20/04/2022


C. LANA,
Certification Manager

 This document is issued by the company under its General Conditions of Certification Services of SGS Belgium NV accessible at http://www.sgs.com/terms_and_conditions.htm.
Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein.

SGS Belgium NV – Division SGS CEBEC
Riverside Business Park
Bld. Internationalelaan, 55 Build. K
BE-1070 Brussels
Tel. +32 2 556 00 20
Fax +32 2 556 00 36


005-PROD