



BC "MOLDINDCONBANK" S.A.

Filiala "Invest"

Republica Moldova, MD-2068
mun. Chișinău, bd. Moscovei, 14/1
Tel. : (373-22) 43-44-81, 43-46-24
Fax : (373-22) 43-44-22
cod: MOLDMD2X329

Data 14. IAN. 2016
Nr. 03/2 - 19/23

Республика Молдова, MD-2068
мун. Кишинэу, бул. Московской, 14/1
Тел. : (373-22) 43-44-81, 43-46-24
Факс : (373-22) 43-44-22
код: MOLDMD2X329

Filiala „Invest” BC „Moldindconbank” SA confirmă existența contului curent în moneda națională al **“BIOSISTEM MLD” S.R.L. (c/f 1010600028048)**, cu **IBAN MD95ML000000002251429243**.

Codul băncii MOLDMD2X329.

Director

Nina Turcan

Director financiar



Nina Balmuş

Ex. Diana Brinza
Tel. 43-45-96

REPUBLICA



MOLDOVA

CERTIFICAT DE ÎNREGISTRARE

Societatea cu Răspundere Limitată "BIOSISTEM MLD"
— ESTE ÎNREGISTRATĂ LA CAMERA ÎNREGISTRĂRII DE STAT —

Numărul de identificare de stat - codul fiscal
1010600028048

Data înregistrării

12.08.2010

Data eliberării

12.08.2010

Svirepova Ludmila, registrator

*Funcția, numele, prenumele persoanei
care a eliberat certificatul*

L. Svirepova
semnătura

MD 0101250





I.P. "AGENȚIA SERVICII PUBLICE"

Departamentul înregistrare și licențiere a unităților de drept

EXTRAS

din Registrul de stat al persoanelor juridice

nr. 8506 din 28.04.2021

Denumirea completă: **Societatea cu Răspundere Limitată «BIOSISTEM MLD».**

Denumirea prescurtată: **«BIOSISTEM MLD» S.R.L.**

Forma juridică de organizare: **Societate cu Răspundere Limitată.**

Numărul de identificare de stat și codul fiscal: **1010600028048.**

Data înregistrării de stat: **12.08.2010.**

Sediul: **MD-2001, str. Albișoara, 16/1, ap.(of.) 7, mun. Chișinău, Republica Moldova.**

Obiectul principal de activitate:

- 1 Activitatea farmaceutică;**
- 2 Importul, fabricarea, comercializarea, asistența tehnică și (sau) reparația dispozitivelor medicale și (sau) a opticii;**
- 3 Acordarea asistenței medicale de către instituțiile medico-sanitare private;**
- 4 Comerțul cu ridicata al calculatoarelor, echipamentelor periferice și software-ului;**
- 5 Întreținerea și repararea mașinilor de birou și a tehnicii de calcul;**
- 6 Consultații în domeniul sistemelor de calcul.**

Capitalul social: **5400 lei.**

Administrator: POIATA VITALIE,

Asociați:

- 1. POIATA VITALIE 33,40 %**
- 2. NASEDCHIN ALEXANDR 33,30 %**
- 3. KOJEVNIKOV DMITRII 33,30 %.**

Prezentul extras este eliberat în temeiul art. 34 al Legii nr. 220-XVI din 19 octombrie 2007 privind înregistrarea de stat a persoanelor juridice și a întreprinzătorilor individuali și confirmă datele din Registrul de stat la data de: 28.04.2021.

Specialist coordonator
tel. 022-207-840



Lazari Aliona



EB 0358735

Lista fondatorilor Biosistem-mld SRL

Nr.	Nume, Prenume	IDNP
1.	Vitalie Poiata	0983103892591
2.	Alexandr Nasedchin	2002001070747
3.	Dmitrii Kojevnikov	0972305012362



Certificate

No. Q5 002596 0002 Rev. 01

Holder of Certificate: **Lansion Biotechnology Co., Ltd.**
No.2 Qiande Road, Science Park, Jiangning District
210000 Nanjing, Jiangsu Province
PEOPLE'S REPUBLIC OF CHINA

Facility(ies): Lansion Biotechnology Co., Ltd.
No.2 Qiande Road, Science Park, Jiangning District, 210000
Nanjing, Jiangsu Province, PEOPLE'S REPUBLIC OF CHINA

Certification Mark:



See Scope of Certificate.

Scope of Certificate: **Design and Development, Production and Distribution of
Dry Fluorescence Immunoassay Analyzer,
Dry Fluorescence Immunoassay test kit,
Coagulation Test Kit(Electrochemistry),
Handheld coagulation Analyzer**

Applied Standard(s): EN ISO 13485:2016
Medical devices - Quality management systems -
Requirements for regulatory purposes
(ISO 13485:2016)
DIN EN ISO 13485:2016

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned above has established and is maintaining a quality management system, which meets the requirements of the listed standard(s). All applicable requirements of the testing and certification regulation of TÜV SÜD Group have to be complied with. For details and certificate validity see:
www.tuvsud.com/ps-cert?q=cert:Q5 002596 0002 Rev. 01

Report No.: SH20126602
Valid from: 2021-04-12
Valid until: 2024-04-02

Date, 2021-04-12

Christoph Dicks
Head of Certification/Notified Body



DECLARATION OF CONFORMITY

According Directive 98/79/EC on in vitro diagnostic medical devices, Annex III.

Manufacturer: Lansion Biotechnology Co., Ltd.

Address: No.2, Qiande Road, Science Park, Jiangning District, 210000 Nanjing, Jiangsu Province, PEOPLE'S REPUBLIC OF CHINA.

European Representative: Lotus NL B.V.

Contact person: Peter E-mail: peter@lotusnl.com

Address: Koningin Julianaplein 10, 1e Verd, 2595AA, The Hague, Netherlands.

In Vitro Diagnostic Directive:

- D-Dimer Test Kit (Dry Fluorescence Immunoassay)

Category: Others.

Conformity assessment route: Declaration of Conformity IVDD Annex III.

Applicable Standards:

ISO 13485:2016	EN ISO 18113-3:2011	EN 13612:2002
ISO 14971:2019	EN 13641:2002	ISO 23640:2015
EN ISO 18113-1:2011	ISO 15223-1:2016	EN 62366-1:2015
EN ISO 18113-2:2011		

We, the manufacturer, here declare with sole responsibility that our product/s mentioned above meet/s the provisions of the Directive 98/79/EC of the European Parliament and of the Council on In Vitro Diagnostic Medical Devices.

We agree to develop, implement and maintain a documented post-production monitoring process.

Signed on:

Place: Nanjing, China

Name of authorized signatory:

Position held in the company: CTO

Date: 07/23/2020

Seal/Stamp:

Lansion Biotechnology Co., Ltd.



D-Dimer Test Kit User Manual

(Dry Fluorescence Immunoassay)

[PRODUCT NAME]

D-Dimer Test Kit (Dry Fluorescence Immunoassay)

[PACKAGE SPECIFICATION]

1 test/kit
5 tests/kit
25 tests/kit
50 tests/kit
100 tests/kit

[INTENDED USE]

D-Dimer Test Kit (Dry Fluorescence Immunoassay) is intended for in-vitro quantitative measurement of D-Dimer in human plasma. This test is used as an aid in the assessment and evaluation of patients suspected of deep vein thrombosis (DVT) and pulmonary embolism (PE), diagnosis of disseminated intravascular coagulation (DIC), effective evaluation and monitoring the effect of thrombolytic therapy, diagnosis and assessment of myocardial infarction and cerebral infarction.

[TEST PRINCIPLE]

D-Dimer Test Kit (Dry Fluorescence Immunoassay) uses the principle of antigen-antibody reaction. The testing sample will migrate forward due to capillary action, then the D-Dimer of the sample will combine with antibody which is attached to fluorescence microspheres. This marked complex is attached to the detection area of immobilized antibody and the other fluorescence microspheres are attached to the control area. When the test strip is inserted into the analyzer, the analyzer automatically scans two ribbons and detected the fluorescence intensity of the composite emission from the testing area and the control area. The ratio of the two fluorescence values was used to calculate the content of the detected substances.

[MAIN COMPONENTS]

1. D-Dimer test strip in a sealed pouch with desiccant.....25 tests
2. Sample diluent.....25 pieces
3. QR code card for calibration.....1 piece
4. User Manual.....1 piece
5. Quantitative suction and dropping tube (Optional).

Note: Do not mix or interchange different batches of kit.

[STORAGE AND VALIDITY]

Store the test kit at 4°C-30°C, with a valid period of 18 months.
Test strip should be used within 30 minutes once the foil pouch is opened.

[APPLICABLE DEVICES]

1. FIC-S100 Dry Fluorescence Immunoassay Analyzer
2. LS-1000 Dry Fluorescence Immunoassay Analyzer
3. LS-2000 Dry Fluorescence Immunoassay Analyzer

4. LS-1100 Dry Fluorescence Immunoassay Analyzer
5. LS-2100 Dry Fluorescence Immunoassay Analyzer
6. LS-4000 Hand-Held Dry Fluorescence Immunoassay Analyzer
7. LS-7000 Microfluidic and Dry Fluorescence Immunoassay Analyzer

[SAMPLE REQUIREMENT]

1. Used for human **plasma**. Other bodily fluids and samples may not get the accurate result.
2. Plasma sample can be anticoagulant with heparin and sodium citrate under aseptic conditions.
3. At room temperature, the test should be performed within 4 hours after the sample collection.
4. Plasma sample can be stored at 2°C-8°C for 3 days at most. If more than 3 days, it should be stored at -20°C. It is suggested to use fresh sample to test. Stale blood sample may not get accurate result.
5. The sample before testing should be recovered to room temperature (15°C-30°C).
6. **Sample Volume: 100μL**

[TEST PROCEDURE]

1. Collect samples according to user manual.
2. Before the test, the sample and test strip should be recovered to room temperature (15°C-30°C).

For LS-1100

3. Perform QR code calibration when necessary. (Details refer to LS-1100 User Manual)
4. On the main interface of LS-1100, press "Test" icon to enter testing interface. Input patient information, sample information, doctor information when necessary. (Details refer to LS-1100 User Manual)
5. Remove test strip from sealed pouch and put it on a clean table, horizontally placed.
6. Using pipette to deliver 100μL sample into one tube of sample diluent. Mix gently and thoroughly. And then drop 100μL of mixed fluid into the sample port in the test strip.

7. Reaction Time: 10 minutes

For panel inside: Insert the test strip into the analyzer immediately after sample dispensing. Then click "Test".

For panel outside: After reaction time 10 minutes is elapsed, insert the test strip into the analyzer and then click "Test".

8. The result will be shown on the screen and printed automatically.

Notes: It is required to perform QR code calibration when starting to use one new batch of kit.

[EXPECTED VALUE]

Cut-Off Value: 0.5μg/mL

D-Dimer concentration is determined using samples obtained from 200 apparently healthy individuals.

It is recommended that each laboratory establish its own reference range for the population it serves.

[INTERPRETATION OF RESULT]

1. If the test result of the sample is more than 10 μ g/mL, the analyzer displays ">10 μ g/mL", and if the result is less than 0.1 μ g/mL, the analyzer displays "<0.1 μ g/mL". Specific data can be exported through related software as needed.
2. When the sample concentration exceeds the detection limit, the maximum dilution ratio is 3 times when the sample is diluted with calf serum or negative sample.

[LIMITATION]

1. This kit is only for plasma.
2. The test result of this kit are only one of the diagnostic aids for the clinicians.
3. The result of the test should be evaluated in the context of all the clinical and laboratory data available. In those instances where the laboratory results do not agree with the clinical evaluation, additional tests should be performed accordingly.

[PRODUCT PERFORMANCE]

1. Measuring Range: 0.1-10 μ g/mL.
2. Lower Detection Limit: \leq 0.1 μ g/mL.
3. Accuracy: Verify with comparison experiments, the relative deviation \leq 15%, the correlation coefficient $r \geq 0.990$.
4. Within-Run Precision: \leq 15%.
5. Between-Run Precision: \leq 15%.
6. Hook Test: No hook effect with high concentration sample. Hook test was conducted with reference material exceeding the upper limit of linear range, and the detection result was greater than the upper limit of detection.

[PRECAUTIONS]

1. Only used for in vitro diagnostics.
2. Do not use the kit beyond the expiration date.
3. After the test strip is removed from the sealed pouch, it should be tested as soon as possible to avoid excessive time in the air, resulting in dampness.
4. Do not reuse the test strip.
5. The damaged test strip or package cannot be used.
6. Do not mix the components of different kits.

[REFERENCES]

1. Xu guanghai, Chen lixia. Experimental evaluation of quantitative determination of plasma d-dimer[J]. Chinese Journal of Misdiagnosis. 2010, 10(09): 2064-2065.
2. Chen jiuyan. Detection of d-dimer and its application in the diagnosis and treatment of thrombotic diseases. Shanxi Journal of Medicine. 2008, 37(24): 1123-1125.



Lansion Biotechnology Co., Ltd.

Add: No.2, Qiande Road, Jiangning District, Nanjing, China

Tel: 86-25-58577600

Fax: 86-25-58758600

E-mail: biz@lansionbio.com

Website: en.lansionbio.com



Llins Service & Consulting GmbH

Address: Am Heiligenhaus 7, 69126, Heidelberg, Germany

Tel: +49 176 63866127

Revision Date: May 31, 2019

Version Number: 0.1

Production date and expiration see the label.